

<b>REF</b> 595-2	6x5mL	<b>EXP</b> 2026-02-05	<b>LOT</b> 38C3CFE	更新日期: 2023-12
				1 / 4

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(ACP)酸性磷酸酶	(AK)阿米卡星	(Alb)白蛋白
(ALP)碱性磷酸酶	(ALT/ SGPT)丙氨酸氨基转移酶	(AMY)淀粉酶
(Apo-A I)载脂蛋白A I	(Apo-B)载脂蛋白B	(AST/ SGOT)天门冬氨酸氨基转移酶
(BUN)尿素氮	(C3)补体 C3	(C4)补体 C4
(Ca)钙	(Cer)铜蓝蛋白	(CHE)胆碱酯酶
(CHOL/ TC)总胆固醇	(CK)肌酸激酶	(Cl)氯
(CO2 (HCO3-))二氧化碳 (碳酸氢根)	(Cortisol)皮质醇	(Cr)肌酐
(CRBM)卡马西平	(DBIL)直接胆红素	(DGXN)地高辛
(EtOH)乙醇	(Fe)铁	(Gentamicin)庆大霉素
(GGT)L-γ-谷氨酰基转移酶	(GLU)葡萄糖	(Hap (HPT))触珠蛋白 (结合珠蛋白)
(HDL-C)高密度脂蛋白胆固醇	(IgA)免疫球蛋白 IgA	(IgG)免疫球蛋白 IgG
(IgM)免疫球蛋白 IgM	(K)钾	(LA)乳酸
(LDH)乳酸脱氢酶	(LDL-C)低密度脂蛋白胆固醇	(Li)锂
(LIP)脂肪酶	(Mg)镁	(Na)钠
(P)磷	(PAB)前白蛋白	(PAMY)胰淀粉酶
(Phenobarbital)苯巴比妥	(PHYT)苯妥英	(PRO/ TP)(总)蛋白
(Salicylate)水杨酸盐	(TBA)总胆汁酸	(TBIL)总胆红素
(TG)甘油三酯	(THEO)茶碱	(Tobramycin)妥布霉素
(TRF)转铁蛋白	(TT4)总T4	(UA)尿酸
(UIBC)不饱和铁结合力	(VALP)丙戊酸	(Vancomycin)万古霉素
(α-1-AG)α-1-酸性糖蛋白	(α-1-AT)α-1-抗胰蛋白酶	(β2M)β-2-微球蛋白

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。

<b>REF</b> 595-2	6x5mL	<b>EXP</b> 2026-02-05	<b>LOT</b> 38C3CFE	更新日期： 2023-12
				2 / 4

批号： 38C3CFE				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
<b>(ACP)酸性磷酸酶</b>				
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	14.5	11.6-17.4	
<b>(AK)阿米卡星</b>				
Beckman AU series Beckman Coulter \ Dipstick	μmol/L	19.2	15.4-23	
<b>(Alb)白蛋白</b>				
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	39.1	31.3-46.9	
<b>(ALP)碱性磷酸酶</b>				
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	168	134-202	
<b>(ALT/ SGPT)丙氨酸氨基转移酶</b>				
Beckman AU series Beckman Coulter \ UV-rate with P5P	U/L	127	102-152	
<b>(AMY)淀粉酶</b>				
Beckman AU series Beckman Coulter \ PNP-G7	U/L	116	92.8-139	
<b>(Apo-A I )载脂蛋白A I</b>				
Beckman AU series Beckman Coulter \ immunoturbidimetric	g/L	1.61	1.29-1.93	
<b>(Apo-B)载脂蛋白B</b>				
Beckman AU series Beckman Coulter \ immunoturbidimetric	g/L	0.849	0.689-1.01	
<b>(AST/ SGOT)天门冬氨酸氨基转移酶</b>				
Beckman AU series Beckman Coulter \ UV-rate with P5P	U/L	115	92-138	
<b>(BUN)尿素氮</b>				
Beckman AU series Beckman Coulter \ Urease-UV-rate	mmol/L	14.3	11.4-17.2	
<b>(C3)补体 C3</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	mg/dL	133	107-159	
<b>(C4)补体 C4</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	mg/dL	25.2	19.2-31.2	
<b>(Ca)钙</b>				
Beckman AU series Beckman Coulter \ BAPTA	mmol/L	2.37	1.89-2.85	
<b>(Cer)铜蓝蛋白</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	g/L	0.28	0.22-0.34	
<b>(CHE)胆碱酯酶</b>				
Beckman AU series Beckman Coulter \ DTNB	U/L	8450	6760-10100	
<b>(CHOL/ TC)总胆固醇</b>				
Beckman AU series Beckman Coulter \ Cholesterol oxidase	mmol/L	4.09	3.27-4.91	
<b>(CK)肌酸激酶</b>				
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	242	194-290	
<b>(Cl)氯</b>				
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	96.9	77.5-116	
<b>(CO2 (HCO3-) )二氧化碳 (碳酸氢根)</b>				
Beckman AU series Beckman Coulter \ PEPC	mmol/L	24.2	19.4-29	
<b>(Cortisol)皮质醇</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	189	151-227	
<b>(Cr)肌酐</b>				
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	169	135-203	

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。

<b>REF</b> 595-2	6x5mL	<b>EXP</b> 2026-02-05	<b>LOT</b> 38C3CFE	更新日期: 2023-12
				3 / 4

<b>(CRBM)卡马西平</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	µg/mL	9.69		7.75-11.6
<b>(DBIL)直接胆红素</b>				
Beckman AU series Beckman Coulter \ Diazotization -Blank Compensation	µmol/L	25.4		20.3-30.5
<b>(DGXN)地高辛</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.41		1.13-1.69
<b>(EtOH)乙醇</b>				
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	16.9		13.5-20.3
<b>(Fe)铁</b>				
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	µmol/L	25.1		20.1-30.1
<b>(Gentamicin)庆大霉素</b>				
Beckman AU series Beckman Coulter \ Dipstick	µg/mL	4.59		3.67-5.51
<b>(GGT)L-γ-谷氨酰基转移酶</b>				
Beckman AU series Beckman Coulter \ γ-glutamyl -4-nitroaniline	U/L	78.8		63-94.6
<b>(GLU)葡萄糖</b>				
Beckman AU series Beckman Coulter \ Hexokinase	mmol/L	6.72		5.38-8.06
<b>(Hap (HPT) )触珠蛋白 (结合珠蛋白)</b>				
Beckman AU series Beckman Coulter \ Enzyme linked immunosorbent assays (ELISA)	g/L	1.05		0.83-1.27
<b>(HDL-C)高密度脂蛋白胆固醇</b>				
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	1.33		1.07-1.59
<b>(IgA )免疫球蛋白 IgA</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	mg/L	2060		1640-2480
<b>(IgG)免疫球蛋白 IgG</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	mg/L	9660		7720-11600
<b>(IgM)免疫球蛋白 IgM</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	mg/L	863		683-1040
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	g/L	0.863		0.683-1.04
<b>(K)钾</b>				
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	4.07		3.25-4.89
<b>(LA)乳酸</b>				
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	1.83		1.47-2.19
<b>(LDH)乳酸脱氢酶</b>				
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	U/L	233		186-280
<b>(LDL-C)低密度脂蛋白胆固醇</b>				
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	2.44		1.96-2.92
<b>(Li)锂</b>				
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	1.16		0.92-1.4
<b>(LIP)脂肪酶</b>				
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	U/L	73.3		58.6-88
<b>(Mg)镁</b>				
Beckman AU series Beckman Coulter \ Xylidyl Blue	mmol/L	0.985		0.785-1.18
<b>(Na)钠</b>				
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	142		114-170

注: 此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。

<b>REF</b> 595-2	6x5mL	<b>EXP</b> 2026-02-05	<b>LOT</b> 38C3CFE	更新日期： 2023-12
				4 / 4

<b>(P)磷</b>				
Beckman AU series Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	1.28		1.02-1.54
<b>(PAB)前白蛋白</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	g/L	0.234		0.194-0.274
<b>(PAMY)胰淀粉酶</b>				
Beckman AU series Beckman Coulter \ Enzymatic	U/L	82		65.6-98.4
<b>(Phenobarbital)苯巴比妥</b>				
Beckman AU series Beckman Coulter \ Enzyme linked immunosorbent assays (ELISA)	μmol/L	105		84-126
<b>(PHYT)苯妥英</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	μg/mL	14.4		11.5-17.3
<b>(PRO/ TP)(总) 蛋白</b>				
Beckman AU series Beckman Coulter \ Biuret-endpoint	g/L	59.8		47.8-71.8
<b>(Salicylate)水杨酸盐</b>				
Beckman AU series Beckman Coulter \ Fluorescence polarization immunoassay (FPIA)	mmol/L	0.885		0.705-1.06
<b>(TBA)总胆汁酸</b>				
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	29.6		23.7-35.5
<b>(TBIL)总胆红素</b>				
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	54.4		43.5-65.3
<b>(TG)甘油三酯</b>				
Beckman AU series Beckman Coulter \ Free glycerol removed	mmol/L	1.55		1.23-1.87
<b>(THEO)茶碱</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	μg/mL	14.5		11.6-17.4
<b>(Tobramycin)妥布霉素</b>				
Beckman AU series Beckman Coulter \ Dipstick	μmol/L	10.9		8.72-13.1
<b>(TRF)转铁蛋白</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	mg/dL	246		196-296
<b>(TT4)总T4</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	nmol/L	115		92-138
<b>(UA)尿酸</b>				
Beckman AU series Beckman Coulter \ Uricase colorimetry	μmol/L	286		226-346
<b>(UIBC)不饱和铁结合力</b>				
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	27.3		21.8-32.8
<b>(VALP)丙戊酸</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	μmol/L	380		304-456
<b>(Vancomycin)万古霉素</b>				
Beckman AU series Beckman Coulter \ Dipstick	μmol/L	12.3		9.84-14.8
<b>(α-1-AG)α-1-酸性糖蛋白</b>				
Beckman AU series Beckman Coulter \ Enzyme linked immunosorbent assays (ELISA)	g/L	0.718		0.578-0.858
<b>(α-1-AT)α-1-抗胰蛋白酶</b>				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	131		105-157
<b>(β2M)β-2-微球蛋白</b>				
Beckman AU series Beckman Coulter \ Enzyme linked immunosorbent assays (ELISA)	mg/L	1.36		1.08-1.64

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。