

<b>REF</b> 8301-5-1	1x5mL	<b>EXP</b> 2023-11-16	<b>LOT</b> 1CB790B	更新日期： 2020-12
				1 / 1

批号：1CB790B				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
<b>(AK)阿米卡星</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	mg/L	4.94	3.98-5.9	
Roche Hitachi/Hitachi Systems Roche \ Chemiluminescence Immunoassay (CLIA )	mg/L	5.17	4.11-6.23	
<b>(caffeine)咖啡因</b>				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	µmol/L	20.7	16.5-24.9	
<b>(CRBM)卡马西平</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	µg/mL	4.49	3.61-5.37	
Roche Hitachi/Hitachi Systems Roche \ Chemiluminescence Immunoassay (CLIA )	µg/mL	5.23	4.19-6.27	
<b>(DGXN)地高辛</b>				
Roche Hitachi/Hitachi Systems Roche \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.09	0.87-1.31	
<b>(Gentamicin)庆大霉素</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	µg/mL	8.09	6.51-9.67	
<b>(Phenobarbital)苯巴比妥</b>				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	µmol/L	38.5	30.9-46.1	
Roche Hitachi/Hitachi Systems Roche \ Chemiluminescence Immunoassay (CLIA )	µmol/L	36.7	29.1-44.3	
<b>(PHYT)苯妥英</b>				
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA )	µg/mL	4.21	3.37-5.05	
Roche Hitachi/Hitachi Systems Roche \ Chemiluminescence Immunoassay (CLIA )	µg/mL	4.84	3.6-6.08	
<b>(Salicylate)水杨酸盐</b>				
Abbott ARCHITECT ci series Abbott \ Enzymatic	mg/dL	7.17	5.75-8.59	
Beckman AU series Beckman Coulter \ Enzymatic	mg/dL	7.26	5.82-8.7	
<b>(THEO)茶碱</b>				
Roche Hitachi/Hitachi Systems Roche \ Chemiluminescence Immunoassay (CLIA )	µg/mL	4.96	3.98-5.94	
Beckman AU series Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	µg/mL	5.06	4.04-6.08	
<b>(Tobramycin)妥布霉素</b>				
Beckman AU series Bayer \ Chemiluminescence Immunoassay (CLIA )	µmol/L	4.26	3.38-5.14	
Roche Hitachi/Hitachi Systems Roche \ Chemiluminescence Immunoassay (CLIA )	µmol/L	4.58	3.66-5.5	
<b>(VALP)丙戊酸</b>				
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA )	µmol/L	929	749-1100	
Roche Hitachi/Hitachi Systems Roche \ Chemiluminescence Immunoassay (CLIA )	µmol/L	977	783-1170	
<b>(Vancomycin)万古霉素</b>				
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA )	µmol/L	5.27	4.22-6.32	
Roche Hitachi/Hitachi Systems Roche \ Chemiluminescence Immunoassay (CLIA )	µmol/L	6.1	4.95-7.25	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。