

REF 595-3	6x5mL	EXP 2026-02-05	LOT 38C640E	更新日期: 2024-06
				1 / 5

本报告含以下项目的示范实验室数据

分析项目(Analyte)	分析项目(Analyte)	分析项目(Analyte)
(ACP)酸性磷酸酶	(a-HBDH) α -羟丁酸脱氢酶	(AK)阿米卡星
(Alb)白蛋白	(ALP)碱性磷酸酶	(ALT/ SGPT)丙氨酸氨基转移酶
(AMY)淀粉酶	(Apo-A I)载脂蛋白A I	(Apo-B)载脂蛋白B
(AST/ SGOT)天门冬氨酸氨基转移酶	(BUN)尿素氮	(C3)补体 C3
(C4)补体 C4	(Ca)钙	(Cer)铜蓝蛋白
(CHE)胆碱酯酶	(CHOL/ TC)总胆固醇	(CK)肌酸激酶
(Cl)氯	(CO2 (HCO3-))二氧化碳 (碳酸氢根)	(Cortisol)皮质醇
(Cr)肌酐	(CRBM)卡马西平	(DBIL)直接胆红素
(DGXN)地高辛	(EtOH)乙醇	(Fe)铁
(Gentamicin)庆大霉素	(GGT)L- γ -谷氨酰基转移酶	(GLU)葡萄糖
(Hap (HPT))触珠蛋白 (结合珠蛋白)	(HDL-C)高密度脂蛋白胆固醇	(IgA)免疫球蛋白 IgA
(IgG)免疫球蛋白 IgG	(IgM)免疫球蛋白 IgM	(K)钾
(LA)乳酸	(LDH)乳酸脱氢酶	(LDL-C)低密度脂蛋白胆固醇
(Li)锂	(LIP)脂肪酶	(Mg)镁
(Na)钠	(P)磷	(PAB)前白蛋白
(PAMY)胰淀粉酶	(Phenobarbital)苯巴比妥	(PHYT)苯妥英
(PRO/ TP)(总)蛋白	(Salicylate)水杨酸盐	(TBA)总胆汁酸
(TBIL)总胆红素	(TG)甘油三酯	(THEO)茶碱
(Tobramycin)妥布霉素	(TRF)转铁蛋白	(TT4)总T4
(UA)尿酸	(UIBC)不饱和铁结合力	(Urea)尿素
(VALP)丙戊酸	(Vancomycin)万古霉素	(α -1-AG) α -1-酸性糖蛋白
(α -1-AT) α -1-抗胰蛋白酶	(β 2M) β -2-微球蛋白	

注: 此报告所提供的分析与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。

REF 595-3	6x5mL	EXP 2026-02-05	LOT 38C640E	更新日期： 2024-06
				2 / 5

批号： 38C640E				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
(ACP)酸性磷酸酶				
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	22	17.6-26.4	
(a-HBDH)α-羟丁酸脱氢酶				
Roche Hitachi/Hitachi Systems chongqiong zybio \ DGKC	U/L	453	362-544	
(AK)阿米卡星				
Beckman AU series Beckman Coulter \ Dipstick	μmol/L	59.6	47.7-71.5	
(Alb)白蛋白				
Roche Hitachi/Hitachi Systems chongqiong zybio \ Bromocresol green	g/L	50.6	40.5-60.7	
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	54.1	43.3-64.9	
(ALP)碱性磷酸酶				
Roche Hitachi/Hitachi Systems chongqiong zybio \ dry chemistry	U/L	382	306-458	
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	351	281-421	
(ALT/ SGPT)丙氨酸氨基转移酶				
Beckman AU series Beckman Coulter \ UV-rate with P5P	U/L	203	162-244	
Roche Hitachi/Hitachi Systems chongqiong zybio \ dry chemistry	U/L	204	163-245	
(AMY)淀粉酶				
Beckman AU series Beckman Coulter \ PNP-G7	U/L	360	288-432	
Roche Hitachi/Hitachi Systems chongqiong zybio \ CNP_G3	U/L	366	293-439	
(Apo-A I)载脂蛋白A I				
Beckman AU series Beckman Coulter \ immunoturbidimetric	g/L	2.65	2.11-3.19	
(Apo-B)载脂蛋白B				
Beckman AU series Beckman Coulter \ immunoturbidimetric	g/L	1.29	1.03-1.55	
(AST/ SGOT)天门冬氨酸氨基转移酶				
Roche Hitachi/Hitachi Systems chongqiong zybio \ dry chemistry	U/L	247	198-296	
Beckman AU series Beckman Coulter \ UV-rate with P5P	U/L	216	173-259	
(BUN)尿素氮				
Beckman AU series Beckman Coulter \ Urease-UV-rate	mmol/L	22.5	18-27	
(C3)补体 C3				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mg/dL	193	155-231	
(C4)补体 C4				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mg/dL	34.9	28.9-40.9	
(Ca)钙				
Roche Hitachi/Hitachi Systems chongqiong zybio \ Arsenazo III	mmol/L	3.53	2.83-4.23	
Beckman AU series Beckman Coulter \ BAPTA	mmol/L	3.43	2.75-4.11	
(Cer)铜蓝蛋白				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	g/L	0.354	0.274-0.434	
(CHE)胆碱酯酶				
Beckman AU series Beckman Coulter \ DTNB	U/L	11800	9430-14100	
Roche Hitachi/Hitachi Systems chongqiong zybio \ Butyrylthiocholine Method	U/L	9930	7940-11900	
(CHOL/ TC)总胆固醇				
Beckman AU series Beckman Coulter \ Cholesterol oxidase	mmol/L	6.21	4.97-7.45	
Roche Hitachi/Hitachi Systems chongqiong zybio \ Cholesterol oxidase	mmol/L	6.48	5.18-7.78	

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。

REF 595-3	6x5mL	EXP 2026-02-05	LOT 38C640E	更新日期： 2024-06
				3 / 5

(CK)肌酸激酶				
Roche Hitachi/Hitachi Systems chongqiong zybio \ DGKC	U/L	440		352-528
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	419		335-503
(Cl)氯				
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	106		84.8-127
Roche Hitachi/Hitachi Systems chongqiong zybio \ Coulometric titration	mmol/L	105		84-126
(CO2 (HCO3-))二氧化碳 (碳酸氢根)				
Beckman AU series Beckman Coulter \ PEPC	mmol/L	32.9		26.3-39.5
(Cortisol)皮质醇				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	291		233-349
(Cr)肌酐				
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	529		423-635
Roche Hitachi/Hitachi Systems chongqiong zybio \ dry chemistry	μmol/L	505		404-606
(CRBM)卡马西平				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μg/mL	12.8		10.2-15.4
(DBIL)直接胆红素				
Roche Hitachi/Hitachi Systems chongqiong zybio \ Diazotization	μmol/L	28.6		22.9-34.3
Beckman AU series Beckman Coulter \ Diazotization -Blank Compensation	μmol/L	47.2		37.8-56.6
(DGXN)地高辛				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.58		2.06-3.1
(EtOH)乙醇				
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	42.3		33.8-50.8
(Fe)铁				
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	41		32.8-49.2
Roche Hitachi/Hitachi Systems chongqiong zybio \ Atomic absorption spectroscopy (AAS)	μmol/L	36.8		29.4-44.2
(Gentamicin)庆大霉素				
Beckman AU series Beckman Coulter \ Dipstick	μg/mL	7.5		6-9
(GGT)L-γ-谷氨酰基转移酶				
Beckman AU series Beckman Coulter \ γ-glutamyl -4-nitroaniline	U/L	151		121-181
Roche Hitachi/Hitachi Systems chongqiong zybio \ dry chemistry	U/L	136		109-163
(GLU)葡萄糖				
Roche Hitachi/Hitachi Systems chongqiong zybio \ Dehydrogenase	mmol/L	16.1		12.9-19.3
Beckman AU series Beckman Coulter \ Hexokinase	mmol/L	16.4		13.1-19.7
(Hap (HPT))触珠蛋白 (结合珠蛋白)				
Beckman AU series Beckman Coulter \ Enzyme linked immunosorbent assays (ELISA)	g/L	1.46		1.16-1.76
(HDL-C)高密度脂蛋白胆固醇				
Roche Hitachi/Hitachi Systems chongqiong zybio \ Direct Measure	mmol/L	1.6		1.28-1.92
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	2.12		1.7-2.54
(IgA)免疫球蛋白 IgA				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mg/L	2970		2370-3570
(IgG)免疫球蛋白 IgG				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mg/L	13900		11100-16600

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。

REF 595-3	6x5mL	EXP 2026-02-05	LOT 38C640E	更新日期： 2024-06
				4 / 5

(IgM)免疫球蛋白 IgM				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mg/L	1230		990-1470
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	g/L	1.23		0.99-1.47
(K)钾				
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	6.99		5.59-8.39
Roche Hitachi/Hitachi Systems chongqiong zybio \ dry chemistry	mmol/L	6.98		5.58-8.38
(LA)乳酸				
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	6.2		4.96-7.44
(LDH)乳酸脱氢酶				
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	U/L	444		355-533
Roche Hitachi/Hitachi Systems chongqiong zybio \ dry chemistry	U/L	435		348-522
(LDL-C)低密度脂蛋白胆固醇				
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	3.97		3.17-4.77
Roche Hitachi/Hitachi Systems chongqiong zybio \ SUR	mmol/L	3.76		3-4.52
(Li)锂				
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	1.96		1.56-2.36
(LIP)脂肪酶				
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	U/L	257		206-308
(Mg)镁				
Roche Hitachi/Hitachi Systems chongqiong zybio \ Arsenazo □	mmol/L	1.88		1.5-2.26
Beckman AU series Beckman Coulter \ Xylidyl Blue	mmol/L	1.79		1.43-2.15
(Na)钠				
Roche Hitachi/Hitachi Systems chongqiong zybio \ dry chemistry	mmol/L	158		126-190
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	160		128-192
(P)磷				
Beckman AU series Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	2.3		1.84-2.76
Roche Hitachi/Hitachi Systems chongqiong zybio \ dry chemistry	mmol/L	2.28		1.82-2.74
(PAB)前白蛋白				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	g/L	0.332		0.272-0.392
(PAMY)胰淀粉酶				
Beckman AU series Beckman Coulter \ Enzymatic	U/L	214		171-257
(Phenobarbital)苯巴比妥				
Beckman AU series Beckman Coulter \ Enzyme linked immunosorbent assays (ELISA)	μmol/L	245		196-294
(PHYT)苯妥英				
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μg/mL	24.6		19.7-29.5
(PRO/ TP)(总) 蛋白				
Beckman AU series Beckman Coulter \ Biuret-endpoint	g/L	86.6		69.3-104
Roche Hitachi/Hitachi Systems chongqiong zybio \ Biuret-endpoint	g/L	89		71.2-107
(Salicylate)水杨酸盐				
Beckman AU series Beckman Coulter \ Fluorescence polarization immunoassay (FPIA)	mmol/L	2.59		2.07-3.11
(TBA)总胆汁酸				
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	65.4		52.3-78.5
Roche Hitachi/Hitachi Systems chongqiong zybio \ Enzymatic Cycling Assays Method	μmol/L	68.4		54.7-82.1

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。

REF 595-3	6x5mL	EXP 2026-02-05	LOT 38C640E	更新日期： 2024-06
				5 / 5

(TBIL)总胆红素				
Beckman AU series				
Beckman Coulter \ Diazotization	μmol/L	91.4		73.1-110
Roche Hitachi/Hitachi Systems				
chongqiong zhibio \ Diazotization	μmol/L	74.2		59.4-89
(TG)甘油三酯				
Roche Hitachi/Hitachi Systems				
chongqiong zhibio \ Acetyl - acetone photometry	mmol/L	2.41		1.93-2.89
Beckman AU series				
Beckman Coulter \ Free glycerol removed	mmol/L	2.25		1.79-2.71
(THEO)茶碱				
Beckman AU series				
Beckman Coulter \ Chemiluminescence	μg/mL	25.6		20.5-30.7
Immunoassay (CLIA)				
(Tobramycin)妥布霉素				
Beckman AU series				
Beckman Coulter \ Dipstick	μmol/L	16.1		12.9-19.3
(TRF)转铁蛋白				
Beckman AU series				
Beckman Coulter \ Chemiluminescence	mg/dL	346		276-416
Immunoassay (CLIA)				
(TT4)总T4				
Beckman AU series				
Beckman Coulter \ Chemiluminescence	nmol/L	185		148-222
Immunoassay (CLIA)				
(UA)尿酸				
Roche Hitachi/Hitachi Systems				
chongqiong zhibio \ dry chemistry	μmol/L	492		394-590
Beckman AU series				
Beckman Coulter \ Uricase colorimetry	μmol/L	493		393-593
(UIBC)不饱和铁结合力				
Beckman AU series				
Beckman Coulter \ Ferrozine colorimetry	μmol/L	33.4		26.7-40.1
(Urea)尿素				
Roche Hitachi/Hitachi Systems				
chongqiong zhibio \ Diacetyl monoxime	mmol/L	20.6		16.5-24.7
(VALP)丙戊酸				
Beckman AU series				
Beckman Coulter \ Chemiluminescence	μmol/L	559		447-671
Immunoassay (CLIA)				
(Vancomycin)万古霉素				
Beckman AU series				
Beckman Coulter \ Dipstick	μmol/L	23.6		18.9-28.3
(α-1-AG)α-1-酸性糖蛋白				
Beckman AU series				
Beckman Coulter \ Enzyme linked immunosorbent	g/L	1		0.8-1.2
assays (ELISA)				
(α-1-AT)α-1-抗胰蛋白酶				
Beckman AU series				
Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	183		147-219
(β2M)β-2-微球蛋白				
Beckman AU series				
Beckman Coulter \ Enzyme linked immunosorbent	mg/L	1.98		1.58-2.38
assays (ELISA)				

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。