

REF	798-N 6x5mL	EXP 2024-12-21	LOT 4C1AF5E	更新日期: 2024-09 1 / 2
------------	----------------	-----------------------	--------------------	------------------------

本报告含以下项目的示范实验室数据

参数 (Parameters)	参数 (Parameters)	参数 (Parameters)
(CMV IgG)抗巨细胞病毒 IgG	(TOX IgG)抗弓形体 IgG	(RV IgG)抗风疹病毒 IgG
(CMV IgM)抗巨细胞病毒 IgM	(TOX IgM)抗弓形体 IgM	(RV IgM)抗风疹病毒 IgM
(HSV-1/2 IgM)抗1/2型单纯疱疹病毒 IgM	(HSV-1/2 IgG)抗1/2型单纯疱疹病毒IgG	

REF	798-N	6x5mL	EXP	2024-12-21	LOT	4C1AF5E	更新日期:	2024-09
								2 / 2

批号: 4C1AF5E			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(CMV IgG)抗巨细胞病毒 IgG			
DiaSorin Liaison /XL			
Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	4.38	2.98-5.78
DiaSorin Liaison /XL			
Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	5	5-5
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	77.3	61.8-92.8
(CMV IgM)抗巨细胞病毒 IgM			
DiaSorin Liaison /XL			
Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	5.46	4.22-6.7
DiaSorin Liaison /XL			
Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	5	4.93-5.07
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	110	88-132
(HSV-1/2 IgG)抗1/2型单纯疱疹病毒IgG			
DiaSorin Liaison /XL			
Diasorin \ Chemiluminescence Immunoassay (CLIA)	(Index)	0.5	0.492-0.508
DiaSorin Liaison /XL			
Dedicated \ Chemiluminescence Immunoassay (CLIA)	(Index)	0.47	0.31-0.63
(HSV-1/2 IgM)抗1/2型单纯疱疹病毒 IgM			
DiaSorin Liaison /XL			
Diasorin \ Chemiluminescence Immunoassay (CLIA)	(Index)	0.5	0.5-0.5
DiaSorin Liaison /XL			
Diasorin \ Electro-Chemiluminescence Immunoassay (ECLIA)	(Index)	0.5	0.5-0.5
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	0.54	0.44-0.64
DiaSorin Liaison /XL			
Dedicated \ Chemiluminescence Immunoassay (CLIA)	(Index)	0.56	0.38-0.74
(RV IgG)抗风疹病毒 IgG			
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	30.6	24.5-36.7
DiaSorin Liaison /XL			
Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.23	0.11-0.35
(RV IgM)抗风疹病毒 IgM			
DiaSorin Liaison /XL			
Dedicated \ Chemiluminescence Immunoassay (CLIA)	AU/mL	9.37	7.63-11.11
DiaSorin Liaison /XL			
Diasorin \ Chemiluminescence Immunoassay (CLIA)	AU/mL	10	10-10
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	133	106-160
(TOX IgG)抗弓形体 IgG			
DiaSorin Liaison /XL			
Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	3.62	2.14-5.1
DiaSorin Liaison /XL			
Diasorin \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	3	3-3
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	18.1	14.5-21.7
(TOX IgM)抗弓形体 IgM			
DiaSorin Liaison /XL			
Dedicated \ Chemiluminescence Immunoassay (CLIA)	AU/mL	2.98	2.57-3.39
DiaSorin Liaison /XL			
Diasorin \ Chemiluminescence Immunoassay (CLIA)	AU/mL	3	3-3
maglumi system			
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	18.3	14.6-22

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。