

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
1 / 14								

		批号：3382806			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(17-OHP)17-羟孕酮					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.568	10.5%	<5	<100
(25-OH-Vit D)25羟基维生素D					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	21.3	9.4%	50	1210
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	53.3	9.4%	50	1210
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	19.6	8.2%	11	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	49.1	8.2%	11	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.8	9.8%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	nmol/L	41.9	9.8%	<5	<100
(AFP)甲胎蛋白					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.68	11.6%	6	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.28	11.6%	59	1552
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.6	10.7%	50	951
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.84	14.3%	5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.36	5.2%	<5	115
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.08	7.9%	<5	105
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.34	9%	33	896
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7	4.9%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.6	8.2%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.36	7.1%	25	188
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.64	3%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.54	4.8%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.29	9.1%	18	163
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.15	5.1%	16	359
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.14	4.9%	13	192
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.37	11.5%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.18	10.2%	<5	<100
(ALD)醛固酮					
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	pg/mL	66.8	15.3%	14	142
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	62.4	9.8%	<5	<100
(Anti-Tg)抗甲状腺球蛋白抗体					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	15.1	23.4%	8	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	18	23.5%	8	137
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	9.65	8.2%	5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	1.74	15.1%	32	802
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	7.68	4.2%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	9.88	11.9%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.2	13.5%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.991	9.4%	24	664
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.846	14%	<5	<100

质控品名： 复合免疫分析非定值质控品

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
2 / 14								

		批号：3382806			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	9.15	20.7%	14	376
Abbott ARCHITECT ci series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	20.2	7.1%	11	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.7	11.5%	11	113
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	1.45	8.3%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	9.91	9.3%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	6.73	3.5%	<5	<100
(Anti-TPO)抗甲状腺过氧化物酶抗体					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.6	23.1%	9	223
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.74	25.5%	6	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.4	23.3%	6	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	8.25	8.6%	46	1298
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	7.25	9%	33	831
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	12.9	4.8%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	18.5	18.6%	29	936
Abbott ARCHITECT ci series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	35.5	2.1%	11	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	13.4	40.4%	10	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.6	9.6%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	8.31	9.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	11.3	5.7%	<5	<100
(ASD)雄烯二酮					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.911	10.1%	<5	<100
(CA 125)糖类抗原 125					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	36.5	9.8%	9	182
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.1	5.6%	59	1540
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	32.9	5.2%	5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.7	4.3%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.7	5.9%	30	550
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.8	5%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34.8	4.9%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.8	7.4%	25	166
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	26.1	1.8%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	30.6	5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	38.7	7.1%	18	168
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	31.1	5.3%	16	313
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.1	3.8%	15	199
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.9	10.6%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	42.1	10.5%	<5	<100
OTHER Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.8	0%	<5	<100
(CA 15-3)糖类抗原 15-3					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27.8	5.3%	62	1608
Siemens ADVIA 1800/2400 Siemens \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	26.2	9.2%	5	<100

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
3 / 14								

批号：3382806					
项目\仪器\试剂方法		单位	均值 \ CV		Lab \ 点数
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)		IU/mL	26.8	7%	5 <100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)		IU/mL	23.6	13%	5 <100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	29.9	10.7%	31 566
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	24.6	7.4%	<5 <100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)		IU/mL	31	3.4%	<5 <100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	28.8	8.9%	<5 <100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)		IU/mL	14.6	3.7%	<5 <100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	14.4	8.3%	25 181
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	28.7	2.9%	<5 <100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence enzyme immunoassay (CLEIA)		IU/mL	27.2	4.9%	<5 <100
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)		IU/mL	29.1	8.5%	15 201
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)		IU/mL	26.6	10.5%	<5 <100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)		IU/mL	30.6	10.5%	<5 <100
OTHER Dedicated \ Chemiluminescence Immunoassay (CLIA)		IU/mL	14.3	0%	<5 <100
(CA 19-9)糖类抗原 19-9					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	24.6	5.8%	59 1526
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	25.3	10.3%	30 532
Siemens ADVIA 1800/2400 Siemens \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	47.8	7.6%	<5 <100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	38	6.4%	22 157
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	26.5	2.1%	<5 <100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence enzyme immunoassay (CLEIA)		IU/mL	31.6	5.1%	<5 <100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)		IU/mL	21.7	4.3%	15 310
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)		IU/mL	35.6	7.7%	13 147
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)		IU/mL	23.6	11.9%	<5 <100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)		IU/mL	20.4	10.8%	<5 <100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)		IU/mL	26.2	10.7%	<5 <100
OTHER Dedicated \ Chemiluminescence Immunoassay (CLIA)		IU/mL	34.5	0%	<5 <100
(CEA)癌胚抗原					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)		ng/mL	4.76	5%	8 131
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)		ng/mL	3.02	14.2%	6 <100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)		ng/mL	4.11	5.5%	6 <100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)		ng/mL	3.97	14.8%	44 827
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)		ng/mL	3.86	7.6%	<5 <100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)		ng/mL	4.54	23.4%	<5 126
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)		ng/mL	4.28	10.2%	33 872
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)		ng/mL	3.29	11.7%	<5 <100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)		ng/mL	3.23	7.7%	<5 <100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)		ng/mL	3.64	11.7%	22 263
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)		ng/mL	2.37	1.7%	<5 <100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)		ng/mL	4.59	11.8%	10 234
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)		ng/mL	4.62	10.4%	<5 <100

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
4 / 14								

		批号：3382806			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(CK-MB)肌酸激酶-MB (活性)					
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.95	8.8%	33	896
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.78	5.8%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.73	5.8%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.96	4%	20	417
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.08	10.3%	13	364
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.02	5.8%	12	290
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.04	8.5%	11	219
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.49	10.2%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.36	9.9%	<5	<100
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.5	2.9%	<5	<100
(Cortisol)皮质醇					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	134	5.5%	8	180
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	113	9.9%	44	1181
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	249	7.1%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	154	7.3%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	143	5.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	141	3.2%	13	362
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	135	8.3%	12	344
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	25.6	119%	11	119
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	158	10.1%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	152	9.5%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	142	9.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	136	3.4%	<5	<100
(C-P)C-肽					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.987	9.1%	9	163
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.02	5.9%	8	107
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.479	70.9%	64	1566
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.877	18.4%	6	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.697	7.6%	5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.572	6.5%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.956	3.4%	17	216
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.609	6.3%	17	190
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.967	5.4%	14	148
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.775	10.2%	13	107
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.93	9.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.54	11.1%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.962	0.8%	<5	<100
(DGXN)地高辛					
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.1	12.9%	14	124
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.09	10.1%	<5	<100

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
5 / 14								

批号：3382806					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.08	8.3%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.906	10.3%	<5	<100
(DHEA-S)硫酸脱氢表雄酮					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3446	28.1%	5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3366	4.7%	15	486
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3548	6.5%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3069	10%	<5	<100
(E2)雌二醇					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	310	7.9%	9	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	322	10.8%	8	191
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	354	16.6%	7	135
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	313	10.2%	5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	322	7.7%	5	<100
Abbott ARCHITECT i series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	366	7.2%	5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	341	12.7%	45	1118
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	344	8.1%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	313	7.7%	32	464
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	323	4.4%	32	1535
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	300	12.1%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	302	14.4%	25	681
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	308	6.3%	24	701
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	275	6.3%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	327	13.6%	16	173
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	266	10.2%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	328	9.2%	<5	<100
(EPO)促红细胞生成素					
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/mL	6.07	8.9%	44	1130
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	6.2	4.4%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	6.76	9.9%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	8.72	10.1%	<5	<100
(FA)叶酸					
maglumi system Snibe \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.42	3.8%	6	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.46	7.2%	5	103
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.87	8.2%	44	1130
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.14	6.7%	24	710
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.32	26.1%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.97	5.1%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.02	9.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.06	15.6%	15	251
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.38	9.3%	14	100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.47	16.9%	10	163

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
6 / 14								

		批号：3382806			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.82	8.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.99	10.1%	<5	<100
(Ferritin)铁蛋白					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	96	8.3%	66	1607
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	83.6	6.2%	5	<100
Siemens ADVIA 1800/2400 Siemens \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	71.1	8.4%	5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	89.4	12.1%	5	100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	61.8	9.9%	44	1158
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	65.8	6.6%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	92	9.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	66.4	13%	15	169
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	91.6	13%	13	158
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	95.4	8.6%	10	166
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	68.1	9.4%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	69.2	9.8%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	75.2	10.1%	<5	<100
OTHER Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	69	0%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.21	12.5%	9	173
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.28	7.2%	6	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.84	5.6%	59	1542
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.06	8.8%	33	899
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.79	5.7%	30	541
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.19	6.1%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.52	6%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.87	10.2%	<5	<100
(FSH)促卵泡刺激素					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.97	5.2%	9	157
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.21	2.2%	8	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.09	8.4%	8	175
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.59	29.6%	7	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.72	14.9%	5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	5.07	10%	44	1097
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.14	3%	31	816
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	4.18	8%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.99	13.4%	27	418
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.9	7.9%	26	714
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.35	5.9%	24	690
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.43	4.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.32	7.1%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.55	10.1%	16	203
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.93	10.5%	<5	<100

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
7 / 14								

		批号：3382806			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	4.05	2.8%	<5	<100
(FT3)游离T3					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.11	4.5%	9	153
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.22	7.6%	8	185
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.69	8%	8	139
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.09	8.6%	50	1354
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.8	10.4%	<5	109
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	5.06	7.6%	35	1652
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.27	5.5%	32	602
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	5.9	3.5%	31	808
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.46	5.3%	17	839
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.13	5.5%	15	214
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.45	4.1%	15	177
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.21	4.5%	13	258
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	5.98	11.1%	12	175
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.49	10.2%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.22	9.8%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	4.52	7.1%	<5	<100
(FT4)游离T4					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	13.3	5.8%	8	185
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	13.1	4.9%	8	136
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	15.4	7.2%	7	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	13.7	4.1%	5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	12.7	3.2%	5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.9	7.5%	49	1354
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	11.3	29.2%	<5	122
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	12.1	7.4%	33	1704
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	15.6	4.1%	32	819
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	17.4	5.6%	31	593
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.5	5.7%	18	837
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	15.8	5.3%	17	207
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	16.8	8.5%	16	239
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	16	8.6%	14	244
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	15.1	5.6%	12	134
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	13.3	11.3%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.4	11.3%	<5	<100
(HCY)同型半胱氨酸					
Beckman AU series Dedicated \ Enzymatic Cycling Assays Method	μmol/L	9.52	15.2%	24	593
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	μmol/L	8.79	5.7%	15	161
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μmol/L	15.7	10.8%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	μmol/L	12.1	9.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μmol/L	11.7	10.3%	<5	<100

质控品名： 复合免疫分析非定值质控品

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
8 / 14								

		批号：3382806			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(HGH)人生长激素					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	μg/L	1.4	13.9%	33	870
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	1.58	3.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μg/L	1.54	10.4%	<5	<100
(IgE)免疫球蛋白 IgE					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	153	5.4%	8	196
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	153	5.3%	50	1273
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	151	6.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	161	9.9%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	69.8	10.3%	<5	<100
(IGF1)胰岛素生长因子1					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	nmol/L	19.4	9.8%	<5	<100
(INS)胰岛素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.5	5%	9	162
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	9.16	8.2%	8	134
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.77	8.3%	62	1545
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	7.12	14.6%	6	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.61	10.6%	5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.67	5.8%	36	814
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	9.52	10.5%	33	719
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	5.72	6.6%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	7.02	4.9%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	7.13	7.1%	18	213
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	6.54	10.1%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.83	10%	<5	<100
(LH)促黄体生成素					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.77	17%	9	162
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.06	8.6%	8	175
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.22	6.5%	7	<100
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.03	4.6%	7	123
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.06	5.4%	5	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	3.77	7.2%	44	1124
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.44	3.8%	31	819
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.64	7.4%	30	417
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	4.99	6.4%	<5	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.82	6%	26	716
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.3	4.2%	24	686
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.84	3.1%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.27	5.4%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.19	11.8%	15	170
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.47	11.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.88	3.5%	<5	<100

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
9 / 14								

		批号：3382806			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	4.49	4.2%	<5	<100
(Myo / Mb)肌红蛋白					
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	μg/L	27.8	8%	33	912
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	26	7.1%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	26.9	4.2%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	29.4	5%	20	416
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	29.7	3.7%	12	285
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	28.5	7.7%	11	244
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μg/L	27.9	9.3%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	μg/L	27.3	9.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μg/L	34.2	10.5%	<5	<100
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	38.9	2.9%	<5	<100
(NT-proBNP)N-末端脑钠肽前体					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	140	11.2%	34	1466
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	16.5	11.2%	34	1466
Roche cobas E601/602 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	160	3.3%	<5	<100
Roche cobas E601/602 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	18.9	3.3%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	148	2.4%	<5	252
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	17.4	2.4%	<5	252
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	167	12.1%	23	837
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	19.7	12.1%	23	837
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	138	3.4%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	16.3	3.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	138	4.5%	14	417
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	16.2	4.5%	14	417
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	166	17.1%	14	161
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	19.5	17.1%	14	161
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	186	11.9%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	21.9	11.9%	<5	<100
(PRL)催/ 泌乳素					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.12	5.2%	8	183
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.7	6.6%	7	120
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.4	27.9%	6	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.4	4.8%	5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.12	9.5%	5	103
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.9	9.6%	44	1155
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.1	3.2%	30	801
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.75	4.8%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.08	6.3%	26	716
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.4	6.7%	24	684
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.33	13.9%	23	432

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
10 / 14								

		批号：3382806			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.56	4.1%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.91	14.6%	16	196
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.53	10.4%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.58	3%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	8.35	4.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.2	0%	<5	<100
OTHER Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.04	0%	<5	<100
(PROG)孕酮					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.43	9.5%	8	178
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.997	11.2%	8	130
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.04	17.4%	7	138
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.086	289.7%	6	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.53	18.4%	45	1140
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.3	17.3%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.01	15.9%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.06	11%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.874	24.5%	29	436
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.42	12.3%	26	725
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.08	9.5%	23	688
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.03	6.1%	22	943
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.5	12.8%	17	190
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.17	11.9%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.02	10.8%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.11	11.7%	<5	<100
(PSA)前列腺特异性抗原					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.27	10.6%	9	166
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.04	4.3%	8	156
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.58	4.6%	6	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5	3.2%	59	1541
Beckman UniCel DxI 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.5	6.5%	33	890
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.04	3.6%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.16	3.5%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.21	3.7%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.47	1.4%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.78	2.7%	<5	<100
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.86	10.4%	14	220
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.76	10.3%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.92	10.5%	<5	<100
(PTH)甲状旁腺素					
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	2.25	6%	8	161
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1.63	20%	8	129

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
11 / 14								

		批号：3382806			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1.98	8.9%	47	1085
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	1.89	8.8%	44	1144
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	2.01	8.4%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1.96	11%	15	166
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1.8	14.3%	13	111
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1.96	9.7%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	2.26	9.8%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	2.08	10.1%	<5	<100
(SHBG)性激素结合球蛋白					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	33	6.1%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	nmol/L	33.2	12%	<5	<100
(TES)睾酮					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.81	7.5%	9	162
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.08	4.3%	8	184
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.45	6.7%	7	133
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.62	5.3%	60	1628
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.19	12.1%	5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.36	3.2%	30	1038
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.24	9.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.93	5.5%	24	688
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.7	11.8%	23	418
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.92	6.6%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.95	12.2%	14	169
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.45	3.5%	11	275
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.1	10.6%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.7	5.8%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.09	5.1%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.72	0%	<5	<100
OTHER Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.7	0%	<5	<100
(Tg)甲状腺球蛋白					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.26	2.5%	9	229
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.1	7.6%	8	154
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.13	4.6%	47	1327
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.59	10.4%	33	842
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.01	8.8%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.6	13.8%	17	273
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.88	9.9%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.04	10.2%	<5	<100
(THEO)茶碱					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	µg/mL	9.66	9.9%	<5	<100

质控品名： 复合免疫分析非定值质控品

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
12 / 14								

		批号：3382806			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TnI)肌钙蛋白I					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.058	4.1%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.066	14.1%	24	705
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.032	12.1%	13	363
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.05	20%	<5	<100
(TnT)肌钙蛋白T					
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.016	5.1%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.016	7.1%	24	625
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.016	13.8%	23	544
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.015	8%	15	313
(TSH)促甲状腺素					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.97	5.1%	8	183
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.79	6%	8	139
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.9	3.3%	7	103
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	3.03	4.8%	50	1353
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.86	4.7%	5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	2.51	6%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.44	5.4%	42	644
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.34	4.3%	40	1087
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	2.8	5.8%	33	1724
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.48	5.2%	31	596
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.31	6.2%	23	409
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.5	4.8%	18	847
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	3.04	3.9%	15	177
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	2.89	7.6%	14	245
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	3.09	10.7%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.08	10.4%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.98	10.7%	<5	<100
OTHER Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.85	0%	<5	<100
(TT3)总三碘甲状原氨酸					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.39	8.3%	9	151
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.16	8.5%	8	191
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.15	4.8%	8	137
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.14	8.3%	49	1317
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.09	9.7%	33	878
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.66	22.4%	22	354
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.46	5.9%	19	359
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.2	4.2%	18	833
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.47	8.1%	15	143
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.61	8.1%	14	257
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.28	9.5%	11	154
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.34	5.5%	10	239

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
13 / 14								

		批号：3382806			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.38	9.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.56	9.6%	<5	<100
(TT4)总T4					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	61.8	6.3%	8	193
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	57.8	19.1%	8	140
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	49.5	6.8%	8	102
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	54.3	9.2%	7	<100
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	47.5	6%	6	227
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	48.5	6.5%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	75.8	17.2%	39	656
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	63.2	8.9%	33	880
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	44.3	11.8%	28	692
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	52	14.3%	19	348
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	54.1	11.4%	14	260
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	53.6	3.7%	10	242
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	51.6	10.1%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	44.9	10%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	50.2	2.9%	<5	<100
OTHER Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	67.9	0%	<5	<100
(VB12)维生素B12					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pg/mL	559	20.4%	8	<100
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	pg/mL	368	7.2%	6	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	376	5.5%	5	104
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pg/mL	282	10.3%	44	1154
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	283	6.9%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	309	7.8%	24	672
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	288	7.5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	588	9.4%	14	215
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pg/mL	451	17%	13	112
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	471	7.4%	10	155
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pg/mL	520	10.2%	<5	<100
(β2M)β-2-微球蛋白					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1158	8.1%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1380	10.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	960	9.4%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.004	5.3%	8	177
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	4.2%	7	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.004	16.7%	<5	<100
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	8.9%	20	492
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.004	2.6%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	12.5%	17	126

REF	540-1	6×3mL	EXP	2023-09-26	LOT	3382806	更新日期：	2024-04
14 / 14								

		批号：3382806			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.004	15.4%	16	172
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	5.2%	10	109
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.004	10.1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。