

REF	512-3	6×3mL	EXP	2024-04-04	LOT	30DBD67	更新日期：	2024-04
1 / 2								

		批号：30DBD67			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	187	3.3%	7	200
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	283	5.7%	6	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	116	2.7%	<5	<100
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	ng/mL	206	5.2%	16	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	282	10%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	272	10%	<5	<100
(CA 125)糖类抗原 125					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	142	3.1%	7	202
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	144	3.6%	6	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	131	1.9%	<5	<100
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	IU/mL	231	4.5%	17	109
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	158	10%	<5	<100
(CA 15-3)糖类抗原 15-3					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	137	6.8%	7	206
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	59.8	3%	<5	<100
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	IU/mL	157	17.6%	17	112
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	134	10%	<5	<100
(CA 19-9)糖类抗原 19-9					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	121	5.5%	8	217
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	157	4.5%	6	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	222	6.7%	<5	<100
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	IU/mL	194	6.2%	16	113
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	155	10%	<5	<100
(CA 72-4)糖类抗原 72-4					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	52.9	4.6%	7	199
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	55	8.4%	6	138
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	IU/mL	66.9	6.4%	17	106
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	63.1	10%	<5	<100
(CEA)癌胚抗原					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	28.4	3.1%	7	202
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	33.5	4.5%	6	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	34.2	3.6%	<5	<100
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	ng/mL	38	8.2%	17	116
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	32.6	10%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	36.7	10%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	64.4	3.1%	7	199
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	79.3	4.3%	6	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	66.1	2.6%	6	145
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	ng/mL	74.9	5.4%	17	108
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	64.7	10%	<5	<100

REF	512-3	6×3mL	EXP	2024-04-04	LOT	30DBD67	更新日期：	2024-04
2 / 2								

批号：30DBD67					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Ferritin)铁蛋白					
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	ng/mL	519	13.9%	7	<100
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	ng/mL	850	10%	<5	<100
(F-PSA)游离前列腺特异性抗原					
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.59	5.5%	9	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.26	4.2%	7	200
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.34	5%	6	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.97	3%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.21	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.09	10%	<5	<100
(NSE)神经元特异性烯醇化酶					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	21.8	4.7%	8	220
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24	5.1%	6	152
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.1	10%	<5	<100
(PSA)前列腺特异性抗原					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.9	3.1%	7	197
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.172	5.1%	6	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.49	4.9%	<5	<100
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.26	9.5%	16	109
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	27.7	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24.8	10%	<5	<100
(SCC)鳞状细胞癌抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	8.21	5.4%	<5	<100
(Tg)甲状腺球蛋白					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	105	10%	<5	<100
(β2M)β-2-微球蛋白					
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	ng/mL	28960	10%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.2	8%	10	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.198	10%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.207	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。