报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-3	6×3mL	EXP	2024-12-07	LOT	33876E5	更新日期：	2024-03

		批号：33876E5			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(25-OH-VD2)25羟基维生素D2					
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	55.1	6%	19	323
(25-OH-Vit D)25羟基维生素D					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	65.8	5.9%	6	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	164	5.9%	6	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	70.2	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	175	12%	<5	<100
(AFP)甲胎蛋白					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	211	8.1%	9	127
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	210	2.8%	8	240
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	228	6%	8	221
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	226	6.4%	7	155
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	253	6.1%	68	1351
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	253	2.7%	6	130
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	250	3.5%	32	882
OTHER Abbott \ OTHER	ng/mL	270	6.1%	29	567
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	207	6.3%	16	107
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	233	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	191	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	233	12%	<5	<100
(ALD)醛固酮					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pg/mL	5360	12%	<5	<100
(Anti-Tg)抗甲状腺球蛋白抗体					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	18.6	7.9%	9	139
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	19.7	9.4%	28	562
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	6.22	8.6%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	31.3	7.6%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	3.68	6.9%	18	217
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	28.7	12%	<5	<100
OTHER Abbott \ OTHER	IU/mL	18.4	0%	<5	<100
(Anti-TPO)抗甲状腺过氧化物酶抗体					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26	11.9%	9	137
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.7	7%	28	560
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	23.6	8.2%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	24	10.8%	19	315
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	34.5	4.8%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	13.8	12%	<5	<100
(CA 125)糖类抗原 125					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	95.2	3.4%	9	130
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	78.8	2.3%	6	131
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	92.9	3.7%	32	883
OTHER Abbott \ OTHER	IU/mL	125	8%	29	568
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	94.5	5.8%	26	488

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-3	6×3mL	EXP	2024-12-07	LOT	33876E5	更新日期：	2024-03

		批号：33876E5			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	108	6.2%	16	102
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	166	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	166	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/mL	130	12%	<5	<100
(CA 15-3)糖类抗原 15-3					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	84.5	12.9%	9	130
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	71.3	2.5%	6	129
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	82.1	4.3%	33	875
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	87.4	7.1%	26	482
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	81.1	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	81.1	12%	<5	<100
(CA 19-9)糖类抗原 19-9					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	133	7%	9	120
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	112	2.7%	6	131
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	104	4.7%	32	877
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	104	6.9%	26	486
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	205	6.6%	16	108
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	78.7	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	78.7	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/mL	194	12%	<5	<100
(CEA)癌胚抗原					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	28.8	5.2%	9	130
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	30.1	6.1%	8	238
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24.5	3.8%	68	1344
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	20.7	1.6%	6	132
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.7	4.3%	33	877
OTHER Abbott \ OTHER	ng/mL	29.9	6.9%	29	567
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23	10.1%	16	112
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	31.5	6.3%	13	327
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	43	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	21.3	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	43	12%	<5	<100
(CK)肌酸激酶					
Roche cobas 4000/6000/8000 Roche \ Urease-UV-rate	U/L	117	12%	<5	<100
(CK-MB mass)肌酸激酶-MB质量					
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	23.9	7.7%	9	157
Beckman AU series BSBE \ OTHER	ng/mL	30.1	6.8%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
Ortho Vitros 3600/5600 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24.1	4.9%	9	252
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	31.2	4.8%	42	847
J&J Vitros ECL Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	23.9	9.5%	18	341
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	24.2	5.8%	17	399
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	46.8	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	46.8	12%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-3	6×3mL	EXP	2024-12-07	LOT	33876E5	更新日期：	2024-03

		批号：33876E5			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	U/L	27.6	12%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	29.5	12%	<5	<100
(Cortisol)皮质醇					
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	507	5.7%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	297	5%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	321	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	321	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	364	12%	<5	<100
(C-P)C-肽					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.34	7.9%	7	100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.23	3%	32	855
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.89	6%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.4	3.7%	10	266
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.98	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.98	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.68	12%	<5	<100
(DGXN)地高辛					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.98	12%	<5	<100
(DHEA-S)硫酸脱氢表雄酮					
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9285	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9875	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9285	12%	<5	<100
(E2)雌二醇					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1633	1.8%	6	143
OTHER Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1684	4.9%	<5	<100
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1417	4.9%	28	566
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1594	4.4%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1850	7.1%	17	246
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	1461	6.6%	13	290
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1497	7.2%	10	125
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1512	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1510	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1850	12%	<5	<100
(EPO)促红细胞生成素					
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	149	12%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	149	12%	<5	<100
(FA)叶酸					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.2	12%	<5	<100
(Ferritin)铁蛋白					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	539	4.5%	32	880
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	ng/mL	448	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	420	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	420	12%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-3	6×3mL	EXP	2024-12-07	LOT	33876E5	更新日期：	2024-03

		批号：33876E5			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(F-PSA)游离前列腺特异性抗原					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.19	4.8%	32	867
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11	5.5%	28	569
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.55	7.3%	26	477
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.83	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.9	12%	<5	<100
OTHER Abbott \ OTHER	ng/mL	10.1	2.9%	<5	<100
(FSH)促卵泡刺激素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	60.1	8%	8	114
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	45	2.3%	6	143
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	57.1	5.7%	28	568
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	46.1	6.8%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/L	72.8	8.7%	19	325
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	67	5.4%	13	320
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	45.5	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	45.5	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	56.7	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/L	39.2	12%	<5	<100
(FT3)游离T3					
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	3.48	9.9%	8	226
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	23.9	2.5%	6	166
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	17.7	6.6%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	23.1	8.4%	22	452
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	20.3	4.5%	19	332
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	19.9	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	pmol/L	11.9	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	19.9	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	26.6	12%	<5	<100
(FT4)游离T4					
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	13.4	6.5%	8	227
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	61.7	4.3%	6	167
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	51.7	5.5%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	47.6	3.7%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	35.1	7.1%	22	442
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	48.8	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	48.8	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	pmol/L	40.8	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	55.9	12%	<5	<100
(hCG)人绒毛膜促性腺激素					
OTHER Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	376	2.6%	<5	<100
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	379	6.4%	23	455
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	618	6.3%	13	313
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/L	435	2.5%	<5	<100
(HCY)同型半胱氨酸					
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μmol/L	40.3	12%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-3	6×3mL	EXP	2024-12-07	LOT	33876E5	更新日期：	2024-03

		批号：33876E5			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μmol/L	40.3	12%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	39.6	12%	<5	<100
(HGH)人生长激素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	9.66	12%	<5	<100
(IgE)免疫球蛋白 IgE					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	322	5.2%	33	843
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	273	12%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	251	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	273	12%	<5	<100
(IGF1)胰岛素生长因子1					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	202	12%	<5	<100
(INS)胰岛素					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	144	5.3%	9	235
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	97.5	8.6%	8	123
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	130	4.5%	32	887
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	128	3.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	118	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	118	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	126	12%	<5	<100
(IntactPTH)全段甲状旁腺激素					
Abbott ARCHITECT i series Abbott \ OTHER	pg/mL	32.7	8.4%	6	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	40.2	4.8%	<5	<100
(LH)促黄体生成素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	44.7	8.8%	8	137
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	54.1	1.6%	6	142
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	45.7	7.2%	28	568
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	43.2	5.4%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/L	60.1	10.9%	18	312
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	74.5	5.7%	13	321
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	56.4	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	47.5	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	47.5	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/L	78.1	12%	<5	<100
(Myo / Mb)肌红蛋白					
Ortho Vitros 3600/5600 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	345	3.9%	9	255
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	274	3.1%	42	851
J&J Vitros ECL Ortho \ Chemiluminescence Immunoassay (CLIA)	μg/L	335	7.5%	18	307
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	μg/L	358	6.9%	17	403
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	351	5.5%	10	155
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	289	12%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	289	12%	<5	<100
(NT-proBNP)N-末端脑钠肽前体					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	2337	5.3%	42	843
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	275	5.3%	42	843

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-3	6×3mL	EXP	2024-12-07	LOT	33876E5	更新日期：	2024-03

		批号：33876E5			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	2320	4.8%	33	903
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	273	4.8%	33	903
Ortho Vitros 3600/5600 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	2664	25.4%	<5	<100
Ortho Vitros 3600/5600 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	313	25.4%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	2215	10.6%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	261	10.6%	<5	<100
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	2577	8%	17	469
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	303	8%	17	469
J&J Vitros ECL Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	2090	0%	<5	<100
J&J Vitros ECL Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	246	0%	<5	<100
(Ost)骨泌素					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12	0%	<5	<100
(PCT)降钙素原					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.97	5%	31	857
(PRL)催/ 泌乳素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	38.3	9.7%	7	109
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	30.6	1.8%	6	144
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	29.4	5.4%	<5	<100
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	33.2	4.5%	28	555
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	36.6	9.6%	18	273
Siemens Immulite1000/2000 Siemens \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	30.3	5.7%	13	315
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	28.8	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	38.2	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	36.9	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	38.2	12%	<5	<100
(PROG)孕酮					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.9	9.8%	8	138
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.5	3.7%	6	142
Ortho Vitros 3600/5600 Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	21.4	2.7%	5	<100
OTHER Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.9	7.9%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.4	5.5%	<5	<100
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	19.5	4.7%	27	570
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	30.5	7.7%	19	318
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	17	5.6%	13	316
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	24.5	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	85.3	2.6%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	24.5	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24.2	12%	<5	<100
(PSA)前列腺特异性抗原					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	23.6	4.8%	9	120
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	22.2	3%	5	123
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	22.1	2.5%	33	875

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-3	6×3mL	EXP	2024-12-07	LOT	33876E5	更新日期：	2024-03

		批号：33876E5			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.9	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	22.4	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	28.5	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.9	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.7	12%	<5	<100
OTHER Abbott \ OTHER	ng/mL	19.4	4.7%	<5	<100
(PTH)甲状旁腺素					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	18.5	6%	18	484
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	20.2	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	25.8	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	25.8	12%	<5	<100
(TES)睾酮					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.31	8.1%	8	106
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.49	5%	6	144
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.43	4.4%	<5	<100
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.27	7.3%	28	565
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.49	6.2%	16	220
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.97	7.7%	13	311
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.49	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.49	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.54	12%	<5	<100
(Tg)甲状腺球蛋白					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.3	12%	<5	<100
(TnI)肌钙蛋白I					
J&J Vitros ECL Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.64	7.2%	18	329
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.07	5.7%	17	404
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.37	6.7%	10	156
(TnT)肌钙蛋白T					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.75	4.1%	42	855
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.75	7%	19	511
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.74	10.4%	17	470
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.98	12%	<5	<100
(TSH)促甲状腺素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	26.2	5.9%	9	132
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	1.97	5.3%	8	226
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	34.3	5.3%	6	163
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	34.2	5.5%	34	645
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	28.4	4.7%	<5	<100
OTHER Abbott \ OTHER	mIU/L	24.3	6.1%	29	567
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	mIU/L	38.6	6.9%	19	326
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	26.5	7%	13	272
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	29.4	12%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	mIU/L	28.9	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	29.4	12%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-3	6×3mL	EXP	2024-12-07	LOT	33876E5	更新日期：	2024-03

		批号：33876E5			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TT3)总三碘甲状原氨酸					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.53	7.2%	9	<100
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.75	8.7%	8	225
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.84	2.7%	6	171
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.155	471.1%	34	616
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.29	7%	<5	<100
OTHER Abbott \ OTHER	ng/mL	0.005	9.4%	29	564
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.59	4.9%	18	325
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.54	12%	<5	<100
(TT4)总T4					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	167	5%	9	<100
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	91.7	4%	8	227
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	163	3.2%	6	169
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	204	9.8%	34	636
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	199	6.2%	<5	<100
OTHER Abbott \ OTHER	ng/mL	194	8%	29	562
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	180	6.4%	18	330
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	133	12%	<5	<100
(VB12)维生素B12					
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	512	6%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pg/mL	1160	6.9%	19	318
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	901	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pg/mL	976	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	976	12%	<5	<100
(β2M)β-2-微球蛋白					
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2810	12%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	ng/mL	2770	12%	<5	<100
Tosoh AIA 900/1800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2810	12%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.38	3.2%	6	139
OTHER ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.389	4.8%	5	105
Ortho Vitros 3600/5600 Ortho \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.553	1.4%	5	101
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.422	3.9%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.307	13.5%	19	308
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.443	6.9%	17	467
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.587	6.9%	10	150
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.413	12%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.522	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.522	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。