报告类别： **汇总实验室数据报告**

质控品名： **特殊肿瘤标志物液体非定值质控品**

REF	513-2	6×3mL	EXP	2024-03-22	LOT	30F1C30	更新日期：	2024-03

		批号：30F1C30			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	62.9	29.1%	29	695
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	52.1	12.9%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	107	7.5%	12	275
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	94.3	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	78.7	10%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	106	6.1%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	94.7	10%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	77.9	3.7%	<5	<100
(CA 125)糖类抗原 125					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	169	13.3%	7	105
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	289	7.6%	32	616
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	341	7.6%	<5	<100
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	202	4%	11	243
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	284	10%	<5	<100
DiaSorin Liaison /XL Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	309	2.7%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	182	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	217	10%	<5	<100
(CA 15-3)糖类抗原 15-3					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	117	8.4%	30	717
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	81.3	10.7%	25	404
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	38.2	8.4%	11	257
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	114	10%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	75.8	10%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	97.8	10%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	116	5%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	125	10%	<5	<100
(CA 19-9)糖类抗原 19-9					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	90.2	6.2%	7	103
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	443	12.4%	30	690
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	163	7.4%	12	280
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	100	10%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	237	10%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	107	4.1%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	482	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	204	10%	<5	<100
(CA 50)糖类抗原 50					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	66.6	5.8%	9	175
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	70.4	1.1%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **特殊肿瘤标志物液体非定值质控品**

REF	513-2	6×3mL	EXP	2024-03-22	LOT	30F1C30	更新日期：	2024-03

		批号：30F1C30			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
maglumi system Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	48.7	1.6%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	68.1	12%	<5	<100
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	49.7	6.7%	<5	<100
(CA 72-4)糖类抗原 72-4					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.9	6.5%	9	174
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.39	5.7%	7	101
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.15	13.6%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	4.99	5.3%	10	216
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.17	7.8%	10	213
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	13.4	10%	<5	<100
(CA242)糖类抗原 242					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	35.3	4.8%	6	120
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/mL	40.5	21.1%	5	114
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/mL	56.8	1.3%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	U/mL	56.6	12%	<5	<100
maglumi system Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	U/mL	24	2.4%	<5	<100
(CEA)癌胚抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	51.1	30.5%	28	581
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.5	15.4%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	29.2	9.5%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	48.8	6.5%	12	262
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	46.7	10%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	29.9	2.8%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	51.1	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	53.9	10%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	68.1	10%	<5	<100
(cPSA)复合前列腺特异性抗原					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/L	8366	3.8%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.7	5%	9	174
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.7	4.5%	5	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.1	2.5%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.3	4.3%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.5	11.6%	13	182
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.6	3.7%	12	237
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.5	4.6%	10	204
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.2	10%	<5	<100
(Ferritin)铁蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	398	8.9%	9	185
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	293	10%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	488	10%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **特殊肿瘤标志物液体非定值质控品**

REF	513-2	6×3mL	EXP	2024-03-22	LOT	30F1C30	更新日期：	2024-03

		批号：30F1C30			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	307	10%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	304	10%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.85	27.3%	28	662
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.59	6.3%	21	338
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.63	8.7%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.5	6.8%	12	277
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.11	10%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.88	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.12	10%	<5	<100
(hCG)人绒毛膜促性腺激素					
Siemens ADVIA Centaur XP/XPT Dedicated \ Enzyme linked immunosorbent assays (ELISA)	IU/L	110	10%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	153	10%	<5	<100
(HE4)人附睾蛋白4					
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	274	5%	9	187
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	412	8.6%	6	120
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	362	3.5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	311	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	318	13.4%	<5	<100
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	320	2.3%	11	252
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	260	5.4%	11	200
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	383	10%	<5	<100
(NSE)神经元特异性烯醇化酶					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.845	1.6%	<5	<100
(PSA)前列腺特异性抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15	10.1%	35	682
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	16.1	13.7%	21	337
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.1	13.8%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.5	5.9%	12	265
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.9	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	16.1	10%	<5	<100
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	20.1	10%	<5	<100
(SCC)鳞状细胞癌抗原					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.5	6.9%	9	174
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17	5.8%	<5	<100
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.2	1.7%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.5	9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.3	7.3%	11	242
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.9	8.4%	11	235
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.3	11.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.2	1.2%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **特殊肿瘤标志物液体非定值质控品**

REF	513-2	6×3mL	EXP	2024-03-22	LOT	30F1C30	更新日期：	2024-03

		批号：30F1C30			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TPA)组织多肽抗原					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	8030	10%	<5	<100
(TPS)组织多肽特异抗原					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	262	10%	<5	<100
(β2M)β-2-微球蛋白					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	941	5.3%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.208	1.1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。