

REF	513-1	6x3mL	EXP	2022-06-15	LOT	30EF64E	更新日期：	2022-01
								1 / 4

		批号：30EF64E			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.31	2.9%	5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.67	6.7%	22	578
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.25	10.5%	18	299
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.28	7.5%	12	125
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.53	4.2%	11	205
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.42	4.8%	10	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.38	8.9%	<5	<100
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.26	8.2%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.55	9.2%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.39	10%	<5	<100
(CA 125)糖类抗原 125					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	32.9	3.6%	5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	22.5	6.4%	31	414
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	23.1	3.4%	19	442
DiaSorin Liaison /XL Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.4	5.5%	13	119
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	35.1	5.1%	11	238
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	23.6	2.8%	10	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26	10%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.7	9.7%	<5	<100
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	26.4	8.3%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.9	8.8%	<5	<100
(CA 15-3)糖类抗原 15-3					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	16.2	13.2%	5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	15.6	7.3%	28	398
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	10.6	7.9%	22	564
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.6	3.8%	10	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.2	7.4%	10	193
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	22.8	10%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.4	7.3%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.2	7.2%	<5	<100
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	17.2	7.6%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	18.3	7.1%	<5	<100
(CA 19-9)糖类抗原 19-9					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	44.6	6.3%	5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.2	11.2%	30	412
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.9	5.8%	22	586
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.1	5.6%	11	217
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.7	10.4%	10	<100

REF	513-1	6x3mL	EXP	2022-06-15	LOT	30EF64E	更新日期：	2022-01
								2 / 4

		批号：30EF64E			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	IU/mL	15.3	7.2%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	47.4	7%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	23.1	9.5%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	20.2	10%	<5	<100
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	27.1	8.1%	<5	<100
(CA 50)糖类抗原 50					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	8.36	6.3%	5	<100
maglumi system Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	10.1	10.5%	13	105
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	10.1	8.9%	10	182
(CA 72-4)糖类抗原 72-4					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	2.37	8.5%	5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	IU/mL	1.76	7.8%	30	396
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	IU/mL	1.57	6.8%	18	409
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	IU/mL	1.51	10.8%	13	247
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	IU/mL	1.58	4.5%	10	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	IU/mL	2.08	8.2%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	1.04	10%	<5	<100
(CA242)糖类抗原 242					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	8.1	5.6%	<5	<100
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/mL	8.63	5.2%	13	265
maglumi system Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	U/mL	8.49	4.9%	13	106
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	U/mL	9.42	10%	<5	<100
(CEA)癌胚抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.34	3.1%	5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.44	6.1%	22	533
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	ng/mL	2.5	9.6%	18	283
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.71	8.1%	13	117
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	ng/mL	2.77	2.9%	11	205
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	ng/mL	2.82	2.6%	10	<100
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.48	7.3%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	ng/mL	3.39	7.7%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.51	8.4%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.96	8.8%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.58	10%	<5	<100
(cPSA)复合前列腺特异性抗原					
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/L	719	5.1%	13	<100
(Cyfra 21-1)细胞角蛋白19片段					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.45	8.4%	5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	ng/mL	1.6	11.2%	30	392
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (CLIA)	ng/mL	1.64	10.1%	19	438

REF	513-1	6x3mL	EXP	2022-06-15	LOT	30EF64E	更新日期：	2022-01
								3 / 4

		批号：30EF64E					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数			
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.46	3.2%	10	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1	4.3%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.92	5.7%	<5	<100		
(Ferritin)铁蛋白							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	47.7	5.9%	5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	47.8	9%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	52.9	9.3%	<5	<100		
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	35.3	9.1%	<5	<100		
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	45.1	9.5%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	43.3	10%	<5	<100		
(F-PSA)游离前列腺特异性抗原							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.557	4.5%	5	<100		
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.803	17.9%	31	416		
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.777	6.6%	22	587		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.896	15.2%	<5	<100		
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.14	7.9%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.01	8.9%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.92	10%	<5	<100		
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.41	6.4%	<5	<100		
(hCG)人绒毛膜促性腺激素							
Siemens ADVIA Centaur XP Dedicated \ Enzyme linked immunosorbent assays (ELISA)	IU/L	14.3	9.8%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	19.2	8.3%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/L	6.03	10%	<5	<100		
(HE4)人附睾蛋白4							
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	25.7	5.5%	7	124		
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	23.7	5.1%	5	<100		
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	28.6	5.6%	30	382		
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	29.5	3.2%	19	454		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	24.7	4.4%	13	244		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	29	3.5%	11	200		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	36.4	7.7%	<5	<100		
(NSE)神经元特异性烯醇化酶							
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.769	5.5%	13	119		
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.625	5%	10	<100		
(PSA)前列腺特异性抗原							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.14	3.4%	5	<100		
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.79	10.1%	31	417		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.15	26.5%	<5	<100		
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.36	8%	22	582		
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.41	9.1%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.82	10%	<5	<100		

REF	513-1	6x3mL	EXP	2022-06-15	LOT	30EF64E	更新日期：	2022-01
								4 / 4

批号：30EF64E					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.48	8.9%	<5	<100
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.28	9.6%	<5	<100
(SCC)鳞状细胞癌抗原					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.29	8.7%	7	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.29	6.9%	6	<100
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.601	8.9%	5	<100
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.19	12.1%	18	412
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.19	5.1%	13	121
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.15	9.5%	10	<100
(TPA)组织多肽抗原					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	489	9.8%	<5	<100
(TPS)组织多肽特异抗原					
Siemens Immulite 1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	36.3	8.8%	<5	<100
(β2M)β-2-微球蛋白					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	786	11.5%	13	113
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.017	4.1%	13	111
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.033	9.4%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。