报告类别： **汇总实验室数据报告**

质控品名： **凝血七项非定值质控品**

REF	671-1	6×1mL	EXP	2023-12-25	LOT	4000CAA	更新日期：	2024-03

		批号：4000CAA			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
BE Thrombolyzer X/XRC/XRM Dedicated \ Light scattering	Seconds	28.7	4%	8	196
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	31.2	5.1%	29	404
Werfen ACL series Dedicated \ Light scattering	Seconds	38.6	3.7%	24	599
Sysmex CS 5100 sysmex \ Light scattering	Seconds	28.2	3.4%	15	276
Stago STA series Stago \ Magnetic Beads	Seconds	38.2	12%	<5	<100
(AT-III)抗凝血酶III					
BE Thrombolyzer X/XRC/XRM Dedicated \ Magnetic Beads	%	113	5.5%	7	189
Sysmex CA 1500/6000/7000 Dedicated \ Nephelometry	%	89	7.7%	28	388
Werfen ACL series Dedicated \ Nephelometry	%	97.3	3.3%	24	596
Sysmex CS 5100 sysmex \ Nephelometry	%	92.5	4.9%	15	257
Stago STA series Stago \ Magnetic Beads	%	96	12%	<5	<100
(D-D(DDU)) D-二聚体(DDU)					
Sysmex CS 5100 SekiSui \ Nephelometry	mg/L	1.6	12%	<5	<100
(FDP)纤维蛋白(原)降解产物					
BE Thrombolyzer X/XRC/XRM Dedicated \ Magnetic Beads	mg/L	2.31	10.3%	8	195
Sysmex CA 1500/6000/7000 Dedicated \ Nephelometry	mg/L	5.1	19.3%	28	347
Werfen ACL series Dedicated \ Nephelometry	mg/L	2.84	11.9%	23	185
Sysmex CS 5100 SekiSui \ Nephelometry	mg/L	3.3	12%	<5	<100
(FIB)纤维蛋白原					
BE Thrombolyzer X/XRC/XRM Dedicated \ Light scattering	g/L	3.03	6.3%	8	196
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	g/L	3.17	9.4%	29	400
Werfen ACL series Dedicated \ Light scattering	g/L	3.55	4.6%	24	597
Sysmex CS 5100 Dedicated \ Magnetic Beads	g/L	3.15	5.7%	15	435
Sysmex CS 5100 sysmex \ Light scattering	g/L	2.35	7.7%	15	276
Sysmex CS 2000i/2100i Dedicated \ Magnetic Beads	g/L	3.22	5.5%	10	222
Stago STA series Stago \ Magnetic Beads	g/L	3.62	12%	<5	<100
(PT INR)PT国际标准化比值					
BE Thrombolyzer X/XRC/XRM Dedicated \ Calculate	INR	1.09	4.7%	8	192
Sysmex CS 2000i/2100i Dedicated \ Calculate	INR	1.05	3.5%	<5	<100
Sysmex CS 5100 Dedicated \ Calculate	INR	1.03	2.8%	<5	220
Sysmex CS 5100 sysmex \ Calculate	INR	1.12	5.8%	15	277
Sysmex CA 1500/6000/7000 Dedicated \ Calculate	INR	1.12	4.5%	11	142
Stago STA series Stago \ Calculate	INR	1.05	12%	<5	<100
(PT)血浆凝血酶原时间					
BE Thrombolyzer X/XRC/XRM Dedicated \ Magnetic Beads	Seconds	12.5	4.7%	8	191
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	12.8	3.6%	29	402
Werfen ACL series Dedicated \ Light scattering	Seconds	11.6	2.2%	24	596
Sysmex CS 5100 sysmex \ Light scattering	Seconds	12.7	5.3%	15	275
Stago STA series Stago \ Magnetic Beads	Seconds	13.4	12%	<5	<100
(TT)凝血酶时间					
BE Thrombolyzer X/XRC/XRM Dedicated \ Light scattering	Seconds	12.6	3.8%	8	194
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	20.9	8.2%	29	403
Werfen ACL series Dedicated \ Light scattering	Seconds	14.2	3.3%	24	595
Sysmex CS 5100 sysmex \ Light scattering	Seconds	19.6	6.8%	15	275
Sysmex CS 5100 Dedicated \ Magnetic Beads	Seconds	20.1	3.2%	14	436
Sysmex CS 2000i/2100i Dedicated \ Magnetic Beads	Seconds	19.4	4.7%	10	224

报告类别： **汇总实验室数据报告**

质控品名： **凝血七项非定值质控品**

REF	671-1	6×1mL	EXP	2023-12-25	LOT	4000CAA	更新日期：	2024-03

		批号：4000CAA			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Stago STA series Stago \ Magnetic Beads	Seconds	16.5	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。