报告类别： **汇总实验室数据报告**

质控品名： **凝血七项非定值质控品**

REF	671-1	6×1mL	EXP	2023-08-26	LOT	4000CA6	更新日期：	2024-03

		批号：4000CA6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
OTHER Mindray \ Magnetic Beads	Seconds	33.8	5.8%	8	153
Werfen ACL series SekiSui \ OTHER	Seconds	38	4.9%	8	117
Sysmex CS 5100 Dedicated \ Light scattering	Seconds	30.5	3.1%	75	2755
Werfen ACL series Werfen \ OTHER	Seconds	37.5	5.4%	71	1603
Mindray EXC810 Mindray \ OTHER	Seconds	33.4	5%	<5	<100
BE Thrombolyzer X/XRC/XRM Dedicated \ Light scattering	Seconds	28	6.5%	<5	<100
Sysmex CS 5100 sysmex \ OTHER	Seconds	31.3	2.2%	26	602
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	37.7	2.3%	23	571
Sysmex CS-1300 sysmex \ Light scattering	Seconds	36.1	4.4%	17	325
Beckman AU series Dedicated \ Light scattering	Seconds	38.4	4.5%	16	321
Sekisui Coapresta 3000 SekiSui \ OTHER	Seconds	30.6	2.6%	14	294
Sysmex CS 5100 Siemens Dade Actin FSL(APTT) \ Light scattering	Seconds	32	2.9%	13	339
OTHER OTHER \ OTHER	Seconds	35.6	3.9%	13	273
Sysmex CS-2400 sysmex \ OTHER	Seconds	30.1	2.3%	12	311
OTHER Dedicated \ OTHER	Seconds	37.1	2.6%	12	266
Sysmex CS 5100 Siemens Dade Actin(APTT) \ Light scattering	Seconds	31.5	5.1%	10	266
Sysmex CS 5100 Siemens \ Light scattering	Seconds	33.7	12%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	33.1	12%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	33.4	12%	<5	<100
(AT-III)抗凝血酶III					
OTHER Mindray \ OTHER	%	87	4.3%	9	176
Werfen ACL series SekiSui \ OTHER	%	96.7	4.9%	8	<100
Sysmex CS-2400 sysmex \ OTHER	%	94.5	2.4%	8	<100
Werfen ACL series Werfen \ OTHER	%	99.4	5.9%	70	1525
BE Thrombolyzer X/XRC/XRM Dedicated \ Magnetic Beads	%	115	6.4%	<5	<100
Sekisui Coapresta 3000 SekiSui \ Nephelometry	%	106	1.8%	<5	<100
Beckman AU series Dedicated \ Nephelometry	%	89.4	6%	25	517
Sysmex CS 5100 Dedicated \ Nephelometry	%	94.6	4.2%	20	466
Sekisui Coapresta 3000 SekiSui \ OTHER	%	89.1	8.6%	13	318
Sysmex CS 5100 OTHER \ Nephelometry	%	73.4	11%	11	200
Sysmex CS 5100 Siemens \ Nephelometry	%	96.9	2.9%	10	270
Werfen ACL series Dedicated \ Nephelometry	%	101	12%	<5	<100
Stago STA series Stago \ Magnetic Beads	%	94	12%	<5	<100
(D-D(DDU)) D-二聚体(DDU)					
OTHER Mindray \ OTHER	mg/L	1.58	11.5%	8	148
Werfen ACL series SekiSui \ OTHER	mg/L	1.38	9.3%	8	120
Sekisui Coapresta 3000 SekiSui \ OTHER	mg/L	2.32	7.6%	7	140
Werfen ACL series Werfen \ OTHER	mg/L	0.488	9.3%	65	1588
Mindray EXC810 Mindray \ OTHER	mg/L	0.778	16.1%	<5	<100
Sysmex CS 5100 Dedicated \ Nephelometry	mg/L	1.45	11.3%	29	907
Sysmex CS 5100 sysmex \ OTHER	mg/L	0.979	11.9%	26	558
OTHER OTHER \ OTHER	mg/L	0.398	8.4%	25	541
Beckman AU series Dedicated \ Nephelometry	mg/L	0.698	23.8%	18	468
Sysmex CS-2400 Siemens \ Nephelometry	mg/L	1.18	9.5%	12	307
Succeeder SF series Succeeder \ Latex agglutination (LA)	mg/L	0.914	1.7%	12	295
OTHER Beckman Coulter \ Latex agglutination (LA)	mg/L	0.427	6.8%	11	236

报告类别： **汇总实验室数据报告**

质控品名： **凝血七项非定值质控品**

REF	671-1	6×1mL	EXP	2023-08-26	LOT	4000CA6	更新日期：	2024-03

		批号：4000CA6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Sysmex CS 5100 Siemens \ Nephelometry	mg/L	1.55	6%	10	269
Sysmex CS 5100 SekiSui \ Nephelometry	mg/L	1.23	12%	<5	<100
(D-D(FEU)) D-二聚体(FEU)					
Sysmex CS 5100 Dedicated \ Nephelometry	mg/L	1.62	8.7%	26	951
Sysmex CS-1300 sysmex \ Nephelometry	mg/L	1.44	11.8%	18	414
Sysmex CS 5100 Siemens Dade Actin FSL(APTT) \ Nephelometry	mg/L	1.62	7.2%	15	340
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	1.91	12%	<5	<100
Stago STA series Dedicated \ Latex agglutination (LA)	mg/L	1.89	12%	<5	<100
(FDP)纤维蛋白(原)降解产物					
OTHER Mindray \ OTHER	mg/L	6.46	8.8%	9	174
Sekisui Coapresta 3000 SekiSui \ OTHER	mg/L	4.45	12.1%	8	133
Werfen ACL series SekiSui \ OTHER	mg/L	2.13	13.6%	8	107
Werfen ACL series Werfen \ OTHER	mg/L	2.93	16.2%	33	819
BE Thrombolyzer X/XRC/XRM Dedicated \ Magnetic Beads	mg/L	2.36	12.8%	<5	<100
Sysmex CS 5100 Dedicated \ Nephelometry	mg/L	6.15	19.4%	26	493
Sysmex CS-1300 sysmex \ Nephelometry	mg/L	6.31	15.4%	16	345
Sysmex CS 5100 sysmex \ Nephelometry	mg/L	5.95	2.6%	10	270
Stago STA series Dedicated \ Nephelometry	mg/L	4.82	11.2%	<5	<100
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	4.98	11.4%	<5	<100
Sysmex CS 5100 SekiSui \ Nephelometry	mg/L	5.2	12%	<5	<100
Werfen ACL series SekiSui \ Nephelometry	mg/L	4.35	11%	<5	<100
(FIB)纤维蛋白原					
OTHER Mindray \ Magnetic Beads	g/L	3.09	9.7%	9	165
Werfen ACL series Werfen \ OTHER	g/L	3.55	8.7%	78	1739
Sysmex CS 5100 Dedicated \ Light scattering	g/L	3.42	7.1%	75	2726
Mindray EXC810 Mindray \ OTHER	g/L	3.2	12.5%	<5	<100
BE Thrombolyzer X/XRC/XRM Dedicated \ Light scattering	g/L	3.48	6.3%	<5	<100
Sysmex CS 5100 sysmex \ OTHER	g/L	3.2	6.6%	26	592
Beckman AU series Dedicated \ Light scattering	g/L	3.5	8.8%	24	612
Succeeder SF series Succeeder \ Magnetic Beads	g/L	3.32	4.8%	23	576
Sysmex CS-1300 sysmex \ Light scattering	g/L	2.96	17.9%	18	419
OTHER OTHER \ OTHER	g/L	3.69	6.5%	18	246
Sysmex CS 5100 sysmex \ Light scattering	g/L	3.46	7.5%	14	339
Sekisui Coapresta 3000 SekiSui \ OTHER	g/L	3.33	9.8%	14	316
Sysmex CS-2400 sysmex \ OTHER	g/L	3.35	5.9%	12	313
OTHER Dedicated \ OTHER	g/L	3.66	5.2%	12	261
OTHER Beckman Coulter \ Magnetic Beads	g/L	3.8	6.2%	12	236
Sysmex CS 5100 Siemens \ Light scattering	g/L	3.14	6.6%	11	321
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	g/L	3.21	12%	<5	<100
Stago STA series Stago \ Magnetic Beads	g/L	3.76	12%	<5	<100
Werfen ACL series Werfen \ Light scattering	g/L	3.47	12%	<5	<100
(PT INR)PT国际标准化比值					
Sysmex CS 5100 Siemens \ Calculate	INR	1.08	13.4%	9	244
OTHER Mindray \ Calculate	INR	1.06	3.8%	9	157
Sysmex CS 5100 Dedicated \ Calculate	INR	1.08	5%	73	2711
Werfen ACL series Werfen \ Calculate	INR	1.04	3.2%	37	785
Mindray EXC810 Mindray \ Calculate	INR	1.12	4.1%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **凝血七项非定值质控品**

REF	671-1	6×1mL	EXP	2023-08-26	LOT	4000CA6	更新日期：	2024-03

		批号：4000CA6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
BE Thrombolyzer X/XRC/XRM Dedicated \ Calculate	INR	1.13	5.5%	<5	<100
Sysmex CS 5100 sysmex \ OTHER	INR	1.1	4.6%	26	599
Beckman AU series Dedicated \ Calculate	INR	1.1	5.3%	22	570
Sysmex CS-1300 sysmex \ Calculate	INR	1.04	11.6%	17	317
Sysmex CS 5100 sysmex \ Calculate	INR	1.06	3.8%	14	347
Sekisui Coapresta 3000 SekiSui \ Calculate	INR	1.11	2.7%	13	306
Sysmex CS-2400 sysmex \ Calculate	INR	1.07	2.4%	12	318
Succeeder SF series Succeeder \ Calculate	INR	1.12	1.5%	12	295
OTHER OTHER \ OTHER	INR	1.06	2.2%	12	263
OTHER OTHER \ Calculate	INR	1.07	2.4%	11	235
Werfen ACL series SekiSui \ Calculate	INR	1.12	3%	10	114
Stago STA series Stago \ Calculate	INR	1.05	12%	<5	<100
(PT)血浆凝血酶原时间					
OTHER Mindray \ Magnetic Beads	Seconds	12.8	3.9%	9	158
Sysmex CS 5100 Dedicated \ Light scattering	Seconds	12.6	4.6%	74	2725
Werfen ACL series Werfen \ OTHER	Seconds	11.5	3.8%	69	1601
Mindray EXC810 Mindray \ OTHER	Seconds	13.5	4%	<5	<100
BE Thrombolyzer X/XRC/XRM Dedicated \ Magnetic Beads	Seconds	12.8	13.6%	<5	<100
Sysmex CS 5100 sysmex \ OTHER	Seconds	12.8	4.8%	26	597
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	14.1	2.5%	22	554
Beckman AU series Dedicated \ Light scattering	Seconds	11.7	3.3%	22	489
Sysmex CS-1300 sysmex \ Light scattering	Seconds	12.1	9.8%	18	355
Sekisui Coapresta 3000 SekiSui \ OTHER	Seconds	13.7	2.4%	16	307
OTHER OTHER \ OTHER	Seconds	12.1	3.1%	15	258
Sysmex CS 5100 sysmex \ Light scattering	Seconds	12.4	3.6%	13	340
Sysmex CS-2400 sysmex \ OTHER	Seconds	12.3	2.3%	12	317
OTHER Dedicated \ OTHER	Seconds	11.7	2.4%	12	261
Sysmex CS 5100 Siemens \ Light scattering	Seconds	12.1	4.5%	11	288
Werfen ACL series SekiSui \ OTHER	Seconds	12.5	2.9%	10	114
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	12.8	12%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	11.4	12%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	13.4	12%	<5	<100
(TT)凝血酶时间					
OTHER Mindray \ Magnetic Beads	Seconds	16.7	8.1%	9	174
Werfen ACL series Werfen \ OTHER	Seconds	14.2	5.7%	78	1715
Sysmex CS 5100 Dedicated \ Light scattering	Seconds	20.4	8.4%	75	2743
Mindray EXC810 Mindray \ OTHER	Seconds	15.5	4.7%	<5	<100
BE Thrombolyzer X/XRC/XRM Dedicated \ Light scattering	Seconds	12.6	3.1%	<5	<100
Sysmex CS 5100 sysmex \ OTHER	Seconds	20	5.5%	26	599
Succeeder SF series Succeeder \ Magnetic Beads	Seconds	17.4	1.4%	23	566
Beckman AU series Dedicated \ Light scattering	Seconds	13	2.2%	21	463
Sysmex CS-1300 sysmex \ Light scattering	Seconds	17.9	19.9%	18	427
OTHER OTHER \ OTHER	Seconds	15.4	6.2%	17	244
Sysmex CS 5100 sysmex \ Light scattering	Seconds	19	4.5%	13	338
Sekisui Coapresta 3000 SekiSui \ OTHER	Seconds	18.5	5.3%	13	296
Sysmex CS-2400 sysmex \ OTHER	Seconds	20.5	5.8%	12	304
OTHER Dedicated \ OTHER	Seconds	14.3	4.3%	11	263

报告类别： **汇总实验室数据报告**

质控品名： **凝血七项非定值质控品**

REF	671-1	6×1mL	EXP	2023-08-26	LOT	4000CA6	更新日期：	2024-03

		批号：4000CA6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
OTHER Beckman Coulter \ Magnetic Beads	Seconds	14.4	4.6%	11	234
Sysmex CS 5100 Siemens \ Light scattering	Seconds	18.8	7.3%	10	270
Stago STA series Stago \ Magnetic Beads	Seconds	16.5	12%	<5	<100
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	18.5	12%	<5	<100
Werfen ACL series SUNBIO \ Light scattering	Seconds	16.2	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。