

REF	671-1	6×1mL	EXP	2024-04-22	LOT	4000CAE	更新日期：	2024-04
1 / 2								

批号：4000CAE					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
BE Thrombolyzer X/XRC/XRM Dedicated \ Light scattering	Seconds	27.9	2.6%	8	<100
Werfen ACL series Werfen \ OTHER	Seconds	37.6	2.7%	8	228
Sysmex CS 5100 Siemens \ Light scattering	Seconds	37.9	13%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	35.7	12.4%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	34.2	11.7%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	34.3	11.1%	<5	<100
(AT-III)抗凝血酶III					
BE Thrombolyzer X/XRC/XRM Dedicated \ Magnetic Beads	%	118	3.8%	8	<100
Werfen ACL series Werfen \ OTHER	%	89	7.3%	8	202
BE Thrombolyzer X/XRC/XRM OTHER \ OTHER	%	86.5	16.2%	16	142
Stago STA series Stago \ Magnetic Beads	%	93.6	10.4%	<5	<100
Werfen ACL series Dedicated \ Nephelometry	%	98.3	10.1%	<5	<100
(D-D(DDU)) D-二聚体(DDU)					
BE Thrombolyzer X/XRC/XRM OTHER \ OTHER	mg/L	0.496	13.8%	85	1355
Werfen ACL series Werfen \ Nephelometry	mg/L	1.43	3.5%	8	225
Sysmex CS 5100 Siemens \ Immunoturbidimetry (ITA)	mg/L	1.61	11.2%	<5	<100
Sysmex CS 5100 SekiSui \ Nephelometry	mg/L	1.12	13.2%	<5	<100
(D-D(FEU)) D-二聚体(FEU)					
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	1.72	11%	<5	<100
Stago STA series Dedicated \ Latex agglutination (LA)	mg/L	1.69	11.2%	<5	<100
(FDP)纤维蛋白(原)降解产物					
BE Thrombolyzer X/XRC/XRM Dedicated \ Magnetic Beads	mg/L	2.58	18.3%	8	<100
Werfen ACL series Werfen \ Nephelometry	mg/L	3.18	10.4%	8	217
BE Thrombolyzer X/XRC/XRM OTHER \ OTHER	mg/L	3.97	13.8%	10	<100
Stago STA series Dedicated \ Nephelometry	mg/L	4.46	12.1%	<5	<100
Werfen ACL series SekiSui \ Nephelometry	mg/L	4.23	11.3%	<5	<100
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	4.67	10.3%	<5	<100
Sysmex CS 5100 SekiSui \ Nephelometry	mg/L	5.03	12.4%	<5	<100
(FIB)纤维蛋白原					
BE Thrombolyzer X/XRC/XRM Dedicated \ Light scattering	g/L	3.24	11.1%	8	<100
Werfen ACL series Werfen \ OTHER	g/L	3.88	6.1%	8	214
Sysmex CS 5100 Siemens \ Light scattering	g/L	3.2	8.6%	<5	<100
BE Thrombolyzer X/XRC/XRM OTHER \ OTHER	g/L	2.83	6.1%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	g/L	3.41	9.7%	<5	<100
Werfen ACL series Werfen \ Light scattering	g/L	3.53	10.2%	<5	<100
Stago STA series Stago \ Magnetic Beads	g/L	3.67	10.1%	<5	<100
(PT INR)PT国际标准化比值					
BE Thrombolyzer X/XRC/XRM Dedicated \ Calculate	INR	1.05	5.1%	8	<100
Werfen ACL series Werfen \ Calculate	INR	1.1	2.7%	8	227
Sysmex CS 5100 Siemens \ Calculate	INR	1.09	3.1%	<5	<100
Stago STA series Stago \ Calculate	INR	1.05	10.5%	<5	<100
BE Thrombolyzer X/XRC/XRM Behnk Elektronik (BE) \ Calculate	INR	1.1	0%	<5	<100
(PT)血浆凝血酶原时间					
BE Thrombolyzer X/XRC/XRM Dedicated \ Magnetic Beads	Seconds	12.1	4.9%	8	<100
Werfen ACL series Werfen \ OTHER	Seconds	11.7	2.7%	8	226
Sysmex CS 5100 Siemens \ Light scattering	Seconds	12.2	7.6%	<5	<100

REF	671-1	6×1mL	EXP	2024-04-22	LOT	4000CAE	更新日期：	2024-04
2 / 2								

		批号：4000CAE			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Stago STA series Stago \ Magnetic Beads	Seconds	12.9	12.5%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	12.2	11.2%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	12.6	10.3%	<5	<100
(TT)凝血酶时间					
BE Thrombolyzer X/XRC/XRM Dedicated \ Light scattering	Seconds	12.2	4.9%	8	<100
Werfen ACL series Werfen \ OTHER	Seconds	13.8	4.3%	8	226
Sysmex CS 5100 Siemens \ Light scattering	Seconds	21.3	3.4%	<5	<100
Werfen ACL series SUNBIO \ Light scattering	Seconds	16.8	11.3%	<5	<100
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	18.5	10.3%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	17.2	12.8%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。