

质控品名： 肿瘤标志物液体非定值质控品

REF	510-2	6×3mL	EXP	2026-01-08	LOT	30A8915	更新日期：	2024-04
1 / 2								

		批号：30A8915			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	69.1	2.3%	8	204
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	69.8	10.9%	<5	<100
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	73.5	11.2%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	80.7	11.3%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	72.8	12%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	75.4	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	87.9	10.2%	<5	<100
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	78.6	0%	<5	<100
(CA 125)糖类抗原 125					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	260	3.6%	8	203
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	263	8.4%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	243	9.5%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	174	10%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	112	9.8%	<5	<100
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	235	0%	<5	<100
(CA 15-3)糖类抗原 15-3					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	67.9	5.2%	8	205
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	88.4	10.2%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	109	10.1%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	108	10.2%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	96.9	9.3%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	59.8	10%	<5	<100
(CA 19-9)糖类抗原 19-9					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	362	5.5%	8	202
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	81.8	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	121	9.1%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	94.2	8.5%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	268	10.8%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	145	11%	<5	<100
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	356	0%	<5	<100
(CA 50)糖类抗原 50					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	72.7	10%	<5	<100
(CA 72-4)糖类抗原 72-4					
Mindray BC-6000 series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	6.92	3.7%	<5	106
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	13	10%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	28.9	10.7%	<5	<100
(CA242)糖类抗原 242					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	23.7	10%	<5	<100
(CEA)癌胚抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	42.6	2.3%	8	204
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	37.6	8.5%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	35.8	10.6%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	26.2	10%	<5	<100

质控品名： 肿瘤标志物液体非定值质控品

REF	510-2	6×3mL	EXP	2026-01-08	LOT	30A8915	更新日期：	2024-04
2 / 2								

批号：30A8915					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	38.6	10.9%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	21.7	10.1%	<5	<100
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	37.7	0%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.3	4.6%	7	202
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.4	10%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.2	9.1%	<5	<100
(Ferritin)铁蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	390	3.4%	7	201
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	266	10.2%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	ng/mL	313	10%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	349	8.3%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.47	3.9%	8	205
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.35	10%	<5	<100
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.26	11.6%	<5	<100
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.65	0%	<5	<100
(PSA)前列腺特异性抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.5	4.4%	8	205
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	11.8	9.3%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.9	10.1%	<5	<100
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	21.3	1%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.9	10%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	15.4	10.4%	<5	<100
(SCC)鳞状细胞癌抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.2	4.1%	7	201
OTHER Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.5	14%	30	602
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.73	10%	<5	<100
Abbott AxSYM Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.9	9.3%	<5	<100
(TPA)组织多肽抗原					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	6047	11.3%	<5	<100
(TPS)组织多肽特异抗原					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	2218	10.5%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.156	9.6%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.163	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。