报告类别： **汇总实验室数据报告**

质控品名： **尿液定量生化液体非定值质控品**

REF	586-1	6×8mL	EXP	2023-04-02	LOT	37E5923	更新日期：	2024-03

		批号：37E5923			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(AMY)淀粉酶					
Roche Hitachi/Hitachi Systems BSBE \ OTHER	U/L	44.5	10.8%	9	226
Beckman AU series Beckman Coulter \ OTHER	U/L	44.9	3.4%	10	483
Roche cobas 4000/6000/8000 Roche \ PNP-G7	U/L	38.5	10%	<5	<100
(Ca)钙					
Roche Hitachi/Hitachi Systems BSBE \ Arsenazo III	mmol/L	1.79	3.4%	8	202
Beckman AU series Reebio \ Arsenazo III	mmol/L	2.02	2.1%	10	528
Roche cobas 4000/6000/8000 Roche \ BAPTA	mmol/L	1.91	10%	<5	<100
(Cl)氯					
Roche Hitachi/Hitachi Systems BSBE \ ISE (indirect)	mmol/L	73.2	10.7%	7	117
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	73.2	10%	<5	<100
(Cortisol)皮质醇					
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	33.3	10%	<5	<100
(Cr)肌酐					
Roche Hitachi/Hitachi Systems Beckman Coulter \ Enzymatic	μmol/L	6557	7.5%	9	229
Beckman AU series Reebio \ OTHER	μmol/L	6190	4.7%	10	523
Roche cobas 4000/6000/8000 Roche \ Enzymatic	μmol/L	6401	10%	<5	<100
(GLU)葡萄糖					
Roche Hitachi/Hitachi Systems BSBE \ OTHER	mmol/L	1.3	8.2%	9	226
Beckman AU series Reebio \ OTHER	mmol/L	1.48	3.8%	10	528
Roche cobas 4000/6000/8000 Roche \ Hexokinase	mmol/L	1.33	10%	<5	<100
(K)钾					
Roche Hitachi/Hitachi Systems BSBE \ ISE (indirect)	mmol/L	23	19.8%	6	144
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	29.9	6.7%	10	475
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	29.2	10%	<5	<100
(MA)微量白蛋白					
Roche Hitachi/Hitachi Systems BSBE \ Immunoturbidimetry (ITA)	mg/L	37.8	11.1%	9	234
TOSHIBA TBA OTHER \ OTHER	mg/L	30.4	18.1%	6	<100
Roche cobas 4000/6000/8000 Roche \ Bromocresol green	mg/L	29.1	10%	<5	<100
(Na)钠					
Roche Hitachi/Hitachi Systems BSBE \ ISE (indirect)	mmol/L	84.2	11.8%	7	135
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	81	4%	11	514
Roche cobas c 501/502/701/702 Roche \ ISE (direct)	mmol/L	79.5	10.1%	<5	<100
(P)磷					
Roche Hitachi/Hitachi Systems BSBE \ OTHER	mmol/L	9.73	6.7%	8	205
Beckman AU series Beckman Coulter \ OTHER	mmol/L	8.41	2.7%	10	497
Roche cobas 4000/6000/8000 Roche \ Phosphomolybdate UV-rate	mmol/L	8.6	10%	<5	<100
(PRO/ TP)(总) 蛋白					
Roche Hitachi/Hitachi Systems BSBE \ OTHER	g/L	0.225	11.4%	8	215
Beckman AU series Dedicated \ Biuret-endpoint	g/L	0.168	12.2%	7	139
TOSHIBA TBA OTHER \ pyrogallol red molybdate complex	g/L	0.228	10.8%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Biuret-endpoint	g/L	0.12	10.6%	<5	<100
(UA)尿酸					
Roche Hitachi/Hitachi Systems BSBE \ OTHER	μmol/L	517	10.4%	9	227
Roche cobas 4000/6000/8000 Roche \ Uricase colorimetry	μmol/L	467	10%	<5	<100
(Urea)尿素					
Roche Hitachi/Hitachi Systems BSBE \ OTHER	mmol/L	150	8.3%	9	229

报告类别： **汇总实验室数据报告**

质控品名： **尿液定量生化液体非定值质控品**

REF	586-1	6×8mL	EXP	2023-04-02	LOT	37E5923	更新日期：	2024-03

		批号：37E5923			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Reebio \ OTHER	mmol/L	143	4.4%	10	542
Roche cobas 4000/6000/8000 Roche \ Urease colorimetry	mmol/L	141	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。