

REF 580-1	6x5mL	EXP 2022-11-25	LOT 3753107	更新日期： 2022-05
				1 / 11

		批号：3753107			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(a-HBDH)α-羟丁酸脱氢酶					
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	119	10.9%	9	135
KHB ZY series DIRUI \ DGKC	U/L	107	3.9%	8	<100
Beckman AU series DiaSys \ Enzymatic Cycling Assays Method	U/L	124	1.5%	7	<100
Abbott ARCHITECT c series MedicalSystem \ OTHER	U/L	135	3.4%	6	<100
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic Cycling Assays Method	U/L	112	3.7%	<5	119
DiruiH-800 XC Bio \ Enzymatic Cycling Assays Method	U/L	117	8%	<5	<100
TOSHIBA TBA chongqiong zybio \ Enzymatic Cycling Assays Method	U/L	107	2.1%	<5	<100
Mindray BS series WEGO \ Enzymatic Cycling Assays Method	U/L	90	1.9%	<5	<100
KHB ZY series KHB \ DGKC	U/L	115	4.8%	<5	<100
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	U/L	114	7%	18	573
Beckman AU series Fosun \ OTHER	U/L	121	5.4%	15	357
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	U/L	102	2.2%	<5	<100
(Alb)白蛋白					
Mindray BS series Mindray \ Bromocresol green	μmol/L	736	6.8%	80	897
Mindray BS series Mindray \ Bromocresol green	g/L	48.6	6.8%	80	897
KHB ZY series DIRUI \ Bromocresol green	μmol/L	656	3.4%	8	<100
KHB ZY series DIRUI \ Bromocresol green	g/L	43.3	3.4%	8	<100
DiruiCS series MedicalSystem \ Bromocresol green	μmol/L	726	5.1%	8	<100
DiruiCS series MedicalSystem \ Bromocresol green	g/L	47.9	5.1%	8	<100
Beckman AU series Fosun \ Bromocresol green	μmol/L	742	3.4%	73	1833
Beckman AU series Fosun \ Bromocresol green	g/L	49	3.4%	73	1833
Beckman AU series OTHER \ Bromocresol green	μmol/L	704	0.3%	7	<100
Beckman AU series OTHER \ Bromocresol green	g/L	46.4	0.3%	7	<100
OTHER KeFang(Guangzhou) \ Bromocresol purple	μmol/L	756	0.3%	7	<100
OTHER KeFang(Guangzhou) \ Bromocresol purple	g/L	49.9	0.3%	7	<100
Beckman AU series KHB \ Bromocresol green	μmol/L	718	5.6%	7	179
Beckman AU series KHB \ Bromocresol green	g/L	47.4	5.6%	7	179
Beckman AU series Reebio \ Bromocresol green	μmol/L	737	6.1%	68	1989
Beckman AU series Reebio \ Bromocresol green	g/L	48.6	6.1%	68	1989
Beckman AU series MedicalSystem \ Bromocresol green	μmol/L	735	3.7%	66	799
Beckman AU series MedicalSystem \ Bromocresol green	g/L	48.5	3.7%	66	799
Abbott ARCHITECT c series MedicalSystem \ Bromocresol green	μmol/L	748	9.4%	6	<100
(ALP)碱性磷酸酶					
SDBK BIOBASE \ Rate -AMP buffer	U/L	81.3	22.1%	9	<100
Beckman AU series Bio-assay \ Rate -AMP buffer	U/L	67.7	4.4%	8	<100
Beckman AU series OTHER \ Rate -AMP buffer	U/L	56.6	1.3%	8	<100
KHB ZY series DIRUI \ Rate -AMP buffer	U/L	65.1	2.1%	8	<100
DiruiCS series DiaSys \ Rate -AMP buffer	U/L	84.1	5.9%	8	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Rate -AMP buffer	U/L	90	2.2%	8	130
SDBK BIOBASE \ dry chemistry	U/L	71.1	10.1%	7	<100
OTHER KeFang(Guangzhou) \ Rate -AMP buffer	U/L	75.8	0.1%	7	<100
Beckman AU series KHB \ Rate -AMP buffer	U/L	81.9	16%	7	197
Beckman AU series Fosun \ Rate -AMP buffer	U/L	87.7	11.5%	64	1625
Abbott ARCHITECT c series MedicalSystem \ Rate -AMP buffer	U/L	72.3	3.8%	6	<100
Beckman AU series Reebio \ dry chemistry	U/L	68.7	12.2%	5	<100
Beckman AU series WanTai \ Rate -AMP buffer	U/L	69.6	6.8%	5	<100

REF	580-1	6x5mL	EXP	2022-11-25	LOT	3753107	更新日期：	2022-05
								2 / 11

		批号：3753107			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Sinnova DS series	U/L	77.2	14.6%	5	<100
Sinnova \ Rate -AMP buffer					
Beckman AU series	U/L	97.1	14.7%	45	678
Beckman Coulter \ Rate -AMP buffer					
Beckman AU series	U/L	72.9	12.4%	44	1060
Reebio \ Rate -AMP buffer					
SDBK	U/L	73.7	1.7%	<5	<100
BIOBASE \ OTHER					
Beckman AU series	U/L	85.7	1.4%	<5	<100
DIRUI \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	109	2.4%	<5	119
MAKER \ Rate -AMP buffer					
(ALT/ SGPT)丙氨酸氨基转移酶					
SDBK	U/L	25.9	9.2%	9	<100
BIOBASE \ OTHER					
Mindray BS series	U/L	29.5	4.7%	9	<100
chongqiong zybio \ UV without P5P					
Beckman AU series	U/L	24.8	3.1%	8	<100
Bio-assay \ UV without P5P					
Beckman AU series	U/L	28.3	0.6%	8	<100
OTHER \ UV without P5P					
KHB ZY series	U/L	24	6.7%	8	<100
DIRUI \ UV without P5P					
DiruiCS series	U/L	28.7	7.2%	8	<100
DiaSys \ UV with P5P					
Roche Hitachi/Hitachi Systems	U/L	26.2	4.1%	8	139
Wako \ UV with P5P					
DiruiCS series	U/L	23.6	9.8%	7	<100
DIRUI \ OTHER					
Beckman AU series	U/L	25.8	2.9%	7	<100
DiaSys \ UV with P5P					
SDBK	U/L	26.8	8.6%	7	<100
BIOBASE \ dry chemistry					
Beckman AU series	U/L	27.7	11.5%	7	197
KHB \ UV without P5P					
Abbott ARCHITECT c series	U/L	24.4	4%	6	<100
MedicalSystem \ UV without P5P					
OTHER	U/L	27.6	0.4%	6	<100
KeFang(Guangzhou) \ UV without P5P					
Roche Hitachi/Hitachi Systems	U/L	25.1	6.2%	6	132
Wako \ UV without P5P					
Beckman AU series	U/L	28.6	14.6%	59	1502
Reebio \ UV with P5P					
Beckman AU series	U/L	27.7	12.7%	5	<100
Reebio \ dry chemistry					
Beckman AU series	U/L	38.8	8.9%	5	<100
XJJC \ dry chemistry					
Beckman AU series	U/L	26.3	3.5%	5	<100
WanTai \ UV without P5P					
Beckman AU series	U/L	29.9	6.6%	40	706
Beckman Coulter \ UV without P5P					
(AMY)淀粉酶					
Mindray BS series	U/L	57.6	1.1%	9	256
Mindray \ PNP-G5					
Beckman AU series	U/L	49.8	1.2%	7	<100
MedicalSystem \ PNP-G7					
Beckman AU series	U/L	56.6	2.3%	7	192
KHB \ PNP-G5					
Beckman AU series	U/L	54.2	4.6%	64	1660
Fosun \ PNP-G7					
Abbott ARCHITECT c series	U/L	53.7	1.8%	6	<100
MedicalSystem \ CNP_G3					
Roche Hitachi/Hitachi Systems	U/L	55.2	6.7%	55	1211
MAKER \ OTHER					
Beckman AU series	U/L	51.1	5.8%	45	1105
Reebio \ CNP_G3					
Mindray BS series	U/L	84.7	6.2%	<5	<100
Mindray \ CNP_G3					
TOSHIBA TBA	U/L	89.7	2.1%	<5	<100
chongqiong zybio \ PNP-G7					
Beckman AU series	U/L	66.3	9.2%	19	389
Beckman Coulter \ PNP-G7					
DiruiH-800	U/L	54.8	6.5%	14	383
XC Bio \ CNP_G3					
Roche Hitachi/Hitachi Systems	U/L	74.3	1.4%	12	192
MedicalSystem \ PNP-G7					
Beckman AU series	U/L	58.1	1.2%	<5	<100
MEDCONN \ CNP_G3					
Roche cobas 4000/6000/8000	U/L	60.3	12%	<5	<100
Roche \ PNP-G7					
Mindray BS series	U/L	68.4	2.4%	<5	<100
WEGO \ PNP-G7					
Roche cobas 4000/6000/8000	U/L	64	2.1%	<5	<100
MAKER \ CNP_G3					
Roche cobas 4000/6000/8000	U/L	67.3	1%	<5	<100
Roche \ Iodine colorimetry					
J&J Vitros 250/350	U/L	50.2	7.1%	<5	<100
Ortho \ dry chemistry					
(Apo-A I)载脂蛋白A I					
Mindray BS series	g/L	1.2	0%	<5	<100
Mindray \ immunoturbidimetric					

REF 580-1	6x5mL	EXP 2022-11-25	LOT 3753107	更新日期： 2022-05	3 / 11
------------------	-------	-----------------------	--------------------	---------------	--------

		批号：3753107			
项目\仪器\试剂方法	单位	均值\ CV		Lab \ 点数	
(Apo-B)载脂蛋白B					
Mindray BS series					
Mindray \ immunoturbidimetric	g/L	0.78	0%	<5	<100
(AST/ SGOT)天门冬氨酸氨基转移酶					
Mindray BS series	U/L	29.3	3.5%	9	<100
Mindray \ OTHER					
Mindray BS series	U/L	27.3	4%	9	<100
chongqiong zybio \ UV without P5P					
Beckman AU series	U/L	26.9	3.8%	8	<100
Bio-assay \ UV without P5P					
Beckman AU series	U/L	27.6	2.1%	8	<100
OTHER \ UV without P5P					
KHB ZY series	U/L	23.5	8.5%	8	<100
DIRUI \ UV without P5P					
DiruiCS series	U/L	26.4	5.8%	8	<100
DiaSys \ UV with P5P					
Beckman AU series	U/L	26.8	3.5%	7	<100
DiaSys \ UV with P5P					
SDBK	U/L	28.8	12.1%	7	<100
BIOBASE \ dry chemistry					
OTHER	U/L	24.5	0.4%	7	<100
KeFang(Guangzhou) \ UV without P5P					
Beckman AU series	U/L	29.5	4.3%	7	184
KHB \ UV without P5P					
Abbott ARCHITECT c series	U/L	27.6	3.6%	6	<100
MedicalSystem \ UV without P5P					
DiruiCS series	U/L	27.1	3.6%	6	<100
DIRUI \ UV with P5P					
Roche Hitachi/Hitachi Systems	U/L	31.3	7.4%	6	132
Wako \ UV without P5P					
Beckman AU series	U/L	29	10%	6	113
Reebio \ dry chemistry					
DiruiCS series	U/L	27.4	5.2%	5	<100
DIRUI \ OTHER					
Beckman AU series	U/L	35.8	5%	5	<100
XJKC \ dry chemistry					
Beckman AU series	U/L	29.7	3.4%	5	<100
WanTai \ UV without P5P					
Beckman AU series	U/L	30.7	6%	40	583
Beckman Coulter \ UV without P5P					
Beckman AU series	U/L	24.8	3.8%	40	1035
Reebio \ UV with P5P					
(BUN)尿素氮					
Mindray BS series	mmol/L	6.18	0%	<5	<100
Mindray \ Dehydrogenase					
(Ca)钙					
Beckman AU series	mmol/L	1.44	3.7%	65	1635
Fosun \ Arsenazo III					
Abbott ARCHITECT c series	mmol/L	1.27	1.4%	6	<100
MedicalSystem \ Arsenazo III					
OTHER	mmol/L	1.45	1.7%	6	160
OTHER \ ISE (direct)					
Roche Hitachi/Hitachi Systems	mmol/L	1.28	1.6%	<5	118
Wako \ MTB colorimetry					
Beckman AU series	mmol/L	1.38	4.9%	38	941
Reebio \ Arsenazo III					
Beckman AU series	mmol/L	1.51	2.4%	20	583
Beckman Coulter \ Arsenazo III					
Abbott ARCHITECT c series	mmol/L	1.16	7.9%	<5	<100
Dedicated \ Arsenazo III					
Mindray BS series	mmol/L	1.55	1.2%	<5	<100
WEGO \ Arsenazo III					
Beckman AU series	mmol/L	1.41	1.4%	<5	<100
DiaSys \ Arsenazo III					
Roche cobas 4000/6000/8000	mmol/L	1.45	11.9%	<5	<100
Roche \ BAPTA					
Roche cobas 4000/6000/8000	mmol/L	1.42	0.8%	<5	<100
Neusoft \ Arsenazo III					
J&J Vitros 250/350	mmol/L	1.43	1.4%	<5	<100
Ortho \ dry chemistry					
Roche cobas 4000/6000/8000	mmol/L	1.44	1%	<5	<100
Roche \ ISE (indirect)					
(CHE)胆碱酯酶					
Beckman AU series	U/L	7673	4.3%	7	<100
MedicalSystem \ Butyrylthiocholine Method					
Abbott ARCHITECT c series	U/L	7421	3.8%	6	<100
MedicalSystem \ Butyrylthiocholine Method					
KHB ZY series	U/L	7369	4.1%	<5	<100
KHB \ Butyrylthiocholine Method					
TOSHIBA TBA	U/L	7826	5.7%	<5	<100
chongqiong zybio \ DTNB					
Beckman AU series	U/L	7822	8.5%	16	414
Fosun \ Butyrylthiocholine Method					
Roche Hitachi/Hitachi Systems	U/L	6803	5.8%	11	260
MAKER \ OTHER					
Roche cobas 4000/6000/8000	U/L	7946	0.5%	<5	<100
Roche \ Butyrylthiocholine Method					
J&J Vitros 250/350	U/L	7214	1.6%	<5	<100
Ortho \ OTHER					

REF	580-1	6x5mL	EXP	2022-11-25	LOT	3753107	更新日期：	2022-05
								4 / 11

		批号：3753107			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(CHOL/ TC)总胆固醇					
Mindray BS series chongqiong zybio \ Cholesterol oxidase	mmol/L	4.95	1.7%	9	<100
KHB ZY series DIRUI \ Cholesterol oxidase	mmol/L	5.34	1.6%	8	<100
Mindray BS series Mindray \ Cholesterol oxidase	mmol/L	5.12	4.2%	78	685
Beckman AU series Reebio \ Cholesterol oxidase	mmol/L	5.11	6.4%	76	1665
Beckman AU series OTHER \ Cholesterol oxidase	mmol/L	5.23	0.6%	7	<100
Beckman AU series KHB \ Cholesterol oxidase	mmol/L	5.02	3.5%	7	192
Abbott ARCHITECT c series MedicalSystem \ Cholesterol oxidase	mmol/L	5.07	10%	6	<100
OTHER KeFang(Guangzhou) \ Cholesterol oxidase	mmol/L	5.18	0.3%	6	<100
Beckman AU series Beckman Coulter \ Cholesterol oxidase	mmol/L	5.43	4.2%	55	931
Beckman AU series XJKC \ Cholesterol oxidase	mmol/L	4.4	4.8%	5	<100
DiruiCS series MedicalSystem \ Cholesterol oxidase	mmol/L	4.82	4.5%	5	<100
Beckman AU series DIRUI \ Cholesterol oxidase	mmol/L	5.2	1.6%	5	<100
Beckman AU series MedicalSystem \ Cholesterol oxidase	mmol/L	5.46	5.5%	46	573
Beckman AU series Fosun \ Cholesterol oxidase	mmol/L	5.34	3.2%	42	1148
Beckman AU series DiaSys \ Cholesterol oxidase	mmol/L	5.03	4.2%	41	327
Roche Modular P series DiaSys \ Cholesterol oxidase	mmol/L	5.16	13.1%	<5	<100
Beckman AU series BIOBASE \ Cholesterol oxidase	mmol/L	5.11	3.2%	<5	<100
Beckman AU series chongqiong zybio \ Cholesterol oxidase	mmol/L	5.05	3.9%	35	300
KHB ZY series KHB \ Cholesterol oxidase	mmol/L	5.04	3.8%	33	246
(CK)肌酸激酶					
Beckman AU series Beckman Coulter \ Rate -other buffers	U/L	125	2.8%	9	256
Beckman AU series DiaSys \ Rate -AMP buffer	U/L	77.9	3.6%	7	<100
Abbott ARCHITECT c series MedicalSystem \ Rate -other buffers	U/L	82.8	2.5%	6	<100
Mindray BS series Mindray \ dry chemistry	U/L	53.9	20.7%	5	<100
Beckman AU series Fosun \ Rate -AMP buffer	U/L	81.8	11.7%	41	1104
KHB ZY series DIRUI \ Rate -other buffers	U/L	101	1.8%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Rate -AMP buffer	U/L	102	2.7%	<5	118
Beckman AU series Fosun \ OTHER	U/L	78.9	11.8%	23	531
Beckman AU series Reebio \ Gelatin particle agglutination (PA)	U/L	85.9	12.5%	20	507
TOSHIBA TBA chongqiong zybio \ Urease-UV-rate	U/L	84.5	6.9%	<5	<100
Beckman AU series Reebio \ dry chemistry	U/L	75.4	5.8%	<5	<100
DiruiH-800 XC Bio \ Rate -other buffers	U/L	79.5	19.4%	15	411
KHB ZY series KHB \ Rate -AMP buffer	U/L	81.8	2.3%	13	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Rate -AMP buffer	U/L	129	1.5%	12	185
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	120	4.7%	10	207
Mindray BS series Mindray \ OTHER	U/L	42.6	1.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	107	12%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Urease-UV-rate	U/L	107	1%	<5	<100
Roche cobas 4000/6000/8000 MAKER \ Rate -AMP buffer	U/L	111	1.3%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
Mindray BS series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	U/L	26.5	0%	<5	<100
(Cl)氯					
OTHER 3V-Diagnostic \ ISE (direct)	mmol/L	98	4.1%	7	188
Abbott ARCHITECT c series Abbott \ ISE (direct)	mmol/L	102	1%	6	<100
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	99.3	2.4%	46	1247
Beckman AU series LEADMAN \ ISE (direct)	mmol/L	82.4	1.2%	<5	119

REF	580-1	6x5mL	EXP	2022-11-25	LOT	3753107	更新日期：	2022-05
								5 / 11

		批号：3753107			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems	mmol/L	101	1.6%	<5	119
Wako \ ISE (indirect)					
Roche cobas 4000/6000/8000	mmol/L	99.1	9.6%	<5	<100
Roche \ ISE (indirect)					
Beckman AU series	mmol/L	102	1.8%	24	640
Fosun \ ISE (direct)					
Beckman AU series	mmol/L	98.5	1.1%	23	522
Dedicated \ ISE (indirect)					
URIT-500B/330/180/55	mmol/L	100	4.9%	20	497
Uritest \ Coulometric titration					
Abbott ARCHITECT c series	mmol/L	98.7	0.3%	<5	<100
Dedicated \ ISE (direct)					
PSD	mmol/L	103	0.3%	<5	<100
OTHER \ ISE (direct)					
TOSHIBA TBA	mmol/L	105	0.3%	<5	<100
chongqiong zybio \ ISE (indirect)					
Beckman AU series	mmol/L	104	3.8%	18	426
Reebio \ Coulometric titration					
Roche Hitachi/Hitachi Systems	mmol/L	91.9	1.7%	12	193
Wako \ ISE (direct)					
Roche Hitachi/Hitachi Systems	mmol/L	97	6.7%	10	245
OTHER \ OTHER					
OTHER	mmol/L	104	0.4%	<5	<100
Dedicated \ dry chemistry					
Roche cobas 4000/6000/8000	mmol/L	102	0.5%	<5	<100
Wako \ ISE (direct)					
J&J Vitros 250/350	mmol/L	104	1.8%	<5	<100
Ortho \ dry chemistry					
(CO2 (HCO3-))二氧化碳 (碳酸氢根)					
Beckman AU series	mmol/L	20	3.4%	7	<100
MedicalSystem \ OTHER					
Abbott ARCHITECT c series	mmol/L	23.6	5.3%	6	<100
MedicalSystem \ PEPC					
Roche Hitachi/Hitachi Systems	mmol/L	21.1	6.9%	36	865
MAKER \ OTHER					
Beckman AU series	mmol/L	18.9	4.4%	24	531
Whitman \ OTHER					
Roche cobas 4000/6000/8000	mmol/L	23.7	10.3%	<5	<100
Roche \ PEPC					
J&J Vitros 250/350	mmol/L	25	3.6%	<5	<100
Ortho \ dry chemistry					
(Cr)肌酐					
Beckman AU series	μmol/L	23.3	22.1%	9	215
Reebio \ Enzymatic					
Roche Hitachi/Hitachi Systems	μmol/L	25.3	3.5%	<5	119
Wako \ Enzymatic					
Beckman AU series	μmol/L	29.6	4.7%	<5	<100
chongqiong zybio \ Enzymatic					
Beckman AU series	μmol/L	24.8	6.6%	<5	<100
Fosun \ OTHER					
Beckman AU series	μmol/L	26	6.9%	16	340
Fosun \ Enzymatic					
Beckman AU series	μmol/L	19.1	2.1%	<5	<100
MEDCONN \ dry chemistry					
Roche cobas 4000/6000/8000	μmol/L	24.7	12%	<5	<100
Roche \ Enzymatic					
J&J Vitros 250/350	μmol/L	39.6	1.3%	<5	<100
Ortho \ dry chemistry					
(DBIL)直接胆红素					
Beckman AU series	μmol/L	5.95	15.6%	9	253
Beckman Coulter \ Diazotization					
KHB ZY series	μmol/L	5.66	6.4%	8	<100
DIRUI \ Vanadate					
SDBK	μmol/L	7.16	6.4%	8	<100
BIOBASE \ Vanadate					
Beckman AU series	μmol/L	7.13	16.7%	8	<100
chongqiong zybio \ Diazotization					
Beckman AU series	μmol/L	7.57	3.7%	7	<100
DiaSys \ Diazotization					
OTHER	μmol/L	6.34	1.8%	7	<100
KeFang(Guangzhou) \ Vanadate					
Roche Hitachi/Hitachi Systems	μmol/L	5.74	5%	7	137
Wako \ Diazotization					
Abbott ARCHITECT c series	μmol/L	6.17	5.6%	6	<100
MedicalSystem \ sodium nitrite oxidase method					
Beckman AU series	μmol/L	7.08	6.9%	6	<100
Beckman Coulter \ Vanadate					
Roche Hitachi/Hitachi Systems	μmol/L	7.99	11%	6	132
Wako \ Vanadate					
Beckman AU series	μmol/L	5.39	9.3%	5	<100
WanTai \ Vanadate					
DiruiCS series	μmol/L	5.47	5.1%	5	<100
DiaSys \ Diazotization					
Beckman AU series	μmol/L	7.22	12.5%	5	<100
MedicalSystem \ Vanadate					
Beckman AU series	μmol/L	7.65	15.7%	49	1189
Fosun \ Vanadate					
Beckman AU series	μmol/L	7.28	6.1%	<5	<100
Reebio \ Vanadate					
Beckman AU series	μmol/L	8.55	0.1%	<5	<100
DIRUI \ Vanadate					

REF	580-1	6x5mL	EXP	2022-11-25	LOT	3753107	更新日期：	2022-05
								6 / 11

		批号：3753107			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series Mindray \ Vanadate	μmol/L	7.79	17.9%	31	201
Beckman AU series KHB \ Diazotization	μmol/L	7.62	3.2%	<5	<100
DiruiH-800 XC Bio \ Vanadate	μmol/L	6.43	9.5%	<5	<100
(Fe)铁					
Roche Hitachi/Hitachi Systems OTHER \ OTHER	μmol/L	52.7	2.7%	32	599
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	48.5	12%	<5	<100
(GGT)L-γ-谷氨酰基转移酶					
Beckman AU series OTHER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	51.3	0.3%	8	<100
KHB ZY series DIRUI \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	47.8	2.1%	8	<100
DiruiCS series DiaSys \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	45.2	7.5%	8	<100
Beckman AU series Reebio \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	46	3%	71	1591
Beckman AU series JS (Shanghai) \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	48.5	6.1%	7	<100
SDBK BIOBASE \ dry chemistry	U/L	43.7	6.1%	7	<100
SDBK BIOBASE \ OTHER	U/L	47.8	2.5%	7	<100
Beckman AU series KHB \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	46	5.4%	7	191
SDBK BIOBASE \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	43.6	10.6%	6	<100
Abbott ARCHITECT c series MedicalSystem \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	45.8	2.8%	6	<100
OTHER KeFang(Guangzhou) \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	43	0.4%	6	<100
Beckman AU series Reebio \ dry chemistry	U/L	47.4	5.5%	5	<100
Beckman AU series XJKC \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	54	6.3%	5	<100
Mindray BS series Mindray \ OTHER	U/L	47.4	1.9%	5	<100
Beckman AU series Beckman Coulter \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	48.9	7.4%	47	741
Mindray BS series Mindray \ dry chemistry	U/L	49.3	10.5%	46	273
Beckman AU series DIRUI \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	50.5	0.2%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	46.2	1.4%	<5	119
KHB ZY series KHB \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	47.6	5.5%	32	240
(GLU)葡萄糖					
Mindray BS series chongqiong zybio \ Dehydrogenase	mmol/L	2.67	2.5%	9	<100
Mindray BS series Mindray \ Hexokinase	mmol/L	2.48	2.4%	9	249
Beckman AU series Bio-assay \ Glucose oxidase (GOD)	mmol/L	2.71	1.9%	8	<100
KHB ZY series DIRUI \ Glucose oxidase (GOD)	mmol/L	3.37	3.7%	8	<100
Beckman AU series OTHER \ Hexokinase	mmol/L	2.8	0.7%	7	<100
Siemens ADVIA 1800/2400 Olympus \ Hexokinase	mmol/L	3.63	0.9%	7	<100
SDBK BIOBASE \ Dehydrogenase	mmol/L	2.86	2.5%	7	<100
Beckman AU series KHB \ Hexokinase	mmol/L	2.43	3.6%	7	181
Beckman AU series Fosun \ Hexokinase	mmol/L	2.75	5.2%	66	1633
Beckman AU series Reebio \ Dehydrogenase	mmol/L	2.63	5.4%	64	1582
DiruiCS series DIRUI \ Hexokinase	mmol/L	2.55	3.7%	6	<100
Abbott ARCHITECT c series MedicalSystem \ Hexokinase	mmol/L	2.6	1.4%	6	<100
Beckman AU series Beckman Coulter \ Glucose oxidase (GOD)	mmol/L	2.77	2.5%	6	<100
OTHER KeFang(Guangzhou) \ Glucose oxidase (GOD)	mmol/L	2.59	0.4%	6	<100
Beckman AU series MedicalSystem \ Hexokinase	mmol/L	2.78	6.4%	6	<100
Beckman AU series Beckman Coulter \ Hexokinase	mmol/L	2.68	9.3%	50	901
Beckman AU series BIOBASE \ Hexokinase	mmol/L	2.65	3.2%	5	<100
Dirui Fus-200/100 MedicalSystem \ Hexokinase	mmol/L	2.56	6%	5	<100
Sinnova DS series Sinnova \ Dehydrogenase	mmol/L	2.66	2%	<5	<100

REF	580-1	6x5mL	EXP	2022-11-25	LOT	3753107	更新日期：	2022-05
								7 / 11

		批号：3753107				
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数		
(HDL-C)高密度脂蛋白胆固醇						
Mindray BS series chongqiong zybio \ Enzymatic	mmol/L	1.59	2.5%	9	<100	
Beckman AU series Beckman Coulter \ Direct Measure	mmol/L	1.27	2.9%	9	243	
Beckman AU series OTHER \ selected suppression	mmol/L	1.54	3.5%	9	142	
KHB ZY series DIRUI \ selected suppression	mmol/L	1.39	2.6%	8	<100	
Beckman AU series MedicalSystem \ Direct Measure	mmol/L	1.51	5.2%	7	<100	
SDBK BIOBASE \ Direct Measure	mmol/L	1.48	11.2%	7	<100	
Beckman AU series KHB \ Direct Measure	mmol/L	1.46	3.7%	7	160	
Beckman AU series MedicalSystem \ Enzymatic	mmol/L	1.72	2.3%	6	<100	
Beckman AU series chongqiong zybio \ Enzymatic	mmol/L	1.51	20%	5	<100	
Beckman AU series Reebio \ Direct Measure	mmol/L	1.58	5%	45	1120	
Beckman AU series Fosun \ Direct Measure	mmol/L	1.49	4.5%	41	1025	
Roche Hitachi/Hitachi Systems Wako \ selected suppression	mmol/L	1.58	2.2%	<5	117	
SDBK BIOBASE \ Enzymatic	mmol/L	1.48	0.9%	<5	<100	
Beckman AU series Fosun \ OTHER	mmol/L	1.48	5%	23	553	
Mindray BS series Getein Biotech \ OTHER	mmol/L	1.25	4.3%	<5	<100	
Abbott ARCHITECT c series Dedicated \ Direct Measure	mmol/L	1.48	7.2%	<5	<100	
Mindray BS series WEGO \ selected suppression	mmol/L	0.95	1.3%	<5	<100	
KHB ZY series KHB \ Enzymatic	mmol/L	2.07	14.3%	<5	<100	
Mindray BS series Mindray \ Direct Measure	mmol/L	1.41	16.2%	16	243	
(K)钾						
Abbott ARCHITECT c series Abbott \ ISE (direct)	mmol/L	2.29	2%	5	<100	
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	2.41	6.2%	44	1117	
Roche Hitachi/Hitachi Systems Wako \ ISE (indirect)	mmol/L	2.33	4%	<5	118	
Beckman AU series Fosun \ ISE (direct)	mmol/L	2.34	2.7%	24	648	
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	2.29	2.2%	23	519	
URIT-500B/330/180/55 Uritest \ ISE (direct)	mmol/L	2.39	2.2%	20	504	
Abbott ARCHITECT c series Dedicated \ ISE (direct)	mmol/L	2.16	0%	<5	<100	
PSD OTHER \ ISE (direct)	mmol/L	2.45	0.9%	<5	<100	
Beckman AU series Reebio \ ISE (direct)	mmol/L	2.37	3.1%	18	411	
OTHER 3V-Diagnostic \ ISE (direct)	mmol/L	2.4	3.9%	12	330	
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	2.33	6%	10	265	
OTHER Dedicated \ dry chemistry	mmol/L	2.41	0.2%	<5	<100	
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	2.42	12%	<5	<100	
Roche cobas 4000/6000/8000 Wako \ ISE (direct)	mmol/L	2.35	1.1%	<5	<100	
J&J Vitros 250/350 Ortho \ dry chemistry	mmol/L	2.46	1.8%	<5	<100	
Roche cobas 4000/6000/8000 Roche \ ISE (direct)	mmol/L	2.49	0.5%	<5	<100	
(LA)乳酸						
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	mmol/L	5.91	12%	<5	<100	
(LDH)乳酸脱氢酶						
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	U/L	10.3	11.9%	<5	<100	
(LDL-C)低密度脂蛋白胆固醇						
Mindray BS series chongqiong zybio \ Enzymatic	mmol/L	2.61	1.4%	9	<100	
Beckman AU series MedicalSystem \ Direct Measure	mmol/L	2.64	1.5%	7	<100	
Beckman AU series KHB \ Direct Measure	mmol/L	2.48	4.3%	7	195	
Beckman AU series Reebio \ detergent	mmol/L	2.75	4.5%	43	1088	
KHB ZY series DIRUI \ detergent	mmol/L	1.88	5.8%	<5	<100	

REF	580-1	6x5mL	EXP	2022-11-25	LOT	3753107	更新日期：	2022-05
								8 / 11

		批号：3753107			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
SDBK BIOBASE \ detergent	mmol/L	2.79	4%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ selected suppression	mmol/L	3.28	3.3%	<5	119
Beckman AU series Fosun \ OTHER	mmol/L	2.28	5%	25	567
Mindray BS series Getein Biotech \ OTHER	mmol/L	3.01	4%	<5	<100
Abbott ARCHITECT c series Dedicated \ Direct Measure	mmol/L	2.46	0.9%	<5	<100
TOSHIBA TBA chongqiong zybio \ Direct Measure	mmol/L	2.48	3.5%	<5	<100
Mindray BS series WEGO \ detergent	mmol/L	2.24	0.8%	<5	<100
SDBK BIOBASE \ Direct Measure	mmol/L	3.61	6.4%	<5	<100
Beckman AU series MedicalSystem \ detergent	mmol/L	2.31	1%	<5	<100
KHB ZY series KHB \ Enzymatic	mmol/L	1.87	4.7%	<5	<100
SDBK BIOBASE \ Enzymatic	mmol/L	2.85	5.1%	<5	<100
Beckman AU series chongqiong zybio \ Enzymatic	mmol/L	2.86	4.9%	18	139
Mindray BS series Mindray \ Direct Measure	mmol/L	2.4	12.8%	16	281
DiruiH-800 XC Bio \ detergent	mmol/L	2.76	3.4%	14	386
(LIP)脂肪酶					
Beckman AU series MedicalSystem \ OTHER	U/L	79.5	5%	6	<100
Roche Hitachi/Hitachi Systems OTHER \ OTHER	U/L	80.6	8.6%	48	760
Roche cobas 4000/6000/8000 DiaSys \ Enzymatic Cycling Assays Method	U/L	53.7	12%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	U/L	48.9	2.7%	<5	<100
(Mg)镁					
Beckman AU series DiaSys \ Xylidyl Blue	mmol/L	0.825	4.1%	7	<100
Beckman AU series Fosun \ Xylidyl Blue	mmol/L	0.7	17.1%	23	606
Abbott ARCHITECT c series Dedicated \ Xylidyl Blue	mmol/L	1.08	6.1%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Xylidyl Blue	mmol/L	0.851	14%	10	278
Beckman AU series Beckman Coulter \ Xylidyl Blue	mmol/L	0.506	2.6%	10	272
Roche cobas 4000/6000/8000 Roche \ Xylidyl Blue	mmol/L	0.49	12.1%	<5	<100
TOSHIBA TBA chongqiong zybio \ Arsenazo □	mmol/L	1.05	1.8%	<5	<100
J&J Vitros 250/350 Ortho \ dry chemistry	mmol/L	0.673	2.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ ICP-MS	mmol/L	0.843	1.4%	<5	<100
(Na)钠					
Abbott ARCHITECT c series Abbott \ ISE (direct)	mmol/L	154	0.9%	6	<100
Beckman AU series LEADMAN \ ISE (direct)	mmol/L	145	0.9%	<5	121
Roche Hitachi/Hitachi Systems Wako \ ISE (indirect)	mmol/L	155	0.4%	<5	118
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	160	2.3%	29	613
Beckman AU series Fosun \ ISE (direct)	mmol/L	154	1.2%	24	596
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	150	0.8%	23	516
URIT-500B/330/180/55 Uritest \ ISE (direct)	mmol/L	158	3%	20	501
Abbott ARCHITECT c series Dedicated \ ISE (direct)	mmol/L	147	0.4%	<5	<100
PSD OTHER \ ISE (direct)	mmol/L	163	0.2%	<5	<100
TOSHIBA TBA chongqiong zybio \ ISE (indirect)	mmol/L	158	0.4%	<5	<100
Beckman AU series Reebio \ ISE (direct)	mmol/L	162	1.2%	18	428
OTHER Dedicated \ ISE (direct)	mmol/L	151	2.4%	15	411
Roche Hitachi/Hitachi Systems Wako \ ISE (direct)	mmol/L	160	1%	12	205
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	172	3%	10	253
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	158	12%	<5	<100
OTHER Dedicated \ dry chemistry	mmol/L	154	0.4%	<5	<100
Roche cobas 4000/6000/8000 Wako \ ISE (direct)	mmol/L	154	0.3%	<5	<100

REF	580-1	6x5mL	EXP	2022-11-25	LOT	3753107	更新日期：	2022-05
								9 / 11

		批号：3753107			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000	mmol/L	158	0.6%	<5	<100
Roche \ ISE (direct)					
J&J Vitros 250/350	mmol/L	165	3.1%	<5	<100
Ortho \ dry chemistry					
(P)磷					
Beckman AU series	mmol/L	0.69	2.2%	9	256
Beckman Coulter \ Phosphomolybdate UV-rate					
Beckman AU series	mmol/L	0.845	10.2%	8	234
Fosun \ Phosphomolybdate UV-rate					
Roche Hitachi/Hitachi Systems	mmol/L	0.761	1.7%	7	132
Wako \ Phosphomolybdate UV-rate					
Abbott ARCHITECT c series	mmol/L	0.597	4%	6	<100
MedicalSystem \ Phosphomolybdate UV-rate					
Roche Hitachi/Hitachi Systems	mmol/L	1.09	1.1%	<5	118
Wako \ Phosphomolybdate-UV-end point					
Abbott ARCHITECT c series	mmol/L	0.812	1.2%	<5	<100
Dedicated \ Phosphomolybdate UV-rate					
DiruiH-800	mmol/L	0.805	0.8%	<5	<100
XC Bio \ Phosphomolybdate-UV-end point					
TOSHIBA TBA	mmol/L	0.669	7.2%	<5	<100
chongqiong zybio \ Phosphomolybdate-UV-end point					
Mindray BS series	mmol/L	0.824	1.2%	<5	<100
WEGO \ Phosphomolybdate-UV-end point					
Roche cobas 4000/6000/8000	mmol/L	0.664	12.9%	<5	<100
Roche \ Phosphomolybdate-UV-end point					
Beckman AU series	mmol/L	0.657	8%	18	494
Beckman Coulter \ Phosphomolybdate-UV-end point					
Beckman AU series	mmol/L	0.834	10.1%	16	416
Fosun \ Molybdenum blue colorimetry					
Beckman AU series	mmol/L	0.831	9.7%	15	378
Fosun \ OTHER					
Beckman AU series	mmol/L	0.766	2.5%	<5	<100
DiaSys \ Phosphomolybdate-UV-end point					
Roche cobas 4000/6000/8000	mmol/L	0.74	12.1%	<5	<100
Roche \ Phosphomolybdate UV-rate					
Roche cobas 4000/6000/8000	mmol/L	0.681	0.8%	<5	<100
Wako \ Molybdenum blue colorimetry					
J&J Vitros 250/350	mmol/L	0.769	1.2%	<5	<100
Ortho \ dry chemistry					
(PAB)前白蛋白					
Beckman AU series	g/L	0.262	0%	<5	<100
Fosun \ Immunoturbidimetry (ITA)					
(PRO/ TP)(总) 蛋白					
KHB ZY series	g/L	81.3	2.7%	8	<100
DIRUI \ Biuret-rate					
DiruiCS series	g/L	79.4	3.3%	8	<100
MedicalSystem \ Biuret-endpoint					
Beckman AU series	g/L	82.2	0.2%	7	<100
OTHER \ Biuret-endpoint					
Beckman AU series	g/L	83.4	2.4%	7	191
KHB \ Biuret-endpoint					
Beckman AU series	g/L	83.6	2.2%	68	1608
Reebio \ Biuret-endpoint					
Beckman AU series	g/L	85.2	3.3%	66	686
MedicalSystem \ Biuret-endpoint					
Beckman AU series	g/L	84.3	3.5%	66	1640
Fosun \ Biuret-endpoint					
Mindray BS series	g/L	84.2	4.4%	63	606
Mindray \ Biuret-endpoint					
Abbott ARCHITECT c series	g/L	83.4	0.9%	6	<100
MedicalSystem \ Biuret-endpoint					
OTHER	g/L	79.7	0.1%	6	<100
KeFang(Guangzhou) \ Biuret-endpoint					
Beckman AU series	g/L	84.8	4.2%	54	932
Beckman Coulter \ Biuret-endpoint					
Beckman AU series	g/L	72.3	4.6%	5	<100
XJKC \ Biuret-endpoint					
Beckman AU series	g/L	83.4	2%	5	<100
WanTai \ Biuret-endpoint					
Beckman AU series	g/L	64.9	1.4%	5	149
LEADMAN \ Biuret-endpoint					
Roche Modular P series	g/L	84.5	7.9%	<5	<100
DiaSys \ Biuret-endpoint					
Beckman AU series	g/L	89.7	0.1%	<5	<100
DIRUI \ Biuret-endpoint					
Beckman AU series	g/L	84.9	4%	35	282
DiaSys \ Biuret-endpoint					
KHB ZY series	g/L	79.8	3.3%	33	246
KHB \ Biuret-endpoint					
Beckman AU series	g/L	83.7	0.6%	<5	<100
BIOSINO \ Biuret-endpoint					
(TBA)总胆汁酸					
Roche Hitachi/Hitachi Systems	μmol/L	8.75	23.8%	8	150
MAKER \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	8.09	7.6%	6	<100
MedicalSystem \ Enzymatic Cycling Assays Method					
Abbott ARCHITECT c series	μmol/L	5.69	9.5%	<5	<100
MedicalSystem \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	10	16.6%	18	346
Fosun \ Enzymatic Cycling Assays Method					

REF	580-1	6x5mL	EXP	2022-11-25	LOT	3753107	更新日期：	2022-05
								10 / 11

		批号：3753107			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000 DiaSys \ Enzymatic Cycling Assays Method	μmol/L	9.06	12%	<5	<100
(TBIL)总胆红素					
KHB ZY series DIRUI \ Vanadate	μmol/L	13.2	3.3%	8	<100
SDBK BIOBASE \ Vanadate	μmol/L	11.9	8.3%	8	<100
DiruiCS series DIRUI \ Diazotization	μmol/L	13.6	10.7%	6	<100
Abbott ARCHITECT c series MedicalSystem \ sodium nitrite oxidase method	μmol/L	11.3	2.1%	6	<100
OTHER KeFang(Guangzhou) \ Vanadate	μmol/L	12.3	0.1%	6	<100
Beckman AU series MedicalSystem \ Vanadate	μmol/L	12.4	6.2%	6	<100
Beckman AU series KHB \ Vanadate	μmol/L	10.7	3%	6	166
Roche Hitachi/Hitachi Systems Wako \ Vanadate	μmol/L	14.7	3.6%	6	132
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	11.8	12.8%	54	1337
Beckman AU series Fosun \ Vanadate	μmol/L	13.1	16%	50	1260
Beckman AU series WanTai \ Diazotization	μmol/L	9.4	8.4%	5	<100
DiruiCS series DiaSys \ Diazotization	μmol/L	14.7	9.2%	5	<100
Sinnowa DS series Sinnowa \ Diazotization	μmol/L	13.5	6.6%	5	<100
Mindray BS series Mindray \ Vanadate	μmol/L	12.1	11.7%	45	350
Beckman AU series Reebio \ Diazotization	μmol/L	11.4	15.3%	45	1128
Beckman AU series DiaSys \ Diazotization	μmol/L	12.6	13.2%	42	337
Beckman AU series Reebio \ Vanadate	μmol/L	11.4	3.1%	<5	<100
Beckman AU series DIRUI \ Vanadate	μmol/L	12.9	1%	<5	<100
Mindray BS series Mindray \ Diazotization	μmol/L	12.3	10.6%	36	208
(TG)甘油三酯					
Mindray BS series chongqiong zybio \ GPO-POD	mmol/L	2.63	2.3%	9	<100
Beckman AU series chongqiong zybio \ Acetyl - acetone photometry	mmol/L	2.72	23.8%	8	<100
KHB ZY series DIRUI \ GPO-POD	mmol/L	2.63	1.9%	8	<100
Beckman AU series Fosun \ GPO-POD	mmol/L	2.8	8%	8	201
Beckman AU series OTHER \ GPO-POD	mmol/L	2.77	0.7%	7	<100
SDBK BIOBASE \ GPO-POD	mmol/L	2.75	6.3%	7	<100
SDBK BIOBASE \ Acetyl - acetone photometry	mmol/L	2.75	0.8%	7	<100
Abbott ARCHITECT c series MedicalSystem \ GPO-POD	mmol/L	2.47	3.4%	6	<100
OTHER KeFang(Guangzhou) \ GPO-POD	mmol/L	2.69	0.4%	6	<100
Beckman AU series BIOBASE \ Dehydrogenase	mmol/L	2.7	5.1%	5	<100
DiruiCS series MedicalSystem \ GPO-POD	mmol/L	2.22	9.4%	5	<100
Beckman AU series chongqiong zybio \ GPO-POD	mmol/L	2.59	4.6%	5	<100
Beckman AU series Beckman Coulter \ GPO-POD	mmol/L	2.89	23.2%	47	923
Roche Modular P series DiaSys \ GPO-POD	mmol/L	2.52	8.1%	<5	<100
SDBK BIOBASE \ Dehydrogenase	mmol/L	2.57	2%	<5	<100
DiruiCS series Dedicated \ GK-PK-LDH	mmol/L	2.7	13%	<5	<100
Beckman AU series DIRUI \ Dehydrogenase	mmol/L	3.79	0.8%	<5	<100
Beckman AU series KHB \ Dehydrogenase	mmol/L	2.9	23.2%	<5	103
Mindray BS series Mindray \ GPO-POD	mmol/L	2.67	10.6%	32	208
(UA)尿酸					
Beckman AU series Bio-assay \ Uricense colorimetry	μmol/L	222	1.3%	8	<100
KHB ZY series DIRUI \ Uricense colorimetry	μmol/L	240	2.9%	8	<100
DiruiCS series MedicalSystem \ Uricense colorimetry	μmol/L	230	7.1%	8	<100
Beckman AU series OTHER \ Uricense colorimetry	μmol/L	230	0.2%	7	<100
SDBK BIOBASE \ Uricense colorimetry	μmol/L	223	4.3%	7	<100

REF	580-1	6x5mL	EXP	2022-11-25	LOT	3753107	更新日期：	2022-05
								11 / 11

		批号：3753107			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series KHB \ Uricase colorimetry	μmol/L	235	2.8%	7	170
DiruiCS series DIRUI \ Uricase-UV	μmol/L	231	6.2%	6	<100
Abbott ARCHITECT c series MedicalSystem \ Uricase colorimetry	μmol/L	234	2.2%	6	<100
Beckman AU series chongqiong zybio \ Uricase colorimetry	μmol/L	234	3.8%	6	<100
Beckman AU series Reebio \ dry chemistry	μmol/L	222	2.7%	5	<100
Beckman AU series WanTai \ Uricase colorimetry	μmol/L	233	3.3%	5	<100
OTHER KeFang(Guangzhou) \ Uricase colorimetry	μmol/L	214	0.1%	5	<100
Beckman AU series Reebio \ Uricase colorimetry	μmol/L	245	6%	43	1001
Roche Modular P series DiaSys \ Uricase-UV	μmol/L	256	4.1%	<5	<100
SDBK BIOBASE \ OTHER	μmol/L	226	1.9%	<5	<100
Beckman AU series DIRUI \ Uricase colorimetry	μmol/L	188	0.1%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ Uricase colorimetry	μmol/L	231	1.2%	<5	118
Beckman AU series Beckman Coulter \ Uricase colorimetry	μmol/L	220	9.6%	36	809
Beckman AU series DiaSys \ Uricase colorimetry	μmol/L	233	1.8%	35	282
(UIBC)不饱和和铁结合力					
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	24.4	14.8%	<5	<100
(Urea)尿素					
Beckman AU series chongqiong zybio \ Urease-UV-rate	mmol/L	6.59	5.8%	9	<100
KHB ZY series DIRUI \ Urease colorimetry	mmol/L	6.16	4.1%	8	<100
Sinnowa DS series Sinnowa \ OTHER	mmol/L	5.44	1.5%	8	<100
Beckman AU series KHB \ Urease colorimetry	mmol/L	6.95	3.7%	7	171
Beckman AU series MedicalSystem \ Urease colorimetry	mmol/L	6.39	6.3%	6	<100
Beckman AU series Fosun \ Urease-UV-rate	mmol/L	7.08	3%	42	1050
SDBK BIOBASE \ Diacetyl monoxime	mmol/L	6.58	3.9%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ Urease-UV-rate	mmol/L	7.89	4.1%	<5	119
Beckman AU series Reebio \ Diacetyl monoxime	mmol/L	7.03	5.4%	38	960
Beckman AU series Beckman Coulter \ Urease-UV-rate	mmol/L	7.43	5.2%	36	683
Roche cobas 4000/6000/8000 Roche \ Urease colorimetry	mmol/L	7.1	10%	<5	<100
SDBK BIOBASE \ Urease colorimetry	mmol/L	6.33	4.2%	<5	<100
Mindray BS series BIOBASE \ Urease colorimetry	mmol/L	6.69	4.8%	<5	<100
DiruiCS series DIRUI \ Urease-UV-rate	mmol/L	7.15	3.9%	<5	<100
Beckman AU series Beckman Coulter \ Urease colorimetry	mmol/L	6.81	4%	<5	<100
Mindray BS series Mindray \ OTHER	mmol/L	7.35	2.4%	<5	<100
DiruiCS series MedicalSystem \ Urease colorimetry	mmol/L	6.93	4.6%	<5	<100
Beckman AU series Fosun \ Urease colorimetry	mmol/L	7.3	5.5%	24	474
Beckman AU series DiaSys \ Urease-UV-rate	mmol/L	7.34	2.1%	24	203
(β2M)β-2-微球蛋白					
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6700	12.1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。