

质控品名： 复合免疫分析非定值质控品

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
1 / 11								

		批号：3382863			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(17-OHP)17-羟孕酮					
Agilent agilent \ MS	ng/mL	0.759	11.9%	<5	<100
(25-OH-VD2)25羟基维生素D2					
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	21.1	8.2%	19	324
(25-OH-VD3)25羟基维生素D3					
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	20	11.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	22.8	7.6%	<5	<100
(25-OH-Vit D)25羟基维生素D					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	19.5	9.1%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	48.7	9.1%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	21.8	11.9%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	nmol/L	54.4	11.9%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	20.7	9.8%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	51.6	9.8%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	19.6	7%	19	253
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	49	7%	19	253
(AFP)甲胎蛋白					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.42	6.6%	9	126
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.11	11.8%	7	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.75	4.1%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.11	10.2%	38	773
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.04	5.4%	<5	<100
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.73	0.8%	24	434
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.96	4.3%	20	401
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.34	11.9%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.65	11.9%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.02	8.4%	17	109
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.79	6.2%	11	209
(ALD)醛固酮					
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	pg/mL	70.3	11.9%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pg/mL	134	9.9%	<5	<100
(AMH)抗缪勒氏管激素					
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.58	5.9%	19	<100
(Anti-Tg)抗甲状腺球蛋白抗体					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	20.2	11.8%	35	690
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.4	51.6%	35	446
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	10.9	7.6%	28	468
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	21.6	23%	25	349
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.4	26.1%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	1.89	11.2%	19	450
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	1.9	7.5%	18	217

质控品名： 复合免疫分析非定值质控品

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
2 / 11								

批号：3382863					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	9.91	9.3%	<5	<100
(Anti-TPO)抗甲状腺过氧化物酶抗体					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.6	24.2%	35	707
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	12.3	5%	30	516
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	28.4	5.1%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	17.5	27.4%	23	341
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.8	14.1%	19	316
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	7.91	10%	18	444
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	25	8.9%	<5	<100
(BAP (Ost))骨碱性磷酸酶 (骨泌素)					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.1	1.4%	<5	<100
(CA 125)糖类抗原 125					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.6	5.3%	9	129
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	37.1	12.1%	6	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	33.2	3.3%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	29.3	8.4%	38	763
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	29.8	3.3%	<5	<100
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	40.8	1%	24	434
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	48.5	11.9%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	28.5	5.1%	19	413
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.6	6%	17	104
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	28.9	13.4%	12	233
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.9	10.6%	<5	<100
(CA 15-3)糖类抗原 15-3					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	32.4	12.9%	9	128
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.8	12.6%	6	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	28.2	5.3%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	28.2	7.9%	37	744
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	35.9	1.2%	24	434
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	29.4	11.9%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.1	0.3%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.7	8.5%	17	113
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27.6	5.2%	12	229
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	16.2	11.7%	<5	<100
(CA 19-9)糖类抗原 19-9					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34.4	17.3%	9	120
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.2	9%	5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	26	3.9%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	26.8	10.8%	38	751
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.6	5%	20	399
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	20.6	11.9%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27	7.9%	<5	<100

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
3 / 11								

批号：3382863					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	42.9	5.7%	17	110
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	25.3	4.1%	12	222
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.2	10.7%	<5	<100
(CA 72-4)糖类抗原 72-4					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	1.84	3.9%	19	382
(CEA)癌胚抗原					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.27	19.6%	9	<100
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.82	4%	9	129
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.55	2.7%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.32	4.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.34	9.4%	37	726
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.89	1.2%	24	432
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.9	11.9%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.55	11.9%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.54	4.3%	19	367
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.51	10.7%	17	114
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.8	20.6%	11	215
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.78	10%	<5	<100
(CK)肌酸激酶					
Roche cobas 4000/6000/8000 Roche \ Urease-UV-rate	U/L	130	10%	<5	<100
(CK-MB mass)肌酸激酶-MB质量					
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.99	6.1%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
J&J Vitros ECL Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.69	6.8%	8	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.47	10.3%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.5	11.9%	<5	101
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.76	7.3%	25	566
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	U/L	9.25	9.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.08	11.9%	<5	<100
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.02	7%	17	405
Beckman UniCel Dxl 600/800 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.06	6.7%	12	333
(Cortisol)皮质醇					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	104	3.3%	22	624
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	115	11.9%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	134	9.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	114	6.3%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	106	6.4%	18	386
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	98.9	9.2%	11	309
(C-P)C-肽					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.348	65%	6	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.478	76.4%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.07	5.6%	31	657

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
4 / 11								

		批号：3382863			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.648	11.9%	<5	<100
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.892	3.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.18	4.9%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.05	2.5%	18	236
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.93	9.7%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.03	10.7%	<5	<100
(DGXN)地高辛					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.88	10%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.87	11.9%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.738	12%	<5	<100
(DHEA-S)硫酸脱氢表雄酮					
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3390	11.9%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3416	9.9%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3152	5.1%	19	597
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3906	10.4%	<5	<100
(E2)雌二醇					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	340	10.8%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	328	4.7%	39	1567
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	322	12.2%	34	861
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	pmol/L	312	9.4%	29	512
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	246	11.9%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	345	6.6%	18	367
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	283	8.9%	16	246
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	306	10.3%	11	275
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	336	3.1%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	346	0%	<5	<100
(EPO)促红细胞生成素					
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	6.53	18.7%	<5	115
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/mL	5.88	7.6%	18	356
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	6.76	9.9%	<5	<100
(FA)叶酸					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.07	15.9%	31	648
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.66	11.9%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.54	18.9%	19	324
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.84	7.4%	18	346
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.36	16.2%	12	345
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.16	10.5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.3	16.6%	<5	<100
(Ferritin)铁蛋白					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	103	5.5%	5	<100

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
5 / 11								

		批号：3382863			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
maglumi system					
Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	91.3	12.5%	5	<100
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	106	7.6%	34	597
YHLO iFlash Analyzer					
YHL \ Chemiluminescence Immunoassay (CLIA)	ng/mL	76.6	8.2%	<5	<100
Roche cobas 4000/6000/8000					
Roche \ Immunoturbidimetry (ITA)	ng/mL	119	9.9%	<5	<100
Tosoh AIA 900/1800					
TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	80.5	11.9%	<5	<100
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	107	2.9%	<5	<100
Beckman UniCel Dxl 600/800					
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	64.4	9.7%	18	377
DiaSorin Liaison /XL					
Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	95.9	12%	<5	<100
Siemens Immulite1000/2000					
Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	69.2	9.8%	<5	<100
Beckman UniCel Dxl 600/800					
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	68.1	9.4%	<5	<100
(F-PSA)游离前列腺特异性抗原					
maglumi system					
Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.69	10.6%	5	<100
Roche cobas E601/602					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.83	3.9%	<5	<100
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.74	8.2%	38	740
YHLO iFlash Analyzer					
YHL \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.95	3.5%	<5	<100
Abbott ARCHITECT ci series					
Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.26	1.7%	24	432
Beckman UniCel Dxl 600/800					
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.06	11.9%	<5	<100
(FSH)促卵泡刺激素					
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.54	9.1%	<5	<100
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.32	2.9%	39	1002
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.33	3.8%	34	809
Siemens ADVIA Centaur XP/XPT					
Siemens \ OTHER	IU/L	6.49	8.1%	29	519
Tosoh AIA 900/1800					
TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.92	11.9%	<5	<100
Mindray CL-2000i/1000i/6000i series					
Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/L	6.28	7.8%	19	326
Beckman UniCel Dxl 600/800					
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	5.36	5.6%	18	376
Beckman UniCel Dxl 600/800					
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.32	11.3%	11	277
Abbott ARCHITECT i series					
Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.88	9.8%	<5	<100
Roche cobas E601/602					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.96	1.5%	<5	<100
Beckman UniCel Dxl 600/800					
Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.6	0%	<5	<100
(FT3)游离T3					
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	5.32	4.1%	6	167
Siemens ADVIA Centaur XP/XPT					
Siemens \ OTHER	pmol/L	6.59	4%	58	1041
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.53	3.9%	38	671
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.5	7.9%	29	528
Abbott ARCHITECT i series					
Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.5	3.9%	25	461
Beckman UniCel Dxl 600/800					
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	4.97	7.8%	20	472
Tosoh AIA 900/1800					
TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.18	11.9%	<5	<100
Mindray CL-2000i/1000i/6000i series					
Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.3	4%	19	333
Roche cobas E601/602					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.77	1.8%	<5	<100
(FT4)游离T4					
Mindray CL-2000i/1000i/6000i series					
Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	17.7	5%	6	<100
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	18.3	4.2%	6	166

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
6 / 11								

		批号：3382863			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	pmol/L	14.5	6.7%	57	1060
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	16.8	3.8%	40	678
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	16.4	7.7%	29	493
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.5	3.3%	25	467
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.2	7.8%	20	485
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	14.1	11.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	18.2	3.9%	<5	<100
(hCG)人绒毛膜促性腺激素					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	3621	3.4%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.22	11.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	3.01	4.8%	17	511
(HCY)同型半胱氨酸					
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	13.9	9.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	μmol/L	12.1	9.9%	<5	<100
TOSHIBA TBA Bio-assay \ Enzymatic Cycling Assays Method	μmol/L	12.1	5.8%	<5	<100
(HE4)人附睾蛋白4					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	62.1	3.1%	19	375
(HGH)人生长激素					
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μg/L	1.51	11.9%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	1.99	9.9%	<5	<100
(HS-CTnI) (高) 超敏肌钙蛋白I					
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.061	11.9%	<5	<100
(IgE)免疫球蛋白 IgE					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	191	3.1%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	160	11.9%	<5	<100
Beckman Image 800 Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	131	9.9%	<5	<100
(IGF1)胰岛素生长因子1					
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	nmol/L	17.7	11.9%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	218	9.6%	<5	<100
(INS)胰岛素					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.93	4.8%	5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.74	5.3%	38	824
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.66	10.1%	31	653
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.38	3.3%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.61	4%	24	731
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	6.95	11.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.36	9.9%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.83	10%	<5	<100
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.08	16.2%	<5	<100
(LH)促黄体生成素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.64	8.6%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.45	3.3%	39	1017

质控品名： 复合免疫分析非定值质控品

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
7 / 11								

		批号：3382863			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000	IU/L	5.54	5.8%	34	803
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Siemens ADVIA Centaur XP/XPT	IU/L	4.07	6.1%	29	520
Siemens \ OTHER					
Tosoh AIA 900/1800	IU/L	3.79	11.9%	<5	<100
TOSOH \ Chemiluminescence Immunoassay (CLIA)					
Beckman UniCel Dxl 600/800	IU/L	3.76	5.8%	18	361
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)					
Mindray CL-2000i/1000i/6000i series	IU/L	5.35	11%	18	310
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Beckman UniCel Dxl 600/800	IU/L	3.7	7.5%	11	259
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	IU/L	4.66	11.6%	<5	<100
Dedicated \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas E601/602	IU/L	5.48	1.2%	<5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Beckman UniCel Dxl 600/800	IU/L	3.79	0%	<5	<100
Dedicated \ Chemiluminescence Immunoassay (CLIA)					
(Myo / Mb)肌红蛋白					
J&J Vitros ECL	μg/L	40.2	3.7%	8	<100
Ortho \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas e411	μg/L	29.1	9.9%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Ortho Vitros 3600/5600	μg/L	41.5	3.1%	<5	<100
Dedicated \ Chemiluminescence Immunoassay (CLIA)					
Beckman UniCel Dxl 600/800	μg/L	26.7	11.9%	<5	118
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas e 801	μg/L	26.9	6.1%	25	566
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Tosoh AIA 900/1800	μg/L	34.2	11.9%	<5	<100
TOSOH \ Chemiluminescence Immunoassay (CLIA)					
Ortho Vitros 3600/5600	μg/L	41.3	4.9%	17	404
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)					
Beckman UniCel Dxl 600/800	μg/L	27.8	6.6%	12	316
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
TOSHIBA TBA	μg/L	42	0%	<5	<100
OTHER \ OTHER					
(NT-proBNP)N-末端脑钠肽前体					
Roche cobas E601/602	pg/mL	172	4.6%	5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas E601/602	pmol/L	20.2	4.6%	5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Ortho Vitros 3600/5600	pg/mL	188	37.9%	<5	<100
Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Ortho Vitros 3600/5600	pmol/L	22.1	37.9%	<5	<100
Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas e411	pg/mL	202	9.9%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas e411	pmol/L	23.8	9.9%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas e 801	pg/mL	191	5.7%	25	570
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas e 801	pmol/L	22.4	5.7%	25	570
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas e411	pg/mL	201	7.8%	17	470
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas e411	pmol/L	23.6	7.8%	17	470
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Beckman UniCel Dxl 600/800	pg/mL	212	3.8%	13	358
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Beckman UniCel Dxl 600/800	pmol/L	25	3.8%	13	358
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(PCT)降钙素原					
Roche cobas c 501/502/701/702	ng/mL	0.469	5%	9	213
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas E601/602	ng/mL	0.521	5.5%	5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(PRL)催/ 泌乳素					
Roche cobas E601/602	ng/mL	8.32	14.1%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas 4000/6000/8000	ng/mL	8.32	3.2%	39	999
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas 4000/6000/8000	ng/mL	8.66	4.9%	34	809
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Siemens ADVIA Centaur XP/XPT	ng/mL	6.38	4.4%	29	518
Siemens \ OTHER					
Tosoh AIA 900/1800	ng/mL	9.01	11.9%	<5	<100
TOSOH \ Chemiluminescence Immunoassay (CLIA)					
Beckman UniCel Dxl 600/800	ng/mL	6.4	6.4%	18	350
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)					

质控品名： 复合免疫分析非定值质控品

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
8 / 11								

		批号：3382863			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.1	8.3%	18	272
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.25	7.4%	11	296
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.34	2.4%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.25	0%	<5	<100
(PROG)孕酮					
Ortho Vitros 3600/5600 Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.96	4%	7	173
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.953	13.4%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.934	10.7%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.847	9.2%	39	1481
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	ng/mL	0.995	14%	29	524
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.22	11.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.01	11.9%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.16	13.3%	19	320
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.29	14.9%	18	358
Beckman UniCel Dxl 600/800 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.813	9.7%	12	359
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.945	45.6%	11	145
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.39	10.8%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.939	4.6%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.71	0%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.24	0%	<5	<100
(PSA)前列腺特异性抗原					
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.47	5.2%	9	120
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.41	11.1%	5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.57	2.7%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.82	2.6%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.56	6.4%	35	688
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.99	1.3%	24	433
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.82	11.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.08	11.9%	<5	<100
YHLO iFlash Analyzer YHLO \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.89	7.6%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.71	10.4%	<5	<100
(PTH)甲状旁腺素					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	2.01	7.3%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1.83	9.8%	27	547
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	2.18	11.6%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1.94	11.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	2.84	3.6%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	2.01	6.9%	18	361
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1.96	9.7%	<5	<100
(SHBG)性激素结合球蛋白					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	nmol/L	33.2	12%	<5	<100

质控品名： 复合免疫分析非定值质控品

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
9 / 11								

		批号：3382863			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TES)睾酮					
Roche cobas E601/602	ng/mL	5.95	11.6%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas 4000/6000/8000	ng/mL	5.87	3.5%	39	993
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas 4000/6000/8000	ng/mL	6.24	7.2%	33	798
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Tosoh AIA 900/1800	ng/mL	6.61	11.9%	<5	<100
TOSOH \ Chemiluminescence Immunoassay (CLIA)					
Beckman UniCel Dxl 600/800	ng/mL	5.35	3.9%	18	362
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)					
Mindray CL-2000i/1000i/6000i series	ng/mL	7.16	7.3%	16	222
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Siemens ADVIA Centaur XP/XPT	ng/mL	5.68	11.6%	15	284
Siemens \ OTHER					
Beckman UniCel Dxl 600/800	ng/mL	5.18	7%	11	263
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	ng/mL	7.01	10.7%	<5	<100
Abbott \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas E601/602	ng/mL	6.35	2%	<5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Beckman UniCel Dxl 600/800	ng/mL	5.32	0%	<5	<100
Dedicated \ Chemiluminescence Immunoassay (CLIA)					
(Tg)甲状腺球蛋白					
Roche cobas 4000/6000/8000	ng/mL	10.6	4.4%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas E601/602	ng/mL	10.7	9.3%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Beckman UniCel Dxl 600/800	ng/mL	8.01	5.9%	19	428
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas 4000/6000/8000	ng/mL	10	2.7%	16	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Abbott ARCHITECT i series	ng/mL	9.88	9.9%	<5	<100
Abbott \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas 4000/6000/8000	ng/mL	9.1	0%	<5	<100
OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(THEO)茶碱					
Agilent	μg/mL	8.7	11.9%	<5	<100
agilent \ MS					
(TnI)肌钙蛋白I					
J&J Vitros ECL	ng/mL	0.076	5.8%	8	<100
Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Beckman UniCel Dxl 600/800	ng/mL	0.066	26.1%	<5	114
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)					
Ortho Vitros 3600/5600	ng/mL	0.073	11.6%	<5	<100
Dedicated \ Chemiluminescence Immunoassay (CLIA)					
Ortho Vitros 3600/5600	ng/mL	0.071	7%	17	404
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)					
(TnT)肌钙蛋白T					
Roche cobas e411	ng/mL	0.02	10.1%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas e 801	ng/mL	0.02	4.6%	25	572
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas E601/602	ng/mL	0.02	7.3%	<5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas e411	ng/mL	0.018	10.2%	17	470
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Beckman UniCel Dxl 600/800	ng/mL	0.018	5.7%	12	331
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(TSH)促甲状腺素					
Roche cobas 4000/6000/8000	mIU/L	3.5	5.6%	6	163
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Siemens ADVIA Centaur XP/XPT	mIU/L	2.88	4.7%	58	1033
Siemens \ OTHER					
Roche cobas 4000/6000/8000	mIU/L	3.41	3.8%	56	1291
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas E601/602	mIU/L	3.49	7.3%	37	642
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Abbott ARCHITECT i series	mIU/L	2.64	4.3%	25	457
Abbott \ Chemiluminescence Immunoassay (CLIA)					
Tosoh AIA 900/1800	mIU/L	2.87	11.9%	<5	<100
TOSOH \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas E601/602	mIU/L	3.39	2%	<5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Beckman UniCel Dxl 600/800	mIU/L	2.92	6%	19	458
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)					
Mindray CL-2000i/1000i/6000i series	mIU/L	3.59	6.1%	19	328
Mindray \ Chemiluminescence Immunoassay (CLIA)					

质控品名： 复合免疫分析非定值质控品

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
10 / 11								

批号：3382863					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TT3)总三碘甲状原氨酸					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.26	3.9%	6	169
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	ng/mL	1.42	10.8%	58	1033
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.39	9.5%	28	490
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.1	9.9%	20	494
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.34	3.3%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.38	5.5%	19	198
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.24	6.1%	18	327
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.56	9.6%	<5	<100
(TT4)总T4					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	63.1	3.1%	6	167
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	ng/mL	67.2	12.3%	56	1045
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	65.6	11.1%	36	583
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	58.3	12.2%	20	496
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	63.1	5.3%	20	209
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	69	7.1%	18	332
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	52.7	11.8%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	71.8	2%	<5	<100
(VB12)维生素B12					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	458	9.7%	31	625
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	pg/mL	310	7.4%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pg/mL	509	11.9%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pg/mL	476	7%	19	320
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pg/mL	303	5.7%	18	352
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pg/mL	299	11.2%	11	299
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	649	2.5%	<5	<100
(β2M)β-2-微球蛋白					
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	ng/mL	1020	9.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1170	11.9%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	ng/mL	1131	4.5%	11	255
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	888	16.2%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Ortho Vitros 3600/5600 Ortho \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.007	4%	6	114
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.003	39%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	13.1%	30	355
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.004	16.4%	<5	<100
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.006	7.7%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	IU/mL	0.004	21.7%	24	453
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.003	11.9%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	0.004	4.5%	19	408
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.003	17.4%	19	305
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	5.8%	19	195

REF	540-1	6×3mL	EXP	2024-03-31	LOT	3382863	更新日期：	2024-04
11 / 11								

		批号：3382863			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	10.9%	17	466
Beckman UniCel DxI 600/800 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	6.4%	13	263
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	10.6%	12	233
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.004	0%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。