

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
1 / 16								

		批号：376DF15			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(a-HBDH)α-羟丁酸脱氢酶					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	279	3.9%	9	140
Siemens ADVIA 1800/2400 Szybio \ DGKC	U/L	284	2.9%	9	101
KHB ZY series KHB \ DGKC	U/L	312	1.9%	8	<100
KHB ZY series KeFang(Guangzhou) \ OTHER	U/L	315	4.4%	8	190
TOSHIBA TBA chongqiong zybjo \ Enzymatic Cycling Assays Method	U/L	292	2.9%	6	145
Roche Hitachi/Hitachi Systems OTHER \ DGKC	U/L	305	3.6%	6	129
Mindray BS series MAKER \ Enzymatic Cycling Assays Method	U/L	273	4%	<5	<100
Mindray BS series Mindray \ OTHER	U/L	314	4.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	U/L	296	3.1%	29	821
Beckman AU series MAKER \ Enzymatic Cycling Assays Method	U/L	268	5.4%	22	607
Beckman AU series Dedicated \ Enzymatic Cycling Assays Method	U/L	298	6.4%	21	930
Beckman AU series Dedicated \ immunoturbidimetric	U/L	293	6.7%	20	910
Beckman AU series HuaChenBio \ DGKC	U/L	310	2.7%	20	266
Beckman AU series Dedicated \ OTHER	U/L	260	2.4%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Enzymatic Cycling Assays Method	U/L	295	2.5%	19	429
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic Cycling Assays Method	U/L	292	2.1%	15	669
Beckman AU series DiaSys \ Enzymatic Cycling Assays Method	U/L	299	2.7%	14	416
Beckman AU series Reebio \ OTHER	U/L	322	2.2%	11	515
Roche Hitachi/Hitachi Systems chongqiong zybjo \ Enzymatic Cycling Assays Method	U/L	293	3.8%	11	290
(Alb)白蛋白					
Beckman AU series KeFang(Guangzhou) \ Bromocresol green	g/L	57.8	1%	8	184
KHB ZY series KHB \ Bromocresol green	g/L	63.7	2.2%	7	<100
Beckman AU series Dedicated \ Bromocresol green	g/L	58.6	1.7%	62	2506
TOSHIBA TBA chongqiong zybjo \ Bromocresol green	g/L	56.5	2.1%	6	144
Roche cobas 4000/6000/8000 Roche \ Bromocresol green	g/L	58.8	10.6%	41	1053
Mindray BS series MAKER \ Bromocresol green	g/L	62.9	1.2%	<5	<100
TOSHIBA TBA MedicalSystem \ Bromocresol green	g/L	43.9	1.6%	<5	<100
Mindray BS series Mindray \ Bromocresol green	g/L	54.5	12.8%	<5	126
Siemens ADVIA 1800/2400 BIOSINO \ Bromocresol green	g/L	59.2	1.8%	22	977
Beckman AU series MAKER \ Bromocresol green	g/L	61.5	3%	22	607
Beckman AU series HuaChenBio \ Bromocresol green	g/L	57.2	1.6%	20	267
Mindray BS series Mindray \ OTHER	g/L	54	1.2%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Bromocresol green	g/L	58.9	0.8%	19	427
Roche Hitachi/Hitachi Systems BIOSINO \ Bromocresol green	g/L	48.4	14%	18	435
Roche cobas c 501/502/701/702 Dedicated \ Bromocresol green	g/L	61.2	1.2%	16	455
Beckman AU series DiaSys \ Bromocresol green	g/L	62.2	1.6%	14	418
OTHER SHLY \ Bromocresol green	g/L	60.9	1.7%	14	324
Siemens ADVIA 1800/2400 Siemens \ Bromocresol green	g/L	61.6	4.8%	11	349
Roche Hitachi/Hitachi Systems chongqiong zybjo \ Bromocresol green	g/L	56.7	2.2%	11	294
(ALP)碱性磷酸酶					
Beckman AU series KeFang(Guangzhou) \ Rate -AMP buffer	U/L	313	1.9%	8	184
KHB ZY series KHB \ Rate -AMP buffer	U/L	327	3%	7	<100
Beckman AU series Dedicated \ Rate -AMP buffer	U/L	326	3%	60	2417
TOSHIBA TBA chongqiong zybjo \ Rate -AMP buffer	U/L	288	2.9%	6	147
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	329	25%	42	1017
TOSHIBA TBA MedicalSystem \ Rate -AMP buffer	U/L	490	1.3%	<5	<100

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
2 / 16								

批号：376DF15					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series	U/L	278	1.9%	<5	<100
Mindray \ OTHER					
Beckman AU series	U/L	304	5.4%	25	344
HuaChenBio \ Rate -AMP buffer					
Siemens ADVIA 1800/2400	U/L	305	2.6%	22	979
BIOSINO \ Rate -AMP buffer					
Beckman AU series	U/L	321	4.5%	22	609
MAKER \ Rate -AMP buffer					
Beckman AU series	U/L	346	2.4%	<5	<100
Dedicated \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	289	3.2%	<5	<100
ABX \ dry chemistry					
Abbott ARCHITECT ci series	U/L	276	3.8%	19	427
BIOSINO \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	124	60.2%	17	406
BIOSINO \ Rate -AMP buffer					
Roche cobas c 501/502/701/702	U/L	260	3.9%	16	463
Dedicated \ Rate -AMP buffer					
Beckman AU series	U/L	275	3.8%	14	419
DiaSys \ Rate -AMP buffer					
OTHER					
SHLY \ Rate -AMP buffer	U/L	297	3.2%	14	325
Siemens ADVIA 1800/2400					
Siemens \ Rate -AMP buffer	U/L	274	5.7%	11	319
Roche Hitachi/Hitachi Systems					
chongqiong zybio \ Rate -AMP buffer	U/L	318	2.5%	11	280

(ALT/ SGPT)丙氨酸氨基转移酶

Roche Hitachi/Hitachi Systems	U/L	39.8	3.1%	9	325
BIOSINO \ UV-rate without P5P					
Beckman AU series	U/L	151	2.3%	8	179
KeFang(Guangzhou) \ OTHER					
Siemens ADVIA 1800/2400	U/L	150	2.6%	7	307
BIOSINO \ UV-rate without P5P					
TOSHIBA TBA	U/L	161	3.6%	7	141
chongqiong zybio \ UV-rate without P5P					
Beckman AU series	U/L	145	1.9%	59	2349
Dedicated \ UV-rate without P5P					
KHB ZY series	U/L	148	2%	<5	<100
KHB \ UV-rate with P5P					
Roche cobas 4000/6000/8000	U/L	144	5.9%	32	868
Roche \ UV-rate without P5P					
TOSHIBA TBA	U/L	189	2%	<5	<100
MedicalSystem \ OTHER					
Mindray BS series	U/L	148	5%	<5	<100
Mindray \ OTHER					
Beckman AU series	U/L	143	4.8%	25	377
HuaChenBio \ OTHER					
Beckman AU series	U/L	157	3.7%	22	606
MAKER \ UV-rate without P5P					
Beckman AU series	U/L	146	1.4%	<5	<100
Dedicated \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	149	13.1%	<5	<100
ABX \ dry chemistry					
Abbott ARCHITECT ci series	U/L	145	1.7%	19	431
BIOSINO \ UV-rate with P5P					
Roche cobas c 501/502/701/702	U/L	135	1.8%	16	459
Dedicated \ UV-rate without P5P					
Beckman AU series	U/L	143	1.3%	14	416
DiaSys \ UV-rate with P5P					
OTHER					
SHLY \ UV-rate without P5P	U/L	131	2%	14	325
Roche Hitachi/Hitachi Systems	U/L	174	4.8%	11	298
chongqiong zybio \ UV-rate without P5P					
Roche cobas 4000/6000/8000	U/L	167	11.3%	11	172
Dedicated \ UV-rate with P5P					

(AMY)淀粉酶

Beckman AU series	U/L	336	2.4%	8	139
KeFang(Guangzhou) \ CNP_G3					
KHB ZY series	U/L	386	1.6%	6	<100
KHB \ CNP_G3					
TOSHIBA TBA	U/L	364	2.3%	6	146
chongqiong zybio \ PNP-G7					
Mindray BS series	U/L	375	1%	<5	<100
MAKER \ PNP-G7					
Roche cobas 4000/6000/8000	U/L	357	2.3%	28	747
Roche \ Iodine colorimetry					
Beckman AU series	U/L	392	2%	24	307
HuaChenBio \ OTHER					
Beckman AU series	U/L	379	1.8%	22	606
MAKER \ CNP_G3					
Beckman AU series	U/L	379	1.5%	<5	<100
Dedicated \ OTHER					
Roche cobas 4000/6000/8000	U/L	376	5.6%	<5	<100
Roche \ PNP-G5					
Roche Hitachi/Hitachi Systems	U/L	348	0.7%	<5	<100
ABX \ Iodine colorimetry					
Mindray BS series	U/L	360	1.7%	<5	<100
Mindray \ OTHER					
Beckman AU series	U/L	323	1.4%	14	419
DiaSys \ PNP-G7					
Siemens ADVIA 1800/2400	U/L	364	4.1%	11	314
Siemens \ PNP-G7					

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
3 / 16								

		批号：376DF15			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems chongqiong zybjo \ PNP-G7	U/L	354	2.6%	11	295
Roche cobas 4000/6000/8000 Roche \ PNP-G7	U/L	362	10.2%	<5	<100
Beckman AU series Beckman Coulter \ Iodine colorimetry	U/L	365	10%	<5	<100
Mindray BS series Mindray \ PNP-G7	U/L	404	2.6%	<5	<100
Beckman AU series Dedicated \ Iodine colorimetry	U/L	370	2.1%	<5	<100
(Apo-A I)载脂蛋白A I					
Roche Hitachi/Hitachi Systems BIOSINO \ immunoturbidimetric	g/L	2.26	3.1%	9	<100
Beckman AU series KeFang(Guangzhou) \ immunoturbidimetric	g/L	2.42	2.1%	8	168
KHB ZY series KHB \ immunoturbidimetric	g/L	2.73	4%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ immunoturbidimetric	g/L	2.3	2.4%	5	<100
Beckman AU series DongOu \ immunoturbidimetric	g/L	2.44	4.8%	40	1795
Roche cobas 4000/6000/8000 Roche \ immunoturbidimetric	g/L	2.55	8%	32	856
TOSHIBA TBA MedicalSystem \ immunoturbidimetric	g/L	1.87	1.4%	<5	<100
Beckman AU series Dedicated \ immunoturbidimetric	g/L	2.5	3.5%	<5	105
Siemens ADVIA 1800/2400 BIOSINO \ immunoturbidimetric	g/L	2.49	3.3%	22	977
Beckman AU series HuaChenBio \ OTHER	g/L	2.55	2.1%	22	264
Mindray BS series Mindray \ immunoturbidimetric	g/L	2.37	4.8%	<5	<100
Siemens ADVIA Centaur CP OTHER \ OTHER	g/L	2.97	2.1%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ immunoturbidimetric	g/L	2.41	2.6%	16	461
Beckman AU series Reebio \ immunoturbidimetric	g/L	2.59	8.4%	11	511
Siemens ADVIA 1800/2400 Siemens \ OTHER	g/L	2.35	3.7%	10	283
Roche cobas 4000/6000/8000 SekiSui \ immunoturbidimetric	g/L	1.89	9.5%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ immunoturbidimetric	g/L	2.31	15%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ immunoturbidimetric	g/L	2.31	3.7%	<5	<100
(Apo-B)载脂蛋白B					
Roche Hitachi/Hitachi Systems BIOSINO \ immunoturbidimetric	g/L	1.06	2.2%	9	<100
Beckman AU series KeFang(Guangzhou) \ immunoturbidimetric	g/L	1.22	2.4%	8	146
KHB ZY series KHB \ immunoturbidimetric	g/L	1.03	2.4%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ immunoturbidimetric	g/L	1.04	4.8%	6	126
Beckman AU series DongOu \ immunoturbidimetric	g/L	1.06	3.4%	40	1823
Roche cobas 4000/6000/8000 Roche \ immunoturbidimetric	g/L	1.24	3.8%	32	869
TOSHIBA TBA MedicalSystem \ immunoturbidimetric	g/L	0.89	1.4%	<5	<100
Beckman AU series Dedicated \ immunoturbidimetric	g/L	1.08	3%	<5	102
Beckman AU series HuaChenBio \ OTHER	g/L	1.17	4.3%	26	372
Siemens ADVIA 1800/2400 BIOSINO \ immunoturbidimetric	g/L	1.12	4.2%	22	979
Mindray BS series Mindray \ immunoturbidimetric	g/L	1.14	1.2%	<5	<100
Siemens ADVIA Centaur CP OTHER \ OTHER	g/L	1.01	6.3%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ immunoturbidimetric	g/L	1.06	2.4%	16	449
Beckman AU series Reebio \ immunoturbidimetric	g/L	1.27	11.3%	11	503
Siemens ADVIA 1800/2400 Siemens \ OTHER	g/L	1.09	4.2%	10	299
Siemens ADVIA 1800/2400 Siemens \ immunoturbidimetric	g/L	1.09	15%	<5	<100
Roche cobas 4000/6000/8000 SekiSui \ immunoturbidimetric	g/L	1.07	11.2%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ immunoturbidimetric	g/L	1.1	3.4%	<5	<100
(AST/ SGOT)天门冬氨酸氨基转移酶					
Roche Hitachi/Hitachi Systems BIOSINO \ UV-rate with P5P	U/L	36.3	4.2%	9	326
TOSHIBA TBA chongqiong zybjo \ UV-rate without P5P	U/L	223	3.4%	7	146
Beckman AU series KeFang(Guangzhou) \ OTHER	U/L	239	2.4%	7	108

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
4 / 16								

批号：376DF15					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
KHB ZY series	U/L	225	2.8%	6	<100
KHB \ UV-rate with P5P					
Beckman AU series	U/L	225	4%	59	2407
Dedicated \ UV-rate without P5P					
Roche cobas 4000/6000/8000	U/L	224	4.6%	32	862
Roche \ UV-rate without P5P					
Mindray BS series	U/L	222	1.8%	<5	<100
MAKER \ OTHER					
TOSHIBA TBA	U/L	243	1.8%	<5	<100
MedicalSystem \ OTHER					
Mindray BS series	U/L	201	4.3%	<5	<100
Mindray \ OTHER					
Siemens ADVIA 1800/2400	U/L	206	4.1%	22	979
BIOSINO \ UV-rate without P5P					
Beckman AU series	U/L	219	3.2%	22	607
MAKER \ UV-rate without P5P					
Beckman AU series	U/L	234	2.2%	<5	<100
Dedicated \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	214	2.3%	<5	<100
ABX \ dry chemistry					
Abbott ARCHITECT ci series	U/L	188	3.1%	19	431
BIOSINO \ UV-rate without P5P					
Beckman AU series	U/L	204	4.6%	19	187
HuaChenBio \ OTHER					
Roche cobas c 501/502/701/702	U/L	199	2.1%	16	453
Dedicated \ UV-rate without P5P					
Beckman AU series	U/L	207	1.8%	14	418
DiaSys \ UV-rate with P5P					
OTHER	U/L	223	2.7%	14	325
SHLY \ UV-rate without P5P					
Roche Hitachi/Hitachi Systems	U/L	237	5.7%	11	292
chongqiong zybio \ UV-rate without P5P					

(Ca)钙					
Roche Hitachi/Hitachi Systems	mmol/L	3.06	6%	8	<100
BIOSINO \ OCPC colorimetry					
Siemens ADVIA 1800/2400	mmol/L	3.05	1.7%	8	358
BIOSINO \ OCPC colorimetry					
Siemens ADVIA 1800/2400	mmol/L	3.06	1.9%	7	306
BIOSINO \ Arsenazo III					
Beckman AU series	mmol/L	3.21	1.3%	62	2502
Dedicated \ Arsenazo III					
Roche cobas 4000/6000/8000	mmol/L	3.23	4.9%	32	865
Roche \ ISE (indirect)					
Mindray BS series	mmol/L	3.11	1.5%	<5	<100
MAKER \ Arsenazo III					
TOSHIBA TBA	mmol/L	2.72	1.5%	<5	<100
MedicalSystem \ Arsenazo III					
Beckman AU series	mmol/L	3.17	3.8%	22	606
MAKER \ Arsenazo III					
Mindray BS series	mmol/L	3.17	1%	<5	<100
Mindray \ OTHER					
Siemens ADVIA Centaur CP	mmol/L	3.03	2%	<5	<100
Dedicated \ Arsenazo III					
Abbott ARCHITECT ci series	mmol/L	3.11	2.7%	19	430
BIOSINO \ OCPC colorimetry					
Roche cobas c 501/502/701/702	mmol/L	3.18	1.4%	16	456
Dedicated \ BAPTA					
Beckman AU series	mmol/L	3.24	2.1%	14	418
DiaSys \ Arsenazo III					
Roche cobas 4000/6000/8000	mmol/L	2.69	4%	11	180
Dedicated \ BAPTA					
Siemens ADVIA 1800/2400	mmol/L	3.17	5%	10	295
Siemens \ Arsenazo III					
Beckman AU series	mmol/L	3.12	9.9%	<5	<100
Beckman Coulter \ BAPTA					
Roche cobas 4000/6000/8000	mmol/L	3.17	10.1%	<5	<100
Roche \ BAPTA					

(CHE)胆碱酯酶					
Roche Hitachi/Hitachi Systems	U/L	10175	3%	9	123
BIOSINO \ Butyrylthiocholine Method					
Beckman AU series	U/L	10822	1.9%	8	178
KeFang(Guangzhou) \ OTHER					
KHB ZY series	U/L	11222	1.6%	7	<100
KHB \ Butyrylthiocholine Method					
TOSHIBA TBA	U/L	9804	2%	6	147
chongqiong zybio \ Butyrylthiocholine Method					
Roche Hitachi/Hitachi Systems	U/L	9.5	1.7%	6	135
OTHER \ DTNB					
Roche cobas 4000/6000/8000	U/L	8477	8.7%	5	120
Roche \ DTNB					
Roche cobas 4000/6000/8000	U/L	10548	9.9%	33	885
Roche \ Butyrylthiocholine Method					
Beckman AU series	U/L	9963	1.7%	<5	<100
Dedicated \ Butyrylthiocholine Method					
Mindray BS series	U/L	10024	1.4%	<5	<100
MAKER \ Butyrylthiocholine Method					
TOSHIBA TBA	U/L	8302	1.2%	<5	<100
MedicalSystem \ Butyrylthiocholine Method					
Mindray BS series	U/L	9831	2.7%	<5	<100
Mindray \ OTHER					
Mindray BS series	U/L	10555	12.9%	<5	124
Mindray \ Butyrylthiocholine Method					

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
5 / 16								

批号：376DF15					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems chongqiong zybio \ Butyrylthiocholine Method	U/L	9738	2.8%	11	293
Beckman AU series HuaChenBio \ Butyrylthiocholine Method	U/L	9964	2.5%	10	<100
Siemens ADVIA 1800/2400 Siemens \ Butyrylthiocholine Method	U/L	12767	5.2%	10	283
Siemens ADVIA 1800/2400 Siemens \ DTNB	U/L	11890	15%	<5	<100
Beckman AU series Beckman Coulter \ Butyrylthiocholine Method	U/L	12059	15%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ DTNB	U/L	11890	1.7%	<5	<100
Roche Hitachi/Hitachi Systems ABX \ DTNB	U/L	11118	0.7%	<5	<100

(CHOL/ TC)总胆固醇					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	5.74	2%	9	213
Beckman AU series KeFang(Guangzhou) \ Cholesterol oxidase	mmol/L	6.14	1.3%	8	189
TOSHIBA TBA chongqiong zybio \ Cholesterol oxidase	mmol/L	6.02	2.5%	6	139
KHB ZY series KHB \ Cholesterol oxidase	mmol/L	6.02	2.1%	5	<100
Beckman AU series Dedicated \ Cholesterol oxidase	mmol/L	5.97	2%	43	1924
TOSHIBA TBA MedicalSystem \ Cholesterol oxidase	mmol/L	5.28	1.2%	<5	<100
Mindray BS series Mindray \ Cholesterol oxidase	mmol/L	5.33	19.8%	<5	118
Beckman AU series HuaChenBio \ Cholesterol oxidase	mmol/L	6.02	2.4%	26	345
Siemens ADVIA 1800/2400 BIOSINO \ Cholesterol oxidase	mmol/L	5.88	1.9%	22	975
Beckman AU series MAKER \ OTHER	mmol/L	6.06	6%	22	607
Mindray BS series Mindray \ OTHER	mmol/L	5.71	2.3%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Cholesterol oxidase	mmol/L	5.6	2.1%	19	430
Roche cobas c 501/502/701/702 Dedicated \ Cholesterol oxidase	mmol/L	5.98	1.7%	16	452
Beckman AU series DiaSys \ Cholesterol oxidase	mmol/L	6.16	1.3%	14	418
OTHER SHLY \ Cholesterol oxidase	mmol/L	5.51	2.3%	14	325
Siemens ADVIA 1800/2400 Siemens \ Cholesterol oxidase	mmol/L	6.3	5.1%	11	321
Roche Hitachi/Hitachi Systems chongqiong zybio \ Cholesterol oxidase	mmol/L	6.07	2.5%	11	284
Roche cobas 4000/6000/8000 Dedicated \ Cholesterol oxidase	mmol/L	5.33	4.5%	11	203
Roche cobas 4000/6000/8000 Roche \ Cholesterol oxidase	mmol/L	5.79	9.8%	<5	<100

(CK)肌酸激酶					
KHB ZY series KHB \ OTHER	U/L	310	3.3%	8	<100
Siemens ADVIA 1800/2400 BIOSINO \ Rate -AMP buffer	U/L	302	4.4%	8	356
Siemens ADVIA 1800/2400 BIOSINO \ Urease-UV-rate	U/L	291	4.3%	7	306
Beckman AU series KeFang(Guangzhou) \ Rate -other buffers	U/L	210	11.2%	7	102
TOSHIBA TBA chongqiong zybio \ Rate -AMP buffer	U/L	277	6.6%	6	146
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	381	39%	40	884
Beckman AU series WEGO \ Rate -AMP buffer	U/L	288	3.7%	40	1755
TOSHIBA TBA MedicalSystem \ Rate -AMP buffer	U/L	647	2.2%	<5	<100
Mindray BS series MAKER \ OTHER	U/L	268	2.8%	<5	<100
Mindray BS series Mindray \ OTHER	U/L	267	4.8%	<5	<100
Beckman AU series MAKER \ Rate -other buffers	U/L	276	7.1%	22	607
Beckman AU series HuaChenBio \ OTHER	U/L	283	6%	22	253
Beckman AU series Dedicated \ OTHER	U/L	267	4.3%	<5	<100
Roche Hitachi/Hitachi Systems ABX \ dry chemistry	U/L	261	1.7%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Rate -AMP buffer	U/L	227	7.2%	19	435
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	132	62.5%	17	431
Beckman AU series DiaSys \ Rate -AMP buffer	U/L	293	3.2%	14	417
OTHER SHLY \ Rate -AMP buffer	U/L	287	3.1%	14	325
Siemens ADVIA 1800/2400 Siemens \ Rate -AMP buffer	U/L	318	5.4%	11	305

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
6 / 16								

		批号：376DF15			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(CK-MB)肌酸激酶-MB (活性)					
Siemens ADVIA 1800/2400 Szybio \ Immune suppression	U/L	46.1	3.1%	9	128
Beckman AU series KeFang(Guangzhou) \ Immune suppression	U/L	35.9	10.7%	8	142
Roche Hitachi/Hitachi Systems OTHER \ DGKC	U/L	42.7	4.8%	6	129
KHB ZY series KHB \ Immune suppression	U/L	55.5	4.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ DGKC	U/L	49.9	7.7%	32	878
Mindray BS series MAKER \ Enzyme immunoassay (EIA)	U/L	50.2	3.8%	<5	<100
Beckman AU series Dedicated \ Immune suppression	U/L	38.7	14.6%	23	908
Roche Hitachi/Hitachi Systems ABX \ DGKC	U/L	52.5	2.8%	<5	<100
Mindray BS series Mindray \ DGKC	U/L	46.8	4.8%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ DGKC	U/L	51.9	7%	15	658
Beckman AU series Reebio \ Immune suppression	U/L	50.4	4.4%	11	525
Roche cobas 4000/6000/8000 Roche \ Immune suppression	U/L	49.9	10%	<5	<100
Beckman AU series Beckman Coulter \ Immune suppression	U/L	55.8	10%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ DGKC	U/L	44.6	15%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ DGKC	U/L	44.6	2.8%	<5	<100
Siemens ADVIA Centaur CP OTHER \ OTHER	U/L	47	17.5%	<5	<100
(Cl)氯					
Roche Hitachi/Hitachi Systems Dedicated \ ICP-MS	mmol/L	115	1.1%	9	326
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	116	1.1%	8	<100
TOSHIBA TBA chongqiong zybio \ ISE (indirect)	mmol/L	132	1.1%	7	145
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	134	1.2%	62	2540
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	131	4.7%	33	911
Mindray BS series MAKER \ ISE (indirect)	mmol/L	134	0.6%	<5	<100
TOSHIBA TBA MedicalSystem \ ISE (indirect)	mmol/L	112	1.3%	<5	<100
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	136	2.5%	23	636
Roche Hitachi/Hitachi Systems ABX \ ICP-MS	mmol/L	131	1.3%	<5	<100
Mindray BS series Mindray \ OTHER	mmol/L	140	1.5%	<5	<100
Abbott ARCHITECT ci series Abbott \ ISE (indirect)	mmol/L	135	0.9%	19	429
Roche cobas c 501/502/701/702 Dedicated \ ISE (indirect)	mmol/L	134	1%	16	451
Siemens ADVIA 1800/2400 Dedicated \ ISE (indirect)	mmol/L	137	1.8%	15	665
Beckman AU series DiaSys \ ISE (indirect)	mmol/L	138	1%	14	419
Siemens ADVIA 1800/2400 Siemens \ ISE (indirect)	mmol/L	136	5.2%	11	317
Roche cobas 4000/6000/8000 Dedicated \ ISE (direct)	mmol/L	108	9.2%	11	201
Siemens ADVIA 1800/2400 Dedicated \ ICP-MS	mmol/L	136	0.3%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ ISE (indirect)	mmol/L	138	1.6%	<5	<100
(Cr)肌酐					
Roche Hitachi/Hitachi Systems BIOSINO \ Enzymatic	μmol/L	78.1	1.8%	9	324
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	μmol/L	260	2.5%	9	250
Beckman AU series KeFang(Guangzhou) \ Enzymatic	μmol/L	252	1.2%	8	165
KHB ZY series KHB \ Enzymatic	μmol/L	255	1.3%	7	<100
TOSHIBA TBA chongqiong zybio \ Enzymatic	μmol/L	255	1.8%	6	146
Mindray BS series MAKER \ Enzymatic	μmol/L	272	1%	6	120
Beckman AU series Dedicated \ Picric acid-No blank Compensation	μmol/L	245	2.3%	39	1483
Roche cobas 4000/6000/8000 Roche \ Enzymatic	μmol/L	264	11.9%	30	857
Roche cobas 4000/6000/8000 Dedicated \ Enzymatic	μmol/L	261	2.8%	<5	<100
Beckman AU series HuaChenBio \ Enzymatic	μmol/L	257	3.2%	26	426

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
7 / 16								

		批号：376DF15			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Dedicated \ OTHER	μmol/L	244	2.8%	22	992
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic	μmol/L	250	1.8%	22	979
Beckman AU series MAKER \ OTHER	μmol/L	259	3%	22	605
Mindray BS series Mindray \ OTHER	μmol/L	278	2.1%	<5	<100
Siemens ADVIA Centaur CP OTHER \ Enzymatic	μmol/L	248	2.6%	<5	<100
Abbott ARCHITECT ci series BIOTEX \ Enzymatic	μmol/L	261	1.5%	19	433
Roche cobas c 501/502/701/702 Dedicated \ Enzymatic	μmol/L	265	1.9%	16	455
Beckman AU series DiaSys \ Enzymatic	μmol/L	257	1.2%	14	416
OTHER SHLY \ Enzymatic	μmol/L	268	2.8%	14	325
(CRP)C反应蛋白					
Roche Hitachi/Hitachi Systems DiaSys \ Immunoturbidimetry (ITA)	mg/dL	0.819	5.3%	9	326
Beckman AU series SekiSui \ Immunonephelometry (INA)	mg/dL	0.226	2.9%	10	168
Mindray BS series SHSN \ Immunonephelometry (INA)	mg/dL	0.223	9.9%	<5	<100
Beckman Immage 800 Roche \ Immunonephelometry (INA)	mg/dL	0.417	9.8%	<5	<100
(Cu)铜					
Beckman AU series BSBE \ OTHER	μmol/L	32.2	2.6%	20	913
Beckman AU series Beckman Coulter \ ICP-MS	μmol/L	33.9	9.8%	<5	<100
(DBIL)直接胆红素					
Roche Hitachi/Hitachi Systems DiaSys \ Diazotization -Blank Compensation	μmol/L	5.15	11.5%	9	326
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	μmol/L	15.6	3.2%	9	124
KHB ZY series KHB \ Vanadate	μmol/L	23	6%	8	<100
Beckman AU series KeFang(Guangzhou) \ Vanadate	μmol/L	14.5	9.4%	8	163
TOSHIBA TBA chongqiong zybio \ Vanadate	μmol/L	24	4%	6	144
Beckman AU series Dedicated \ Diazotization -Blank Compensation	μmol/L	11.6	3.4%	59	2391
Beckman AU series MAKER \ Vanadate	μmol/L	20.3	7.2%	48	1044
Roche cobas 4000/6000/8000 Roche \ Vanadate	μmol/L	20.4	27.5%	31	792
Mindray BS series Mindray \ OTHER	μmol/L	22.5	2.9%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Vanadate	μmol/L	15.3	5.3%	22	980
Beckman AU series Dedicated \ Diazotization -No blank Compensation	μmol/L	10.8	4%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Diazotization	μmol/L	22.5	4.5%	19	429
Roche cobas c 501/502/701/702 Dedicated \ Diazotization -No blank Compensation	μmol/L	18.6	5.6%	16	462
Beckman AU series DiaSys \ Diazotization -Blank Compensation	μmol/L	11.3	2.9%	14	419
OTHER SHLY \ Vanadate	μmol/L	20	4.2%	14	325
Siemens ADVIA 1800/2400 Siemens \ Vanadate	μmol/L	20.9	7%	11	317
Roche Hitachi/Hitachi Systems chongqiong zybio \ Vanadate	μmol/L	31.2	8.4%	10	265
Roche cobas 4000/6000/8000 Dedicated \ Diazotization	μmol/L	42.8	18%	10	187
Roche cobas 4000/6000/8000 Roche \ Diazotization -Blank Compensation	μmol/L	17.8	9.6%	<5	<100
(DGXN)地高辛					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.23	8.2%	<5	<100
(E2)雌二醇					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	218	9.6%	<5	<100
(FA)叶酸					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.35	10.1%	<5	<100
(Fe)铁					
Roche Hitachi/Hitachi Systems BIOSINO \ Ferrozine colorimetry	μmol/L	33.3	1.3%	8	<100
Siemens ADVIA 1800/2400 BIOSINO \ Ferrozine colorimetry	μmol/L	33	2.4%	22	976
Beckman AU series Dedicated \ Ferrozine colorimetry	μmol/L	33.6	1.7%	20	916

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
8 / 16								

		批号：376DF15			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Dedicated \ TPTZ	μmol/L	33.7	1.8%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	34.2	9.9%	<5	<100
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	36.7	10%	<5	<100
(Ferritin)铁蛋白					
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	ng/mL	99.3	10%	<5	<100
(FT3)游离T3					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	3.12	9.9%	<5	<100
(FT4)游离T4					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	17.5	10.2%	<5	<100
(GGT)L-γ-谷氨酰基转移酶					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	154	2.9%	9	446
Roche Hitachi/Hitachi Systems BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	38	2.3%	9	323
Beckman AU series KeFang(Guangzhou) \ OTHER	U/L	160	1.3%	8	169
Roche cobas 4000/6000/8000 Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	188	8%	7	<100
KHB ZY series KHB \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	151	1.4%	7	<100
Beckman AU series Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	154	1.9%	63	2469
TOSHIBA TBA chongqiong zybio \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	150	2.8%	6	142
Beckman AU series HuaChenBio \ OTHER	U/L	133	1.3%	<5	<100
Roche cobas 4000/6000/8000 Roche \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	156	10.4%	33	904
TOSHIBA TBA MedicalSystem \ γ-glutamyl -4-nitroaniline	U/L	188	1.5%	<5	<100
Mindray BS series Mindray \ OTHER	U/L	144	2%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	149	2.1%	22	976
Beckman AU series MAKER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	163	3.3%	22	607
Beckman AU series Dedicated \ OTHER	U/L	150	2%	<5	<100
Roche Hitachi/Hitachi Systems ABX \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	151	0.9%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	162	1.5%	19	429
Roche cobas c 501/502/701/702 Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	150	2.3%	16	465
Beckman AU series DiaSys \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	150	1.3%	14	418
OTHER SHLY \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	143	1.6%	14	325
(GLDH)谷氨酸脱氢酶					
Roche Hitachi/Hitachi Systems OTHER \ DGKC	U/L	8.94	5.9%	6	136
Beckman AU series Roche \ Direct Measure	U/L	9.48	6.3%	39	1424
Beckman AU series Roche \ OTHER	U/L	9.61	5.6%	20	905
Roche cobas 4000/6000/8000 DiaSys \ Colorimetry	U/L	8.73	10%	<5	<100
Beckman AU series Beckman Coulter \ DGKC	U/L	9.82	9.9%	<5	<100
(GLU)葡萄糖					
Roche Hitachi/Hitachi Systems BIOSINO \ Hexokinase	mmol/L	10.3	2%	9	397
Beckman AU series KeFang(Guangzhou) \ Glucose oxidase (GOD)	mmol/L	11	1.7%	8	192
KHB ZY series KHB \ Glucose oxidase (GOD)	mmol/L	9.7	2.2%	7	<100
TOSHIBA TBA chongqiong zybio \ Hexokinase	mmol/L	10.5	1.8%	7	147
Beckman AU series Dedicated \ Hexokinase	mmol/L	10.3	2.4%	62	2492
Beckman AU series HuaChenBio \ Glucose oxidase (GOD)	mmol/L	9.46	1.4%	5	<100
Roche cobas 4000/6000/8000 Roche \ Hexokinase	mmol/L	10.2	8.2%	31	868
Mindray BS series MAKER \ Hexokinase	mmol/L	10	0.8%	<5	<100
Mindray BS series Mindray \ OTHER	mmol/L	9.72	1.9%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Hexokinase	mmol/L	10.4	1.8%	22	977
Beckman AU series MAKER \ Glucose oxidase (GOD)	mmol/L	10.6	3.7%	22	607

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
9 / 16								

		批号：376DF15			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur CP Dedicated \ OTHER	mmol/L	10.2	2.5%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Hexokinase	mmol/L	10.3	1.9%	19	427
Roche cobas c 501/502/701/702 Dedicated \ Hexokinase	mmol/L	10.2	1.4%	16	455
Beckman AU series DiaSys \ Glucose oxidase (GOD) OTHER	mmol/L	10.3	2.3%	14	417
SHLY \ Hexokinase	mmol/L	10.5	1.4%	14	323
Siemens ADVIA 1800/2400 Siemens \ Hexokinase	mmol/L	10.4	5.4%	11	317
Roche Hitachi/Hitachi Systems chongqiong zybjo \ Hexokinase	mmol/L	10.4	3.3%	11	294
Beckman AU series Beckman Coulter \ Dehydrogenase	mmol/L	9.89	9.7%	<5	<100
(HCY)同型半胱氨酸					
Roche Hitachi/Hitachi Systems BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	18	6.9%	9	230
Siemens ADVIA 1800/2400 Szybio \ Immunonephelometry (INA)	μmol/L	18.8	3.4%	9	144
KHB ZY series KHB \ Enzymatic Cycling Assays Method	μmol/L	21.6	5%	8	<100
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	18.1	3.1%	8	355
Beckman AU series DongOu \ OTHER	μmol/L	18.5	3.8%	40	1808
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	16.5	28.5%	37	940
Beckman AU series REEGEN \ Enzymatic Cycling Assays Method	μmol/L	18	10.8%	22	879
Mindray BS series Mindray \ Enzymatic Cycling Assays Method	μmol/L	14.4	19.7%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Immunoturbidimetry (ITA)	μmol/L	17	3.9%	14	615
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	19.9	4.2%	11	539
Beckman AU series HuaChenBio \ OTHER	μmol/L	21.5	3.4%	10	112
Beckman UniCel Dx C 600/800 Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	18.4	15%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ Immunonephelometry (INA)	μmol/L	18.4	2%	<5	<100
Siemens ADVIA Centaur CP OTHER \ Enzymatic Cycling Assays Method	μmol/L	19.8	2.9%	<5	<100
TOSHIBA TBA MedicalSystem \ Enzymatic Cycling Assays Method	μmol/L	16.2	0%	<5	<100
(HDL-C)高密度脂蛋白胆固醇					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	2.09	2.5%	9	252
Beckman AU series KeFang(Guangzhou) \ Direct Measure	mmol/L	1.62	1.7%	8	187
Roche cobas 4000/6000/8000 Dedicated \ Direct Measure	mmol/L	1.11	13%	8	113
KHB ZY series KHB \ Direct Measure	mmol/L	1.78	3.5%	7	<100
Beckman AU series WEGO \ Direct Measure	mmol/L	1.63	2.4%	40	1827
Roche cobas 4000/6000/8000 Roche \ selected suppression	mmol/L	1.39	7.7%	32	815
TOSHIBA TBA MedicalSystem \ Direct Measure	mmol/L	1.6	1.7%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Direct Measure	mmol/L	2	2.7%	22	976
Beckman AU series MAKER \ OTHER	mmol/L	1.69	5.9%	22	606
Beckman AU series Dedicated \ Direct Measure	mmol/L	1.58	2.7%	<5	<100
Mindray BS series Mindray \ Enzymatic	mmol/L	1.74	5.9%	<5	<100
Siemens ADVIA Centaur CP OTHER \ Direct Measure	mmol/L	1.58	13.8%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Enzymatic	mmol/L	2.05	3.5%	19	428
Roche cobas c 501/502/701/702 Dedicated \ Direct Measure	mmol/L	1.47	2.2%	16	457
Beckman AU series DiaSys \ Enzymatic	mmol/L	1.66	1.8%	14	418
OTHER SHLY \ Enzymatic	mmol/L	1.47	2.9%	14	325
Beckman AU series HuaChenBio \ OTHER	mmol/L	1.94	2.5%	14	245
Siemens ADVIA 1800/2400 Siemens \ OTHER	mmol/L	1.77	1.6%	10	302
Siemens ADVIA 1800/2400 Siemens \ Direct Measure	mmol/L	1.77	15%	<5	<100
(IgA)免疫球蛋白 IgA					
Beckman Image 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	308	10.1%	<5	<100
Abbott ARCHITECT ci series MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	255	2.4%	<5	<100

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
10 / 16								

批号：376DF15					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(IgG)免疫球蛋白 IgG					
Abbott ARCHITECT ci series MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	1528	2.9%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	1450	9.7%	<5	<100
(IgM)免疫球蛋白 IgM					
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	149	9.4%	<5	<100
Abbott ARCHITECT ci series MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	128	1.4%	<5	<100
(K)钾					
TOSHIBA TBA chongqiong zybio \ ISE (indirect)	mmol/L	5.94	0.9%	7	140
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	5.82	1.1%	60	2461
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	6.21	5.8%	33	932
Mindray BS series MAKER \ ISE (indirect)	mmol/L	5.93	0.7%	<5	<100
TOSHIBA TBA MedicalSystem \ ISE (indirect)	mmol/L	6.96	1.3%	<5	<100
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	5.95	4.6%	23	652
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	5.97	1.2%	<5	<100
Roche Hitachi/Hitachi Systems ABX \ ICP-MS	mmol/L	5.8	4.1%	<5	<100
Mindray BS series Mindray \ OTHER	mmol/L	6.18	1.3%	<5	<100
Abbott ARCHITECT ci series Abbott \ ISE (indirect)	mmol/L	6	1.2%	19	424
Roche cobas c 501/502/701/702 Dedicated \ ISE (indirect)	mmol/L	6.05	0.7%	16	454
Siemens ADVIA 1800/2400 Dedicated \ ISE (indirect)	mmol/L	6.16	1.6%	15	666
Beckman AU series DiaSys \ ISE (indirect)	mmol/L	6.03	0.8%	14	418
Siemens ADVIA 1800/2400 Siemens \ ISE (indirect)	mmol/L	6.08	5.1%	11	310
Roche cobas 4000/6000/8000 Dedicated \ ISE (direct)	mmol/L	6.6	6.8%	10	<100
Siemens ADVIA 1800/2400 Dedicated \ ISE (direct)	mmol/L	6.18	0.5%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ ISE (indirect)	mmol/L	6.21	1.1%	<5	<100
(LA)乳酸					
TOSHIBA TBA WanTai \ Lactate to pyruvate (L-P)	mmol/L	5.04	2.3%	<5	<100
Roche Hitachi/Hitachi Systems ABX \ dry chemistry	mmol/L	5.63	0.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	mmol/L	5.68	9.9%	<5	<100
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	5.88	9.9%	<5	<100
(LDH)乳酸脱氢酶					
Roche Hitachi/Hitachi Systems BIOSINO \ Lactate to pyruvate (L-P)	U/L	124	5.3%	9	325
KHB ZY series KHB \ Lactate to pyruvate (L-P)	U/L	323	2%	8	<100
Beckman AU series KeFang(Guangzhou) \ OTHER	U/L	331	2.2%	8	185
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	327	4.1%	8	103
Roche cobas 4000/6000/8000 Dedicated \ Lactate to pyruvate (L-P)	U/L	450	8.9%	7	<100
Beckman AU series Dedicated \ Lactate to pyruvate (L-P)	U/L	326	2.7%	60	2389
TOSHIBA TBA chongqiong zybio \ Lactate to pyruvate (L-P)	U/L	310	4.1%	6	143
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	U/L	353	13.3%	31	857
Mindray BS series MAKER \ Lactate to pyruvate (L-P)	U/L	316	3.2%	<5	<100
Roche cobas 4000/6000/8000 chongqiong zybio \ Lactate to pyruvate (L-P)	U/L	340	2.2%	<5	<100
TOSHIBA TBA MedicalSystem \ Lactate to pyruvate (L-P)	U/L	485	1.4%	<5	<100
Mindray BS series Mindray \ OTHER	U/L	303	2.7%	<5	<100
Beckman AU series HuaChenBio \ OTHER	U/L	319	2.4%	25	336
Beckman AU series MAKER \ Lactate to pyruvate (L-P)	U/L	303	3.2%	22	607
Beckman AU series Dedicated \ OTHER	U/L	318	3.5%	<5	<100
Roche Hitachi/Hitachi Systems ABX \ dry chemistry	U/L	297	1.8%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Lactate to pyruvate (L-P)	U/L	307	1.7%	19	427

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
11 / 16								

		批号：376DF15			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA 1800/2400 BIOSINO \ Lactate to pyruvate (L-P)	U/L	333	3.6%	15	668
Beckman AU series DiaSys \ Lactate to pyruvate (L-P)	U/L	339	2.1%	14	417
(LDL-C)低密度脂蛋白胆固醇					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	3.3	3.9%	9	229
Beckman AU series KeFang(Guangzhou) \ Direct Measure	mmol/L	2.89	3.5%	8	156
KHB ZY series KHB \ Direct Measure	mmol/L	3.14	1.9%	6	<100
TOSHIBA TBA chongqiong zybio \ Enzymatic	mmol/L	4.45	4.4%	6	141
Roche cobas 4000/6000/8000 Dedicated \ Direct Measure	mmol/L	2.88	4.9%	6	106
Beckman AU series WEGO \ Direct Measure	mmol/L	3.03	2.9%	40	1826
Roche cobas 4000/6000/8000 Roche \ SUR	mmol/L	3.53	5.8%	32	865
TOSHIBA TBA MedicalSystem \ Direct Measure	mmol/L	3.27	1.3%	<5	<100
Mindray BS series Mindray \ Direct Measure	mmol/L	2.98	20.3%	<5	124
Beckman AU series HuaChenBio \ SUR	mmol/L	3.67	3.4%	24	320
Siemens ADVIA 1800/2400 BIOSINO \ Direct Measure	mmol/L	3.2	2.6%	22	979
Beckman AU series MAKER \ OTHER	mmol/L	3.66	5.8%	22	610
Beckman AU series Dedicated \ Direct Measure	mmol/L	4	1.8%	<5	<100
Mindray BS series Mindray \ Enzymatic	mmol/L	3.45	1.4%	<5	<100
Siemens ADVIA Centaur CP OTHER \ Direct Measure	mmol/L	3.49	12.4%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ SUR	mmol/L	2.98	1.4%	19	426
Roche cobas c 501/502/701/702 Dedicated \ Direct Measure	mmol/L	3.55	1.8%	16	457
Beckman AU series DiaSys \ selected suppression	mmol/L	3.36	1.6%	14	419
OTHER SHLY \ SUR	mmol/L	2.92	2.4%	14	325
(Li)锂					
Beckman AU series Beckman Coulter \ Colorimetry	mmol/L	1.79	10.1%	<5	<100
(LIP)脂肪酶					
Siemens ADVIA 1800/2400 HOMA \ OTHER	U/L	112	3.1%	8	127
Roche Hitachi/Hitachi Systems OTHER \ Glucose oxidase (GOD)	U/L	127	3.9%	6	140
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	U/L	122	12.8%	31	800
Beckman AU series Autobio \ OTHER	U/L	123	14.4%	27	300
Beckman AU series Beckman Coulter \ OTHER	U/L	137	3.9%	11	548
Roche cobas 4000/6000/8000 DiaSys \ Enzymatic Cycling Assays Method	U/L	93.7	9.8%	<5	<100
Beckman AU series Beckman Coulter \ Glucose oxidase (GOD)	U/L	119	9.2%	<5	<100
(Mg)镁					
Beckman AU series HuaChenBio \ Arsenazo □	mmol/L	1.36	2.7%	7	<100
TOSHIBA TBA chongqiong zybio \ Xylidyl Blue	mmol/L	1.58	3.7%	6	135
Beckman AU series Dedicated \ Calmagite colorimetry	mmol/L	1.47	1.4%	59	2412
Roche Hitachi/Hitachi Systems BIOSINO \ Xylidyl Blue	mmol/L	1.58	12.7%	5	<100
Roche cobas 4000/6000/8000 Roche \ ICP-MS	mmol/L	1.48	4.2%	32	873
TOSHIBA TBA MedicalSystem \ Xylidyl Blue	mmol/L	1.19	4.2%	<5	<100
Mindray BS series Mindray \ Xylidyl Blue	mmol/L	1.29	21.9%	<5	124
Beckman AU series Dedicated \ Xylidyl Blue	mmol/L	1.45	1.6%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ Calmagite colorimetry	mmol/L	1.45	4.8%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Calmagite colorimetry	mmol/L	1.44	2%	16	458
Siemens ADVIA 1800/2400 BIOSINO \ Xylidyl Blue	mmol/L	1.45	4.6%	15	667
Beckman AU series DiaSys \ Xylidyl Blue	mmol/L	1.49	3.3%	14	421
Roche Hitachi/Hitachi Systems chongqiong zybio \ Xylidyl Blue	mmol/L	1.43	3.7%	11	289
Roche cobas 4000/6000/8000 Dedicated \ Calmagite colorimetry	mmol/L	1.48	3%	11	187

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
12 / 16								

		批号：376DF15			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA 1800/2400 Siemens \ Xylidyl Blue	mmol/L	1.47	7.7%	10	294
Beckman AU series Beckman Coulter \ Xylidyl Blue	mmol/L	1.43	9.8%	<5	<100
(Na)钠					
Roche Hitachi/Hitachi Systems Dedicated \ ICP-MS	mmol/L	145	1%	9	325
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	143	0.9%	8	<100
TOSHIBA TBA chongqiong zyblio \ ISE (indirect)	mmol/L	156	0.8%	7	139
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	153	0.9%	62	2542
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	158	2.7%	33	941
Mindray BS series MAKER \ ISE (indirect)	mmol/L	156	0.8%	<5	<100
TOSHIBA TBA MedicalSystem \ ISE (indirect)	mmol/L	152	1.5%	<5	<100
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	155	1.3%	22	607
Roche Hitachi/Hitachi Systems ABX \ ICP-MS	mmol/L	154	1%	<5	<100
Mindray BS series Mindray \ OTHER	mmol/L	161	1.3%	<5	<100
Abbott ARCHITECT ci series Abbott \ ISE (indirect)	mmol/L	161	0.8%	19	427
Roche cobas c 501/502/701/702 Dedicated \ ISE (indirect)	mmol/L	157	0.8%	16	454
Siemens ADVIA 1800/2400 Dedicated \ ISE (indirect)	mmol/L	156	1.2%	15	667
Beckman AU series DiaSys \ ISE (indirect)	mmol/L	159	1%	14	417
Siemens ADVIA 1800/2400 Siemens \ ISE (indirect)	mmol/L	156	5.4%	11	330
Roche cobas 4000/6000/8000 Dedicated \ ISE (direct)	mmol/L	152	1.3%	11	105
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	156	9%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ ICP-MS	mmol/L	156	0.2%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ ISE (indirect)	mmol/L	161	1%	<5	<100
(NEFA)游离脂肪酸					
Roche Hitachi/Hitachi Systems OTHER \ ACS - ACOD	mmol/L	0.776	1.6%	6	135
Beckman AU series DongOu \ ACS - ACOD	mmol/L	0.78	3.2%	40	1831
Siemens ADVIA 1800/2400 OTHER \ ACS - ACOD	mmol/L	0.789	4.5%	22	981
Roche cobas 4000/6000/8000 DiaSys \ ACS - ACOD	mmol/L	0.72	11.1%	<5	<100
(P)磷					
Siemens ADVIA 1800/2400 Siemens \ Phosphomolybdate-UV-end point	mmol/L	3.13	2.8%	9	256
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	3.07	1.3%	8	<100
Roche cobas 4000/6000/8000 Dedicated \ Phosphomolybdate-UV-end point	mmol/L	2.73	4.1%	7	<100
TOSHIBA TBA chongqiong zyblio \ Phosphomolybdate-UV-end point	mmol/L	3.1	1.9%	6	144
Beckman AU series Dedicated \ Phosphomolybdate UV-rate	mmol/L	3.11	1.5%	59	2408
TOSHIBA TBA MedicalSystem \ Phosphomolybdate-UV-end point	mmol/L	2.6	1.1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Phosphomolybdate-UV-end point	mmol/L	3.24	3.2%	32	820
Beckman AU series HuaChenBio \ Phosphomolybdate UV-rate	mmol/L	2.85	7.2%	29	478
Beckman AU series MAKER \ OTHER	mmol/L	3.17	4.4%	22	607
Beckman AU series Dedicated \ OTHER	mmol/L	3.04	1.9%	<5	<100
Mindray BS series Mindray \ OTHER	mmol/L	2.6	1.4%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ Phosphomolybdate UV-rate	mmol/L	3.25	1.3%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Phosphomolybdate-UV-end point	mmol/L	3.04	1.6%	16	450
Siemens ADVIA 1800/2400 BIOSINO \ Phosphomolybdate-UV-end point	mmol/L	2.79	2.8%	15	669
Beckman AU series DiaSys \ Phosphomolybdate UV-rate	mmol/L	2.9	2%	14	418
Roche Hitachi/Hitachi Systems chongqiong zyblio \ Phosphomolybdate-UV-end point	mmol/L	2.95	1.9%	11	282
Roche cobas 4000/6000/8000 Roche \ Phosphomolybdate UV-rate	mmol/L	3.09	10%	<5	<100
Beckman AU series Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	3.12	9.3%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Phosphomolybdate UV-rate	mmol/L	3.23	15%	<5	<100

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
13 / 16								

		批号：376DF15			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(PAMY)胰淀粉酶					
Beckman AU series Reebio \ OTHER	U/L	185	2.7%	11	507
Beckman AU series Beckman Coulter \ Enzymatic	U/L	184	10%	<5	<100
(PRO/ TP)(总) 蛋白					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	g/L	97.5	1.6%	9	<100
Roche Hitachi/Hitachi Systems BIOSINO \ Biuret-endpoint	g/L	67.8	1.5%	9	324
KHB ZY series KHB \ Biuret-endpoint	g/L	103	2.4%	8	<100
Beckman AU series KeFang(Guangzhou) \ Biuret-rate	g/L	99.2	1.6%	8	185
Beckman AU series Dedicated \ Biuret-endpoint	g/L	97.6	1.6%	60	2423
TOSHIBA TBA chongqiong zybio \ Biuret-endpoint	g/L	98.7	1.5%	6	146
Roche cobas 4000/6000/8000 Roche \ Biuret-endpoint	g/L	93	8.3%	33	906
Mindray BS series MAKER \ Biuret-endpoint	g/L	99.3	2.3%	<5	<100
TOSHIBA TBA MedicalSystem \ Biuret-endpoint	g/L	69.6	0.9%	<5	<100
Mindray BS series Mindray \ OTHER	g/L	97.6	2%	<5	<100
Mindray BS series Mindray \ Biuret-endpoint	g/L	92.5	17.1%	<5	120
Beckman AU series HuaChenBio \ Biuret-endpoint	g/L	97.6	2.8%	23	299
Siemens ADVIA 1800/2400 BIOSINO \ Biuret-rate	g/L	95.6	1.9%	22	978
Beckman AU series MAKER \ Biuret-endpoint	g/L	101	2.4%	22	606
Beckman AU series Dedicated \ OTHER	g/L	96.4	1.4%	<5	<100
Roche Hitachi/Hitachi Systems ABX \ dry chemistry	g/L	103	0.7%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Biuret-endpoint	g/L	94.7	1.3%	19	427
Roche cobas c 501/502/701/702 Dedicated \ Biuret-endpoint	g/L	94.2	1.6%	16	464
Beckman AU series DiaSys \ Biuret-endpoint	g/L	101	1.5%	14	419
(PROG)孕酮					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.1	10%	<5	<100
(TBA)总胆汁酸					
Roche Hitachi/Hitachi Systems chongqiong zybio \ Enzymatic Cycling Assays Method	μmol/L	89.9	4.5%	9	215
Siemens ADVIA 1800/2400 Szybio \ Enzymatic Cycling Assays Method	μmol/L	72.5	4.3%	9	136
Roche Hitachi/Hitachi Systems BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	80.9	6.8%	9	130
Beckman AU series KeFang(Guangzhou) \ Enzymatic Cycling Assays Method	μmol/L	72.9	6.6%	8	142
KHB ZY series KHB \ Enzymatic Cycling Assays Method	μmol/L	88.6	3.7%	7	<100
TOSHIBA TBA chongqiong zybio \ Enzymatic Cycling Assays Method	μmol/L	89.6	4.5%	6	143
Roche Hitachi/Hitachi Systems OTHER \ Enzymatic Cycling Assays Method	μmol/L	85.4	10.5%	6	134
Beckman AU series DongOu \ Enzymatic Cycling Assays Method	μmol/L	90.3	2.7%	59	2400
TOSHIBA TBA MedicalSystem \ Enzymatic Cycling Assays Method	μmol/L	61.4	2.9%	<5	<100
Beckman AU series HuaChenBio \ Enzymatic Cycling Assays Method	μmol/L	53.7	16.2%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Enzymatic Cycling Assays Method	μmol/L	92.3	4%	<5	<100
Beckman AU series Dedicated \ Enzymatic Cycling Assays Method	μmol/L	69	4.5%	<5	106
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	95.5	3.5%	22	976
Mindray BS series Mindray \ OTHER	μmol/L	89.2	5.9%	<5	<100
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	72.8	3%	11	526
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	142	10%	<5	<100
Roche cobas 4000/6000/8000 DiaSys \ Enzymatic Cycling Assays Method	μmol/L	76.4	9.9%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Enzymatic Cycling Assays Method	μmol/L	78.3	15%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ Enzymatic Cycling Assays Method	μmol/L	71.4	4.3%	<5	<100
(TBIL)总胆红素					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	μmol/L	75.8	3.4%	9	129

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
14 / 16								

		批号：376DF15			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series	μmol/L	92.2	2.9%	8	170
KeFang(Guangzhou) \ Vanadate					
Roche cobas 4000/6000/8000	μmol/L	107	17.6%	8	127
Dedicated \ Diazotization -Blank Compensation					
KHB ZY series	μmol/L	72.7	1%	6	<100
KHB \ Vanadate					
TOSHIBA TBA	μmol/L	91.8	4.4%	6	144
chongqiong zybio \ Vanadate					
Beckman AU series	μmol/L	71.5	3.9%	47	988
MAKER \ OTHER					
Beckman AU series	μmol/L	77.6	2%	39	1482
Dedicated \ Diazotization					
Roche cobas 4000/6000/8000	μmol/L	74	12.5%	32	842
Roche \ Vanadate					
TOSHIBA TBA	μmol/L	118	1.7%	<5	<100
MedicalSystem \ Vanadate					
Siemens ADVIA 1800/2400	μmol/L	87.1	3.3%	22	979
BIOSINO \ Vanadate					
Beckman AU series	μmol/L	78.9	1.4%	<5	<100
Dedicated \ OTHER					
Mindray BS series	μmol/L	73.6	1.3%	<5	<100
Mindray \ OTHER					
Abbott ARCHITECT ci series	μmol/L	76.2	2.6%	19	423
BIOSINO \ Diazotization					
Roche cobas c 501/502/701/702	μmol/L	72.6	2.3%	16	455
Dedicated \ Diazotization					
Beckman AU series	μmol/L	78	1.5%	14	419
DiaSys \ Diazotization -Blank Compensation					
OTHER					
SHLY \ Vanadate	μmol/L	68.8	1.7%	14	325
Siemens ADVIA 1800/2400					
Siemens \ Vanadate	μmol/L	86.7	5.1%	11	303
Roche Hitachi/Hitachi Systems	μmol/L	93.2	2%	11	291
chongqiong zybio \ Vanadate					
Beckman AU series	μmol/L	74.1	10%	<5	<100
Beckman Coulter \ Diazotization					
(TES)睾酮					
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.77	10.2%	<5	<100
(TG)甘油三酯					
Roche Hitachi/Hitachi Systems	mmol/L	5.06	2.1%	9	346
BIOSINO \ OTHER					
Beckman AU series	mmol/L	4.9	1.9%	8	178
KeFang(Guangzhou) \ GPO-POD					
KHB ZY series	mmol/L	5.31	2.9%	7	<100
KHB \ GPO-POD					
TOSHIBA TBA	mmol/L	4.5	4%	6	143
chongqiong zybio \ GPO-POD					
Beckman AU series	mmol/L	4.96	1.7%	40	1828
Dedicated \ Dehydrogenase					
Roche cobas 4000/6000/8000	mmol/L	4.83	1.6%	<5	<100
Dedicated \ GPO-POD					
Beckman AU series	mmol/L	4.98	1.7%	<5	108
Dedicated \ GPO-POD					
Roche cobas 4000/6000/8000	mmol/L	5.66	2.1%	28	774
Roche \ GPO-POD					
Siemens ADVIA 1800/2400	mmol/L	4.67	2.4%	22	971
BIOSINO \ GPO-POD					
Beckman AU series	mmol/L	5.19	3.9%	22	606
MAKER \ OTHER					
Beckman AU series	mmol/L	4.95	1.5%	20	266
HuaChenBio \ OTHER					
Mindray BS series	mmol/L	4.47	1%	<5	<100
Mindray \ OTHER					
Siemens ADVIA Centaur CP	mmol/L	5.17	1.6%	<5	<100
Dedicated \ GPO-POD					
Abbott ARCHITECT ci series	mmol/L	4.89	2.1%	19	431
BIOSINO \ GPO-POD					
Roche cobas c 501/502/701/702	mmol/L	4.95	1.7%	16	452
Dedicated \ GPO-POD					
Beckman AU series	mmol/L	4.86	1.5%	14	418
DiaSys \ GK-PK-LDH					
OTHER					
SHLY \ GPO-POD	mmol/L	5.19	2%	14	325
Siemens ADVIA 1800/2400					
Siemens \ GPO-POD	mmol/L	5.03	4.8%	11	326
Roche Hitachi/Hitachi Systems	mmol/L	4.92	2.3%	11	291
chongqiong zybio \ GPO-POD					
(THEO)茶碱					
Beckman UniCel DxI 600/800					
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μg/mL	21.4	9.2%	<5	<100
(TRF)转铁蛋白					
Mindray BS series	mg/dL	386	0.8%	<5	<100
Mindray \ Immunoturbidimetry (ITA)					
Roche cobas 4000/6000/8000	mg/dL	389	9.8%	<5	<100
Roche \ Immunoturbidimetry (ITA)					
(TSH)促甲状腺素					
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	2.59	9.7%	<5	<100

质控品名： 生化与激素非定值质控品

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
15 / 16								

批号：376DF15					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TT3)总三碘甲状原氨酸					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.599	9.8%	<5	<100
(TT4)总T4					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.08	9.4%	<5	<100
(UA)尿酸					
Roche Hitachi/Hitachi Systems BIOSINO \ Uricase-UV	μmol/L	289	4.5%	9	326
Roche Hitachi/Hitachi Systems BIOSINO \ Uricase colorimetry	μmol/L	657	1.7%	9	246
Beckman AU series KeFang(Guangzhou) \ Uricase-UV	μmol/L	678	2.3%	8	154
KHB ZY series KHB \ Uricase colorimetry	μmol/L	702	1.5%	7	<100
Roche cobas 4000/6000/8000 Dedicated \ Uricase colorimetry	μmol/L	509	16.2%	7	100
Beckman AU series Dedicated \ Uricase colorimetry	μmol/L	669	1.5%	60	2430
TOSHIBA TBA chongqiong zybio \ Uricase colorimetry	μmol/L	648	1.6%	6	143
Mindray BS series Mindray \ Uricase colorimetry	μmol/L	645	1.6%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Uricase colorimetry	μmol/L	642	10%	33	931
TOSHIBA TBA MedicalSystem \ Uricase colorimetry	μmol/L	478	1%	<5	<100
Beckman AU series HuaChenBio \ Uricase colorimetry	μmol/L	684	2%	23	225
Siemens ADVIA 1800/2400 BIOSINO \ Uricase colorimetry	μmol/L	659	1.7%	22	980
Beckman AU series MAKER \ Uricase colorimetry	μmol/L	683	2.3%	22	607
Beckman AU series Dedicated \ OTHER	μmol/L	671	2.1%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ Uricase colorimetry	μmol/L	696	1.7%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Uricase colorimetry	μmol/L	666	4.5%	19	423
Roche cobas c 501/502/701/702 Dedicated \ Uricase colorimetry	μmol/L	653	1.6%	16	455
OTHER DiaSys \ Uricase-UV	μmol/L	668	1.7%	14	418
OTHER SHLY \ Uricase colorimetry	μmol/L	612	1.5%	14	325
(UIBC)不饱和铁结合力					
Siemens ADVIA 1800/2400 DiaSys \ Ferrozine colorimetry	μmol/L	59.1	2.2%	7	311
Roche cobas c 501/502/701/702 Dedicated \ Ferrozine colorimetry	μmol/L	55.9	2.8%	6	119
Beckman AU series Dedicated \ Ferrozine colorimetry	μmol/L	56.5	2.3%	20	909
Beckman AU series Dedicated \ OTHER	μmol/L	56.6	1.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	53.5	9.9%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Ferrozine colorimetry	μmol/L	54.6	10%	<5	<100
(Urea)尿素					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	12.8	1.9%	9	180
Beckman AU series KeFang(Guangzhou) \ Urease-UV-rate	mmol/L	13.5	3.1%	8	169
Roche cobas 4000/6000/8000 Dedicated \ Urease-UV-rate	mmol/L	13.1	3.8%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Urease colorimetry	mmol/L	13.5	6.3%	29	806
Beckman AU series HuaChenBio \ OTHER	mmol/L	13.3	6.9%	26	457
Beckman AU series Dedicated \ OTHER	mmol/L	13	3.3%	<5	<100
Abbott ARCHITECT ci series BIOTEX \ Urease-UV-rate	mmol/L	12.9	2.6%	19	433
Roche cobas c 501/502/701/702 Dedicated \ Urease-UV-rate	mmol/L	13.1	2.3%	16	470
Beckman AU series DiaSys \ Urease-UV-rate	mmol/L	13.8	2.2%	14	419
OTHER SHLY \ Urease colorimetry	mmol/L	13	2.6%	14	325
Siemens ADVIA 1800/2400 Siemens \ Urease-UV-rate	mmol/L	13.8	5.5%	11	317
Beckman AU series Beckman Coulter \ Urease-UV-rate	mmol/L	12.4	10%	<5	<100
Beckman AU series Dedicated \ Urease colorimetry	mmol/L	14.3	5.7%	<5	<100
(VB12)维生素B12					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	593	9.6%	<5	<100

REF	581-2	6×5mL	EXP	2023-05-06	LOT	376DF15	更新日期：	2024-04
16 / 16								

		批号：376DF15			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Zn)锌					
Roche Hitachi/Hitachi Systems WanTai DRD \ OTHER	μmol/L	34.9	4.6%	8	103
Beckman AU series BSBE \ OTHER	μmol/L	31.7	4%	20	895
Beckman AU series Beckman Coulter \ ICP-MS	μmol/L	33.2	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。