报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(17-OHP)17-羟孕酮					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.568	10.5%	<5	<100
(25-OH-Vit D)25羟基维生素D					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	52.7	9.5%	52	1259
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	48.6	8.1%	12	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.8	5.5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.2	4.3%	<5	<100
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.6	9.9%	<5	<100
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	nmol/L	41.3	9.9%	<5	<100
(AFP)甲胎蛋白					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.99	4.8%	55	1478
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.1	6.1%	40	892
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.04	4.8%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.3	7.1%	25	188
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.35	4.8%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.78	9.4%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.97	5.1%	16	343
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.2	4.9%	15	215
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.46	9.8%	12	242
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.91	7.5%	11	196
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.37	11.5%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.14	10.3%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.44	11.3%	<5	<100
(ALD)醛固酮					
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	pg/mL	53	12.1%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	62.4	9.8%	<5	<100
(Anti-Tg)抗甲状腺球蛋白抗体					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	1.66	9.8%	51	1439
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	20.9	9.1%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	18.4	27%	32	965
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	16.4	8.8%	22	224
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	10.5	11.6%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	16.6	13.7%	14	300
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	9.91	9.3%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	11.4	0%	<5	<100
(Anti-TPO)抗甲状腺过氧化物酶抗体					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	7.88	8.5%	50	1385
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	17.8	18.5%	30	948
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	7.69	41.7%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	17.4	17.4%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	11.2	7.9%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.6	9.6%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	8.31	9.9%	<5	<100
(ASD)雄烯二酮					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.911	10.1%	<5	<100
(CA 125)糖类抗原 125					
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	32.8	4.9%	7	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.9	3.8%	53	1422
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	33.2	4.9%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	33.2	7.4%	25	166
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	31.4	5.3%	16	313
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.4	3.7%	16	216
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34.9	5.2%	14	283
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34	8.5%	11	232
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	36.1	9.3%	11	200
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	31.7	5.1%	10	223
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	42.1	10.5%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.9	10.6%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.3	2.7%	<5	<100
(CA 15-3)糖类抗原 15-3					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	24.2	10.9%	9	209
Siemens ADVIA 1800/2400 Siemens \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27.4	11.3%	7	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27.5	5.5%	54	1418
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	30.5	4.9%	5	138
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.6	7.9%	<5	<100
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	28.8	8.1%	15	213
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.3	6.8%	14	282
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	25.9	5.3%	10	216
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16	4.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	30.6	10.5%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.6	10.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	26.2	6.1%	<5	<100
(CA 19-9)糖类抗原 19-9					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	25.5	5.6%	53	1418
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.4	5%	5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	22.5	4.3%	15	310
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	23.9	8.4%	10	205
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.2	10.7%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	23.6	11.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	20.4	10.8%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	25.2	4.8%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(CEA)癌胚抗原					
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.11	6.8%	6	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.42	4.7%	55	1473
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.52	5%	40	876
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.57	12.3%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3	8.1%	25	176
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.67	5.4%	16	346
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.52	14.4%	16	303
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.79	7.5%	14	196
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.58	10.1%	11	186
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.62	10.4%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.78	8.3%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.58	0%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.5	8.5%	5	163
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.82	5.2%	30	662
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.52	5.8%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.98	3.7%	<5	<100
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.93	10.5%	15	425
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.46	10.3%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.36	9.9%	<5	<100
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.5	13.1%	<5	<100
(Cortisol)皮质醇					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	114	2%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	109	3.7%	26	695
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	105	7.4%	25	719
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	110	5.5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	106	3.5%	17	234
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	120	7.9%	14	271
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	97.5	5.7%	12	301
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	102	5.9%	11	311
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	141	10.6%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	130	11.1%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	152	9.9%	<5	<100
(C-P)C-肽					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.947	9.7%	8	159
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.657	7.3%	7	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.453	70.7%	66	1614
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.568	6.9%	24	287
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.913	3.8%	22	358
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.561	8.5%	14	275

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.808	18.1%	11	255
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.9	2.2%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.54	12%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.96	11.5%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.905	0.8%	<5	<100
(DGXN)地高辛					
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.03	12.8%	19	164
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.99	11.1%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.922	10.2%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.08	8.3%	<5	<100
(DHEA-S)硫酸脱氢表雄酮					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3298	4.7%	15	486
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3069	10%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3548	6.5%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4274	9.5%	<5	<100
(E2)雌二醇					
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	385	15.5%	8	172
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	348	7%	8	141
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	349	16.5%	7	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	321	10.6%	7	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	296	7.9%	6	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	293	10%	5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	346	4.4%	32	1424
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	330	4.4%	25	747
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	321	13.3%	24	683
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	351	8.7%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	322	7.5%	17	271
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	385	13.6%	12	305
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	343	9.7%	11	322
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	332	8.1%	10	<100
Abbott ARCHITECT i series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	388	7.3%	10	205
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	266	10.5%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	328	10.4%	<5	<100
(EPO)促红细胞生成素					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	6.28	5.3%	5	121
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/mL	5.84	6.4%	12	298
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	8.72	10.1%	<5	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	6.45	10.4%	<5	<100
(FA)叶酸					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.86	9.7%	5	<100
Beckman AU series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.92	5.6%	5	162

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.05	6.6%	26	745
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.22	6.6%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.33	9.5%	19	150
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.06	17.2%	17	284
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.32	17.6%	13	205
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.12	7.3%	12	282
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.82	8.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.99	10.1%	<5	<100
(Ferritin)铁蛋白					
Siemens ADVIA 1800/2400 Siemens \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	76.2	8%	7	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	99	8.2%	69	1651
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	90.8	10.8%	5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	65.1	6.7%	5	164
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	68.5	12.5%	21	250
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	76.9	7.3%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	81.8	9.1%	14	281
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	98.9	9%	14	233
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	56.6	8.6%	12	304
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	108	9.2%	12	191
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	68.1	9.4%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	70.2	9.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	77.2	10%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.87	5.4%	53	1431
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.19	5.4%	5	136
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.25	11.7%	11	242
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.87	5.4%	10	223
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.88	10.1%	<5	<100
(FSH)促卵泡刺激素					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.43	9.3%	7	<100
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.97	13.7%	7	<100
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.33	28.5%	7	110
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.21	6.9%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	3.93	3%	34	886
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.06	4.5%	25	721
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.65	7.9%	25	714
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.21	4.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.68	14.2%	17	272
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	3.85	6.5%	14	268
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	4.47	6.7%	12	304
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.81	7.1%	11	307
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.82	5.4%	11	239

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	4.01	2.1%	10	107
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.93	10.5%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	3.84	2.8%	<5	<100
(FT3)游离T3					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.16	7.6%	50	1390
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.04	3.6%	36	897
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.53	4.6%	22	270
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.33	2.4%	20	629
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.27	10.3%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.03	9.8%	19	282
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.52	5.8%	18	767
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.3	4.1%	18	362
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	6.36	4.4%	16	341
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.32	7.6%	11	314
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.48	9.2%	10	202
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	5.47	9.9%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	6.18	9.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	4.6	7.1%	<5	<100
(FT4)游离T4					
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	13.5	4.3%	7	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.6	7.3%	50	1392
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	15.9	12.9%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	15.5	5.8%	24	306
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	15.4	4.6%	22	471
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	16.8	2.8%	19	628
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	16.3	7.5%	19	245
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.2	6%	18	760
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	15.6	6.1%	18	365
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	15.6	8.4%	16	319
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	12.3	3.5%	14	176
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.9	5.6%	11	314
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.3	2.2%	10	204
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	13.3	11.3%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	13.1	10.7%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.4	11.3%	<5	<100
(HCY)同型半胱氨酸					
Beckman AU series Dedicated \ Enzymatic Cycling Assays Method	μmol/L	9.71	15.5%	26	546
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	μmol/L	8.94	6%	22	245
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μmol/L	15.7	10.8%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	μmol/L	12.1	9.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μmol/L	11.7	10.3%	<5	<100
(HGH)人生长激素					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	1.55	3.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μg/L	1.51	10.6%	<5	<100

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(IgE)免疫球蛋白 IgE					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	180	5.3%	52	1309
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	177	6.7%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	182	5.5%	12	357
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	165	10.3%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	69.8	10.3%	<5	<100
(IGF1)胰岛素生长因子1					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	nmol/L	19.4	9.8%	<5	<100
(INS)胰岛素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.81	5.3%	8	158
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.23	13.5%	7	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.93	6.9%	58	1506
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	9.49	11.9%	36	905
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.86	4.9%	30	776
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	7.24	8%	24	307
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	7.32	5.2%	14	255
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	6.93	9.4%	11	232
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	10.3	3.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	6.54	10.1%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.83	10%	<5	<100
(LH)促黄体生成素					
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	3.92	4.9%	8	153
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.07	4.8%	7	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.22	6.5%	6	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.12	3.9%	25	729
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.72	6%	25	715
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.72	3.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.77	3.5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.4	13.2%	17	293
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.24	4.1%	14	348
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	4.58	6.7%	14	267
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	3.82	5.1%	12	303
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.85	9.3%	11	308
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.86	11.5%	11	221
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.07	3.2%	10	105
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.71	11.5%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.47	11.8%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	4.35	4.2%	<5	<100
(Myo / Mb)肌红蛋白					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	27.9	5.7%	5	165
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	30.6	4.7%	30	664
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	28.1	4.2%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	30.4	3.9%	<5	<100
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	28	3.4%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μg/L	34.1	10%	<5	<100
Roche cobas E601/602 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	27.3	9.5%	<5	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	27.9	9.3%	<5	<100
(NT-proBNP)N-末端脑钠肽前体					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	176	13.5%	68	1833
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	20.7	13.5%	67	1831
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	214	25.1%	6	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	174	10.7%	30	647
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	20.4	10.7%	30	647
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	19.9	12.1%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	179	2.4%	<5	126
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	21	2.4%	<5	126
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	168	3.4%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	19.7	3.4%	<5	<100
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	184	6%	16	443
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	21.7	6%	16	443
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	186	11.9%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	21.9	11.9%	<5	<100
(PRL)催/ 泌乳素					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.56	16.1%	7	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.39	6.6%	6	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.8	4.6%	6	133
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.84	5.8%	6	122
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.13	16.4%	5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.78	3.1%	33	871
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.09	5%	25	726
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.09	5.5%	25	714
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.99	5.8%	16	265
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.95	4.7%	10	285
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.81	4.6%	10	261
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.53	9.9%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.77	0%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.26	4.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.88	0%	<5	<100
(PROG)孕酮					
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.12	18.1%	8	173
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.09	7.9%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.54	14%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.76	11.3%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.17	6.6%	25	737
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.54	11.8%	25	723
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.14	6.9%	25	1019
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.979	15.2%	17	272
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.2	10.4%	14	274
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.56	9%	11	308
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.04	15.6%	11	225
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.17	11.9%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.11	10.8%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.21	11.7%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.25	0%	<5	<100
(PSA)前列腺特异性抗原					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.76	2.7%	53	1389
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.92	5.2%	5	126
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.98	3.7%	<5	<100
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.64	10.4%	15	232
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.9	4.1%	14	282
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.37	6.6%	11	242
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.78	4.5%	10	211
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.96	10.4%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.76	10.3%	<5	<100
(PTH)甲状旁腺素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1.57	18.4%	9	168
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1.93	9%	50	1163
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1.77	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1.78	15.8%	20	165
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1.96	8.4%	<5	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	1.73	4.1%	12	296
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	2.15	6.8%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	2.04	10.3%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1.96	9.7%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1.83	9.8%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	2.2	10%	<5	<100
(TES)睾酮					
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.78	7.4%	8	170
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.21	13.5%	7	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.62	8.2%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.68	3.2%	33	1108
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.35	5.4%	26	728
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.24	4.3%	25	716

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.07	5.6%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.5	8.2%	17	278
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.46	10%	14	269
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.88	3.5%	12	305
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.44	4.7%	11	315
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.21	5.7%	11	232
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.1	10.6%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.69	10.8%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.53	5.1%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.1	0%	<5	<100
(Tg)甲状腺球蛋白					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.2	4.6%	51	1414
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.5	4.8%	5	137
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.86	11.2%	13	316
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.82	11.2%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.89	10.4%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.5	9.6%	<5	<100
(THEO)茶碱					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/mL	9.66	9.9%	<5	<100
(TnI)肌钙蛋白I					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.054	7%	5	166
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.032	11.8%	15	424
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.05	20%	<5	<100
(TnT)肌钙蛋白T					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.019	8.3%	31	697
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.019	5.1%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.019	4.9%	29	740
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.019	9.8%	17	445
(TSH)促甲状腺素					
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.76	5.2%	7	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.93	4.8%	50	1389
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.55	2.3%	5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.24	5.2%	45	1188
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.95	4.1%	22	271
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.01	16.6%	20	302
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.42	5.1%	19	796
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.41	4.4%	19	636
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.79	5%	18	378
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	2.79	6.9%	16	318
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	2.46	6.2%	14	180
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.85	5%	11	314

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.65	8.7%	11	229
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	3.09	10.7%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	3.08	10.4%	<5	<100
(TT3)总三碘甲状原氨酸					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.07	7.6%	50	1362
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.36	9.6%	22	230
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.13	4.2%	19	768
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.27	7.8%	19	273
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.37	8.6%	18	357
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.52	8.3%	16	328
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.27	5.1%	15	332
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.39	3.6%	14	409
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.09	7.7%	12	343
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.03	3.5%	10	197
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.56	9.6%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.38	9.9%	<5	<100
(TT4)总T4					
Siemens ADVIA 1800/2400 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	63.2	6.2%	7	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	56.2	7.9%	50	1391
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	76.8	6.5%	5	135
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	72.6	8.8%	22	223
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	60.8	4.9%	19	755
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	62	14.9%	19	284
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	63.4	6.2%	18	365
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	69.6	11%	16	340
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	67.3	7.7%	15	333
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	62.1	3.2%	14	415
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	56.8	10.5%	11	238
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	51.8	10.2%	<5	<100
(VB12)维生素B12					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pg/mL	366	7.1%	6	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	286	7.2%	5	162
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	276	6.9%	25	716
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pg/mL	405	16.2%	21	191
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	320	3.4%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	519	10.8%	16	227
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	420	8.4%	14	200
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pg/mL	228	4.5%	12	302
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pg/mL	523	10.1%	<5	<100
(β2M)β-2-微球蛋白					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1095	8.1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	960	9.4%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1380	10.9%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-1	6×3mL	EXP	2023-01-31	LOT	338298D	更新日期：	2024-03

		批号：338298D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.003	4.3%	9	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.004	18.5%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.004	5%	16	428
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.004	7%	16	414
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.005	16%	12	249
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.004	5.5%	11	316
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.004	10.1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。