

<b>REF</b>	500-3	6x3mL	<b>EXP</b>	2022-03-15	<b>LOT</b>	2FB6E4B	更新日期：	2022-01
								1 / 3

		批号：2FB6E4B					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
<b>(17-OHP)17-羟孕酮</b>							
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.9	10.1%	<5	<100		
<b>(ACTH)促肾上腺皮质激素</b>							
maglumi system Snibe \ Chemiluminescence enzyme immunoassay (CLEIA)	pg/mL	41810	6.7%	<5	<100		
<b>(ADA)腺苷脱氨酶</b>							
maglumi system Snibe \ Peroxidase	U/L	1	0%	<5	<100		
AB Sciex API Series AllergScreen \ Peroxidase	U/L	1	0%	<5	<100		
<b>(AFU)<math>\alpha</math>-L-岩藻糖苷酶</b>							
ABX pentre80 MediRox \ CNPF	U/L	122	0%	<5	<100		
<b>(<math>\alpha</math>-HBDH)<math>\alpha</math>-羟丁酸脱氢酶</b>							
maglumi system Snibe \ DGKC	U/L	122	0.5%	<5	<100		
<b>(Anti-Tg)抗甲状腺球蛋白抗体</b>							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	66.3	7.2%	<5	<100		
<b>(Anti-TPO)抗甲状腺过氧化物酶抗体</b>							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	38.9	18%	<5	<100		
<b>(Cortisol)皮质醇</b>							
Beckman UniCel Dx C 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	303	6.2%	<5	<100		
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	346	10.4%	<5	<100		
<b>(C-P)C-肽</b>							
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.68	19.3%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.55	9.9%	<5	<100		
<b>(DGXN)地高辛</b>							
ABX Pentra Series ImmunoCAP \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.781	0%	<5	<100		
<b>(E2)雌二醇</b>							
Beckman UniCel Dx C 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	pmol/L	1988	11%	<5	<100		
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1786	9.9%	<5	<100		
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	pmol/L	1472	9.7%	<5	<100		
<b>(FA)叶酸</b>							
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.64	10.5%	<5	<100		
<b>(FSH)促卵泡刺激素</b>							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	41.5	3.2%	6	<100		
Beckman UniCel Dx C 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/L	35.1	9%	<5	<100		
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	42.7	9.1%	<5	<100		
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	IU/L	43.8	9.8%	<5	<100		
<b>(FT3)游离T3</b>							
Beckman UniCel Dx C 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	27.6	4.5%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	41.5	2.6%	<5	<100		
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	pmol/L	31.3	10.3%	<5	<100		
Beckman UniCel Dx I 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	pmol/L	52.1	10%	<5	<100		
<b>(FT4 )游离T4</b>							
Beckman UniCel Dx C 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	44.5	3.8%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	54.4	7.8%	<5	<100		
Beckman UniCel Dx I 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	pmol/L	31.2	9.9%	<5	<100		

<b>REF</b> 500-3	6x3mL	<b>EXP</b> 2022-03-15	<b>LOT</b> 2FB6E4B	更新日期： 2022-01
				2 / 3

批号：2FB6E4B					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	pmol/L	53.2	9.9%	<5	<100
<b>(hCG)人绒毛膜促性腺激素</b>					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	IU/L	621	3.3%	9	173
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	454	8.8%	6	<100
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	IU/L	756	9.8%	<5	<100
<b>(IgA )免疫球蛋白 IgA</b>					
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	19800	10.1%	<5	<100
<b>(INS)胰岛素</b>					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	mIU/L	99.2	7.7%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	125	9.6%	<5	<100
<b>(LH)促黄体生成素</b>					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	55.2	1.3%	<5	<100
Beckman UniCel Dx C 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/L	37.4	4.8%	<5	<100
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	IU/L	58.6	9.6%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	40.2	10.4%	<5	<100
DiaSorin Liaison /XL Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	65.5	10.1%	<5	<100
<b>(PRL)催/ 泌乳素</b>					
Beckman UniCel Dx C 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	23.3	4.5%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	35.8	10%	<5	<100
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	ng/mL	24.7	10.5%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	35	10%	<5	<100
<b>(PROG)孕酮</b>					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	20.6	5.8%	9	172
Beckman UniCel Dx C 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	22.5	8.2%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24	3.7%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	17.2	10.5%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	23.7	10.5%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.3	10.1%	<5	<100
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	ng/mL	15.6	10.3%	<5	<100
<b>(TES)睾酮</b>					
Beckman UniCel Dx C 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.48	4.1%	<5	<100
J&J Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.98	6.1%	<5	<100
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.91	10%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.46	9.8%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.41	10%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.97	10%	<5	<100
<b>(TSH)促甲状腺素</b>					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	27.7	4%	8	<100
Beckman UniCel Dx C 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	23.1	7.1%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	mIU/L	26.5	10.2%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	25.8	10.1%	<5	<100
<b>(TT3)总三碘甲状腺原氨酸</b>					
Beckman UniCel Dx C 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.02	8.4%	<5	<100

<b>REF</b>	500-3	6x3mL	<b>EXP</b>	2022-03-15	<b>LOT</b>	2FB6E4B	更新日期：	2022-01
								3 / 3

批号：2FB6E4B					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.16	4.6%	<5	<100
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	ng/mL	989	9.9%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	5.64	9%	<5	<100
<b>(TT4)总T4</b>					
Beckman UniCel Dx C 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	162	9.3%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	179	5.5%	<5	<100
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA )	ng/mL	181	9.9%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	123	10.1%	<5	<100
<b>(VB12)维生素B12</b>					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	865	0%	<5	<100
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	0.36	11.1%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.658	9.4%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。