

质控品名： 生化与激素非定值质控品

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
1 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(5'NT)5`核苷酸酶					
Beckman AU series OTHER \ Enzymatic	U/L	3.28	18.9%	7	189
Beckman AU series LEADMAN \ Enzymatic	U/L	1.74	5.6%	<5	<100
Beckman AU series BSBE \ Enzymatic	U/L	6.26	47.7%	35	600
Mindray BS series Dedicated \ Enzymatic	U/L	6.14	5.8%	18	415
TOSHIBA TBA simes-sikma \ Enzymatic	U/L	7.24	9.2%	12	293
(ACE)血管紧张素转化酶					
Abbott ARCHITECT c series BSBE \ OTHER	U/L	53.1	4.5%	25	685
Beckman AU series BSBE \ OTHER	U/L	49.7	3.4%	25	373
(ADA)腺苷脱氨酶					
Abbott ARCHITECT c series BSBE \ Peroxidase	U/L	8.07	23%	8	<100
Beckman AU series KUAKE \ Peroxidase	U/L	9.98	4.5%	5	<100
Beckman AU series OTHER \ Peroxidase	U/L	6.95	25.3%	34	847
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	6.77	11.4%	22	987
Beckman AU series simes-sikma \ Peroxidase	U/L	7.51	14.4%	20	480
Mindray BS series Dedicated \ OTHER	U/L	7.31	3.4%	18	361
TOSHIBA TBA OTHER \ Peroxidase	U/L	7.76	3.8%	11	351
Roche Hitachi/Hitachi Systems chongqiong zybjo \ Peroxidase	U/L	5.02	50.8%	11	226
Beckman AU series SaiKe(Ningbo) \ Trinder reaction	U/L	7.12	3.1%	10	295
Beckman AU series OTHER \ OTHER	U/L	9.05	2.5%	<5	<100
(AFU)α-L-岩藻糖苷酶					
Beckman AU series MAKER \ CNPF	U/L	15.8	4.4%	<5	<100
Beckman AU series Autobio \ OTHER	U/L	9.47	28.6%	38	1009
Beckman AU series BSBE \ OTHER	U/L	18.1	1.7%	25	383
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	18.4	2.5%	24	1041
TOSHIBA TBA OTHER \ CNPF	U/L	15.1	7.5%	11	352
Beckman AU series Reebio \ CNPF	U/L	10.4	4.1%	10	294
TOSHIBA TBA simes-sikma \ CNPF	U/L	7.07	0%	<5	<100
(a-HBDH)α-羟丁酸脱氢酶					
Mindray BS series Mindray \ DGKC	U/L	121	3.2%	8	<100
Beckman AU series OTHER \ OTHER	U/L	125	9.9%	50	1304
Mindray BS series Elikan \ Enzymatic Cycling Assays Method	U/L	115	3.2%	5	<100
Abbott ARCHITECT c series BSBE \ OTHER	U/L	116	5.3%	26	739
Beckman AU series Szybio \ DGKC	U/L	131	6.8%	26	584
Beckman AU series Dedicated \ Enzymatic Cycling Assays Method	U/L	102	3.1%	20	500
Roche Hitachi/Hitachi Systems OTHER \ OTHER	U/L	134	9.9%	20	473
Beckman AU series Reebio \ OTHER	U/L	126	3.5%	20	1158
Mindray BS series OTHER \ OTHER	U/L	137	4.8%	<5	<100
Mindray BS series MAKER \ OTHER	U/L	115	3.8%	<5	<100
Mindray BS series Dedicated \ Enzymatic Cycling Assays Method	U/L	116	7.6%	18	335
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	U/L	114	2.9%	16	442
Beckman AU series LEADMAN \ DGKC	U/L	115	3.9%	15	340
Beckman AU series simes-sikma \ OTHER	U/L	116	4.1%	14	374
Beckman AU series Beckman Coulter \ OTHER	U/L	106	3.1%	14	324
TOSHIBA TBA simes-sikma \ immunoturbidimetric	U/L	127	9.2%	12	300
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	U/L	121	11.6%	12	245
TOSHIBA TBA OTHER \ OTHER	U/L	115	2.9%	11	349

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
2 / 21								

		批号：376B869			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series Elikan \ OTHER	U/L	115	3.2%	11	252
(Alb)白蛋白					
Mindray BS series	g/L	43.6	1.3%	8	<100
Mindray \ Bromocresol green					
Roche Hitachi/Hitachi Systems	g/L	42.8	3.4%	7	101
BIOSINO \ Bromocresol green					
Beckman AU series	g/L	44.3	3.6%	64	2119
Beckman Coulter \ Bromocresol green					
TOSHIBA TBA	g/L	43.2	3.7%	6	146
chongqiong zybio \ OTHER					
Roche Hitachi/Hitachi Systems	g/L	42.8	4.1%	59	1515
MAKER \ Bromocresol green					
Beckman AU series	g/L	42.9	2%	51	1306
Dedicated \ Bromocresol green					
Roche Hitachi/Hitachi Systems	g/L	44	3.4%	34	688
OTHER \ Bromocresol green					
Roche Hitachi/Hitachi Systems	g/L	44.1	1.4%	32	862
LEADMAN \ Bromocresol green					
Abbott ARCHITECT c series	g/L	43.8	1.4%	25	721
BSBE \ Bromocresol green					
Beckman AU series	g/L	44.2	1.7%	25	588
Szybio \ Bromocresol green					
Mindray BS series	g/L	44.4	2.6%	<5	<100
OTHER \ Bromocresol green					
Mindray BS series	g/L	42.1	1.5%	<5	<100
MAKER \ Bromocresol green					
Mindray BS series	g/L	42.4	1.9%	18	383
Dedicated \ Bromocresol green					
Mindray BS series	%	42.8	1.9%	15	890
Dedicated \ Bromocresol green					
Beckman AU series	g/L	43.9	6.3%	15	408
BSBE \ Bromocresol green					
Mindray BS series	g/L	44.6	2%	15	338
Elikan \ Bromocresol green					
Beckman AU series	g/L	43.5	2.4%	15	336
BIOSINO \ Bromocresol green					
Beckman AU series	g/L	44.2	5.3%	14	396
OTHER \ Bromocresol green					
Roche Hitachi/Hitachi Systems	g/L	41.9	1.5%	13	341
chongqiong zybio \ Bromocresol green					
(ALP)碱性磷酸酶					
Mindray BS series	U/L	89.1	2%	8	<100
Mindray \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	78.6	18%	7	150
BIOSINO \ Rate -AMP buffer					
TOSHIBA TBA	U/L	93.3	4.1%	6	147
chongqiong zybio \ OTHER					
Beckman AU series	U/L	95.9	5.5%	51	1330
Dedicated \ Rate -AMP buffer					
Beckman AU series	U/L	95.7	4.8%	45	1050
Beckman Coulter \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	90.7	6.8%	33	699
OTHER \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	97.6	2.2%	32	871
LEADMAN \ Rate -AMP buffer					
Beckman AU series	U/L	93.5	10.2%	27	749
OTHER \ OTHER					
Abbott ARCHITECT c series	U/L	92.1	3.9%	25	725
BSBE \ OTHER					
Beckman AU series	U/L	99.5	4.1%	25	608
Szybio \ OTHER					
Mindray BS series	U/L	92.9	3.9%	<5	<100
OTHER \ Rate -AMP buffer					
Mindray BS series	U/L	89.1	3.1%	<5	<100
MAKER \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	92.4	5.3%	19	518
BSBE \ Rate -AMP buffer					
Mindray BS series	U/L	91	2.5%	18	385
Dedicated \ Rate -AMP buffer					
Beckman AU series	U/L	97.7	2.8%	17	1010
Beckman Coulter \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	92.2	2.4%	15	423
MAKER \ Rate -AMP buffer					
Mindray BS series	U/L	93	3%	15	344
Elikan \ Rate -other buffers					
Beckman AU series	U/L	91.4	4.1%	15	333
BIOSINO \ Rate -AMP buffer					
Roche Hitachi/Hitachi Systems	U/L	94.6	4.1%	15	320
MAKER \ OTHER					
(ALT/ SGPT)丙氨酸氨基转移酶					
TOSHIBA TBA	U/L	42.5	3.1%	6	150
chongqiong zybio \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	44.6	4.2%	57	1498
MAKER \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	44	2.2%	48	1289
LEADMAN \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	40.4	8%	38	790
OTHER \ OTHER					
Beckman AU series	U/L	41.2	4.3%	36	976
Dedicated \ UV-rate without P5P					

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
3 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series	U/L	43.3	4.1%	34	1269
Dedicated \ UV-rate without P5P					
Mindray BS series	U/L	44.4	10%	<5	<100
OTHER \ OTHER					
Beckman AU series	U/L	41.5	5.2%	28	681
Beckman Coulter \ UV-rate with P5P					
Abbott ARCHITECT c series	U/L	41.8	4.3%	26	687
BSBE \ OTHER					
Beckman AU series	U/L	45.5	5%	25	588
Szybio \ OTHER					
Beckman AU series	U/L	39.2	9%	21	562
OTHER \ OTHER					
Mindray BS series	U/L	46	3.2%	<5	<100
MAKER \ OTHER					
Beckman AU series	U/L	41.5	4.5%	18	517
Beckman Coulter \ UV-rate without P5P					
Beckman AU series	U/L	40	6.3%	17	457
BSBE \ OTHER					
Beckman AU series	U/L	40	3.7%	17	1009
Beckman Coulter \ OTHER					
Beckman AU series	U/L	41.5	5%	15	351
Dedicated \ UV-rate with P5P					
Beckman AU series	U/L	43.2	4.3%	15	349
BIOSINO \ UV-rate without P5P					
Mindray BS series	U/L	45	3.4%	15	348
Elikan \ UV-rate without P5P					
Roche Hitachi/Hitachi Systems	U/L	45.7	3.3%	13	340
chongqiong zybio \ UV-rate without P5P					
(ammo)血氨					
Beckman AU series	mmol/L	9	24.3%	<5	<100
LEADMAN \ OTHER					
(AMY)淀粉酶					
Beckman AU series	U/L	94.2	1.4%	7	187
Elikan \ PNP-G7					
Roche Hitachi/Hitachi Systems	U/L	110	0.7%	7	163
LEADMAN \ PNP-G7					
Beckman AU series	U/L	99.8	2.7%	51	1333
Dedicated \ PNP-G7					
Abbott ARCHITECT ci series	U/L	93.4	1.4%	5	249
SaiKe(Ningbo) \ PNP-G7					
Beckman AU series	U/L	61.5	1.9%	5	116
chongqiong zybio \ OTHER					
Mindray BS series	U/L	69	3.3%	<5	<100
Mindray \ PNP-G7					
Beckman AU series	U/L	94.2	5.2%	28	764
Beckman Coulter \ PNP-G7					
Beckman AU series	U/L	92.8	22.1%	20	456
OTHER \ OTHER					
Mindray BS series	U/L	96.1	1.6%	<5	<100
MAKER \ OTHER					
Mindray BS series	U/L	96	1.8%	<5	<100
OTHER \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	92.1	2.1%	19	518
WanTai DRD \ PNP-G7					
Mindray BS series	U/L	103	2.9%	19	360
Dedicated \ PNP-G7					
Beckman AU series	U/L	105	2.1%	17	992
Beckman Coulter \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	100	3.1%	16	371
OTHER \ OTHER					
Mindray BS series	U/L	96.6	12.2%	15	892
Dedicated \ Iodine colorimetry					
Beckman AU series	U/L	93	3.1%	15	345
BIOSINO \ PNP-G7					
Roche Hitachi/Hitachi Systems	U/L	92.7	2.8%	14	377
MAKER \ PNP-G7					
Abbott ARCHITECT c series	U/L	103	1.7%	14	313
BSBE \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	91.3	1.7%	13	343
chongqiong zybio \ OTHER					
(APO E)载脂蛋白E					
Beckman AU series	mg/dL	2661	5.4%	5	144
SaiKe(Ningbo) \ immunoturbidimetric					
(Apo-A I)载脂蛋白A I					
Roche Hitachi/Hitachi Systems	g/L	2.16	2.1%	7	<100
BIOBASE \ immunoturbidimetric					
Mindray BS series	g/L	1.69	1.2%	6	<100
Mindray \ immunoturbidimetric					
Beckman AU series	g/L	1.81	3.1%	<5	<100
Dedicated \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	1.61	12.8%	33	691
OTHER \ immunoturbidimetric					
Mindray BS series	g/L	0.804	1.9%	<5	<100
Mindray \ OTHER					
Roche Hitachi/Hitachi Systems	g/L	1.67	2.9%	24	1003
MAKER \ OTHER					
Beckman AU series	g/L	1.34	23.5%	21	562
OTHER \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	1.77	7.6%	20	529
MAKER \ immunoturbidimetric					

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
4 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series	g/L	1.78	4%	20	1163
Reebio \ immunoturbidimetric					
Mindray BS series	g/L	1.59	2.9%	<5	<100
OTHER \ immunoturbidimetric					
Beckman AU series	g/L	1.54	17%	19	469
BSBE \ immunoturbidimetric					
Mindray BS series	g/L	1.71	3.9%	18	387
Dedicated \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	1.88	1.4%	16	435
LEADMAN \ immunoturbidimetric					
Abbott ARCHITECT c series	g/L	1.7	2.8%	15	<100
BSBE \ immunoturbidimetric					
Mindray BS series	g/L	1.76	2%	15	349
Elkan \ immunoturbidimetric					
Beckman AU series	g/L	1.69	5%	15	339
BIOSINO \ immunoturbidimetric					
Beckman AU series	g/L	1.61	1.4%	13	286
Szybio \ immunoturbidimetric					
TOSHIBA TBA	g/L	1.88	9.2%	12	284
DENUO \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	1.19	32.8%	12	220
BIOSINO \ immunoturbidimetric					
(Apo-B)载脂蛋白B					
Mindray BS series	g/L	0.746	6.5%	8	<100
Mindray \ immunoturbidimetric					
Beckman AU series	g/L	0.645	5.2%	<5	<100
Dedicated \ immunoturbidimetric					
Beckman AU series	g/L	0.632	21.9%	38	1004
OTHER \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.77	11%	33	679
OTHER \ immunoturbidimetric					
Mindray BS series	g/L	0.414	3.1%	<5	<100
Mindray \ OTHER					
Abbott ARCHITECT c series	g/L	0.727	2.8%	26	684
BSBE \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.74	3.1%	24	1047
MAKER \ OTHER					
Roche Hitachi/Hitachi Systems	g/L	0.809	5.6%	20	527
MAKER \ immunoturbidimetric					
Beckman AU series	g/L	0.75	4.7%	20	1148
Reebio \ immunoturbidimetric					
Mindray BS series	g/L	0.782	2.4%	<5	<100
OTHER \ immunoturbidimetric					
Beckman AU series	g/L	0.72	17.7%	<5	<100
BSBE \ immunoturbidimetric					
Mindray BS series	g/L	0.752	8.4%	18	365
Dedicated \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.643	3%	16	435
LEADMAN \ immunoturbidimetric					
Mindray BS series	g/L	0.708	2.3%	15	351
Elkan \ immunoturbidimetric					
Beckman AU series	g/L	0.728	4.5%	15	326
BIOSINO \ immunoturbidimetric					
Beckman AU series	g/L	0.854	2.9%	13	299
Szybio \ immunoturbidimetric					
TOSHIBA TBA	g/L	0.817	7.2%	12	297
DENUO \ immunoturbidimetric					
Roche Hitachi/Hitachi Systems	g/L	0.546	24.2%	12	212
BIOSINO \ immunoturbidimetric					
TOSHIBA TBA	g/L	0.705	2.2%	11	350
OTHER \ OTHER					
(ASO)抗链球菌溶血素 O 抗体					
Beckman AU series	IU/mL	121	21.7%	38	973
OTHER \ Immunonephelometry (INA)					
Beckman AU series	IU/mL	127	3.9%	14	233
LEADMAN \ Immunoturbidimetry (ITA)					
(AST/ SGOT)天门冬氨酸氨基转移酶					
TOSHIBA TBA	U/L	38.3	7%	6	143
chongqiong zybjo \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	38.9	3.3%	58	1480
MAKER \ OTHER					
Beckman AU series	U/L	40	3.7%	51	1327
Dedicated \ UV-rate without P5P					
Abbott ARCHITECT ci series	U/L	31.4	3.3%	5	251
SaiKe(Ningbo) \ UV-rate without P5P					
Beckman AU series	U/L	39.6	8.8%	44	1222
Beckman Coulter \ UV-rate without P5P					
Roche Hitachi/Hitachi Systems	U/L	36.6	8.6%	38	779
OTHER \ OTHER					
Mindray BS series	U/L	38.3	7.3%	33	1256
Dedicated \ UV-rate without P5P					
Roche Hitachi/Hitachi Systems	U/L	35.8	2.7%	32	866
LEADMAN \ OTHER					
Mindray BS series	U/L	36.9	7.4%	<5	<100
OTHER \ OTHER					
Beckman AU series	U/L	37.7	4.4%	28	609
Szybio \ OTHER					
Abbott ARCHITECT c series	U/L	39.1	3%	26	682
BSBE \ OTHER					
Beckman AU series	U/L	39.5	4.6%	20	536
OTHER \ OTHER					

质控品名： 生化与激素非定值质控品

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
5 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series	U/L	36	5.1%	<5	<100
MAKER \ OTHER					
Beckman AU series	U/L	39.8	3.6%	18	989
Beckman Coulter \ OTHER					
Beckman AU series	U/L	38.1	4.9%	17	455
BSBE \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	36.7	2.8%	16	415
LEADMAN \ UV-rate with P5P					
Mindray BS series	U/L	39	4.1%	15	347
Elikan \ UV-rate without P5P					
Beckman AU series	U/L	39.9	4.3%	15	342
BIOSINO \ UV-rate without P5P					
Roche Hitachi/Hitachi Systems	U/L	36.1	5.4%	13	341
chongqiong zybjo \ UV-rate without P5P					
(BUN)尿素氮					
Abbott ARCHITECT ci series	mmol/L	4.23	3.4%	5	250
SaiKe(Ningbo) \ Urease-UV-rate					
Beckman AU series	mmol/L	4.25	4.2%	16	469
Dedicated \ Urease-UV-rate					
Mindray BS series	mmol/L	4.06	2.6%	15	350
Elikan \ Urease-UV-rate					
Beckman AU series	mmol/L	4.22	3%	15	330
BIOSINO \ Urease-UV-rate					
TOSHIBA TBA	mmol/L	4.12	2.7%	11	275
simes-sikma \ Dehydrogenase					
Siemens ADVIA Chemistry XPT	mmol/L	4.36	2.3%	<5	<100
DENUO \ Urease-UV-rate					
(C1q)补体C1q					
Beckman AU series	mg/L	170	3.8%	20	490
BeiJia(Shanghai) \ Immunoturbidimetry (ITA)					
Beckman AU series	mg/L	174	5.5%	13	232
OTHER \ Immunoturbidimetry (ITA)					
(C3)补体 C3					
Beckman AU series	mg/dL	138	1.8%	<5	<100
OTHER \ Immunonephelometry (INA)					
Abbott ARCHITECT c series	mg/dL	136	3.8%	13	309
BSBE \ OTHER					
(C4)补体 C4					
Beckman AU series	mg/dL	23.4	14.9%	17	456
OTHER \ Immunonephelometry (INA)					
Abbott ARCHITECT c series	mg/dL	22.4	5.2%	13	338
BSBE \ OTHER					
(Ca)钙					
Roche Hitachi/Hitachi Systems	mmol/L	2.3	2.6%	59	1990
MAKER \ Arsenazo III					
Abbott ARCHITECT ci series	mmol/L	2.21	2.5%	5	250
LEADMAN \ Arsenazo III					
Beckman AU series	mmol/L	2.23	1.7%	5	155
LEADMAN \ OCPC colorimetry					
Beckman AU series	mmol/L	2.26	2.8%	36	904
Beckman Coulter \ Arsenazo III					
Beckman AU series	mmol/L	2.34	1.7%	35	752
Dedicated \ Arsenazo III					
Beckman AU series	mmol/L	2.19	5.8%	28	606
OTHER \ Arsenazo III					
Abbott ARCHITECT c series	mmol/L	2.21	1.7%	27	743
BSBE \ Arsenazo III					
Beckman AU series	mmol/L	2.32	1.9%	25	605
Szybio \ Arsenazo III					
Beckman AU series	mmol/L	2.29	3.4%	20	479
LEADMAN \ Arsenazo III					
Roche Hitachi/Hitachi Systems	mmol/L	2.21	4.6%	18	422
OTHER \ Arsenazo III					
Mindray BS series	mmol/L	2.34	1.8%	18	333
Dedicated \ Arsenazo III					
Beckman AU series	mmol/L	2.33	1.4%	17	998
Reebio \ Arsenazo III					
Beckman AU series	mmol/L	2.38	2.1%	16	472
Dedicated \ ISE (direct)					
Roche Hitachi/Hitachi Systems	mmol/L	2.31	1.5%	16	433
LEADMAN \ OTHER					
HORRON-H900	mmol/L	2.29	7.3%	15	270
HORRON \ ISE (indirect)					
Roche Hitachi/Hitachi Systems	mmol/L	2.32	4.2%	13	342
LEADMAN \ Arsenazo III					
Mindray BS series	mmol/L	2.15	8.6%	11	<100
Mindray \ Arsenazo III					
TOSHIBA TBA	mmol/L	2.31	1.4%	11	348
MedicalSystem \ Arsenazo III					
Roche cobas 4000/6000/8000	mmol/L	2.27	10.1%	<5	<100
Roche \ BAPTA					
(CG)甘胆酸					
Beckman AU series	mg/L	127	4.2%	5	155
OTHER \ Latex agglutination (LA)					
Beckman AU series	mg/L	1330	7.2%	5	141
OTHER \ OTHER					
Beckman AU series	mg/L	1375	59.1%	<5	<100
BIOSINO \ Latex agglutination (LA)					

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
6 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(CHE)胆碱酯酶					
Mindray BS series	U/L	8047	1.6%	8	<100
Mindray \ Butyrylthiocholine Method					
Roche Hitachi/Hitachi Systems	U/L	7368	11.2%	7	<100
BIOSINO \ Butyrylthiocholine Method					
Beckman AU series	U/L	8678	2.8%	6	159
Elikan \ Butyrylthiocholine Method					
TOSHIBA TBA	U/L	7795	2.4%	6	151
chongqiong zybio \ OTHER					
Abbott ARCHITECT ci series	U/L	7758	1.7%	5	249
BSBE \ Butyrylthiocholine Method					
Beckman AU series	U/L	7647	2.4%	5	153
SaiKe(Ningbo) \ DTNB					
Beckman AU series	U/L	7495	3%	5	145
SaiKe(Ningbo) \ Butyrylthiocholine Method					
Beckman AU series	U/L	8116	7.3%	42	1103
Autobio \ Butyrylthiocholine Method					
Beckman AU series	U/L	8517	3.5%	<5	<100
BSBE \ Butyrylthiocholine Method					
Abbott ARCHITECT c series	U/L	8252	2.2%	25	700
BSBE \ Butyrylthiocholine Method					
Beckman AU series	U/L	7861	2.6%	25	610
Szybio \ Butyrylthiocholine Method					
Beckman AU series	U/L	7927	2.4%	20	501
Dedicated \ Butyrylthiocholine Method					
Mindray BS series	U/L	7898	2.9%	<5	<100
OTHER \ Butyrylthiocholine Method					
Mindray BS series	U/L	7254	1.7%	<5	<100
MAKER \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	8094	2.3%	19	519
BSBE \ OTHER					
Mindray BS series	U/L	7912	3.4%	19	366
Dedicated \ OTHER					
Beckman AU series	U/L	7823	1.8%	18	1007
Beckman Coulter \ Butyrylthiocholine Method					
Roche Hitachi/Hitachi Systems	U/L	7838	1.5%	17	304
OTHER \ Butyrylthiocholine Method					
Roche Hitachi/Hitachi Systems	U/L	7675	0.8%	16	434
LEADMAN \ Butyrylthiocholine Method					
(CHOL/ TC)总胆固醇					
Mindray BS series	mmol/L	3.85	1.3%	8	<100
Mindray \ Cholesterol oxidase					
Roche Hitachi/Hitachi Systems	mmol/L	3.83	6.6%	8	228
BIOSINO \ OTHER					
Beckman AU series	mmol/L	3.8	1.7%	7	186
OTHER \ Cholesterol oxidase					
TOSHIBA TBA	mmol/L	3.78	3.7%	6	177
chongqiong zybio \ OTHER					
Roche Hitachi/Hitachi Systems	mmol/L	3.92	2.1%	45	1189
LEADMAN \ Cholesterol oxidase					
Roche Hitachi/Hitachi Systems	mmol/L	3.87	3.8%	34	934
MAKER \ Cholesterol oxidase					
Beckman AU series	mmol/L	3.88	2.5%	31	1368
Beckman Coulter \ OTHER					
Abbott ARCHITECT c series	mmol/L	3.82	1.6%	26	717
BSBE \ OTHER					
Mindray BS series	mmol/L	3.93	2.9%	<5	<100
OTHER \ OTHER					
Mindray BS series	mmol/L	3.7	1.3%	<5	<100
MAKER \ Cholesterol oxidase					
Beckman AU series	mmol/L	3.91	3.1%	19	374
Dedicated \ Cholesterol oxidase					
Mindray BS series	mmol/L	3.82	2.2%	18	381
Dedicated \ Cholesterol oxidase					
Roche Hitachi/Hitachi Systems	mmol/L	3.87	4.3%	17	371
OTHER \ OTHER					
Roche Hitachi/Hitachi Systems	mmol/L	3.84	1.6%	17	304
OTHER \ Cholesterol oxidase					
Mindray BS series	mmol/L	3.94	1.5%	15	348
Elikan \ Cholesterol oxidase					
Beckman AU series	mmol/L	3.83	4.4%	15	346
LEADMAN \ Cholesterol oxidase					
Beckman AU series	mmol/L	3.83	5.8%	14	395
BSBE \ Cholesterol oxidase					
Beckman AU series	mmol/L	3.97	6%	14	381
OTHER \ OTHER					
Beckman AU series	mmol/L	3.91	1.5%	13	316
Szybio \ OTHER					
(CK)肌酸激酶					
Mindray BS series	U/L	83.8	7.7%	8	<100
Mindray \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	87.6	6.3%	72	1876
MAKER \ OTHER					
Roche Hitachi/Hitachi Systems	U/L	111	4%	6	<100
BIOBASE \ Rate -AMP buffer					
Abbott ARCHITECT ci series	U/L	91.5	5%	5	249
OTHER \ Rate -other buffers					
Roche Hitachi/Hitachi Systems	U/L	96.7	10.7%	33	597
OTHER \ OTHER					
Beckman AU series	U/L	91.2	5%	28	572
Beckman Coulter \ OTHER					

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
7 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT c series BSBE \ OTHER	U/L	84.2	5.5%	25	616
Beckman AU series Szybio \ OTHER	U/L	93.7	6.8%	24	541
Beckman AU series Dedicated \ Rate -other buffers	U/L	109	4.9%	20	499
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	U/L	98	4.5%	<5	<100
Mindray BS series OTHER \ OTHER	U/L	84.8	5.2%	<5	<100
Mindray BS series MAKER \ Rate -AMP buffer	U/L	104	2.8%	<5	<100
Beckman AU series Reebio \ OTHER	U/L	87.9	3.6%	17	987
Beckman AU series Dedicated \ Rate -AMP buffer	U/L	88.1	5.1%	16	457
Mindray BS series Dedicated \ Rate -AMP buffer	U/L	75.2	9.5%	16	321
Mindray BS series Dedicated \ Rate -other buffers	U/L	87.1	4.7%	15	855
Beckman AU series OTHER \ OTHER	U/L	100	12.6%	15	410
Mindray BS series Elikan \ Rate -other buffers	U/L	104	4%	15	340
Beckman AU series Dedicated \ OTHER	U/L	88.8	5.2%	15	320
(CK-MB mass)肌酸激酶-MB质量					
Abbott ARCHITECT ci series BSBE \ OTHER	ng/mL	9.68	7.5%	5	251
Beckman AU series BSBE \ OTHER	ng/mL	10	22.8%	26	679
Mindray BS series BSBE \ OTHER	ng/mL	14	9.2%	13	773
Roche Hitachi/Hitachi Systems G-CELL \ OTHER	ng/mL	6.75	28.6%	12	276
(CK-MB)肌酸激酶-MB (活性)					
Mindray BS series Mindray \ Immune suppression	U/L	14.7	14.7%	8	<100
Beckman AU series Dedicated \ Immune suppression	U/L	17.4	13.5%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Immune suppression	U/L	14.1	9%	38	1440
Roche Hitachi/Hitachi Systems OTHER \ Immune suppression	U/L	19.1	22.3%	33	668
Roche Hitachi/Hitachi Systems LEADMAN \ Immune suppression	U/L	16.2	4.9%	<5	<100
Mindray BS series BIOSINO \ Immune suppression	U/L	16.1	5.4%	<5	<100
Mindray BS series Mindray \ Immune suppression	ng/mL	4.4	10.5%	<5	<100
Beckman AU series OTHER \ Immune suppression	U/L	11.5	12.8%	29	614
Beckman AU series Szybio \ Immune suppression	U/L	15.6	14.6%	29	593
Abbott ARCHITECT c series BSBE \ OTHER	ng/mL	10.1	19%	25	666
Beckman AU series BSBE \ Immune suppression	U/L	11.3	22.2%	24	346
Beckman AU series Reebio \ Immune suppression	U/L	18.3	7.7%	20	1150
Beckman AU series BSBE \ OTHER	ng/mL	11.2	7%	<5	<100
Mindray BS series MAKER \ Immune suppression	U/L	15	11.9%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	14.3	11.1%	19	519
Mindray BS series Dedicated \ Immune suppression	U/L	9.52	22.8%	18	364
Roche Hitachi/Hitachi Systems MAKER \ DGKC	U/L	12	13.6%	16	411
TOSHIBA TBA simes-sikma \ Immune suppression	U/L	14.3	10.2%	12	311
TOSHIBA TBA BSBE \ OTHER	ng/mL	10.9	10.1%	11	289
(CI)氯					
Mindray BS series Mindray \ ISE (indirect)	mmol/L	118	1.2%	8	<100
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	117	1.9%	68	2227
Beckman AU series Dedicated \ ISE (direct)	mmol/L	117	1.4%	51	1382
TOSHIBA TBA SHXD \ ISE (indirect)	mmol/L	114	4%	5	<100
Abbott ARCHITECT ci series Abbott \ ISE (direct)	mmol/L	119	0.8%	5	250
Roche Hitachi/Hitachi Systems OTHER \ ISE (direct)	mmol/L	113	3.7%	31	730
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	117	5.7%	27	783
Beckman AU series OTHER \ ISE (direct)	mmol/L	119	4.8%	27	593

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
8 / 21								

		批号：376B869			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT c series BSBE \ ISE (direct)	mmol/L	118	1%	26	753
Beckman AU series Szybio \ ISE (indirect)	mmol/L	117	0.7%	25	597
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	111	1.5%	24	606
Mindray BS series Dedicated \ OTHER	mmol/L	118	1.5%	18	404
Roche Hitachi/Hitachi Systems OTHER \ ISE (indirect)	mmol/L	113	2.3%	17	303
Roche Hitachi/Hitachi Systems Wako \ ISE (indirect)	mmol/L	112	1%	16	430
HORRON-H900 HORRON \ ISE (indirect)	mmol/L	120	10.6%	16	249
Mindray BS series Dedicated \ ISE (indirect)	mmol/L	118	1.1%	15	881
Roche Hitachi/Hitachi Systems LEADMAN \ Mercuric thiocyanate colorimetry	mmol/L	113	2.1%	13	340
TOSHIBA TBA OTHER \ ISE (direct)	mmol/L	118	2.6%	12	278
TOSHIBA TBA OTHER \ ISE (indirect)	mmol/L	117	1.1%	11	348
(CO2 (HCO3-))二氧化碳 (碳酸氢根)					
Beckman AU series Dedicated \ PEPC	mmol/L	0.66	0%	<5	<100
(Cr)肌酐					
Abbott ARCHITECT ci series OTHER \ Enzymatic	μmol/L	71.4	3.2%	5	249
Beckman AU series chongqiong zybio \ OTHER	μmol/L	65.4	6.2%	5	118
Beckman AU series OTHER \ OTHER	μmol/L	66.9	13.4%	44	955
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	μmol/L	69.9	3.4%	32	851
Beckman AU series Dedicated \ Enzymatic	μmol/L	73	4.3%	31	796
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	73.6	5%	29	639
Beckman AU series Szybio \ OTHER	μmol/L	75.8	2.5%	28	598
Abbott ARCHITECT c series BSBE \ OTHER	μmol/L	76.3	2.3%	26	699
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	71.8	3.2%	24	560
Beckman AU series Randox \ Enzymatic	μmol/L	72.2	4.6%	20	486
Roche Hitachi/Hitachi Systems OTHER \ Enzymatic	μmol/L	78.7	7.8%	20	484
Mindray BS series OTHER \ Enzymatic	μmol/L	69.7	7%	<5	<100
Mindray BS series MAKER \ OTHER	μmol/L	77	1.3%	<5	<100
Roche Hitachi/Hitachi Systems Whitman \ Enzymatic	μmol/L	77.6	3.4%	19	518
Mindray BS series Dedicated \ Picric acid-Blank Compensation	μmol/L	75.8	2.8%	18	355
Beckman AU series OTHER \ Enzymatic	μmol/L	75.1	2.3%	17	486
Beckman AU series BSBE \ Enzymatic	μmol/L	76	3.8%	17	452
Roche Hitachi/Hitachi Systems OTHER \ OTHER	μmol/L	72.2	2.3%	17	304
Beckman AU series Reebio \ OTHER	μmol/L	73.7	1.9%	17	1026
(CRP)C反应蛋白					
Beckman AU series DiaSys \ Immunoturbidimetry (ITA)	mg/dL	0.586	5.1%	7	170
Beckman AU series OTHER \ Immunonephelometry (INA)	mg/dL	0.672	13.8%	13	264
Beckman Image 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	0.638	8.3%	<5	<100
Roche cobas E601/602 Roche \ Immunonephelometry (INA)	mg/dL	0.923	10%	<5	<100
(Cu)铜					
Beckman AU series LEADMAN \ PAESA	μmol/L	17.9	4.3%	17	163
Roche Hitachi/Hitachi Systems Roche \ ICP-MS	μmol/L	19.6	9.7%	<5	<100
(Cys C)半胱氨酸蛋白酶抑制剂C/ 胱抑素C					
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mg/L	0.088	16.5%	7	125
Beckman AU series Szybio \ Immunoturbidimetry (ITA)	mg/L	0.468	18.3%	6	182
Beckman AU series chongqiong zybio \ OTHER	mg/L	0.097	37.1%	5	<100
Mindray BS series Elisan \ Immunoturbidimetry (ITA)	mg/L	0.418	11.3%	5	<100
Beckman AU series LEADMAN \ Immunoturbidimetry (ITA)	mg/L	0.272	6.5%	5	<100

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
9 / 21								

批号：376B869					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series MedicalSystem \ Immunoturbidimetry (ITA)	mg/L	0.334	7.2%	5	143
Beckman AU series OTHER \ OTHER	mg/L	0.225	43.1%	27	639
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	mg/L	0.277	14.2%	24	1018
Mindray BS series Dedicated \ Immunonephelometry (INA)	mg/L	0.217	12.3%	19	399
Beckman AU series Szybio \ Chemiluminescence enzyme immunoassay (CLEIA)	mg/L	0.439	4.6%	12	294
(DBIL)直接胆红素					
TOSHIBA TBA Szybio \ Vanadate	μmol/L	5	30.8%	9	543
Roche Hitachi/Hitachi Systems BIOBASE \ Vanadate	μmol/L	6.76	2.2%	8	<100
Roche Hitachi/Hitachi Systems MAKER \ Vanadate	μmol/L	4.88	10.5%	35	936
Roche Hitachi/Hitachi Systems OTHER \ Vanadate	μmol/L	5.82	11.3%	34	695
Mindray BS series Dedicated \ Vanadate	μmol/L	5.83	14.2%	33	1279
Beckman AU series Dedicated \ Vanadate	μmol/L	3.97	10%	31	775
Beckman AU series Beckman Coulter \ Vanadate	μmol/L	3.43	10.3%	29	512
Abbott ARCHITECT c series BSBE \ Vanadate	μmol/L	6.55	4.3%	26	697
Beckman AU series Szybio \ OTHER	μmol/L	5.94	14%	25	571
Beckman AU series OTHER \ OTHER	μmol/L	6.5	9.1%	20	538
Beckman AU series Dedicated \ Diazotization -Blank Compensation	μmol/L	4.01	4.6%	20	499
Mindray BS series OTHER \ Vanadate	μmol/L	7.42	2.4%	<5	<100
Mindray BS series MAKER \ OTHER	μmol/L	7.56	1.6%	<5	<100
Beckman AU series Beckman Coulter \ Diazotization -Blank Compensation	μmol/L	3.81	14.2%	18	523
Beckman AU series BSBE \ Vanadate	μmol/L	6.44	5.4%	18	461
Beckman AU series Reebio \ OTHER	μmol/L	4.29	6.1%	17	998
Roche Hitachi/Hitachi Systems LEADMAN \ Vanadate	μmol/L	7.12	2.1%	16	438
Mindray BS series Elikan \ Vanadate	μmol/L	4.44	5.1%	15	345
Beckman AU series LEADMAN \ Vanadate	μmol/L	6.59	5.1%	15	292
(DGXN)地高辛					
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.734	11.7%	<5	<100
(E2)雌二醇					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	113	11.4%	<5	<100
Mindray BS series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	127	0%	<5	<100
(Fe)铁					
Mindray BS series Mindray \ Ferrozine colorimetry	μmol/L	17.8	1.7%	7	<100
Beckman AU series OTHER \ TPTZ	μmol/L	18.3	0.8%	5	154
Abbott ARCHITECT c series LEADMAN \ Ferrozine colorimetry	μmol/L	18.7	7.8%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	μmol/L	18.7	1.9%	25	1202
Beckman AU series BSBE \ Ferrozine colorimetry	μmol/L	18.7	2.4%	23	348
Beckman AU series Autobio \ Ferrozine colorimetry	μmol/L	19	7.9%	22	583
Beckman AU series Dedicated \ TPTZ	μmol/L	18.7	2.1%	20	400
Beckman AU series Reebio \ Ferrozine colorimetry	μmol/L	17.6	6.2%	<5	<100
Mindray BS series Dedicated \ OTHER	μmol/L	18.2	7.4%	18	376
Roche Hitachi/Hitachi Systems LEADMAN \ Ferrozine colorimetry	μmol/L	18.4	1.1%	16	434
Abbott ARCHITECT c series BSBE \ OTHER	μmol/L	18.6	4.4%	15	355
Beckman AU series Dedicated \ Ferrozine colorimetry	μmol/L	17.1	3.3%	15	348
Beckman AU series Szybio \ Ferrozine colorimetry	μmol/L	18.3	1.5%	13	313
TOSHIBA TBA XJKC \ Ferrozine colorimetry	μmol/L	18.7	1.9%	12	351
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	19.7	9.6%	<5	<100

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
10 / 21								

批号：376B869					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	19.5	9.7%	<5	<100
(Ferritin)铁蛋白					
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	ng/mL	51.7	6.2%	10	293
(FN)纤维结合蛋白					
Beckman AU series OTHER \ Immunoturbidimetry (ITA)	mg/L	211	2.1%	5	143
(FRU)果糖					
TOSHIBA TBA simes-sikma \ Enzymatic	mmol/L	2.33	8.8%	12	304
(FT3)游离T3					
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	3.37	11.3%	<5	<100
Mindray BS series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	2.64	5.3%	<5	<100
(FT4)游离T4					
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	17.3	9.2%	<5	<100
Mindray BS series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	15.6	0.5%	<5	<100
(GA)糖化血清白蛋白					
Beckman AU series OTHER \ OTHER	g/dL	17.1	5%	13	346
(GGT)L-γ-谷氨酰基转移酶					
Mindray BS series Mindray \ OTHER	U/L	37.8	1%	7	<100
TOSHIBA TBA chongqiong zybjo \ OTHER	U/L	34.9	4.1%	6	146
Roche Hitachi/Hitachi Systems OTHER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	36.9	8.8%	38	774
Beckman AU series OTHER \ OTHER	U/L	36.1	9%	37	1003
Beckman AU series Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	38.3	2.8%	35	848
Roche Hitachi/Hitachi Systems MAKER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	35.9	2.6%	34	941
Beckman AU series Beckman Coulter \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.2	4.3%	34	925
Mindray BS series Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	38.6	1.8%	33	1271
Roche Hitachi/Hitachi Systems LEADMAN \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	34.6	1.5%	32	878
Mindray BS series Mindray \ γ-glutamyl -4-nitroaniline	U/L	33.8	2.8%	<5	<100
Beckman AU series Szybio \ OTHER	U/L	35.6	5.4%	28	566
Abbott ARCHITECT c series BSBE \ OTHER	U/L	38.3	2.2%	26	724
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	38.7	2.4%	24	580
Mindray BS series OTHER \ OTHER	U/L	36.7	4.5%	<5	<100
Mindray BS series MAKER \ OTHER	U/L	37.1	1.7%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	U/L	39	2.1%	17	1009
Beckman AU series Dedicated \ γ-glutamyl -4-nitroaniline	U/L	39.1	2.9%	16	468
Mindray BS series Elikan \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.5	1.8%	15	338
Beckman AU series BJBJ \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	36.7	1.9%	15	334
(GLDH)谷氨酸脱氢酶					
Beckman AU series Beckman Coulter \ Direct Measure	U/L	5.28	11.9%	<5	<100
(GLU)葡萄糖					
Roche Hitachi/Hitachi Systems BIOBASE \ Glucose oxidase (GOD)	mmol/L	4.38	1.5%	9	143
Beckman AU series OTHER \ Hexokinase	mmol/L	4.05	3.7%	63	1449
Beckman AU series Dedicated \ Hexokinase	mmol/L	4.12	2.5%	51	1309
Abbott ARCHITECT ci series SaiKe(Ningbo) \ Glucose oxidase (GOD)	mmol/L	4.2	1%	5	250
Beckman AU series chongqiong zybjo \ Hexokinase	mmol/L	4.04	1.6%	5	118
TOSHIBA TBA chongqiong zybjo \ Hexokinase	mmol/L	3.87	3.1%	5	112
Roche Hitachi/Hitachi Systems MAKER \ Hexokinase	mmol/L	3.88	2%	39	992
Beckman AU series Beckman Coulter \ Hexokinase	mmol/L	4.06	3.5%	35	917

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
11 / 21								

批号：376B869					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	mmol/L	4.01	1.4%	32	867
Roche Hitachi/Hitachi Systems LEADMAN \ Hexokinase	mmol/L	3.98	1.8%	30	758
Beckman AU series Szybio \ Hexokinase	mmol/L	4.01	1.7%	28	601
Abbott ARCHITECT c series BSBE \ Glucose oxidase (GOD)	mmol/L	3.98	1.4%	26	684
Roche Hitachi/Hitachi Systems OTHER \ Glucose oxidase (GOD)	mmol/L	4.17	8.3%	20	489
Mindray BS series OTHER \ Glucose oxidase (GOD)	mmol/L	3.76	2.7%	<5	<100
Mindray BS series MAKER \ Glucose oxidase (GOD)	mmol/L	3.79	1.5%	<5	<100
Roche Hitachi/Hitachi Systems WanTai DRD \ Hexokinase	mmol/L	3.97	1.4%	19	515
Mindray BS series Dedicated \ Glucose oxidase (GOD)	mmol/L	3.88	1.5%	18	360
Roche Hitachi/Hitachi Systems OTHER \ Hexokinase	mmol/L	4.04	1.8%	17	301
Beckman AU series Reebio \ OTHER	mmol/L	4.2	1.5%	17	1005
(GR)谷胱甘肽还原酶					
Beckman AU series BSBE \ OTHER	U/L	42.3	16.2%	38	954
Mindray BS series Dedicated \ Enzymatic	U/L	49.4	9.9%	17	314
Beckman AU series chongqiong zybjo \ Enzymatic	U/L	52.3	5.9%	11	102
(HCY)同型半胱氨酸					
Mindray BS series Mindray \ Enzymatic Cycling Assays Method	μmol/L	13.5	4.2%	8	<100
Beckman AU series OTHER \ Enzymatic Cycling Assays Method	μmol/L	11.8	17.9%	53	1238
Beckman AU series MedicalSystem \ Immunoturbidimetry (ITA)	μmol/L	13.6	3.9%	5	141
Beckman AU series BSBE \ Enzymatic Cycling Assays Method	μmol/L	11.7	5.9%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic Cycling Assays Method	μmol/L	13.9	7.2%	34	823
Roche Hitachi/Hitachi Systems LEADMAN \ Enzymatic Cycling Assays Method	μmol/L	15	4.1%	32	769
Abbott ARCHITECT c series BSBE \ Enzymatic Cycling Assays Method	μmol/L	12.6	3.1%	26	676
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	14.1	5.4%	21	1193
Beckman AU series Szybio \ Enzymatic Cycling Assays Method	μmol/L	12.9	7.2%	20	504
Roche Hitachi/Hitachi Systems OTHER \ Enzymatic Cycling Assays Method	μmol/L	15.5	16.8%	17	301
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	14.1	4.9%	15	541
Roche Hitachi/Hitachi Systems BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	12.4	18%	12	335
TOSHIBA TBA WanTai \ OTHER	μmol/L	14.6	3.8%	11	354
Roche Hitachi/Hitachi Systems BIOBASE \ OTHER	μmol/L	14	3.9%	10	178
Roche Hitachi/Hitachi Systems Roche \ Enzymatic Cycling Assays Method	μmol/L	15.3	10.5%	<5	<100
Mindray BS series Mindray \ Enzyme linked immunosorbent assays (ELISA)	μmol/L	12.7	10.2%	<5	<100
Siemens ADVIA Centaur CP Siemens \ Enzymatic Cycling Assays Method	μmol/L	14.6	10.3%	<5	<100
Beckman AU series SHRKSU \ Enzymatic Cycling Assays Method	μmol/L	12.8	7.4%	<5	<100
TOSHIBA TBA simes-sikma \ Enzymatic Cycling Assays Method	μmol/L	15.6	2.7%	<5	<100
(HDL-C)高密度脂蛋白胆固醇					
Beckman AU series Elikan \ Direct Measure	mmol/L	1.12	4.3%	6	185
Roche Hitachi/Hitachi Systems LEADMAN \ selected suppression	mmol/L	1.11	2.2%	48	1285
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	0.985	7.5%	43	1088
Beckman AU series OTHER \ selected suppression	mmol/L	0.895	19.7%	39	1039
Roche Hitachi/Hitachi Systems OTHER \ selected suppression	mmol/L	1.13	13.5%	38	785
Beckman AU series Beckman Coulter \ OTHER	mmol/L	1.01	4.2%	31	1343
TOSHIBA TBA chongqiong zybjo \ Enzymatic	mmol/L	0.909	6.4%	<5	<100
Abbott ARCHITECT c series BSBE \ selected suppression	mmol/L	1.06	1.6%	26	731
Mindray BS series MAKER \ Enzymatic	mmol/L	1.22	1.5%	<5	<100
Beckman AU series Dedicated \ Direct Measure	mmol/L	1.02	6%	19	377
Mindray BS series Dedicated \ Direct Measure	mmol/L	1.11	2.4%	18	387

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
12 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems MAKER \ Direct Measure	mmol/L	1.06	3.2%	15	404
Mindray BS series Elikan \ Direct Measure	mmol/L	1.09	2.6%	15	349
Beckman AU series BJBJ \ selected suppression	mmol/L	1.07	5.6%	15	337
Roche Hitachi/Hitachi Systems LEADMAN \ Enzymatic	mmol/L	1.09	2.2%	13	339
Beckman AU series Szybio \ selected suppression	mmol/L	0.943	2%	13	314
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	1.06	24%	12	389
TOSHIBA TBA simes-sikma \ selected suppression	mmol/L	1.21	9.8%	12	317
Roche Hitachi/Hitachi Systems BIOBASE \ Direct Measure	mmol/L	1.07	3.3%	12	181
(H-FABP)心型脂肪酸结合蛋白					
Roche Hitachi/Hitachi Systems chongqiong zybio \ immunoturbidimetric	ng/mL	1.88	41.7%	9	182
Beckman AU series OTHER \ immunoturbidimetric	ng/mL	1.35	14.2%	20	497
(hs-CRP)超敏C反应蛋白					
Beckman AU series OTHER \ Immunonephelometry (INA)	mg/L	7.13	4.8%	5	154
Beckman AU series OTHER \ Immunoturbidimetry (ITA)	mg/L	7.28	4%	5	148
Abbott ARCHITECT c series BSBE \ OTHER	mg/L	7.06	4.2%	26	686
Beckman AU series Beckman Coulter \ OTHER	mg/L	6.82	3.9%	17	1039
Roche Hitachi/Hitachi Systems MAKER \ Immunonephelometry (INA)	mg/L	8.07	4.3%	12	281
(IgA)免疫球蛋白 IgA					
Beckman AU series OTHER \ OTHER	mg/dL	204	7.3%	9	248
Roche Hitachi/Hitachi Systems OTHER \ Immunoturbidimetry (ITA)	mg/dL	203	5.6%	17	302
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	207	3.5%	12	319
TOSHIBA TBA DiaSys \ Immunoturbidimetry (ITA)	mg/dL	211	2.2%	11	347
Beckman Image 800 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	193	9.8%	<5	<100
(IgE)免疫球蛋白 IgE					
Beckman AU series Autobio \ Immunonephelometry (INA)	IU/mL	93.9	26.8%	17	445
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	IU/mL	91.9	8.8%	11	336
(IgG)免疫球蛋白 IgG					
Roche Hitachi/Hitachi Systems OTHER \ Immunoturbidimetry (ITA)	mg/dL	1005	9%	17	300
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	1067	4.4%	13	323
TOSHIBA TBA DiaSys \ Immunoturbidimetry (ITA)	mg/dL	998	3.8%	11	348
Beckman AU series OTHER \ OTHER	mg/dL	867	17.3%	11	285
Beckman Image 800 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	1050	12%	<5	<100
(IgM)免疫球蛋白 IgM					
Beckman AU series OTHER \ OTHER	mg/dL	87.9	9.1%	9	244
Roche Hitachi/Hitachi Systems OTHER \ Immunoturbidimetry (ITA)	mg/dL	86.8	5%	17	302
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	90.2	4.2%	13	324
TOSHIBA TBA DiaSys \ Immunoturbidimetry (ITA)	mg/dL	85.7	2.7%	11	344
Beckman Image 800 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	89	9%	<5	<100
(IMA)缺血修饰白蛋白					
Beckman AU series Autobio \ OTHER	IU/mL	88.2	9.2%	21	480
Roche Hitachi/Hitachi Systems OTHER \ OTHER	IU/mL	68.6	6.4%	19	512
Beckman AU series OTHER \ OTHER	IU/mL	64.1	5.2%	17	991
(IV-Col)IV型胶原					
Beckman AU series Bio-assay \ OTHER	μg/L	22.4	24.1%	31	805
(K)钾					
Mindray BS series Mindray \ ISE (indirect)	mmol/L	3.7	1.5%	8	<100
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	3.67	2%	52	1950

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
13 / 21								

		批号：376B869			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Dedicated \ ISE (direct)	mmol/L	3.69	1.4%	51	1414
TOSHIBA TBA SHXD \ ISE (direct)	mmol/L	3.7	1.1%	5	<100
Abbott ARCHITECT ci series Abbott \ ISE (direct)	mmol/L	3.68	1%	5	251
Roche Hitachi/Hitachi Systems OTHER \ ISE (direct)	mmol/L	3.64	2%	32	739
Beckman AU series OTHER \ ISE (direct)	mmol/L	3.7	3%	28	659
Abbott ARCHITECT c series BSBE \ ISE (direct)	mmol/L	3.65	1.6%	26	768
Beckman AU series Szybio \ ISE (indirect)	mmol/L	3.76	3.4%	25	577
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	3.64	1.3%	24	602
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	3.73	1.5%	19	511
Mindray BS series Dedicated \ OTHER	mmol/L	3.79	1.6%	18	416
Roche Hitachi/Hitachi Systems OTHER \ ISE (indirect)	mmol/L	3.63	1.2%	17	305
Roche Hitachi/Hitachi Systems Wako \ ISE (indirect)	mmol/L	3.6	1.1%	16	433
HORRON-H900 HORRON \ ISE (indirect)	mmol/L	3.81	7.9%	16	298
Mindray BS series Dedicated \ ISE (indirect)	mmol/L	3.89	1.4%	15	877
Beckman AU series LEADMAN \ ISE (indirect)	mmol/L	3.64	1%	15	333
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	mmol/L	3.7	3.2%	13	342
TOSHIBA TBA OTHER \ ISE (direct)	mmol/L	3.92	2.8%	12	265
(LA)乳酸					
Beckman AU series Dedicated \ Lactate to pyruvate (L-P)	mmol/L	1.98	3.3%	36	865
Beckman AU series Beckman Coulter \ OTHER	mmol/L	1.88	3.6%	27	733
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	mmol/L	1.89	2.1%	19	515
Beckman AU series Dedicated \ OTHER	mmol/L	1.97	1.9%	15	353
Roche cobas E601/602 Roche \ Lactate to pyruvate (L-P)	mmol/L	2.21	9.5%	<5	<100
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	1.94	9.8%	<5	<100
(LAP)亮氨酸氨基肽酶					
Roche Hitachi/Hitachi Systems OTHER \ OTHER	U/L	32.1	3%	24	1046
(LDH)乳酸脱氢酶					
Mindray BS series Mindray \ OTHER	U/L	138	2.4%	8	<100
Abbott ARCHITECT ci series SaiKe(Ningbo) \ Lactate to pyruvate (L-P)	U/L	136	1.5%	5	249
Beckman AU series Beckman Coulter \ OTHER	U/L	131	5.2%	45	1667
Roche Hitachi/Hitachi Systems OTHER \ Lactate to pyruvate (L-P)	U/L	137	7.3%	37	776
Beckman AU series Dedicated \ Lactate to pyruvate (L-P)	U/L	134	6.8%	36	977
Mindray BS series Dedicated \ Lactate to pyruvate (L-P)	U/L	143	5.5%	33	1214
Mindray BS series OTHER \ Lactate to pyruvate (L-P)	U/L	130	6.2%	<5	<100
Mindray BS series Mindray \ Lactate to pyruvate (L-P)	U/L	152	4.1%	<5	<100
Abbott ARCHITECT c series BSBE \ OTHER	U/L	133	1.9%	25	680
Beckman AU series Szybio \ OTHER	U/L	133	3.5%	25	582
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	136	2.2%	24	1055
Beckman AU series OTHER \ OTHER	U/L	139	4.7%	21	499
Mindray BS series MAKER \ OTHER	U/L	142	2.4%	<5	<100
Roche Hitachi/Hitachi Systems Whitman \ Lactate to pyruvate (L-P)	U/L	137	2.5%	19	515
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	U/L	132	2.5%	16	440
Roche Hitachi/Hitachi Systems MAKER \ Lactate to pyruvate (L-P)	U/L	132	6.5%	15	421
Mindray BS series Elikan \ Lactate to pyruvate (L-P)	U/L	129	3.8%	15	351
Beckman AU series LEADMAN \ Lactate to pyruvate (L-P)	U/L	131	3.7%	15	325
Beckman AU series Dedicated \ OTHER	U/L	130	5%	15	323

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
14 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(LDH-I)乳酸脱氢酶同工酶1					
Roche Hitachi/Hitachi Systems WanTai DRD \ OTHER	U/L	65.3	10.3%	<5	107
Beckman AU series Reebio \ OTHER	U/L	72.9	3.9%	17	1003
(LDL-C)低密度脂蛋白胆固醇					
Beckman AU series OTHER \ Direct Measure	mmol/L	2.26	2%	7	187
Roche Hitachi/Hitachi Systems LEADMAN \ SUR	mmol/L	2.15	6%	61	1629
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	2.22	8.3%	55	1405
Beckman AU series OTHER \ SUR	mmol/L	1.86	24.6%	38	995
Abbott ARCHITECT c series BSBE \ SUR	mmol/L	1.94	1.8%	26	701
Mindray BS series MAKER \ Enzymatic	mmol/L	1.96	1.8%	<5	<100
Mindray BS series Dedicated \ Direct Measure	mmol/L	2.31	2.9%	19	392
Beckman AU series Dedicated \ Direct Measure	mmol/L	2.54	3.3%	19	378
Roche Hitachi/Hitachi Systems OTHER \ SUR	mmol/L	2.07	3.6%	17	302
Beckman AU series Reebio \ OTHER	mmol/L	2.11	2.1%	17	1004
Mindray BS series Elkan \ Direct Measure	mmol/L	2.26	2%	15	350
Beckman AU series BIOSINO \ selected suppression	mmol/L	2.19	5.8%	15	338
Beckman AU series Beckman Coulter \ OTHER	mmol/L	2.61	2.8%	14	377
Beckman AU series Szybio \ Direct Measure	mmol/L	2.13	2.5%	13	303
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	mmol/L	1.6	29.8%	12	401
TOSHIBA TBA simes-sikma \ SUR	mmol/L	1.86	4.9%	12	316
Roche Hitachi/Hitachi Systems BIOBASE \ Direct Measure	mmol/L	2.16	2.4%	12	218
Mindray BS series Mindray \ OTHER	mmol/L	1.88	31.8%	11	<100
TOSHIBA TBA OTHER \ SUR	mmol/L	2.28	1.5%	11	351
(Li)锂					
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	0.82	11.6%	<5	<100
(LIP)脂肪酶					
Beckman AU series Beckman Coulter \ OTHER	U/L	59.6	4%	6	228
Beckman AU series chongqiong zybio \ OTHER	U/L	22.4	4.5%	5	118
Beckman AU series Reebio \ OTHER	U/L	57.2	11.1%	31	1206
Mindray BS series Mindray \ OTHER	U/L	29.4	5%	<5	<100
Beckman AU series OTHER \ OTHER	U/L	49.7	35.2%	23	524
Beckman AU series Dedicated \ Enzymatic Cycling Assays Method	U/L	59.4	3%	20	495
Mindray BS series Elkan \ OTHER	U/L	65.7	6.3%	<5	<100
Mindray BS series Dedicated \ Dehydrogenase	U/L	44	14.8%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ OTHER	U/L	79.5	4%	19	518
Abbott ARCHITECT c series BSBE \ OTHER	U/L	80.6	3.9%	13	371
TOSHIBA TBA Reebio \ OTHER	U/L	53	4.4%	12	308
Beckman AU series Szybio \ Dehydrogenase	U/L	42.2	8.5%	11	272
Mindray BS series OTHER \ OTHER	U/L	57.2	8.6%	10	219
Roche cobas E601/602 Roche \ Enzymatic Cycling Assays Method	U/L	43.9	10.1%	<5	<100
Roche Hitachi/Hitachi Systems BIOBASE \ OTHER	U/L	67.6	1.6%	<5	<100
Beckman AU series Beckman Coulter \ Glucose oxidase (GOD)	U/L	52.1	10.1%	<5	<100
(Lp(a))脂蛋白a					
Beckman AU series MedicalSystem \ Immunoturbidimetry (ITA)	mg/L	34.1	6.1%	5	154
Beckman AU series MedicalSystem \ Immunonephelometry (INA)	mg/L	38.8	5.6%	5	140
Beckman AU series BSBE \ Immunoturbidimetry (ITA)	mg/L	95.4	7.6%	<5	<100
Beckman AU series OTHER \ OTHER	mg/L	56.2	12.5%	38	998

质控品名： 生化与激素非定值质控品

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
15 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series BSBE \ OTHER	mg/L	66.3	10.8%	24	343
Beckman AU series Szybio \ Immunoturbidimetry (ITA)	mg/L	69.2	6.4%	16	322
Roche Hitachi/Hitachi Systems chongqiong zybjo \ OTHER	mg/L	48.1	11%	12	265
(Lp-PLA2)脂蛋白磷脂酶A2					
Beckman AU series OTHER \ OTHER	ng/mL	93.1	13%	5	105
Beckman AU series BJBJ \ Rate -other buffers	ng/mL	105	14.8%	12	141
(MAO)单胺氧化酶					
Beckman AU series OTHER \ OTHER	U/L	1.02	28.7%	6	121
Beckman AU series HOMA \ Colorimetry	U/L	0.755	23.9%	<5	<100
(Mg)镁					
Beckman AU series Dedicated \ Xylidyl Blue	mmol/L	0.887	3.6%	51	1232
Beckman AU series chongqiong zybjo \ Xylidyl Blue	mmol/L	0.679	3.2%	5	118
Beckman AU series OTHER \ Xylidyl Blue	mmol/L	0.804	14%	44	1090
Beckman AU series Beckman Coulter \ Xylidyl Blue	mmol/L	0.877	4.9%	30	743
Beckman AU series BSBE \ Xylidyl Blue	mmol/L	0.861	5.7%	23	392
Roche Hitachi/Hitachi Systems MAKER \ Calmagite colorimetry	mmol/L	0.903	4.1%	19	505
Mindray BS series Dedicated \ Xylidyl Blue	mmol/L	0.982	10.5%	18	382
Beckman AU series Reebio \ Arsenazo □	mmol/L	0.96	5.3%	17	1006
Roche Hitachi/Hitachi Systems MAKER \ Xylidyl Blue	mmol/L	0.869	3.8%	16	417
Abbott ARCHITECT c series BSBE \ Xylidyl Blue	mmol/L	0.991	16%	14	387
Beckman AU series Szybio \ OTHER	mmol/L	0.875	2.4%	14	317
TOSHIBA TBA MedicalSystem \ Arsenazo □	mmol/L	0.928	5.2%	11	349
Mindray BS series Mindray \ Xylidyl Blue	mmol/L	0.82	12.3%	11	100
Beckman AU series SekiSui \ Xylidyl Blue	mmol/L	0.905	2.8%	10	298
Roche cobas 4000/6000/8000 Roche \ Xylidyl Blue	mmol/L	0.93	9.7%	<5	<100
Beckman UniCel DxC 600/800 Beckman Coulter \ Atomic absorption spectroscopy (AAS)	mmol/L	0.84	10.7%	<5	<100
(Myo / Mb)肌红蛋白					
Abbott ARCHITECT ci series BSBE \ immunoturbidimetric	μg/L	29	5.2%	5	249
Beckman AU series BSBE \ OTHER	μg/L	26.8	8.6%	5	154
Roche Hitachi/Hitachi Systems OTHER \ Immunoturbidimetry (ITA)	μg/L	40.1	5.8%	23	893
Beckman AU series OTHER \ OTHER	μg/L	36.4	19.4%	21	542
Roche Hitachi/Hitachi Systems BIOSINO \ OTHER	μg/L	36.8	21.5%	12	247
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	μg/L	39.1	6.7%	11	345
(Na)钠					
Mindray BS series Mindray \ ISE (indirect)	mmol/L	141	1%	8	<100
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	142	1.4%	54	1848
Beckman AU series Dedicated \ ISE (direct)	mmol/L	141	1.1%	51	1401
TOSHIBA TBA SHXD \ ISE (indirect)	mmol/L	141	1.7%	5	<100
Abbott ARCHITECT ci series Abbott \ ISE (direct)	mmol/L	143	1%	5	252
Roche Hitachi/Hitachi Systems OTHER \ ISE (direct)	mmol/L	142	2.8%	32	750
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	144	4.6%	29	835
Beckman AU series OTHER \ ISE (direct)	mmol/L	142	2.6%	28	654
Abbott ARCHITECT c series BSBE \ ISE (direct)	mmol/L	140	1.2%	26	767
Beckman AU series Szybio \ ISE (indirect)	mmol/L	142	1.1%	25	591
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	141	1%	24	591
Mindray BS series Dedicated \ OTHER	mmol/L	144	1.2%	18	406
Roche Hitachi/Hitachi Systems OTHER \ ISE (indirect)	mmol/L	142	0.9%	17	305

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
16 / 21								

		批号：376B869			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems Wako \ ISE (indirect)	mmol/L	142	0.7%	16	435
HORRON-H900 HORRON \ ISE (indirect)	mmol/L	144	7%	16	271
Mindray BS series Dedicated \ ISE (indirect)	mmol/L	139	1.2%	15	880
Beckman AU series LEADMAN \ ISE (indirect)	mmol/L	140	0.9%	15	326
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	mmol/L	134	2%	13	341
TOSHIBA TBA OTHER \ ISE (direct)	mmol/L	150	1.8%	12	277
(NEFA)游离脂肪酸					
Beckman AU series SekiSui \ ACS - ACOD	μEq/L	472	2.1%	5	145
Roche cobas E601/602 Roche \ ACS - ACOD	mmol/L	0.53	11.3%	<5	<100
(NGAL)中性粒细胞明胶酶相关脂质运载蛋白					
Roche Hitachi/Hitachi Systems chongqiong zyb10 \ immunoturbidimetric	ng/mL	14	0%	<5	<100
(NO)一氧化氮					
Beckman AU series chongqiong zyb10 \ Colorimetry	mmol/L	0.039	0%	<5	<100
(P)磷					
Mindray BS series Mindray \ OTHER	mmol/L	1.01	3.5%	8	<100
Beckman AU series Elikan \ Molybdenum blue colorimetry	mmol/L	1.01	3.5%	7	191
Beckman AU series Dedicated \ Phosphomolybdate-UV-end point	mmol/L	1.04	4.5%	31	815
Beckman AU series Beckman Coulter \ Phosphomolybdate-UV-end point	mmol/L	1.02	4.4%	27	641
Abbott ARCHITECT c series BSBE \ OTHER	mmol/L	1.04	3.7%	26	706
Beckman AU series Szybio \ Phosphomolybdate-UV-end point	mmol/L	1	3.4%	25	608
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	1.02	2.9%	22	916
Beckman AU series Dedicated \ Phosphomolybdate UV-rate	mmol/L	1.05	2.5%	20	401
Roche Hitachi/Hitachi Systems MAKER \ Phosphomolybdate-UV-end point	mmol/L	1.12	8%	19	508
Mindray BS series Dedicated \ Phosphomolybdate-UV-end point	mmol/L	1.01	5.4%	19	337
Beckman AU series Beckman Coulter \ OTHER	mmol/L	1.03	2.2%	17	987
Roche Hitachi/Hitachi Systems MAKER \ Phosphomolybdate UV-rate	mmol/L	0.985	4.6%	15	408
Beckman AU series BSBE \ Phosphomolybdate-UV-end point	mmol/L	1.02	5.5%	15	397
Beckman AU series LEADMAN \ Phosphomolybdate-UV-end point	mmol/L	0.996	2.2%	15	346
TOSHIBA TBA Reebio \ Phosphomolybdate-UV-end point	mmol/L	1.07	6.9%	12	316
Beckman AU series OTHER \ OTHER	mmol/L	1.04	4.6%	12	257
TOSHIBA TBA MedicalSystem \ Phosphomolybdate-UV-end point	mmol/L	0.967	2.4%	11	348
Beckman AU series SaiKe(Ningbo) \ Phosphomolybdate-UV-end point	mmol/L	1.05	1.5%	10	296
Roche cobas 4000/6000/8000 Roche \ Phosphomolybdate UV-rate	mmol/L	1.08	11.1%	<5	<100
(PAB)前白蛋白					
Beckman AU series OTHER \ OTHER	g/L	0.285	18.5%	38	1026
Beckman AU series Dedicated \ Immunoturbidimetry (ITA)	g/L	0.282	1.6%	20	498
Beckman AU series BSBE \ OTHER	g/L	0.286	2.9%	20	303
Mindray BS series Dedicated \ Immunoturbidimetry (ITA)	g/L	0.271	2.2%	18	353
Beckman AU series Reebio \ OTHER	g/L	0.254	2.8%	17	1001
Beckman AU series BIOSINO \ Immunoturbidimetry (ITA)	g/L	0.258	2.5%	14	245
TOSHIBA TBA simes-sikma \ Immunoturbidimetry (ITA)	g/L	0.269	3.7%	12	316
Roche Hitachi/Hitachi Systems BIOSINO \ Immunonephelometry (INA)	g/L	0.194	35.4%	12	204
TOSHIBA TBA OTHER \ Immunoturbidimetry (ITA)	g/L	0.262	5.1%	10	332
(PAMY)胰淀粉酶					
Beckman AU series Reebio \ OTHER	U/L	45.1	2.4%	20	1154
Siemens ADVIA Chemistry XPT Siemens \ Enzymatic	U/L	45.9	10.5%	<5	<100

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
17 / 21								

批号：376B869					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(PCT)降钙素原					
Abbott ARCHITECT c series BSBE \ OTHER	ng/mL	0.235	27.9%	12	212
(PRO/ TP)(总) 蛋白					
Mindray BS series	g/L	70.6	1.3%	8	<100
Mindray \ Biuret-endpoint					
TOSHIBA TBA	g/L	68.4	4%	6	149
chongqiong zybio \ OTHER					
Roche Hitachi/Hitachi Systems	g/L	67.1	16.5%	6	133
BIOSINO \ OTHER					
Beckman AU series	g/L	67.9	3.2%	36	941
Dedicated \ Biuret-endpoint					
Roche Hitachi/Hitachi Systems	g/L	66.9	3.7%	34	679
OTHER \ Biuret-endpoint					
Roche Hitachi/Hitachi Systems	g/L	67.9	1.2%	32	870
LEADMAN \ Biuret-endpoint					
Beckman AU series	g/L	65.3	2.6%	29	666
Beckman Coulter \ Biuret-rate					
Abbott ARCHITECT c series	g/L	69.1	1.2%	26	715
BSBE \ Biuret-endpoint					
Beckman AU series	g/L	69.8	1.6%	25	577
Szybio \ Biuret-endpoint					
Roche Hitachi/Hitachi Systems	g/L	68.9	1.7%	24	583
MAKER \ OTHER					
Mindray BS series	g/L	69.3	1.2%	<5	<100
OTHER \ Biuret-endpoint					
Mindray BS series	g/L	70.5	1.2%	<5	<100
MAKER \ Biuret-rate					
Roche Hitachi/Hitachi Systems	g/L	68.5	1.6%	19	515
MAKER \ Biuret-endpoint					
Beckman AU series	g/L	68.1	2.6%	18	509
Beckman Coulter \ Biuret-endpoint					
Mindray BS series	g/L	68.8	1.5%	18	374
Dedicated \ Biuret-rate					
Beckman AU series	g/L	69.9	2.5%	17	1015
Beckman Coulter \ OTHER					
Mindray BS series	g/L	67.2	1.7%	15	879
Dedicated \ Biuret-endpoint					
Roche Hitachi/Hitachi Systems	g/L	69.2	1.8%	15	413
MAKER \ Biuret-rate					
Beckman AU series	g/L	67.7	5.5%	15	402
BSBE \ Biuret-endpoint					
(PROG)孕酮					
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.415	11.7%	<5	<100
Mindray BS series					
Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.6	0%	<5	<100
(RBP)视黄醇结合蛋白					
Beckman AU series					
OTHER \ Immunoturbidimetry (ITA)	mg/dL	4.18	4.7%	5	144
Beckman AU series					
BIOSINO \ Immunoturbidimetry (ITA)	mg/dL	4.26	7.2%	14	250
(RF)类风湿因子					
Beckman AU series					
OTHER \ Immunonephelometry (INA)	IU/mL	7.96	23.7%	38	1154
(SA)唾液酸					
Abbott ARCHITECT c series					
BSBE \ Enzymatic	mg/L	640	2%	10	275
(SAA)血清淀粉样蛋白A					
Abbott ARCHITECT c series					
BSBE \ OTHER	mg/L	4.01	24.3%	24	556
Beckman AU series					
PRB \ immunoturbidimetric	mg/L	5.14	7.9%	10	295
(sd LDL-C)小而密低密度脂蛋白胆固醇					
Beckman AU series					
OTHER \ Peroxidase	mmol/L	0.756	7%	28	787
(SOD)超氧化物歧化酶					
TOSHIBA TBA					
OTHER \ OTHER	U/L	84.6	3.8%	11	351
(sTfR)可溶性转铁蛋白受体					
Beckman AU series					
JingYuan(Shanghai) \ Immunoturbidimetry (ITA)	mg/L	37.4	10.1%	5	157
(TBA)总胆汁酸					
Roche Hitachi/Hitachi Systems					
BIOBASE \ Enzymatic Cycling Assays Method	μmol/L	12.9	8.2%	8	144
Beckman AU series					
Szybio \ Enzymatic Cycling Assays Method	μmol/L	10.2	4%	6	188
Mindray BS series					
Elikan \ Enzymatic Cycling Assays Method	μmol/L	13.5	2.5%	5	<100
Beckman AU series					
Dedicated \ Enzymatic Cycling Assays Method	μmol/L	12.2	12%	5	<100

质控品名： 生化与激素非定值质控品

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
18 / 21								

		批号：376B869			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series	μmol/L	13	2.6%	5	152
Ausbio \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	13.7	14%	39	1010
OTHER \ Enzymatic Cycling Assays Method					
Mindray BS series	μmol/L	11.2	3.9%	<5	<100
Mindray \ Enzymatic Cycling Assays Method					
Abbott ARCHITECT c series	μmol/L	15.5	1.9%	26	700
BSBE \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	14.1	8.2%	20	428
simes-sikma \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	11.2	2.7%	20	1169
Beckman Coulter \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	14.2	4.8%	<5	<100
sysmex \ Enzymatic Cycling Assays Method					
Mindray BS series	μmol/L	11.9	8.2%	<5	<100
OTHER \ Enzymatic Cycling Assays Method					
Mindray BS series	μmol/L	16.3	6.3%	<5	<100
MAKER \ Enzymatic Cycling Assays Method					
Roche Hitachi/Hitachi Systems	μmol/L	16.3	19.5%	19	516
BSBE \ Enzymatic Cycling Assays Method					
Mindray BS series	μmol/L	13.3	5.7%	18	405
Dedicated \ Enzymatic Cycling Assays Method					
Roche Hitachi/Hitachi Systems	μmol/L	12.7	7.8%	17	301
OTHER \ Enzymatic Cycling Assays Method					
Roche Hitachi/Hitachi Systems	μmol/L	15.2	6.4%	16	431
LEADMAN \ Enzymatic Cycling Assays Method					
Beckman AU series	μmol/L	12.9	4.8%	15	332
BIOSINO \ Enzymatic Cycling Assays Method					
TOSHIBA TBA	μmol/L	17.9	8.3%	12	303
simes-sikma \ Enzymatic Cycling Assays Method					
(TBIL)总胆红素					
TOSHIBA TBA	μmol/L	17.5	3.1%	8	496
Szybio \ Vanadate					
Mindray BS series	μmol/L	16.4	2.6%	7	<100
Mindray \ Vanadate					
Roche Hitachi/Hitachi Systems	μmol/L	20.7	2%	46	1195
LEADMAN \ Vanadate					
Roche Hitachi/Hitachi Systems	μmol/L	17.9	10.8%	36	754
OTHER \ Vanadate					
Roche Hitachi/Hitachi Systems	μmol/L	16.5	10.8%	35	903
MAKER \ Vanadate					
Beckman AU series	μmol/L	18	6.6%	31	812
Dedicated \ Vanadate					
Mindray BS series	μmol/L	13.6	4.8%	<5	<100
Mindray \ OTHER					
Beckman AU series	μmol/L	16.5	2.9%	29	571
Beckman Coulter \ Vanadate					
Abbott ARCHITECT c series	μmol/L	16	6.2%	28	734
BSBE \ Vanadate					
Beckman AU series	μmol/L	16.1	6%	24	556
Szybio \ OTHER					
Beckman AU series	μmol/L	17.4	12%	20	538
OTHER \ OTHER					
Beckman AU series	μmol/L	19.7	2%	20	501
Dedicated \ Diazotization -Blank Compensation					
Mindray BS series	μmol/L	18.7	9.1%	<5	<100
OTHER \ Vanadate					
Mindray BS series	μmol/L	16.6	1.9%	<5	<100
MAKER \ OTHER					
Beckman AU series	μmol/L	17.4	12.9%	18	463
BSBE \ Vanadate					
Mindray BS series	μmol/L	16.4	2.4%	18	408
Dedicated \ Vanadate					
Beckman AU series	μmol/L	18.9	5.7%	17	501
Beckman Coulter \ Diazotization -Blank Compensation					
Beckman AU series	μmol/L	16.3	3.1%	17	1006
Reebio \ OTHER					
Mindray BS series	μmol/L	18.1	1.7%	15	338
Elikan \ Vanadate					
(TCO2)二氧化碳总量					
HORRON-H900	mmol/L	22	0%	<5	<100
HORRON \ OTHER					
(TES)睾酮					
Roche cobas E601/602	ng/mL	3.72	9.3%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(TG)甘油三酯					
Beckman AU series	mmol/L	1.02	2.1%	7	179
OTHER \ GPO-POD					
Roche Hitachi/Hitachi Systems	mmol/L	0.981	2.4%	61	1616
LEADMAN \ GPO-POD					
TOSHIBA TBA	mmol/L	1.03	3.9%	6	138
chongqiong zybio \ OTHER					
Roche Hitachi/Hitachi Systems	mmol/L	1.04	3%	43	1088
MAKER \ OTHER					
Roche Hitachi/Hitachi Systems	mmol/L	1.1	0%	<5	<100
BIOBASE \ OTHER					
Beckman AU series	mmol/L	1.05	2%	<5	<100
Dedicated \ GPO-POD					
Beckman AU series	mmol/L	0.997	12.8%	38	1027
OTHER \ OTHER					

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
19 / 21								

		批号：376B869			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Beckman Coulter \ OTHER	mmol/L	1.03	3.8%	31	1320
Abbott ARCHITECT c series BSBE \ OTHER	mmol/L	1.08	2.8%	26	709
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	1.02	9.9%	21	482
Mindray BS series OTHER \ OTHER	mmol/L	1.07	9.4%	<5	<100
Mindray BS series MAKER \ Dehydrogenase	mmol/L	1.21	6.7%	<5	<100
Mindray BS series Dedicated \ GPO-POD	mmol/L	1.03	1.9%	18	360
Roche Hitachi/Hitachi Systems OTHER \ Free glycerol removed	mmol/L	1.08	2%	17	303
Roche Hitachi/Hitachi Systems MAKER \ GPO-POD	mmol/L	1.06	2.4%	15	379
Beckman AU series BIOSINO \ GPO-POD	mmol/L	0.989	6.1%	15	349
Mindray BS series Elikan \ GPO-POD	mmol/L	1.02	1.6%	15	346
Beckman AU series Dedicated \ OTHER	mmol/L	1.04	4%	15	334
Beckman AU series Szybio \ OTHER	mmol/L	1.03	1.1%	13	315
(THEO)茶碱					
J&J Vitros Fusion 5,1 FS Ortho \ Chemiluminescence Immunoassay (CLIA)	µg/mL	14.2	12.2%	<5	<100
(TIBC)总铁结合力					
Beckman AU series Dedicated \ Ferrozine colorimetry	µmol/L	58	0%	<5	<100
(TRF)转铁蛋白					
Beckman AU series OTHER \ Immunoturbidimetry (ITA)	mg/dL	275	1.9%	5	152
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	289	1.4%	23	587
Roche Hitachi/Hitachi Systems LEADMAN \ Immunoturbidimetry (ITA)	mg/dL	327	2.1%	16	440
TOSHIBA TBA OTHER \ Immunoturbidimetry (ITA)	mg/dL	265	1.5%	11	350
Roche cobas E601/602 Roche \ Immunonephelometry (INA)	mg/dL	285	9.8%	<5	<100
(TSH)促甲状腺素					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.43	9.1%	<5	<100
Mindray BS series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	1.53	2%	<5	<100
(TT3)总三碘甲状原氨酸					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.755	9.5%	<5	<100
Mindray BS series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.42	2.4%	<5	<100
(TT4)总T4					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	84.7	10.1%	<5	<100
Mindray BS series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	49.8	0.6%	<5	<100
(UA)尿酸					
Mindray BS series Mindray \ OTHER	µmol/L	295	1.2%	8	<100
Roche Hitachi/Hitachi Systems BIOSINO \ Uricase colorimetry	µmol/L	299	7.3%	8	207
TOSHIBA TBA chongqiong zybio \ OTHER	µmol/L	292	2.7%	6	142
Beckman AU series Beckman Coulter \ Uricase colorimetry	µmol/L	305	4.7%	50	1246
Roche Hitachi/Hitachi Systems LEADMAN \ Uricase colorimetry	µmol/L	313	1.8%	48	1270
Beckman AU series OTHER \ OTHER	µmol/L	297	6.4%	42	1109
Roche Hitachi/Hitachi Systems OTHER \ Uricase colorimetry	µmol/L	303	2.7%	33	690
Abbott ARCHITECT c series BSBE \ OTHER	µmol/L	297	2.4%	26	706
Beckman AU series Szybio \ OTHER	µmol/L	304	1.6%	25	599
Roche Hitachi/Hitachi Systems MAKER \ OTHER	µmol/L	301	2%	24	577
Mindray BS series Autobio \ Uricase colorimetry	µmol/L	304	1.3%	<5	<100
Roche Hitachi/Hitachi Systems Whitman \ Uricase colorimetry	µmol/L	303	3.5%	19	514
Mindray BS series Dedicated \ Uricase-UV	µmol/L	295	1.8%	18	356
Beckman AU series Reebio \ OTHER	µmol/L	296	1.9%	17	996

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
20 / 21								

		批号：376B869			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series OTHER \ Uricase colorimetry	μmol/L	313	3.7%	17	491
Beckman AU series Dedicated \ Uricase-UV	μmol/L	296	2.5%	16	477
Mindray BS series Dedicated \ Uricase colorimetry	μmol/L	289	2.1%	15	868
Roche Hitachi/Hitachi Systems MAKER \ Uricase colorimetry	μmol/L	298	3.2%	15	425
Beckman AU series Dedicated \ Uricase colorimetry	μmol/L	294	3.7%	15	354
(UIBC)不饱和铁结合力					
Abbott ARCHITECT c series LEADMAN \ OTHER	μmol/L	42.2	5.8%	8	198
Beckman AU series Beckman Coulter \ OTHER	μmol/L	44.4	5.3%	8	176
Beckman AU series Dedicated \ Ferrozine colorimetry	μmol/L	44.1	2.1%	20	414
Roche Hitachi/Hitachi Systems LEADMAN \ Ferrozine colorimetry	μmol/L	43.8	4.1%	16	440
TOSHIBA TBA XJKC \ Ferrozine colorimetry	μmol/L	48.7	3.9%	11	342
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	43.6	11.3%	<5	<100
Roche cobas E601/602 Roche \ Ferrozine colorimetry	μmol/L	32.9	11%	<5	<100
(Urea)尿素					
Roche Hitachi/Hitachi Systems BIOBASE \ OTHER	mmol/L	3.96	2.7%	9	137
Beckman AU series OTHER \ Urease colorimetry	mmol/L	4.16	3.4%	7	173
Beckman AU series OTHER \ OTHER	mmol/L	4.39	9.7%	64	1319
TOSHIBA TBA chongqiong zybio \ OTHER	mmol/L	4.02	6.2%	6	139
Roche Hitachi/Hitachi Systems LEADMAN \ OTHER	mmol/L	4.41	2.9%	48	1240
Beckman AU series chongqiong zybio \ OTHER	mmol/L	4.94	6.7%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	4.37	2.9%	39	1491
Beckman AU series Beckman Coulter \ OTHER	mmol/L	4.25	4.5%	29	614
Abbott ARCHITECT c series BSBE \ OTHER	mmol/L	4.1	3%	26	736
Beckman AU series Szybio \ OTHER	mmol/L	4.21	3.4%	26	604
Beckman AU series Dedicated \ Urease-UV-rate	mmol/L	4.38	2.2%	20	503
Roche Hitachi/Hitachi Systems OTHER \ Urease-UV-rate	mmol/L	4.15	8.7%	20	471
Mindray BS series OTHER \ OTHER	mmol/L	4.21	10.5%	<5	<100
Mindray BS series MAKER \ OTHER	mmol/L	4.55	4.1%	<5	<100
Roche Hitachi/Hitachi Systems Whitman \ Urease-UV-rate	mmol/L	4.2	3.4%	19	515
Mindray BS series Dedicated \ Urease-UV-rate	mmol/L	4.21	4.7%	18	381
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	4.14	3.5%	17	304
Beckman AU series Reebio \ OTHER	mmol/L	4.39	2.5%	17	1016
Beckman AU series Dedicated \ OTHER	mmol/L	4.32	3.2%	15	323
(Zn)锌					
Beckman AU series BSBE \ Colorimetry	μmol/L	23.6	6.5%	5	<100
Beckman AU series Szybio \ OTHER	μmol/L	23.4	3.4%	16	322
Roche Hitachi/Hitachi Systems Roche \ ICP-MS	μmol/L	26.4	9.8%	<5	<100
(β2M)β-2-微球蛋白					
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	ng/mL	988	5%	6	<100
Beckman AU series chongqiong zybio \ OTHER	ng/mL	608	4.1%	5	<100
Beckman AU series JP SY \ Immunoturbidimetry (ITA)	ng/mL	1079	3.8%	5	142
Beckman AU series Dedicated \ Immunoturbidimetry (ITA)	ng/mL	960	4.5%	<5	<100
Beckman AU series OTHER \ OTHER	ng/mL	1065	15.4%	38	995
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	ng/mL	810	6.1%	24	1019
Mindray BS series Dedicated \ Immunoturbidimetry (ITA)	ng/mL	950	8.4%	18	393
Beckman AU series Reebio \ OTHER	ng/mL	1070	9.6%	17	1025
Beckman AU series BJBJ \ Immunoturbidimetry (ITA)	ng/mL	952	6.5%	12	157

报告类别： 汇总实验室数据报告



质控品名： 生化与激素非定值质控品

REF	581-1	6×5mL	EXP	2024-04-25	LOT	376B869	更新日期：	2024-05
								21 / 21

		批号：376B869	
项目\仪器\试剂\方法	单位	均值 \ CV	Lab \ 点数

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。