

REF	539-2	6×3mL	EXP	2024-02-22	LOT	336C936	更新日期：	2024-04
1 / 3								

批号：336C936					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Mindray BC-1000 series	ng/mL	124	2.1%	<5	<100
Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Mindray CL-2000i/1000i/6000i series	ng/mL	122	1.8%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	ng/mL	120	4.2%	18	483
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)					
Roche cobas E601/602	ng/mL	134	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(ALD)醛固酮					
Autolumo A2000/A2000 PLUS	pg/mL	278	10%	<5	<100
Autobio \ Chemiluminescence Immunoassay (CLIA)					
Mindray CL-2000i/1000i/6000i series	pg/mL	218	8.8%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
(Anti-Tg)抗甲状腺球蛋白抗体					
Roche cobas E601/602	IU/mL	67.6	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(Anti-TPO)抗甲状腺过氧化物酶抗体					
Roche cobas E601/602	IU/mL	33.3	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(CA 125)糖类抗原 125					
Mindray CL-2000i/1000i/6000i series	IU/mL	85.8	4.4%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	IU/mL	69.3	4.4%	18	468
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)					
Roche cobas E601/602	IU/mL	52.7	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(CA 15-3)糖类抗原 15-3					
Mindray CL-2000i/1000i/6000i series	IU/mL	27.4	4.7%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	IU/mL	71.3	11.3%	18	448
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)					
Roche cobas E601/602	IU/mL	52.8	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(CA 19-9)糖类抗原 19-9					
Mindray CL-2000i/1000i/6000i series	IU/mL	43.3	4.3%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas E601/602	IU/mL	55.7	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(CEA)癌胚抗原					
Abbott ARCHITECT i series	ng/mL	12.7	4.7%	18	490
Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)					
Roche cobas E601/602	ng/mL	11.1	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(Cortisol)皮质醇					
Roche cobas E601/602	ng/mL	240	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(C-P)C-肽					
Roche cobas E601/602	ng/mL	2.31	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Mindray CL-2000i/1000i/6000i series	ng/mL	1.48	6.8%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
(DGXN)地高辛					
Roche cobas e411	ng/mL	2.08	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(DHEA-S)硫酸脱氢表雄酮					
Roche cobas E601/602	ng/mL	5941	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(E2)雌二醇					
Abbott ARCHITECT i series	pmol/L	866	3.5%	5	<100
Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)					
Abbott ARCHITECT i series	pmol/L	846	2.6%	14	<100
OTHER \ Chemiluminescence enzyme immunoassay (CLEIA)					
Roche cobas E601/602	pmol/L	1009	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(FA)叶酸					
Roche cobas E601/602	ng/mL	6.94	10%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
(Ferritin)铁蛋白					
Mindray CL-2000i/1000i/6000i series	ng/mL	202	3.9%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas 4000/6000/8000	ng/mL	190	10%	<5	<100
Roche \ Immunoturbidimetry (ITA)					

质控品名： 多专业免疫分析非定值质控品

REF	539-2	6×3mL	EXP	2024-02-22	LOT	336C936	更新日期：	2024-04
2 / 3								

批号：336C936					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(F-PSA)游离前列腺特异性抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.22	3.9%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.56	5.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.59	10%	<5	<100
(FSH)促卵泡刺激素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	33.6	2%	5	<100
Abbott ARCHITECT i series OTHER \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	33.7	2.8%	11	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	33.3	10%	<5	<100
(FT3)游离T3					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	10.7	5.2%	5	123
OTHER chongqiong zybio \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	9.35	4%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.7	1.4%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	11.1	3.9%	14	360
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	13.9	10%	<5	<100
(FT4)游离T4					
OTHER chongqiong zybio \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	29.4	9.3%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	24.3	10.3%	<5	103
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	23	3.9%	14	380
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	31.5	10%	<5	<100
(hCG)人绒毛膜促性腺激素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	145	2.7%	5	119
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	148	4%	15	341
Ortho Vitros 3600/5600 Ortho \ Chemiluminescence Immunoassay (CLIA)	IU/L	235	12%	<5	<100
(HCY)同型半胱氨酸					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	μmol/L	18.8	2.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	22.8	10%	<5	<100
(HGH)人生长激素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	4.2	10%	<5	<100
(IGF1)胰岛素生长因子1					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	234	10%	<5	<100
(INS)胰岛素					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	33.3	10%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	mIU/L	29.2	3.6%	<5	<100
(LH)促黄体生成素					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	16.2	4.1%	18	103
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	22.9	10%	<5	<100
(PRL)催/ 泌乳素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	18.2	8.2%	5	<100
Abbott ARCHITECT i series OTHER \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	18.6	2.7%	10	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.6	10%	<5	<100
(PROG)孕酮					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	8.8	5.4%	5	104
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	9.04	5.2%	15	326
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.6	10%	<5	<100

REF	539-2	6×3mL	EXP	2024-02-22	LOT	336C936	更新日期：	2024-04
3 / 3								

批号：336C936					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(PSA)前列腺特异性抗原					
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	10.6	4.8%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11	3.9%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14	10%	<5	<100
(TES)睾酮					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.442	4.1%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.474	10%	<5	<100
(Tg)甲状腺球蛋白					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.1	10%	<5	<100
(TSH)促甲状腺素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	9.33	3%	5	122
OTHER chongqiong zybio \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	8.25	5%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	mIU/L	11.7	4.7%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	8.73	3.6%	13	361
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	11.7	10%	<5	<100
(TT3)总三碘甲状腺原氨酸					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.22	6.6%	5	127
OTHER chongqiong zybio \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.45	3%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.25	5.5%	14	376
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.43	10%	<5	<100
(TT4)总T4					
OTHER chongqiong zybio \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	119	4.5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	124	4.1%	<5	103
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	130	4.7%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	121	3.9%	14	373
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	121	10%	<5	<100
(VB12)维生素B12					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	718	10%	<5	<100
(β2M)β-2-微球蛋白					
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	ng/mL	1355	10%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.122	3.5%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.154	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。