报告类别： **汇总实验室数据报告**

质控品名： **凝血七项非定值质控品**

REF	671-2	6×1mL	EXP	2023-07-02	LOT	400335C	更新日期：	2024-03

		批号：400335C			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Sysmex CS 2000i/2100i Dedicated \ Light scattering	Seconds	85.5	2.7%	<5	126
Werfen ACL series Werfen \ Light scattering	Seconds	58.6	10.1%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	102	8.8%	<5	<100
Sysmex CS 5100 Siemens \ Light scattering	Seconds	88.9	9.8%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	Seconds	76.4	10.2%	<5	<100
(AT-III)抗凝血酶III					
Sysmex CS 2000i/2100i Dedicated \ Nephelometry	%	42.5	5.1%	<5	133
Werfen ACL series Dedicated \ Nephelometry	%	42.9	9.8%	<5	<100
(D-D(DDU)) D-二聚体(DDU)					
Werfen ACL series Dedicated \ Nephelometry	mg/L	0.756	9.7%	<5	<100
(D-D(FEU)) D-二聚体(FEU)					
Sysmex CS 2000i/2100i Dedicated \ Nephelometry	mg/L	3.47	6.2%	<5	126
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	3.28	9.5%	<5	<100
Stago STA series Dedicated \ Latex agglutination (LA)	mg/L	1.63	8.6%	<5	<100
Sysmex CS 5100 Dedicated \ Nephelometry	mg/L	2.86	10.1%	<5	<100
(FDP)纤维蛋白(原)降解产物					
Stago STA series Dedicated \ Nephelometry	mg/L	5.56	8.8%	<5	<100
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	5.96	9.7%	<5	<100
Werfen ACL series SekiSui \ Nephelometry	mg/L	5.33	9.2%	<5	<100
(FIB)纤维蛋白原					
Sysmex CS 2000i/2100i Dedicated \ Light scattering	g/L	1.98	4.3%	<5	129
Werfen ACL series Werfen \ Light scattering	g/L	2.38	9.7%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	g/L	2.19	9.6%	<5	<100
Sysmex CS 5100 Siemens \ Light scattering	g/L	1.72	9.3%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	g/L	2.35	9.8%	<5	<100
(PT)血浆凝血酶原时间					
Sysmex CS 2000i/2100i Dedicated \ Light scattering	Seconds	29.3	5.6%	<5	131
Sysmex CS 5100 Siemens \ Light scattering	Seconds	34.9	9.2%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	31.2	10.3%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	Seconds	32.6	9.5%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	33.4	9.6%	<5	<100
(TT)凝血酶时间					
Sysmex CS 2000i/2100i Dedicated \ Light scattering	Seconds	19.5	4.8%	<5	127
Werfen ACL series SUNBIO \ Light scattering	Seconds	17.3	9.2%	<5	<100
Stago STA series Dedicated \ Magnetic Beads	Seconds	16.8	9.5%	<5	<100
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	17.2	9.9%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。