报告类别： **汇总实验室数据报告**

质控品名： **血气与电解质非定值质控品**

REF	600-2	30×1.7mL	EXP	2024-07-27	LOT	393DCF8	更新日期：	2024-03

		批号：393DCF8			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Ca)钙					
Radiometer ABL Series Dedicated \ BAPTA	mmol/L	1.05	1.7%	7	136
Radiometer ABL Series Radiometer \ ISE (indirect)	mmol/L	1.16	1.4%	10	<100
Radiometer ABL Series Dedicated \ ISE (indirect)	mmol/L	1.09	12%	<5	<100
(Cl)氯					
Radiometer ABL Series Dedicated \ ICP-MS	mmol/L	87.2	1.6%	7	136
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	93.9	12%	<5	<100
(K)钾					
Radiometer ABL Series Dedicated \ ICP-MS	mmol/L	4.03	2.8%	7	134
Radiometer ABL Series Radiometer \ ISE (indirect)	mmol/L	4.2	1.2%	18	<100
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	4.2	12%	<5	<100
(Na)钠					
Radiometer ABL Series Dedicated \ ICP-MS	mmol/L	135	0.9%	7	129
Radiometer ABL Series Radiometer \ ISE (indirect)	mmol/L	133	0.8%	16	<100
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	135	12%	<5	<100
(PCO2)二氧化碳分压					
Radiometer ABL Series Dedicated \ Electrode	mmHg	39.5	10.3%	8	155
Radiometer ABL Series Radiometer \ Electrode	mmHg	40.5	3.1%	19	105
Abbott FreeStyle Abbott \ Electrode	mmHg	26	10.3%	15	<100
(pH)酸碱度					
Radiometer ABL Series Dedicated \ Electrode	No Units	7.4	5.4%	8	163
Abbott FreeStyle Abbott \ Dipstick	Qualitative	NA	NA	NA	NA
Radiometer ABL Series Radiometer \ Electrode	No Units	7.39	0.1%	12	<100
(PO2)氧分压					
Radiometer ABL Series Dedicated \ Electrode	mmHg	104	8.4%	8	181
Radiometer ABL Series Radiometer \ Electrode	mmHg	112	2.9%	19	106
Abbott FreeStyle Abbott \ Electrode	mmHg	102	2.3%	15	<100
(TCO2)二氧化碳总量					
Radiometer ABL Series Radiometer \ OTHER	mmol/L	24.7	0%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。