报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-2	6×3mL	EXP	2025-02-15	LOT	3384FD6	更新日期：	2024-03

		批号：3384FD6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(17-OHP)17-羟孕酮					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.78	10.4%	<5	<100
(25-OH-Vit D)25羟基维生素D					
OTHER Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	44.4	9%	<5	<100
OTHER Siemens \ Chemiluminescence Immunoassay (CLIA)	nmol/L	111	9%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	30.9	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	77.1	10%	<5	<100
(AFP)甲胎蛋白					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	101	5.7%	6	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	110	6.5%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	108	8.1%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	113	13.2%	<5	<100
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	111	3.1%	15	427
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	119	5.5%	10	204
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	135	5.2%	10	198
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	134	10%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	117	9.4%	<5	<100
(ALD)醛固酮					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	167	9.6%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pg/mL	278	10%	<5	<100
(Anti-Tg)抗甲状腺球蛋白抗体					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	7.2	24.5%	7	112
Beckman UniCel DxI 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.92	6.7%	6	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	58.2	5.8%	5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	65.6	6.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	73.7	11.7%	22	299
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	83.8	3.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	64.6	9.9%	<5	<100
(Anti-TPO)抗甲状腺过氧化物酶抗体					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	23.3	9.3%	7	121
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.3	9.7%	6	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	21.1	6.4%	6	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	74.4	3.8%	5	<100
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34.2	8.3%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	38.1	27.3%	22	298
(ASD)雄烯二酮					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.12	11%	<5	<100
(CA 125)糖类抗原 125					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	40.2	28.7%	8	149
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	53.7	2.9%	5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/mL	77.8	16.2%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	72.3	18.6%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	51.2	3.9%	10	183
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	64.9	9.7%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-2	6×3mL	EXP	2025-02-15	LOT	3384FD6	更新日期：	2024-03

		批号：3384FD6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	85.7	9.5%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	52.7	10%	<5	<100
(CA 15-3)糖类抗原 15-3					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	27.8	82%	9	222
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	59.1	5.2%	9	169
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	53.3	6.4%	5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	47.9	9.3%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	58.1	10.5%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	50.6	10.1%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	52.8	10%	<5	<100
(CA 19-9)糖类抗原 19-9					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	94.5	118.2%	9	220
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	72.9	5.3%	5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	40.5	18.3%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/mL	84.9	13.2%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	53.3	10%	13	211
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	55.7	10%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	39.4	9.9%	<5	<100
(CEA)癌胚抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.8	9.4%	9	206
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.3	2.6%	5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.6	15.3%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.3	9.3%	<5	<100
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12	5.6%	15	433
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.8	8.7%	10	197
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.1	10%	<5	<100
Abbott ARCHITECT i series Abbott \ Enzyme linked immunosorbent assays (ELISA)	ng/mL	11.5	9.6%	<5	<100
(CK)肌酸激酶					
Roche cobas 4000/6000/8000 Roche \ Urease-UV-rate	U/L	61.3	10%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	11	9.3%	7	<100
Roche Hitachi/Hitachi Systems MAKER \ Immune suppression	U/L	10.7	8.8%	6	<100
TOSHIBA TBA MAKER \ Immune suppression	U/L	12.8	23.9%	10	134
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.51	1.3%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	U/L	12	10%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.3	10%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.8	10.1%	<5	<100
(Cortisol)皮质醇					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	174	5%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	198	16.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	225	9.7%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	240	10%	<5	<100
(C-P)C-肽					
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.1	4.9%	10	210

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-2	6×3mL	EXP	2025-02-15	LOT	3384FD6	更新日期：	2024-03

		批号：3384FD6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.31	10%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.74	11.5%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.35	9.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.46	9.6%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.28	9.6%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.52	12.4%	6	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.932	5.9%	10	193
(DGXN)地高辛					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.08	10%	<5	<100
(DHEA-S)硫酸脱氢表雄酮					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5941	10%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6338	10.5%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5711	10.3%	<5	<100
(E2)雌二醇					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	959	5%	6	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	998	3.6%	5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	978	7.6%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	996	3.1%	<5	105
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	808	10.1%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1009	10%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1042	9.9%	<5	<100
(EPO)促红细胞生成素					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/mL	46.2	5.3%	6	<100
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	45.9	9.4%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	33.1	10%	<5	<100
(FA)叶酸					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.89	7.7%	6	144
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.5	6.6%	6	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.37	10.8%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.73	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.94	10%	<5	<100
(Ferritin)铁蛋白					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	129	6.8%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	131	6%	<5	142
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	194	9.3%	<5	<100
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	ng/mL	159	6.1%	11	168
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	53.2	4.6%	10	194
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	210	12%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	ng/mL	190	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	153	10.5%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	137	10.2%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-2	6×3mL	EXP	2025-02-15	LOT	3384FD6	更新日期：	2024-03

		批号：3384FD6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	182	12%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.33	5.8%	6	<100
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.64	3.4%	11	206
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.74	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.59	10%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.92	10%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.43	10.4%	<5	<100
(FSH)促卵泡刺激素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	33.7	6.2%	7	148
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	26.9	5.9%	6	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	33.8	3.6%	5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	28.1	10.4%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	30.4	4.4%	<5	109
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/L	38.9	9.8%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	33.3	10%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/L	23.6	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	26.8	10.8%	<5	<100
(FT3)游离T3					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	14.9	5.4%	7	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	9.76	11.2%	7	101
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	8.65	7%	6	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	13.3	9.2%	25	431
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	10.7	4.2%	15	435
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	pmol/L	8.36	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	11.3	9.7%	<5	<100
OTHER Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	12.8	2.2%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	15	9.9%	<5	<100
(FT4)游离T4					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	24.8	6.5%	7	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	27.8	7.4%	7	101
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	32.4	3.7%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	32.6	6.7%	25	430
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	21.7	7%	15	435
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	26.5	9.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	34.2	10.5%	<5	<100
OTHER Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	30.6	5.2%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	32.2	9.9%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	pmol/L	25	12%	<5	<100
(hCG)人绒毛膜促性腺激素					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	155	4.5%	7	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	152	9.2%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-2	6×3mL	EXP	2025-02-15	LOT	3384FD6	更新日期：	2024-03

		批号：3384FD6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	147	9.5%	<5	<100
(HCY)同型半胱氨酸					
TOSHIBA TBA Bio-assay \ Enzymatic Cycling Assays Method	μmol/L	20.6	5.3%	8	168
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	22.8	10%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	μmol/L	24.3	9.5%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μmol/L	22.1	10.9%	<5	<100
(HE4)人附睾蛋白4					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	28.6	8.1%	10	186
(HGH)人生长激素					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	μg/L	2.86	5.2%	6	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	4.2	10%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μg/L	3.53	9.6%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μg/L	3.97	10.1%	<5	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	3.53	9.6%	<5	<100
(HS-CTnT)超敏肌钙蛋白T					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.865	4%	5	<100
(IgE)免疫球蛋白 IgE					
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	80.6	10.2%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	79.1	10%	<5	<100
(IGF1)胰岛素生长因子1					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	220	11.1%	<5	<100
(INS)胰岛素					
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	mIU/L	28.3	2.4%	10	196
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	32.1	10.6%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	33.3	10%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	34.7	10.4%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	25.1	9.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	31.2	10.3%	<5	<100
(LH)促黄体生成素					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	16.2	5.7%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	16.8	9%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	22.5	2.1%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	15.1	7.7%	<5	103
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	16	12.1%	10	237
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/L	32.1	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	22.9	10%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/L	28.5	9.8%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	17.9	10.6%	<5	<100
(Myo / Mb)肌红蛋白					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	μg/L	127	5.5%	7	<100
Roche Hitachi/Hitachi Systems Bio-assay \ immunoturbidimetric	μg/L	93.2	4.2%	<5	<100
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	189	7.6%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μg/L	126	7.2%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-2	6×3mL	EXP	2025-02-15	LOT	3384FD6	更新日期：	2024-03

		批号：3384FD6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
TOSHIBA TBA OTHER \ OTHER	μg/L	87.3	7.5%	10	174
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μg/L	143	9.8%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	152	10%	<5	<100
(NT-proBNP)N-末端脑钠肽前体					
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	pg/mL	1442	2.6%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	170	2.6%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	1276	10%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	150	10%	<5	<100
(PRL)催/ 泌乳素					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	13.6	4.6%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.4	7.7%	<5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	19.1	4.8%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	22	5.4%	<5	103
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.2	9.9%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16	12%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.9	8.8%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.6	10%	<5	<100
(PROG)孕酮					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.6	2.8%	8	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	10.8	10.2%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.5	4.9%	<5	102
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.6	10%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.2	9.8%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.6	10.6%	<5	<100
Siemens Immulite1000/2000 Siemens Dade Actin FS(APTT) \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	8.23	10%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.6	10.7%	<5	<100
(proGRP)胃泌素释放肽前体					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/L	8712	7.8%	9	173
(PSA)前列腺特异性抗原					
Beckman UniCel DxI 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.5	4.3%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.1	8.8%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.19	101%	11	256
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.6	9.1%	10	199
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.9	12%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.5	11.3%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14	10%	<5	<100
(PTH)甲状旁腺素					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	9.38	5.2%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	10.1	4.4%	<5	103
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	8.46	10.3%	<5	<100
OTHER Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	10.9	4.7%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	9.12	10%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-2	6×3mL	EXP	2025-02-15	LOT	3384FD6	更新日期：	2024-03

		批号：3384FD6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(SCC)鳞状细胞癌抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.46	4.4%	9	186
(SHBG)性激素结合球蛋白					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	nmol/L	78.8	12%	<5	<100
(TES)睾酮					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.464	6.3%	7	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.654	6.9%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.758	12.8%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.732	5.4%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.59	10.2%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.474	10%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.42	9.5%	<5	<100
(Tg)甲状腺球蛋白					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.16	9.4%	7	113
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	8.02	4.8%	6	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.37	9.7%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.1	10%	<5	<100
(TnI)肌钙蛋白I					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.442	14%	<5	123
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.345	6.6%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.47	12.8%	<5	<100
(TnT)肌钙蛋白T					
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.914	10%	<5	<100
(TSH)促甲状腺素					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	9.58	4.5%	7	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	10.2	4.4%	6	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	11.6	2.2%	5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	mIU/L	11.2	2.1%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	11.6	8.3%	25	414
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.46	4.6%	15	435
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	10.7	11.2%	<5	<100
OTHER Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	9.94	5.1%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	9.78	8.8%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	mIU/L	10	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	8.76	10.3%	<5	<100
(TT3)总三碘甲状原氨酸					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.22	7.1%	6	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.89	10.4%	6	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.8	6.4%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.47	9.2%	26	435
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.18	4.2%	15	436

报告类别： **汇总实验室数据报告**

质控品名： **复合免疫分析非定值质控品**

REF	540-2	6×3mL	EXP	2025-02-15	LOT	3384FD6	更新日期：	2024-03

		批号：3384FD6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.11	7.3%	10	234
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.64	12%	<5	<100
OTHER Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.64	4.9%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.82	9.9%	<5	<100
(TT4)总T4					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	117	13.6%	7	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	115	7.3%	7	122
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	129	5.6%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	123	8.2%	26	431
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	119	4.5%	15	431
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	129	7.7%	10	182
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	ng/mL	103	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	136	9.6%	<5	<100
(VB12)维生素B12					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pg/mL	479	5.8%	6	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pg/mL	533	11%	<5	<100
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	pg/mL	487	12%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	528	4.3%	<5	103
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	718	10%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pg/mL	659	10.3%	<5	<100
(β2M)β-2-微球蛋白					
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1440	11.1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	ng/mL	1355	10%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.163	6.5%	5	<100
Roche cobas e411 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.165	5.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.132	13.6%	26	436
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.207	7%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.154	12.1%	<5	<100
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.166	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。