

REF	522-2	6x2mL	EXP	2023-02-10	LOT	31CD832	更新日期：	2022-01
								1 / 1

		批号：31CD832			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(PCT)降钙素原					
OTHER					
OTHER \ Immunofluorescence (IFA)	ng/mL	9.5	7.1%	<5	<100
Sysmex XS series					
Dedicated \ OTHER	ng/mL	5.76	25.7%	22	659
OTHER					
OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.3	19.2%	20	370
maglumi system					
Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.99	10%	<5	<100
Roche cobas e411					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.5	0.9%	<5	<100
Bio-Merieux VIDAS					
BioMerieux \ Dot-immuogold filtration assay (DIGFA)	ng/mL	11.4	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。