报告类别： **汇总实验室数据报告**

质控品名： **凝血七项非定值质控品**

| REF | 671-3 | 6×1mL | EXP | 2024-01-01 | LOT | 4005ACA | 更新日期： | 2024-03 |
|---|---|---|---|---|---|---|---|---|



| | | 批号：4005ACA | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| **(APTT)活化部分凝血活酶时间** | | | | | |
| Succeeder SA series<br>Succeeder \ Magnetic Beads | Seconds | 84.5 | 1.5% | 8 | 175 |
| Werfen ACL series<br>Dedicated \ Light scattering | Seconds | 93.8 | 4.3% | 24 | 598 |
| Succeeder SF series<br>Succeeder \ Magnetic Beads | Seconds | 102 | 5% | 23 | 585 |
| Werfen ACL series<br>Werfen \ Light scattering | Seconds | 101 | 10.9% | <5 | <100 |
| Sysmex CA 1500/6000/7000<br>Siemens \ Light scattering | Seconds | 112 | 14.3% | <5 | <100 |
| Stago STA series<br>Dedicated \ Magnetic Beads | Seconds | 105 | 10.5% | <5 | <100 |
| **(AT-III)抗凝血酶III** | | | | | |
| Werfen ACL series<br>Dedicated \ Nephelometry | % | 26.4 | 9.5% | <5 | <100 |
| Sysmex CA 1500/6000/7000<br>Siemens \ Nephelometry | % | 26.9 | 9.7% | <5 | <100 |
| **(D-D(DDU)) D-二聚体(DDU)** | | | | | |
| Succeeder SA series<br>Succeeder \ Immunoturbidimetry (ITA) | mg/L | 1.72 | 1.3% | 8 | 176 |
| Werfen ACL series<br>Dedicated \ Latex agglutination (LA) | mg/L | 1.14 | 3.2% | 24 | 598 |
| Succeeder SF series<br>Succeeder \ Latex agglutination (LA) | mg/L | 1.91 | 2.5% | 12 | 292 |
| Werfen ACL series<br>Dedicated \ Nephelometry | mg/L | 2.88 | 10.1% | <5 | <100 |
| **(D-D(FEU)) D-二聚体(FEU)** | | | | | |
| Sysmex CA 1500/6000/7000<br>Siemens \ Nephelometry | mg/L | 5.26 | 10.3% | <5 | <100 |
| Stago STA series<br>Dedicated \ Latex agglutination (LA) | mg/L | 4.38 | 9.8% | <5 | <100 |
| **(FDP)纤维蛋白(原)降解产物** | | | | | |
| Werfen ACL series<br>Dedicated \ Nephelometry | mg/L | 6.67 | 11.8% | 24 | 235 |
| Sysmex CA 1500/6000/7000<br>SekiSui \ Nephelometry | mg/L | 8.23 | 10.4% | <5 | <100 |
| Stago STA series<br>Dedicated \ Nephelometry | mg/L | 7.89 | 10.3% | <5 | <100 |
| **(FIB)纤维蛋白原** | | | | | |
| Succeeder SA series<br>Succeeder \ Magnetic Beads | g/L | 1.33 | 2.4% | 8 | 175 |
| Werfen ACL series<br>Dedicated \ Light scattering | g/L | 1.26 | 4.5% | 24 | 597 |
| Succeeder SF series<br>Succeeder \ Magnetic Beads | g/L | 1.32 | 10.1% | 21 | 512 |
| Stago STA series<br>Dedicated \ Magnetic Beads | g/L | 1.46 | 12.3% | <5 | <100 |
| Sysmex CA 1500/6000/7000<br>Siemens \ Light scattering | g/L | 1.32 | 12.1% | <5 | <100 |
| Werfen ACL series<br>Werfen \ Light scattering | g/L | 1.35 | 9.6% | <5 | <100 |
| **(PT INR)PT国际标准化比值** | | | | | |
| Succeeder SF series<br>Succeeder \ Calculate | INR | 3.83 | 2.3% | 12 | 295 |
| **(PT)血浆凝血酶原时间** | | | | | |
| Succeeder SA series<br>Succeeder \ Magnetic Beads | Seconds | 50 | 4.1% | 8 | 176 |
| Werfen ACL series<br>Dedicated \ Light scattering | Seconds | 49.2 | 4.5% | 24 | 596 |
| Succeeder SF series<br>Succeeder \ Magnetic Beads | Seconds | 49.3 | 3.2% | 23 | 569 |
| Stago STA series<br>Dedicated \ Light scattering | Seconds | 43.6 | 9.9% | <5 | <100 |
| Werfen ACL series<br>Werfen \ Light scattering | Seconds | 42.5 | 10.4% | <5 | <100 |
| Sysmex CA 1500/6000/7000<br>Siemens \ Light scattering | Seconds | 42.3 | 10.2% | <5 | <100 |
| **(TT)凝血酶时间** | | | | | |
| Succeeder SA series<br>Succeeder \ Magnetic Beads | Seconds | 16.3 | 2.4% | 8 | 176 |
| Werfen ACL series<br>Dedicated \ Light scattering | Seconds | 22.8 | 4% | 24 | 594 |
| Succeeder SF series<br>Succeeder \ Magnetic Beads | Seconds | 15.4 | 1.6% | 23 | 564 |
| Werfen ACL series<br>Werfen \ Light scattering | Seconds | 23.5 | 9.4% | <5 | <100 |
| Sysmex CA 1500/6000/7000<br>Dedicated \ Light scattering | Seconds | 16.8 | 10.1% | <5 | <100 |

**注**：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。