

REF	580-2	6x5mL	EXP	2022-07-27	LOT	3755814	更新日期：	2022-05
								1 / 6

		批号：3755814			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(a-HBDH)α-羟丁酸脱氢酶					
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	U/L	287	1.9%	14	392
(Alb)白蛋白					
Beckman AU series Fosun \ Bromocresol green	μmol/L	469	3.1%	70	1636
Beckman AU series Fosun \ Bromocresol green	g/L	31	3.1%	70	1636
Beckman AU series WanTai \ Bromocresol green	μmol/L	430	2.2%	5	<100
Beckman AU series WanTai \ Bromocresol green	g/L	28.4	2.2%	5	<100
Beckman AU series DIRUI \ Bromocresol green	μmol/L	510	0.4%	<5	<100
Beckman AU series DIRUI \ Bromocresol green	g/L	33.6	0.4%	<5	<100
DiruiCS series DIRUI \ Bromocresol green	μmol/L	465	3.1%	<5	<100
DiruiCS series DIRUI \ Bromocresol green	g/L	30.7	3.1%	<5	<100
KHB ZY series OTHER \ Bromocresol green	μmol/L	540	7%	25	143
KHB ZY series OTHER \ Bromocresol green	g/L	35.7	7%	25	143
Beckman AU series Beckman Coulter \ Bromocresol green	μmol/L	492	1.2%	14	392
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	32.5	1.2%	14	392
Roche Hitachi/Hitachi Systems Wako \ Bromocresol green	μmol/L	424	2.4%	13	203
Roche Hitachi/Hitachi Systems Wako \ Bromocresol green	g/L	28	2.4%	13	203
Roche cobas 4000/6000/8000 Roche \ Bromocresol green	μmol/L	417	9.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Bromocresol green	g/L	27.5	9.5%	<5	<100
Beckman UniCel Dx C 600/800 Beckman Coulter \ Bromocresol green	μmol/L	423	9.7%	<5	<100
Beckman UniCel Dx C 600/800 Beckman Coulter \ Bromocresol green	g/L	27.9	9.7%	<5	<100
(ALP)碱性磷酸酶					
Beckman AU series Fosun \ Rate -AMP buffer	U/L	362	3%	70	1861
Roche Hitachi/Hitachi Systems MedicalSystem \ Rate -AMP buffer	U/L	413	0.7%	6	<100
Beckman AU series WanTai \ Rate -AMP buffer	U/L	284	4%	5	<100
DiruiCS series DIRUI \ Rate -AMP buffer	U/L	291	11.3%	<5	<100
KHB ZY series OTHER \ Rate -AMP buffer	U/L	264	1.5%	25	144
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	282	4.4%	14	392
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	319	10.3%	<5	<100
Abbott ARCHITECT c series Abbott \ Rate -other buffers	U/L	326	9.8%	<5	<100
Beckman AU series DIRUI \ Rate -AMP buffer	U/L	256	0%	<5	<100
(ALT/ SGPT)丙氨酸氨基转移酶					
Beckman AU series WanTai \ UV without P5P	U/L	272	5.2%	5	<100
Beckman AU series DIRUI \ UV without P5P	U/L	244	0%	<5	<100
DiruiCS series DIRUI \ UV without P5P	U/L	284	4.6%	<5	<100
KHB ZY series OTHER \ UV without P5P	U/L	268	1.5%	25	143
Beckman AU series Fosun \ OTHER	U/L	268	1.6%	24	639
Beckman AU series Fosun \ dry chemistry	U/L	268	1.3%	22	612
Beckman AU series Fosun \ UV without P5P	U/L	274	2.5%	21	466
Beckman AU series Beckman Coulter \ UV with P5P	U/L	252	1.6%	14	392
Beckman UniCel Dx C 600/800 HuaChenBio \ UV with P5P	U/L	243	9.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ UV without P5P	U/L	254	10.2%	<5	<100
(AMY)淀粉酶					
Beckman AU series Fosun \ PNP-G7	U/L	217	2.4%	70	1871
Beckman AU series Beckman Coulter \ PNP-G7	U/L	204	2.1%	14	392
(Apo-A I)载脂蛋白A I					
Beckman AU series Fosun \ immunoturbidimetric	g/L	0.923	3.3%	13	347

REF	580-2	6x5mL	EXP	2022-07-27	LOT	3755814	更新日期：	2022-05
								2 / 6

		批号：3755814			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(Apo-B)载脂蛋白B					
Beckman AU series Fosun \ immunoturbidimetric	g/L	0.395	3.8%	13	347
(AST/ SGOT)天门冬氨酸氨基转移酶					
Beckman AU series WanTai \ UV without P5P	U/L	254	5.4%	5	<100
Beckman AU series DIRUI \ UV without P5P	U/L	244	0%	<5	<100
DiruiCS series DIRUI \ UV with P5P	U/L	261	6.9%	<5	<100
KHB ZY series OTHER \ UV without P5P	U/L	253	2.5%	25	144
Beckman AU series Fosun \ OTHER	U/L	262	2%	24	637
Beckman AU series Fosun \ dry chemistry	U/L	262	1.6%	22	638
Beckman AU series Fosun \ UV without P5P	U/L	281	2.1%	22	474
Beckman AU series Beckman Coulter \ UV with P5P	U/L	241	2%	14	392
Roche Hitachi/Hitachi Systems Wako \ UV with P5P	U/L	195	1.5%	13	160
Roche cobas 4000/6000/8000 Roche \ UV without P5P	U/L	278	9.4%	<5	<100
Beckman UniCel DxC 600/800 Beckman Coulter \ UV with P5P	U/L	269	9.7%	<5	<100
(BUN)尿素氮					
Beckman AU series DIRUI \ Dehydrogenase	mmol/L	22.6	0.6%	7	<100
Beckman AU series WanTai \ Dehydrogenase	mmol/L	21.9	2.7%	5	<100
KHB ZY series OTHER \ Urease colorimetry	mmol/L	15.6	2.5%	25	144
Roche Hitachi/Hitachi Systems MedicalSystem \ Urease colorimetry	mmol/L	16.7	2.9%	13	157
(Ca)钙					
Beckman AU series Fosun \ Arsenazo III	mmol/L	2.69	3%	71	1812
Roche Hitachi/Hitachi Systems MedicalSystem \ Arsenazo III	mmol/L	2.94	1.6%	7	107
Beckman AU series Beckman Coulter \ Arsenazo III	mmol/L	2.89	1.3%	14	392
Beckman UniCel DxC 600/800 Beckman Coulter \ ISE (direct)	mmol/L	2.82	9.9%	<5	<100
Roche Hitachi/Hitachi Systems SekiSui \ ISE (direct)	mmol/L	2.88	9.7%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Arsenazo III	mmol/L	2.79	9.7%	<5	<100
(CHE)胆碱酯酶					
Beckman AU series Fosun \ Butrylthiocholine Method	U/L	3811	3.2%	24	625
(CHOL/ TC)总胆固醇					
Beckman AU series DIRUI \ Cholesterol oxidase	mmol/L	2.36	0.8%	6	<100
Beckman AU series BIOBASE \ Cholesterol oxidase	mmol/L	2.53	3.6%	5	<100
Beckman AU series Fosun \ Cholesterol oxidase	mmol/L	2.61	2.9%	47	1254
DiruiCS series DIRUI \ Cholesterol oxidase	mmol/L	2.8	1.6%	<5	<100
Beckman AU series Fosun \ OTHER	mmol/L	2.58	1.6%	24	638
KHB ZY series OTHER \ Cholesterol oxidase	mmol/L	1.98	5.8%	24	136
Beckman AU series Beckman Coulter \ Cholesterol oxidase	mmol/L	2.45	2%	14	392
Roche Hitachi/Hitachi Systems Wako \ Cholesterol oxidase	mmol/L	2.8	1.5%	13	190
Beckman UniCel DxC 600/800 Beckman Coulter \ Cholesterol oxidase	mmol/L	2.41	9.5%	<5	<100
(CK)肌酸激酶					
Beckman AU series Fosun \ Rate -AMP buffer	U/L	248	5.4%	45	1184
Beckman AU series Fosun \ OTHER	U/L	242	5.6%	24	613
Beckman AU series Beckman Coulter \ Rate -other buffers	U/L	331	3.3%	14	392
Roche Hitachi/Hitachi Systems MedicalSystem \ Rate -AMP buffer	U/L	427	1%	13	189
Beckman UniCel DxC 600/800 HuaChenBio \ Gelatin particle agglutination (PA)	U/L	257	9.3%	<5	<100
(Cl)氯					
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	94.5	0.9%	36	1014
Beckman AU series Fosun \ ISE (direct)	mmol/L	96.6	1.1%	24	639

REF	580-2	6x5mL	EXP	2022-07-27	LOT	3755814	更新日期：	2022-05
								3 / 6

		批号：3755814			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	94.1	0.7%	24	605
Roche Hitachi/Hitachi Systems Wako \ ISE (direct)	mmol/L	88.3	1.7%	13	152
Beckman UniCel DxC 600/800 Beckman Coulter \ ISE (direct)	mmol/L	97.3	9.7%	<5	<100
(CO2 (HCO3-))二氧化碳 (碳酸氢根)					
Beckman AU series Whitman \ OTHER	mmol/L	11.1	5%	23	588
Beckman UniCel DxC 600/800 Beckman Coulter \ PEPC	mmol/L	13.8	10.1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ PEPC	mmol/L	15.9	10.1%	<5	<100
(Cr)肌酐					
Beckman AU series Fosun \ Enzymatic	μmol/L	475	3.3%	44	1156
Beckman AU series WanTai \ Enzymatic	μmol/L	403	2.9%	<5	<100
Beckman AU series DIRUI \ Enzymatic	μmol/L	419	0.1%	<5	<100
DiruiCS series DIRUI \ Enzymatic	μmol/L	437	1.6%	<5	<100
KHB ZY series OTHER \ Enzymatic	μmol/L	334	0.9%	25	143
Beckman AU series Fosun \ OTHER	μmol/L	476	1.8%	24	639
Beckman AU series LEADMAN \ Enzymatic	μmol/L	406	1.4%	14	392
Roche Hitachi/Hitachi Systems Wako \ Enzymatic	μmol/L	481	0.7%	13	183
Beckman UniCel DxC 600/800 Dedicated \ Enzymatic	μmol/L	474	9.7%	<5	<100
(Cu)铜					
Beckman UniCel DxC 600/800 Beckman Coulter \ ICP-MS	μmol/L	12.7	9.4%	<5	<100
(DBIL)直接胆红素					
Beckman AU series Fosun \ Vanadate	μmol/L	36.1	2.2%	70	1723
Beckman AU series WanTai \ Vanadate	μmol/L	45.2	3.4%	5	<100
Beckman AU series DIRUI \ Vanadate	μmol/L	32.6	0.3%	<5	<100
DiruiCS series DIRUI \ Vanadate	μmol/L	46.6	7.3%	<5	<100
KHB ZY series OTHER \ Vanadate	μmol/L	44.5	6.1%	25	144
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	33.6	5.3%	14	391
Roche Hitachi/Hitachi Systems Wako \ Diazotization	μmol/L	28.4	3.5%	13	204
Beckman UniCel DxC 600/800 Beckman Coulter \ Diazotization -Blank Compensation	μmol/L	27.7	9.9%	<5	<100
(Fe)铁					
Roche cobas 4000/6000/8000 Roche \ Chromazuro colorimetry	μmol/L	11.2	9.9%	<5	<100
(GGT)L-γ-谷氨酰基转移酶					
Beckman AU series JS (Shanghai) \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	232	3.7%	<5	<100
Beckman AU series DIRUI \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	212	0%	<5	<100
DiruiCS series DIRUI \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	235	3.9%	<5	<100
KHB ZY series OTHER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	203	1.4%	25	139
Beckman AU series Fosun \ OTHER	U/L	222	1.6%	24	615
Beckman AU series Fosun \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	222	1.5%	22	638
Beckman AU series Fosun \ γ-glutamyl -4-nitroaniline	U/L	218	1.5%	17	397
Beckman AU series Beckman Coulter \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	211	1.6%	14	391
(GLDH)谷氨酸脱氢酶					
Roche cobas 4000/6000/8000 Roche \ Direct Measure	U/L	116	10.3%	<5	<100
(GLU)葡萄糖					
Beckman AU series Fosun \ Hexokinase	mmol/L	15.8	2.9%	68	1752
Beckman AU series BIOBASE \ Hexokinase	mmol/L	17.4	3.2%	5	<100
Beckman AU series DIRUI \ Glucose oxidase (GOD)	mmol/L	16.4	0.6%	<5	<100
DiruiCS series DIRUI \ Glucose oxidase (GOD)	mmol/L	17.4	2.5%	<5	<100
KHB ZY series OTHER \ Hexokinase	mmol/L	15.7	2.9%	24	136

REF	580-2	6x5mL	EXP	2022-07-27	LOT	3755814	更新日期：	2022-05
								4 / 6

		批号：3755814			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Beckman Coulter \ Hexokinase	mmol/L	16.6	1.6%	14	392
Roche Hitachi/Hitachi Systems Wako \ Hexokinase	mmol/L	15.8	1.2%	13	199
Beckman UniCel DxC 600/800 HuaChenBio \ Dehydrogenase	mmol/L	14.8	9.5%	<5	<100
Beckman UniCel DxC 600/800 Beckman Coulter \ Hexokinase	mmol/L	14.2	9.4%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Dehydrogenase	mmol/L	15.3	9.4%	<5	<100
Abbott ARCHITECT c series Abbott \ Hexokinase	mmol/L	13	9.8%	<5	<100
(HDL-C)高密度脂蛋白胆固醇					
Beckman AU series Fosun \ Direct Measure	mmol/L	0.739	5.4%	46	1104
Beckman AU series Fosun \ OTHER	mmol/L	0.743	3.6%	24	638
Beckman AU series Beckman Coulter \ Direct Measure	mmol/L	0.685	3.8%	14	383
Roche Hitachi/Hitachi Systems Wako \ Direct Measure	mmol/L	0.714	3.1%	13	191
Roche cobas 4000/6000/8000 Roche \ Enzymatic	mmol/L	0.82	11%	<5	<100
Beckman UniCel DxC 600/800 Beckman Coulter \ Enzymatic	mmol/L	0.799	9.4%	<5	<100
(K)钾					
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	5.07	3.5%	36	1035
Beckman AU series Fosun \ ISE (direct)	mmol/L	5.06	1.3%	24	641
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	4.93	0.9%	24	621
Roche Hitachi/Hitachi Systems Wako \ ISE (direct)	mmol/L	6.17	0.7%	13	203
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	5.62	9.3%	<5	<100
Beckman UniCel DxC 600/800 Beckman Coulter \ ISE (indirect)	mmol/L	5.34	9.9%	<5	<100
(LA)乳酸					
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	mmol/L	1.91	9.9%	<5	<100
Beckman UniCel DxC 600/800 Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	1.88	10.1%	<5	<100
(LDH)乳酸脱氢酶					
Roche Hitachi/Hitachi Systems MedicalSystem \ Lactate to pyruvate (L-P)	U/L	406	1.1%	9	110
Beckman AU series Fosun \ Lactate to pyruvate (L-P)	U/L	335	4.7%	47	1195
Beckman AU series Fosun \ OTHER	U/L	347	1.6%	22	562
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	U/L	306	2.2%	14	393
Beckman UniCel DxC 600/800 Beckman Coulter \ Lactate to pyruvate (L-P)	U/L	302	9.3%	<5	<100
(LDL-C)低密度脂蛋白胆固醇					
Beckman AU series Fosun \ OTHER	mmol/L	1.45	3.1%	22	586
Beckman UniCel DxC 600/800 Beckman Coulter \ Enzymatic	mmol/L	1.5	8.8%	<5	<100
Roche cobas 4000/6000/8000 KHB \ Enzymatic	mmol/L	1.54	9.7%	<5	<100
(Li)锂					
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	1.82	9.9%	<5	<100
Beckman UniCel DxC 600/800 Beckman Coulter \ ISE (indirect)	mmol/L	1.73	9.8%	<5	<100
(LIP)脂肪酶					
Roche Hitachi/Hitachi Systems Sekisui \ Enzymatic Cycling Assays Method	U/L	273	9.5%	<5	<100
(Mg)镁					
Beckman AU series Fosun \ Xylidyl Blue	mmol/L	2.08	2%	45	1170
Abbott ARCHITECT c series Wako \ Xylidyl Blue	mmol/L	1.57	10.2%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Calmagite colorimetry	mmol/L	1.62	9.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ ICP-MS	mmol/L	1.79	0%	<5	<100
(Na)钠					
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	115	0.9%	24	616
Beckman AU series Fosun \ ISE (direct)	mmol/L	117	1.1%	24	589
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	111	0.9%	14	392

REF	580-2	6x5mL	EXP	2022-07-27	LOT	3755814	更新日期：	2022-05
								5 / 6

		批号：3755814			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems Wako \ ISE (direct)	mmol/L	126	0.8%	13	204
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	121	7.4%	<5	<100
Beckman UniCel Dx C 600/800 Beckman Coulter \ ISE (indirect)	mmol/L	120	7.5%	<5	<100
(P)磷					
Beckman AU series Fosun \ OTHER	mmol/L	2.32	2%	24	617
Beckman AU series Fosun \ Molybdenum blue colorimetry	mmol/L	2.37	2.8%	24	605
Beckman AU series Fosun \ Phosphomolybdate UV-rate	mmol/L	2.32	2.1%	22	635
Beckman AU series Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	2.63	2%	14	392
Roche Hitachi/Hitachi Systems Wako \ Phosphomolybdate UV-rate	mmol/L	2.43	1.4%	13	144
Beckman UniCel Dx C 600/800 Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	2.63	9.9%	<5	<100
(PAB)前白蛋白					
Beckman AU series Fosun \ Immunoturbidimetry (ITA)	g/L	0.151	6.9%	22	596
(PRO/ TP)(总) 蛋白					
Beckman AU series Fosun \ Biuret-endpoint	g/L	47.3	2%	70	1831
Beckman AU series WanTai \ Biuret-endpoint	g/L	42.8	2.5%	5	<100
Beckman AU series DIRUI \ Biuret-endpoint	g/L	49.5	0.3%	<5	<100
DiruiCS series DIRUI \ Biuret-endpoint	g/L	41.8	3.5%	<5	<100
KHB ZY series OTHER \ Biuret-endpoint	g/L	44.4	5.1%	25	144
Beckman AU series Beckman Coulter \ Biuret-endpoint	g/L	49.8	1.6%	14	392
Roche Hitachi/Hitachi Systems Wako \ Biuret-endpoint	g/L	41.7	3.5%	13	203
Roche cobas 4000/6000/8000 Roche \ Biuret-rate	g/L	45.9	9.4%	<5	<100
Beckman UniCel Dx C 600/800 Beckman Coulter \ Biuret-endpoint	g/L	42.7	10.1%	<5	<100
(TBA)总胆汁酸					
Beckman AU series Fosun \ Enzymatic Cycling Assays Method	μmol/L	76	6.2%	44	875
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	95.3	9.8%	<5	<100
Roche Hitachi/Hitachi Systems SekiSui \ Enzymatic Cycling Assays Method	μmol/L	114	9.6%	<5	<100
Beckman UniCel Dx C 600/800 HuaChenBio \ Enzymatic Cycling Assays Method	μmol/L	115	9.6%	<5	<100
Abbott ARCHITECT c series Randox \ Enzymatic Cycling Assays Method	μmol/L	112	9.8%	<5	<100
(TBIL)总胆红素					
Beckman AU series WanTai \ Diazotization	μmol/L	82.2	2%	5	<100
Beckman AU series Fosun \ Vanadate	μmol/L	71.1	1.8%	46	1228
Beckman AU series DIRUI \ Vanadate	μmol/L	82.6	0.1%	<5	<100
DiruiCS series DIRUI \ Diazotization	μmol/L	71.8	2.4%	<5	<100
Beckman AU series Fosun \ Diazotization -No blank Compensation	μmol/L	68.6	8.1%	25	556
KHB ZY series OTHER \ sodium nitrite oxidase method	μmol/L	87	1.7%	25	144
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	88.4	2.9%	14	390
Roche Hitachi/Hitachi Systems Wako \ Diazotization	μmol/L	67.9	1.9%	13	187
Beckman UniCel Dx C 600/800 Beckman Coulter \ Diazotization	μmol/L	73.9	10%	<5	<100
(TG)甘油三酯					
Beckman AU series BIOBASE \ Dehydrogenase	mmol/L	1.08	4.8%	5	<100
Beckman AU series DIRUI \ Dehydrogenase	mmol/L	1.08	0.8%	<5	<100
DiruiCS series DIRUI \ GPO-POD	mmol/L	1.08	4.3%	<5	<100
Beckman AU series Fosun \ OTHER	mmol/L	0.929	3.1%	24	637
Beckman AU series Fosun \ GPO-POD	mmol/L	0.92	1.5%	20	484
Beckman AU series Fosun \ GK-PK-LDH	mmol/L	0.893	4.2%	20	381
KHB ZY series OTHER \ GPO-POD	mmol/L	1.02	16.3%	<5	<100
Beckman AU series Beckman Coulter \ GPO-POD	mmol/L	1.13	1.8%	14	392

REF	580-2	6x5mL	EXP	2022-07-27	LOT	3755814	更新日期：	2022-05
								6 / 6

		批号：3755814			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems Wako \ Free glycerol removed	mmol/L	1.05	1.9%	10	111
(TIBC)总铁结合力					
Beckman UniCel Dx C 600/800 Beckman Coulter \ Ferrozine colorimetry	μmol/L	37.8	10%	<5	<100
(UA)尿酸					
Beckman AU series WanTai \ Uricase colorimetry	μmol/L	530	2.7%	5	<100
Beckman AU series DIRUI \ Uricase colorimetry	μmol/L	511	0%	<5	<100
DiruiCS series DIRUI \ Uricase colorimetry	μmol/L	549	2%	<5	<100
Beckman AU series Fosun \ Phosphotungstate colorimetry	μmol/L	497	3.7%	25	487
KHB ZY series OTHER \ Uricase colorimetry	μmol/L	465	1.1%	25	144
Beckman AU series Fosun \ OTHER	μmol/L	493	1.6%	24	638
Beckman AU series Fosun \ Uricase colorimetry	μmol/L	492	1.2%	22	557
Beckman AU series Beckman Coulter \ Uricase colorimetry	μmol/L	545	1.3%	14	392
Roche Hitachi/Hitachi Systems MedicalSystem \ Uricase colorimetry	μmol/L	549	1.6%	13	193
Beckman UniCel Dx C 600/800 Beckman Coulter \ Uricase-UV	μmol/L	450	9%	<5	<100
Beckman UniCel Dx C 600/800 Beckman Coulter \ Uricase colorimetry	μmol/L	418	10.1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Phosphotungstate colorimetry	μmol/L	376	9.8%	<5	<100
Beckman UniCel Dx C 600/800 Wako \ Uricase-UV	μmol/L	431	9.4%	<5	<100
(UIBC)不饱和铁结合力					
Beckman UniCel Dx C 600/800 Beckman Coulter \ Ferrozine colorimetry	μmol/L	28.3	10.1%	<5	<100
(Urea)尿素					
Beckman AU series Fosun \ Urease-UV-rate	mmol/L	20.8	2.4%	44	1198
KHB ZY series OTHER \ Urease colorimetry	mmol/L	15.8	1.9%	25	140
Beckman AU series Fosun \ Urease colorimetry	mmol/L	21	5.3%	24	458
DiruiCS series DIRUI \ Urease colorimetry	mmol/L	21.9	4.1%	<5	<100
Beckman AU series Beckman Coulter \ Urease-UV-rate	mmol/L	22.2	1.8%	14	392
(Zn)锌					
Beckman UniCel Dx C 600/800 Beckman Coulter \ Colorimetry	μmol/L	19.5	10.8%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。