报告类别： **汇总实验室数据报告**

质控品名： **凝血五项非定值质控品**

REF	670-2	6×1mL	EXP	2026-04-06	LOT	3FEADE6	更新日期：	2024-05

		批号：3FEADE6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Sysmex XN series sysmex \ Light scattering	Seconds	63.9	3.4%	6	269
Mindray CX Series Mindray \ Light scattering	Seconds	99.1	4.4%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	93.2	21.4%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	74.2	13.4%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Dade Actin FS(APTT) \ Light scattering	Seconds	76.2	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Pathrombin SL(APTT) \ Light scattering	Seconds	105	10%	<5	<100
Werfen ACL series Werfen Hemosil SynthAsil APTT \ Light scattering	Seconds	63.9	10%	<5	<100
Mindray EXC810 Mindray \ OTHER	Seconds	96	4.7%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Dade Actin(APTT) \ Light scattering	Seconds	96.8	10%	<5	<100
Werfen ACL series Werfen Hemosil APTT-SP \ Light scattering	Seconds	68	10%	<5	<100
Mindray CX Series Mindray \ OTHER	Seconds	98.9	3.3%	<5	<100
(AT-III)抗凝血酶III					
Werfen ACL series Werfen IL Antithrombin III Chromogenic（AT□）\ Nephelometry	%	42	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Berichrom Antithrombin III（AT□）\ Nephelometry	%	43	10%	<5	<100
Stago STA series Stago \ Nephelometry	%	43	10%	<5	<100
Sysmex CS 5100 sysmex \ Nephelometry	%	36.8	2.6%	<5	<100
Werfen ACL series Werfen HemosIL Liquid Antithrombin（AT□）\ Nephelometry	%	43	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Innovance Antithrombin（AT□）\ Nephelometry	%	45	10%	<5	<100
(FIB)纤维蛋白原					
Sysmex XN series sysmex \ Light scattering	g/L	2.07	4.7%	6	271
Mindray CX Series Mindray \ clauss	g/L	1.93	4.3%	<5	<100
Stago STA series Stago \ Magnetic Beads	g/L	1.83	22.8%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	g/L	0.715	142%	<5	<100
Werfen ACL series Werfen HemosIL Fib-C \ Light scattering	g/L	1.9	10%	<5	<100
Werfen ACL series Werfen HemosIL QFA Thrombin（FIB）\ Light scattering	g/L	2.01	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Multifibren U（FIB）\ Light scattering	g/L	2.61	10%	<5	<100
Mindray EXC810 Mindray \ OTHER	g/L	1.95	1.1%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Thrombin（FIB）\ Light scattering	g/L	2	10%	<5	<100
Mindray CX Series Mindray \ OTHER	g/L	1.76	2.2%	<5	<100
(PT INR)PT国际标准化比值					
Sysmex XN series sysmex \ Calculate	INR	2.37	2.9%	6	273
Mindray CX Series Mindray \ Calculate	INR	2.04	3%	<5	100
Mindray EXC810 Mindray \ Calculate	INR	1.85	2.6%	<5	<100
Stago STA series Stago \ Calculate	INR	3.67	10%	<5	<100
(PT)血浆凝血酶原时间					
Sysmex XN series sysmex \ Light scattering	Seconds	27	2.9%	6	273
Mindray CX Series Mindray \ OTHER	Seconds	23.1	3%	<5	100
Stago STA series Stago \ Magnetic Beads	Seconds	35.4	9.9%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	26.2	16.4%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin \ Light scattering	Seconds	32	10%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Light scattering	Seconds	32.8	10%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Light scattering	INR	2.48	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Thromborel S（PT）\ Light scattering	INR	2.59	10%	<5	<100
Mindray EXC810 Mindray \ OTHER	Seconds	21.8	2.5%	<5	<100
Stago STA series Stago \ Magnetic Beads	INR	2.96	10%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin \ Light scattering	INR	2.52	10%	<5	<100

报告类别： **汇总实验室数据报告**

质控品名： **凝血五项非定值质控品**

REF	670-2	6×1mL	EXP	2026-04-06	LOT	3FEADE6	更新日期：	2024-05

		批号：3FEADE6			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Sysmex CA 1500/6000/7000 Siemens Thromborel S（PT）\ Light scattering	Seconds	34.2	10%	<5	<100
Werfen ACL series Werfen Hemosil PT-Fib HS Plus \ Light scattering	Seconds	41.6	10%	<5	<100
Werfen ACL series Werfen Hemosil PT-Fib HS Plus \ Light scattering	INR	2.75	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens PT Innovin \ Light scattering	Seconds	29.9	10%	<5	<100
Sysmex CA 1500/6000/7000 Siemens PT Innovin \ Light scattering	INR	2.46	10%	<5	<100
(TT)凝血酶时间					
Sysmex XN series sysmex \ Light scattering	Seconds	16.6	2.7%	6	274
Mindray CX Series Mindray \ OTHER	Seconds	21.3	5.7%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	18.1	6.3%	<5	<100
Mindray EXC810 Mindray \ OTHER	Seconds	17.1	1.6%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。