

质控品名： 特殊肿瘤标志物液体非定值质控品

REF	513-2	6×3mL	EXP	2027-04-30	LOT	30F1D5B	更新日期：	2024-05
1 / 2								

批号：30F1D5B					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	51.6	3%	9	183
OTHER OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	53.5	6.9%	5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	53.7	4.9%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	54.5	12%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.01	6.7%	<5	<100
(CA 125)糖类抗原 125					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	170	3.3%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	314	7.4%	10	209
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	249	12%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.9	7.3%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	231	12%	<5	<100
(CA 15-3)糖类抗原 15-3					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	47.9	5.5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	115	7%	10	231
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/mL	3.79	27.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	114	12%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	103	12%	<5	<100
(CA 19-9)糖类抗原 19-9					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	445	9.5%	8	205
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	91.7	7.4%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	171	12%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/mL	7.42	11.6%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	67	12%	<5	<100
(CA 50)糖类抗原 50					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	39.3	4%	12	277
(CA 72-4)糖类抗原 72-4					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	44.1	2.6%	12	305
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.9	12%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/mL	2.4	16.2%	<5	<100
(CA242)糖类抗原 242					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	22.7	3.2%	12	279
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	U/mL	38	12%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	U/mL	13	4.8%	<5	<100
(CEA)癌胚抗原					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	40	3.8%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	58.5	5.9%	10	211
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.95	18.1%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	45.9	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	38.3	12%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.1	5.5%	8	177
Beckman UniCel Dxl 600/800 OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.9	6.7%	<5	<100

REF	513-2	6×3mL	EXP	2027-04-30	LOT	30F1D5B	更新日期：	2024-05
2 / 2								

		批号：30F1D5B			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	15.7	12%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.88	10.6%	<5	<100
(Ferritin)铁蛋白					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	243	16.6%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	426	6.7%	10	222
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	340	12%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.65	11%	8	202
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.06	5.3%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.89	12%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.518	13.5%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.55	12%	<5	<100
(hCG)人绒毛膜促性腺激素					
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	128	12%	<5	<100
(HE4)人附睾蛋白4					
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	253	6.6%	9	180
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	289	12%	<5	<100
(PSA)前列腺特异性抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.8	9.9%	9	219
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.3	3.3%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.9	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	18.5	12%	<5	<100
(SCC)鳞状细胞癌抗原					
Beckman UniCel DxI 600/800 OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.4	6.1%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.3	6.2%	10	208
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.25	9.4%	<5	<100
(TPA)组织多肽抗原					
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	U/L	2445	12%	<5	<100
(TPS)组织多肽特异抗原					
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	U/L	83	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。