

质控品名： 复合免疫分析非定值质控品

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
1 / 13								

		批号：3387621			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(17-OHP)17-羟孕酮					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.46	10.4%	<5	<100
(A II)血管紧张素II					
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/L	188	10.1%	<5	<100
(AFP)甲胎蛋白					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	256	6.9%	71	1339
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	283	18.2%	7	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	261	7.3%	40	974
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	227	4.1%	<5	114
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	232	6.9%	<5	107
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	248	4%	35	717
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	201	7.2%	33	894
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	257	4.2%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	266	3.5%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	216	6.2%	26	198
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	224	4.7%	24	664
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	252	4.1%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	229	5%	19	177
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	222	5.7%	13	191
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	247	8.2%	10	170
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	234	9.8%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	248	10.1%	<5	<100
(ALD)醛固酮					
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	pg/mL	348	11.5%	15	147
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	264	9.8%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	351	10.1%	<5	<100
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	pg/mL	232	24.9%	<5	<100
(Anti-Tg)抗甲状腺球蛋白抗体					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	36.8	11.9%	8	139
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	47.9	7.6%	7	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	56.1	44.7%	68	876
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	43.7	9.6%	37	1148
Beckman UniCel Dxl 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	7.14	13.7%	31	771
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	20.2	4.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	44.2	11%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	20.1	4.3%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	17.3	9.2%	<5	<100
(Anti-TPO)抗甲状腺过氧化物酶抗体					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27.6	35.3%	9	222
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	45.5	16.7%	7	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	29.1	21.7%	42	1363

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
2 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	30.2	5.3%	<5	104
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	18.1	10%	33	833
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33	2.8%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	20.6	9.7%	<5	<100
(ASD)雄烯二酮					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.53	10.4%	<5	<100
(CA 125)糖类抗原 125					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	114	4.1%	9	153
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	108	9%	6	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	90.7	8.2%	41	825
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	104	5.4%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	121	11.6%	<5	111
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	105	7.9%	38	960
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	96.1	6%	35	725
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	122	4.4%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	110	7.4%	27	182
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	98.9	7.7%	19	525
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	133	6.1%	19	184
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	98.3	4.4%	15	197
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	103	9.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	167	9.6%	<5	<100
(CA 15-3)糖类抗原 15-3					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	71.6	9.1%	9	151
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	76	10.1%	6	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	66.8	11.5%	42	1040
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	74.8	12.3%	40	782
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	64.8	7.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	64.4	7.1%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	86.9	3.7%	<5	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	88.8	11.1%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	35.8	10.4%	26	190
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	94.4	24.9%	21	502
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	66.9	9.8%	15	197
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	67.3	9.8%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	69.5	11.1%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	66.9	9.9%	<5	<100
(CA 19-9)糖类抗原 19-9					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	97.4	8.3%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	95.3	9.2%	41	796
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	95.6	5.6%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	96.2	6.6%	39	976
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	91.8	4.8%	35	708

质控品名： 复合免疫分析非定值质控品

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
3 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	132	6.1%	15	185
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	102	9.8%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	94.5	9.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	79.4	9.6%	<5	<100
(CEA)癌胚抗原					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24.6	4.5%	70	1282
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	23.6	11.3%	7	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	25.5	3.9%	40	967
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	25.2	4.3%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	30.6	11.3%	<5	106
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	26.2	14%	36	754
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	28.4	5.9%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24.5	6.4%	26	184
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	31.1	5.7%	23	278
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	31.2	3.7%	22	562
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	32.1	10.2%	13	180
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	26.6	7.1%	10	167
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	25.9	10%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
J&J Vitros ECL Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	22.3	11%	8	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	35.5	9.6%	8	260
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	40.6	6.3%	7	125
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	33.9	7.9%	33	903
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	29.6	4.8%	30	511
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	27.4	8.5%	28	627
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	30	6.5%	14	320
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	26	9.6%	13	364
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	31	7.8%	11	218
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	37.8	12.2%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	33.5	10.1%	<5	<100
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	19.8	3.7%	<5	<100
(Cortisol)皮质醇					
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	282	22.2%	44	1174
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	279	9.1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	287	3.7%	39	1134
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	310	2.3%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	291	4.1%	<5	<100
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	83.6	128.6%	15	153
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	270	5.9%	11	311
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	335	10.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	283	11.3%	<5	<100

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
4 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	355	10.1%	<5	<100
(C-P)C-肽					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.74	2%	9	<100
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.69	5.6%	8	108
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.8	37.8%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.77	5.9%	38	766
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.53	37%	31	813
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.73	5.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.57	8.7%	24	278
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.7	4%	12	314
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.02	10%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.12	9.9%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.89	10.3%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.01	5.6%	<5	<100
(DGXN)地高辛					
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.74	6.6%	19	166
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.38	10.2%	<5	<100
(DHEA-S)硫酸脱氢表雄酮					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10681	31.1%	5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9427	18.6%	28	879
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13117	9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14259	9.8%	<5	<100
(E2)雌二醇					
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1715	12.2%	8	171
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1831	6.4%	74	3158
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1758	4.1%	5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	1587	13.9%	45	1123
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1720	5.1%	43	1249
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1539	8.9%	<5	118
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	2339	5.2%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	pmol/L	1669	5.1%	29	507
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1125	9.5%	18	353
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1810	6.8%	17	244
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1701	8.8%	16	160
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1707	4.9%	12	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1662	7.2%	11	316
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1478	9.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	1516	9.9%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1820	3.7%	<5	<100
(EPO)促红细胞生成素					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	117	5.3%	7	170

质控品名： 复合免疫分析非定值质控品

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
5 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/mL	117	8.5%	33	872
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	120	5.2%	26	559
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/mL	141	9.9%	<5	<100
(FA)叶酸					
Beckman AU series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	14	5.2%	7	236
maglumi system Snibe \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	11.9	6%	6	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	13.5	23%	44	1122
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.7	8.1%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	13.5	8.5%	38	717
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.6	6.5%	28	592
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.3	15.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16	11.9%	18	145
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.8	9.3%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.6	10%	<5	<100
(Ferritin)铁蛋白					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	546	7%	9	165
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	407	6.1%	7	241
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	600	8.3%	6	135
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	598	10.3%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	546	7.7%	46	821
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	327	36%	44	1159
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	518	6.9%	31	802
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	390	6.9%	27	657
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	457	13.5%	21	245
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	579	1.9%	<5	<100
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	584	7.9%	15	186
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	452	10%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	90.8	2.4%	<5	<100
(F-PSA)游离前列腺特异性抗原					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.78	9%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.89	9.5%	41	804
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.6	18%	<5	125
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.95	5.2%	38	964
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	11.6	7.2%	33	894
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.9	5.3%	<5	<100
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.5	6%	21	528
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.15	10.7%	<5	<100
(FSH)促卵泡刺激素					
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	71.2	13.7%	8	117
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	52.6	8.5%	61	1574
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	38.1	24.6%	44	1101
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	56.6	4.4%	43	1207

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
6 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	57.1	7.6%	<5	109
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	53.3	6.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	IU/L	81.3	6.2%	29	505
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	65.1	5.2%	<5	<100
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	IU/L	39.4	5.4%	18	361
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	47.1	10.9%	16	197
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	52.1	11.9%	13	183
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	57.8	2.5%	12	103
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	46.8	6.5%	11	307
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	42.7	10.5%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	76.4	12.8%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	52.9	7%	<5	<100
(FT3)游离T3					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	39.8	16.5%	9	203
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	21.1	5.1%	9	153
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	34.8	1.2%	7	138
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	pmol/L	22	6.9%	58	1028
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	25.1	17.5%	5	113
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	28	3%	46	1219
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	22	9.1%	<5	112
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	27.5	2.8%	35	766
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	16.9	6.9%	35	1662
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	22.2	8.5%	22	259
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	3.9	13.7%	20	539
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	21.6	6.8%	18	769
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	20.7	7.6%	16	333
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	18.2	6.8%	11	314
Sysmex XP Series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	21.6	9.7%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	19.5	10.8%	<5	<100
(FT4)游离T4					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	47.5	4.1%	7	<100
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	53.3	2.7%	7	137
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	pmol/L	43.4	7%	56	1041
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	58	4.7%	47	1223
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	36.2	15.5%	<5	119
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	28.1	23.3%	<5	104
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	58.1	7.3%	33	753
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	46.7	4.9%	33	1716
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	45.7	6.2%	24	292
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	48.2	10.1%	<5	<100
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	31	5.7%	18	761
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	13.6	4.3%	18	518

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
7 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	45.4	8.4%	16	313
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	46.1	4.3%	11	315
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	43.6	10.6%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	55.8	1.9%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	52.4	10.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	46.7	9.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	31.4	2.6%	<5	<100
(HCY)同型半胱氨酸					
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	μmol/L	34.3	12%	22	233
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μmol/L	49.4	10.1%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μmol/L	48.9	10%	<5	<100
(HGH)人生长激素					
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	μg/L	8.49	9.3%	33	876
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	9.06	4.7%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	μg/L	9.23	10%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μg/L	8.3	10.4%	<5	<100
(IgE)免疫球蛋白 IgE					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	318	6%	30	744
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	317	8.2%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	329	5.8%	12	358
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	224	11.2%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/mL	289	10.4%	<5	<100
(IGF1)胰岛素生长因子1					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	nmol/L	22.6	9.3%	<5	<100
(INS)胰岛素					
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	134	6.1%	8	131
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	132	4.1%	52	1366
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	126	4.5%	51	1459
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	127	6.2%	39	765
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	132	3.5%	31	806
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	112	5.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	114	7.7%	24	296
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	mIU/L	104	14.6%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mIU/L	132	9.8%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	119	10.1%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	119	10.9%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	168	1.2%	<5	<100
(LH)促黄体生成素					
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	52.1	6.1%	8	153
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	51.6	7%	73	1910
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	27.1	26.5%	44	1136
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	55.8	6.1%	43	1217

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
8 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	46.1	13.1%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	72	10.6%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	IU/L	51.4	5.5%	29	515
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	42.1	3.4%	<5	<100
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	IU/L	47.5	4.1%	18	362
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	54.4	5.3%	16	217
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	56	2.7%	13	130
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	32.2	8.1%	11	307
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	39.5	22.1%	10	115
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	50.3	10.1%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	IU/L	47.2	10.4%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/L	44.4	4.1%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	68.6	1.5%	<5	<100
(Myo / Mb)肌红蛋白					
J&J Vitros ECL Ortho \ Chemiluminescence Immunoassay (CLIA)	μg/L	351	3.6%	8	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	234	7%	8	280
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	μg/L	229	7.6%	33	920
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	273	3.2%	30	510
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	365	3.7%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	249	3.9%	28	627
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	μg/L	229	6.3%	25	605
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	278	5.3%	15	364
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	μg/L	348	7.6%	13	363
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	270	5%	11	244
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	μg/L	266	10.5%	<5	<100
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	μg/L	350	3.3%	<5	<100
(NT-proBNP)N-末端脑钠肽前体					
J&J Vitros ECL Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	3843	9.2%	7	<100
J&J Vitros ECL Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	452	9.2%	7	<100
J&J Vitros 250/350 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	3569	1.8%	7	194
J&J Vitros 250/350 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	420	1.8%	7	194
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	pg/mL	2350	2.3%	5	<100
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	276	2.3%	5	<100
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	3641	4.5%	5	103
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	428	4.5%	5	103
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	2087	7.7%	40	1597
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	245	7.7%	40	1597
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	1947	10.5%	30	555
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	229	10.5%	30	555
Roche cobas E601/602 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	1884	3%	<5	<100
Roche cobas E601/602 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	222	3%	<5	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	2084	6.3%	28	687

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
9 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	245	6.3%	28	687
Ortho Vitros 3600/5600 Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	4015	18%	28	420
Ortho Vitros 3600/5600 Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	472	18%	28	420
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	2388	7.4%	24	899
(PRL)催/ 泌乳素					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	34.9	6.3%	9	187
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	29.6	8.3%	8	158
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	36.3	4.2%	72	1871
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	23.5	22.8%	44	1158
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	37	5.1%	43	1218
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	35.4	6.6%	<5	114
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	32.8	3.7%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	ng/mL	28.2	7.2%	29	520
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	29.9	5.1%	<5	<100
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	27.9	5.3%	18	353
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	30.4	14.4%	16	182
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	27.7	5.8%	11	315
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	36.3	10.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	31	4.2%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	32.4	11.2%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	36.1	0%	<5	<100
(PROG)孕酮					
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	16.1	8.3%	8	172
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	25.9	13.1%	60	2351
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.1	18.1%	5	<100
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	19.8	18.6%	45	1136
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	22.7	13%	<5	114
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	17.8	5.1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24.2	18.2%	29	761
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	ng/mL	14.4	6%	29	520
Ortho Vitros 3600/5600 Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	26.9	2.8%	<5	<100
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18	7.3%	18	346
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	22.1	11.4%	17	200
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	21.3	28.4%	15	208
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.06	247.7%	13	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	21.1	6.8%	11	304
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	27.1	10.3%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	16.9	1.6%	<5	<100
(PSA)前列腺特异性抗原					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.7	10.2%	5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	22.1	6%	41	792
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	20.7	12.1%	<5	113

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
10 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24.7	3.9%	38	935
Beckman UniCel DxI 600/800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24	6.4%	33	888
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	20.1	4.6%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	21.2	4.3%	<5	<100
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.1	4.9%	20	514
Siemens ADVIA Centaur XP/XPT Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	21.3	4.9%	<5	<100
Autolumo A2000/A2000 PLUS Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	30	7.1%	14	218
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	21.7	9.7%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	24.5	9.8%	<5	<100
(PTH)甲状旁腺素					
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	21	8%	9	196
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	17.9	28.4%	44	1143
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	18.5	7.7%	35	718
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	15.2	4%	23	605
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pmol/L	21.7	5.9%	21	172
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	26.1	5.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	19.5	8.2%	12	132
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	17.2	10.4%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Enzyme linked immunosorbent assays (ELISA)	pmol/L	31.4	9.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pmol/L	20.6	10.7%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	pmol/L	17.8	10.1%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	23.1	0.2%	<5	<100
(Renin)肾素					
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	μIU/mL	17.8	9.6%	<5	<100
(rT3)反三碘甲状腺原氨酸					
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.62	12%	<5	<100
(SHBG)性激素结合球蛋白					
Beckman UniCel DxI 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	nmol/L	67.2	8.8%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	nmol/L	68.1	12%	<5	<100
(TES)睾酮					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.39	6.6%	9	206
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.32	5.3%	72	2105
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.94	7.6%	7	146
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.27	5%	43	1190
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.25	9%	<5	116
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.35	6%	33	914
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.96	9.2%	<5	<100
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.81	8%	17	298
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	ng/mL	1.89	6.5%	14	264
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.25	4.5%	11	315
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.33	9.5%	10	133

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
11 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.76	10.2%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.03	10.5%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.93	5.3%	<5	<100
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.82	2.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.15	0%	<5	<100
(Tg)甲状腺球蛋白					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.6	8.2%	8	152
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	8.28	11.8%	33	843
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.2	2.3%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.72	7.9%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.6	9.6%	23	350
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.91	8.8%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Enzyme linked immunosorbent assays (ELISA)	ng/mL	4.23	10.9%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.3	8.9%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.3	10.7%	<5	<100
Roche cobas 4000/6000/8000 OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11	0%	<5	<100
(THEO)茶碱					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	µg/mL	30.1	10.2%	<5	<100
(TnI)肌钙蛋白I					
J&J Vitros ECL Ortho \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.61	3.9%	8	<100
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.9	11.4%	8	108
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.65	5.5%	7	239
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.41	11.9%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	ng/mL	15.4	11.1%	25	454
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.56	5%	24	702
Beckman AU series OTHER \ OTHER	ng/mL	1.36	1.4%	20	479
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.005	4.6%	<5	<100
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.06	3.7%	13	361
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.23	10.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	20.2	2.5%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.96	9.9%	<5	<100
Sysmex XP Series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	19.3	9.8%	<5	<100
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.005	3.5%	<5	<100
(TnT)肌钙蛋白T					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.98	6.2%	31	616
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2	4.6%	29	699
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.9	6.8%	28	630
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.95	8.3%	25	576
(TSH)促甲状腺素					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	32.4	4.6%	73	1921
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	mIU/L	25.3	10.3%	7	142

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
12 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	28.6	3.8%	7	104
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	mIU/L	29.9	4.8%	57	1019
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	32.9	3%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	29.9	6.7%	<5	109
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	32.6	4.6%	33	759
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	26.9	5.9%	33	1724
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	mIU/L	29.6	4.9%	22	262
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	24.2	4.5%	19	797
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	27.5	7.6%	16	311
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mIU/L	26.8	6.3%	11	315
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	32.8	8.1%	10	211
Beckman UniCel Dxl 600/800 Beckman Coulter \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	31.2	10.3%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	28.9	9.7%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	mIU/L	31.2	9.3%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	25.5	0.3%	<5	<100
(TT3)总三碘甲状腺氨酸					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.07	7.3%	9	151
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.65	3.6%	7	145
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	ng/mL	4.33	9%	57	1014
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.89	3%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.92	14.3%	<5	122
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.39	9.6%	33	880
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.39	9.7%	23	220
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.17	4.2%	21	470
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.006	2.6%	<5	<100
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.44	6.5%	19	768
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.828	15.3%	19	528
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.64	9%	16	323
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.32	7.8%	12	343
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.63	14%	10	194
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.71	10.2%	<5	<100
Sysmex XP Series Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.69	10%	<5	<100
(TT4)总T4					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	190	7%	8	101
Ortho Vitros 3600/5600 OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	185	10.8%	7	142
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	ng/mL	197	9.7%	55	1027
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	158	3.4%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	172	13.4%	<5	117
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	161	7%	33	880
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	214	10.2%	24	230
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	170	3.9%	21	461
Abbott ARCHITECT ci series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	88	10.8%	20	516

REF	540-3	6×3mL	EXP	2023-04-29	LOT	3387621	更新日期：	2024-05
13 / 13								

		批号：3387621			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	202	3.3%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	175	9.5%	<5	<100
Abbott Alinity ci-series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	172	5.1%	19	756
Siemens ADVIA Centaur XP/XPT ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	211	10.7%	16	333
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	208	6.6%	11	316
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	149	9.9%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	193	9.8%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	211	10.4%	<5	<100
(VB12)维生素B12					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	611	6.6%	7	238
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	pg/mL	751	5.8%	6	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	737	6.8%	5	103
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	pg/mL	554	12.5%	44	1153
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	880	9.4%	39	737
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	pg/mL	597	7%	26	561
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	pg/mL	848	16.5%	20	180
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	pg/mL	846	10.1%	<5	<100
(β2M)β-2-微球蛋白					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2584	9.6%	<5	<100
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1880	11.1%	<5	<100
Tosoh AIA 900/1800 TOSOH \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2820	10.3%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2560	10.5%	<5	<100
Beckman AU series XC Bio \ Immunonephelometry (INA)	ng/mL	2888	1.8%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.395	2.6%	7	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.312	5.6%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.392	4.1%	36	1197
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.329	5.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ OTHER	IU/mL	0.339	5.5%	27	513
Roche cobas e411 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.408	7.1%	22	586
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.385	10%	13	151
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.395	2.4%	11	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.378	4.7%	11	314
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.402	9%	<5	<100
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.404	1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。