



REF	782-A	6×1mL	EXP	2023-09-29	LOT	4A96C0D	更新日期：	2024-05
1 / 1								

		批号：4A96C0D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(HBeAg)乙肝e抗原					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	Sample/Cutoff	1.84	5.7%	9	247
YHLO iFlash Analyzer YHL \ Chemiluminescence Immunoassay (CLIA)	Sample/Cutoff	2.61	7.5%	7	193
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	Sample/Cutoff	6.55	23.1%	6	135
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	PEIU/ml	0.906	5.2%	5	<100
Abbott ARCHITECT i series OTHER \ Chemiluminescence Immunoassay (CLIA)	Sample/Cutoff	3.41	16.1%	40	930
WANTAI Biopharm Wan200 WanTai \ Chemiluminescence Immunoassay (CLIA)	PEIU/ml	0.276	18%	<5	<100
Sysmex HISCL-5000 sysmex \ Chemiluminescence Immunoassay (CLIA)	Sample/Cutoff	10.6	17.3%	16	301
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	Sample/Cutoff	3.49	18%	15	218
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	PEIU/ml	0.578	22.1%	14	129
Abbott ARCHITECT ci series Dedicated \ Chemiluminescence Immunoassay (CLIA)	Sample/Cutoff	5.66	18.6%	10	274
Caris200 WanTai \ Chemiluminescence Immunoassay (CLIA)	PEIU/ml	0.294	10.4%	10	139
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	NCU/ml	1.83	1.3%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	Sample/Cutoff	3.89	10%	<5	<100
Maccura IS1200 MAKER \ Chemiluminescence Immunoassay (CLIA)	Sample/Cutoff	5.66	9.9%	<5	<100
Manual KHB \ Enzyme linked immunosorbent assays (ELISA)	Sample/Cutoff	1.36	16.9%	<5	<100
Manual InTec \ Enzyme linked immunosorbent assays (ELISA)	Sample/Cutoff	1.89	14.8%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	Sample/Cutoff	11.6	11.2%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。