

REF	671-1	6×1mL	EXP	2026-06-03	LOT	4000D78	更新日期：	2024-04
1 / 1								

批号：4000D78					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Sysmex CA 1500/6000/7000	Seconds	38.1	10%	<5	<100
Siemens Pathrombin SL(APTT) \ Light scattering					
Werfen ACL series	Seconds	28.8	10%	<5	<100
Werfen Hemosil APTT-SP \ Light scattering					
Stago STA series	Seconds	34	10%	<5	<100
Stago \ Magnetic Beads					
Sysmex CA 1500/6000/7000	Seconds	34	10%	<5	<100
Siemens Dade Actin FS(APTT) \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	35	10%	<5	<100
Siemens Dade Actin(APTT) \ Light scattering					
Werfen ACL series	Seconds	33	10%	<5	<100
Werfen Hemosil SynthAsil APTT \ Light scattering					
(AT-III)抗凝血酶III					
Sysmex CA 1500/6000/7000	%	98	10%	<5	<100
Siemens Berichrom Antithrombin III (AT□) \ Nephelometry					
Werfen ACL series	%	108	10%	<5	<100
Werfen Hemosil Liquid Antithrombin (AT□) \ Nephelometry					
Werfen ACL series	%	99	10%	<5	<100
Werfen IL Antithrombin III Chromogenic (AT□) \ Nephelometry					
Sysmex CA 1500/6000/7000	%	101	9.9%	<5	<100
Siemens Innovance Antithrombin (AT□) \ Nephelometry					
(D-D(DDU)) D-二聚体(DDU)					
Werfen ACL series	mg/L	0.552	10%	<5	<100
Werfen Hemosil D-Dimer HS \ Nephelometry					
Werfen ACL series	mg/L	0.577	10%	<5	<100
Wenfen Hemosil D-Dimer \ Nephelometry					
(D-D(FEU)) D-二聚体(FEU)					
Sysmex CA 1500/6000/7000	mg/L	1.6	10%	<5	<100
Siemens Innovance Antithrombin (AT□) \ Nephelometry					
Stago STA series	mg/L	0.95	10%	<5	<100
Stago \ Nephelometry					
(FIB)纤维蛋白原					
Werfen ACL series	g/L	3.47	10%	<5	<100
Werfen Hemosil Fib-C \ Light scattering					
Stago STA series	g/L	3.74	10%	<5	<100
Stago \ Magnetic Beads					
Werfen ACL series	g/L	3.8	10%	<5	<100
Werfen Hemosil QFA Thrombin (FIB) \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	3.69	10%	<5	<100
Siemens Multifibren U (FIB) \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	3.4	10%	<5	<100
Siemens Thrombin (FIB) \ Light scattering					
(PT)血浆凝血酶原时间					
Werfen ACL series	INR	1	10%	<5	<100
Werfen Hemosil PT-Fib HS Plus \ Light scattering					
Stago STA series	INR	0.98	10%	<5	<100
Stago \ Magnetic Beads					
Werfen ACL series	INR	0.98	9.7%	<5	<100
Werfen Hemosil PT Recombiplastin 2G \ Light scattering					
Werfen ACL series	Seconds	15	10%	<5	<100
Werfen Hemosil PT-Fib HS Plus \ Light scattering					
Sysmex CA 1500/6000/7000	INR	1.08	10.2%	<5	<100
Siemens PT Innovin \ Light scattering					
Werfen ACL series	Seconds	11	10%	<5	<100
Werfen Hemosil PT Recombiplastin \ Light scattering					
Werfen ACL series	Seconds	10.9	10%	<5	<100
Werfen Hemosil PT Recombiplastin 2G \ Light scattering					
Stago STA series	Seconds	13.5	10%	<5	<100
Stago \ Magnetic Beads					
Werfen ACL series	INR	0.98	10%	<5	<100
Werfen Hemosil PT Recombiplastin \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	12.6	10%	<5	<100
Siemens Thromborel S (PT) \ Light scattering					
Sysmex CA 1500/6000/7000	INR	1	10%	<5	<100
Siemens Thromborel S (PT) \ Light scattering					
(TT)凝血酶时间					
Werfen ACL series	Seconds	14.3	10%	<5	<100
Werfen \ Light scattering					

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。