

REF	580-2	6×5mL	EXP	2025-02-19	LOT	375593A	更新日期：	2024-05
1 / 7								

		批号：375593A			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(5'NT)5`核苷酸酶					
Roche Hitachi/Hitachi Systems OTHER \ OTHER	U/L	6.68	5.2%	12	415
(a-HBDH)α-羟丁酸脱氢酶					
Beckman AU series Fosun \ OTHER	U/L	362	1.8%	8	211
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic Cycling Assays Method	U/L	318	2.3%	6	114
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	342	1.9%	12	427
(Alb)白蛋白					
Beckman AU series LEADMAN \ Bromocresol green	g/L	28.6	1.7%	7	165
Roche Hitachi/Hitachi Systems Wako \ Bromocresol green	g/L	30.6	1.5%	6	114
Beckman AU series Fosun \ Bromocresol green	g/L	29.8	2.4%	23	633
Beckman AU series MedicalSystem \ Bromocresol green	g/L	22.7	2.6%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Bromocresol green	g/L	28.9	0.5%	<5	<100
Mindray BS series Mindray \ Bromocresol green	g/L	29.2	1.2%	<5	<100
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	28.2	1.6%	16	335
KHB ZY series KHB \ Bromocresol green	g/L	27.8	3%	11	112
Roche cobas 4000/6000/8000 Roche \ Bromocresol green	g/L	29.4	12%	<5	<100
Beckman AU series Genrui \ Bromocresol green	g/L	29.6	0%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ Bromocresol green	g/L	27.6	0%	<5	<100
DiruiCS series DIRUI \ Bromocresol green	g/L	30.3	0%	<5	<100
(ALP)碱性磷酸酶					
Beckman AU series LEADMAN \ Rate -AMP buffer	U/L	322	3.1%	7	166
Roche Hitachi/Hitachi Systems MAKER \ Rate -AMP buffer	U/L	291	2.5%	6	114
Beckman AU series Fosun \ Rate -AMP buffer	U/L	297	2.4%	23	612
Beckman AU series MedicalSystem \ Rate -AMP buffer	U/L	208	3.4%	<5	<100
Mindray BS series Mindray \ dry chemistry	U/L	294	3.3%	<5	<100
Beckman AU series Beckman Coulter \ dry chemistry	U/L	357	19.4%	<5	<100
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	310	1.8%	14	351
KHB ZY series KHB \ Rate -AMP buffer	U/L	336	4%	11	112
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	225	12%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ dry chemistry	U/L	291	0%	<5	<100
DiruiCS series DIRUI \ dry chemistry	U/L	283	0%	<5	<100
Beckman AU series Genrui \ dry chemistry	U/L	330	0%	<5	<100
(ALT/ SGPT)丙氨酸氨基转移酶					
Beckman AU series Fosun \ OTHER	U/L	279	1.2%	8	211
Beckman AU series Fosun \ UV-rate without P5P	U/L	289	2.9%	8	211
Beckman AU series Fosun \ dry chemistry	U/L	278	1.1%	7	181
Beckman AU series LEADMAN \ UV-rate without P5P	U/L	258	2.4%	7	165
Roche Hitachi/Hitachi Systems Wako \ UV-rate without P5P	U/L	268	1%	6	113
Beckman AU series MedicalSystem \ UV-rate without P5P	U/L	257	2%	<5	<100
Beckman AU series Beckman Coulter \ dry chemistry	U/L	283	0.4%	<5	<100
Mindray BS series Mindray \ dry chemistry	U/L	287	3.2%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ dry chemistry	U/L	294	3.4%	<5	<100
Beckman AU series Beckman Coulter \ UV-rate without P5P	U/L	259	2.1%	13	334
KHB ZY series KHB \ UV-rate without P5P	U/L	279	4%	11	116
Roche cobas 4000/6000/8000 Roche \ UV-rate with P5P	U/L	256	12%	<5	<100
Beckman AU series Genrui \ dry chemistry	U/L	296	0%	<5	<100
KHB ZY series YTZY \ dry chemistry	U/L	223	0%	<5	<100



REF	580-2	6×5mL	EXP	2025-02-19	LOT	375593A	更新日期：	2024-05
2 / 7								

		批号：375593A			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
DiruiCS series DIRUI \ dry chemistry	U/L	308	0%	<5	<100
(AMY)淀粉酶					
Beckman AU series LEADMAN \ PNP-G7	U/L	239	6.6%	7	166
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	253	1.6%	7	160
Beckman AU series Fosun \ PNP-G7	U/L	260	1.4%	15	412
Roche cobas 4000/6000/8000 Roche \ PNP-G7	U/L	242	12%	<5	<100
(AST/ SGOT)天门冬氨酸氨基转移酶					
Beckman AU series Fosun \ OTHER	U/L	236	2.2%	8	211
Beckman AU series Fosun \ dry chemistry	U/L	236	2.2%	7	202
Beckman AU series Fosun \ UV-rate without P5P	U/L	243	2.1%	7	169
Beckman AU series LEADMAN \ UV-rate without P5P	U/L	232	2.5%	7	164
Roche Hitachi/Hitachi Systems Wako \ UV-rate without P5P	U/L	246	2%	6	114
Beckman AU series MAKER \ OTHER	U/L	228	2%	<5	153
Beckman AU series MedicalSystem \ UV-rate without P5P	U/L	211	2.9%	<5	<100
Beckman AU series Beckman Coulter \ dry chemistry	U/L	238	3%	<5	<100
Mindray BS series Mindray \ dry chemistry	U/L	249	1.6%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ dry chemistry	U/L	230	21.8%	<5	<100
Beckman AU series Beckman Coulter \ UV-rate without P5P	U/L	238	1.7%	13	318
KHB ZY series KHB \ dry chemistry	U/L	279	3.3%	11	107
Roche cobas 4000/6000/8000 Roche \ UV-rate with P5P	U/L	246	12%	<5	<100
KHB ZY series YTZY \ dry chemistry	U/L	193	0%	<5	<100
DiruiCS series DIRUI \ dry chemistry	U/L	225	0%	<5	<100
(Ca)钙					
Roche Hitachi/Hitachi Systems Wako \ MTB colorimetry	mmol/L	2.8	1.1%	6	114
Beckman AU series Beckman Coulter \ Arsenazo III	mmol/L	2.79	1.6%	22	500
Beckman AU series Fosun \ Arsenazo III	mmol/L	2.71	1.6%	15	409
Roche cobas 4000/6000/8000 Roche \ BAPTA	mmol/L	2.82	12%	<5	<100
Beckman AU series Genrui \ Arsenazo III	mmol/L	2.83	0%	<5	<100
Mindray BS series Mindray \ Arsenazo III	mmol/L	2.8	0%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ Arsenazo III	mmol/L	2.83	0%	<5	<100
(CHE)胆碱酯酶					
Roche Hitachi/Hitachi Systems MAKER \ OTHER	U/L	3040	2.6%	12	462
KHB ZY series KHB \ Butyrylthiocholine Method	U/L	3338	2.7%	11	115
(CHOL/ TC)总胆固醇					
Beckman AU series Fosun \ OTHER	mmol/L	2.41	1.5%	8	211
Beckman AU series LEADMAN \ Cholesterol oxidase	mmol/L	2.38	3.1%	7	166
Roche Hitachi/Hitachi Systems Wako \ Cholesterol oxidase	mmol/L	2.66	1.2%	6	114
Beckman AU series MAKER \ OTHER	mmol/L	2.66	0.9%	<5	<100
Beckman AU series MedicalSystem \ Cholesterol oxidase	mmol/L	2.05	3.6%	<5	<100
Beckman AU series Beckman Coulter \ Cholesterol oxidase	mmol/L	2.57	10.5%	<5	<100
Mindray BS series Mindray \ Cholesterol oxidase	mmol/L	2.4	5.9%	<5	<100
Beckman AU series Fosun \ Cholesterol oxidase	mmol/L	2.37	2.7%	13	295
KHB ZY series KHB \ Cholesterol oxidase	mmol/L	2.62	6.2%	11	114
Roche cobas 4000/6000/8000 Roche \ Cholesterol oxidase	mmol/L	2.37	12%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ Cholesterol oxidase	mmol/L	2.54	0%	<5	<100
DiruiCS series DIRUI \ Cholesterol oxidase	mmol/L	2.6	0%	<5	<100
KHB ZY series YTZY \ Cholesterol oxidase	mmol/L	2.87	0%	<5	<100

REF	580-2	6×5mL	EXP	2025-02-19	LOT	375593A	更新日期：	2024-05
3 / 7								

		批号：375593A			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Genrui \ Cholesterol oxidase	mmol/L	3.09	0%	<5	<100
(CK)肌酸激酶					
Beckman AU series Fosun \ OTHER	U/L	204	3.4%	8	213
Beckman AU series Fosun \ DGKC	U/L	229	3.1%	8	160
Beckman AU series Fosun \ Rate -AMP buffer	U/L	205	3.6%	7	190
Beckman AU series LEADMAN \ Rate -other buffers	U/L	215	3.9%	7	167
Roche Hitachi/Hitachi Systems MAKER \ Rate -AMP buffer	U/L	239	2.4%	6	113
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	202	12%	<5	<100
(Cl)氯					
Beckman AU series Fosun \ ISE (direct)	mmol/L	84.4	0.9%	8	213
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	84.3	0.6%	8	210
Roche Hitachi/Hitachi Systems Wako \ ISE (indirect)	mmol/L	82.2	1%	6	113
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	81.7	1.7%	5	113
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	85.1	1.3%	28	694
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	78.7	12%	<5	<100
Beckman AU series Beckman Coulter \ Coulometric titration	mmol/L	84.7	0%	<5	<100
Beckman AU series Genrui \ Coulometric titration	mmol/L	81	0%	<5	<100
Mindray BS series Mindray \ Coulometric titration	mmol/L	80	0%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ Coulometric titration	mmol/L	78	0%	<5	<100
(CO2 (HCO3-))二氧化碳 (碳酸氢根)					
Beckman AU series Whitman \ OTHER	mmol/L	10.1	4.8%	8	211
Roche Hitachi/Hitachi Systems MAKER \ OTHER	mmol/L	11.4	7.2%	12	455
Roche cobas 4000/6000/8000 Roche \ PEPC	mmol/L	14.1	12%	<5	<100
(Cr)肌酐					
Beckman AU series Fosun \ OTHER	μmol/L	399	1.8%	8	211
Beckman AU series LEADMAN \ Enzymatic	μmol/L	421	1.8%	7	164
Roche Hitachi/Hitachi Systems Wako \ Enzymatic	μmol/L	401	2.4%	6	114
Beckman AU series MedicalSystem \ Enzymatic	μmol/L	356	1.8%	<5	<100
Beckman AU series Beckman Coulter \ dry chemistry	μmol/L	414	4.1%	<5	<100
Mindray BS series Mindray \ dry chemistry	μmol/L	395	2.3%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ dry chemistry	μmol/L	418	10%	<5	<100
Beckman AU series Fosun \ Enzymatic	μmol/L	404	2.4%	15	389
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	431	3.8%	14	348
KHB ZY series KHB \ Enzymatic	μmol/L	410	2.6%	11	115
Roche cobas 4000/6000/8000 Roche \ Enzymatic	μmol/L	398	12%	<5	<100
Beckman AU series Genrui \ dry chemistry	μmol/L	384	0%	<5	<100
DiruiCS series DIRUI \ dry chemistry	μmol/L	410	0%	<5	<100
(DBIL)直接胆红素					
Beckman AU series LEADMAN \ Vanadate	μmol/L	47.2	2.9%	7	167
Roche Hitachi/Hitachi Systems Wako \ Vanadate	μmol/L	44.7	0.9%	6	113
Beckman AU series MedicalSystem \ Diazotization	μmol/L	41.6	2.1%	<5	<100
Beckman AU series Fosun \ Vanadate	μmol/L	45.2	1.9%	15	371
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	49.2	1.7%	14	329
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	39.6	2.7%	12	293
KHB ZY series KHB \ Vanadate	μmol/L	45.4	4.9%	11	113
Roche cobas 4000/6000/8000 Roche \ Diazotization -Blank Compensation	μmol/L	49.6	12%	<5	<100

REF	580-2	6×5mL	EXP	2025-02-19	LOT	375593A	更新日期：	2024-05
4 / 7								

		批号：375593A			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Fe)铁					
Roche Hitachi/Hitachi Systems OTHER \ OTHER	μmol/L	13.3	1.7%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	13.2	12%	<5	<100
(GGT)L-γ-谷氨酰基转移酶					
Beckman AU series Fosun \ OTHER	U/L	215	1.3%	8	211
Beckman AU series LEADMAN \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	199	5%	7	167
Roche Hitachi/Hitachi Systems MAKER \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	203	0.7%	6	114
Beckman AU series MedicalSystem \ γ-glutamyl -4-nitroaniline	U/L	209	3.4%	<5	<100
Beckman AU series Beckman Coulter \ dry chemistry	U/L	217	7.2%	<5	<100
Mindray BS series Mindray \ dry chemistry	U/L	206	3.1%	<5	<100
Beckman AU series Fosun \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	210	3.1%	14	356
Beckman AU series Beckman Coulter \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	219	1.3%	14	333
KHB ZY series KHB \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	206	2.2%	11	116
Roche cobas 4000/6000/8000 Roche \ γ-glutamyl -4-nitroaniline	U/L	219	12%	<5	<100
Beckman AU series Genrui \ dry chemistry	U/L	209	0%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ dry chemistry	U/L	182	0%	<5	<100
DiruiCS series DIRUI \ dry chemistry	U/L	215	0%	<5	<100
(GLU)葡萄糖					
Beckman AU series LEADMAN \ Glucose oxidase (GOD)	mmol/L	16.5	2.1%	7	167
Roche Hitachi/Hitachi Systems Wako \ Hexokinase	mmol/L	16.8	0.9%	6	114
Beckman AU series Fosun \ Hexokinase	mmol/L	17.1	1.7%	22	605
Roche Hitachi/Hitachi Systems MAKER \ Hexokinase	mmol/L	16.5	4.1%	<5	<100
Beckman AU series Beckman Coulter \ Dehydrogenase	mmol/L	17.5	4.3%	<5	<100
Mindray BS series Mindray \ Dehydrogenase	mmol/L	16.7	1.8%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ Dehydrogenase	mmol/L	17.1	3.1%	<5	<100
Beckman AU series Beckman Coulter \ Hexokinase	mmol/L	16.9	1.6%	14	328
KHB ZY series KHB \ Glucose oxidase (GOD)	mmol/L	15.7	3.7%	11	115
Roche cobas 4000/6000/8000 Roche \ Hexokinase	mmol/L	16.9	12%	<5	<100
Beckman AU series Genrui \ Dehydrogenase	mmol/L	16.6	0%	<5	<100
DiruiCS series DIRUI \ Dehydrogenase	mmol/L	17.3	0%	<5	<100
(HDL-C)高密度脂蛋白胆固醇					
Beckman AU series Fosun \ OTHER	mmol/L	0.677	2.8%	8	211
Beckman AU series LEADMAN \ OTHER	mmol/L	0.687	2.8%	7	155
Roche Hitachi/Hitachi Systems Wako \ selected suppression	mmol/L	0.683	1.2%	6	114
Beckman AU series Fosun \ Direct Measure	mmol/L	0.614	13.1%	11	239
KHB ZY series KHB \ Direct Measure	mmol/L	0.62	9.1%	11	113
Roche cobas 4000/6000/8000 Roche \ Enzymatic	mmol/L	0.7	12%	<5	<100
(K)钾					
Beckman AU series Fosun \ ISE (direct)	mmol/L	5.29	1%	8	212
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	5.31	0.7%	8	210
Roche Hitachi/Hitachi Systems Wako \ ISE (indirect)	mmol/L	5.42	0.5%	6	114
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	5.35	1.4%	5	139
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	5.3	2.2%	<5	112
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	5.35	1.2%	28	701
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	5.57	12%	<5	<100
Abbott ARCHITECT c series Abbott \ ISE (direct)	mmol/L	6.06	0.4%	<5	<100
Beckman AU series Beckman Coulter \ dry chemistry	mmol/L	5.39	0%	<5	<100

REF	580-2	6×5mL	EXP	2025-02-19	LOT	375593A	更新日期：	2024-05
5 / 7								

		批号：375593A			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Genrui \ dry chemistry	mmol/L	5.09	0%	<5	<100
Mindray BS series Mindray \ dry chemistry	mmol/L	5.4	0%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ dry chemistry	mmol/L	5.54	0%	<5	<100
(LA)乳酸					
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	mmol/L	1.83	12%	<5	<100
(LDH)乳酸脱氢酶					
Beckman AU series Fosun \ OTHER	U/L	310	2%	8	211
Beckman AU series LEADMAN \ Lactate to pyruvate (L-P)	U/L	307	2.3%	7	167
Roche Hitachi/Hitachi Systems MAKER \ Lactate to pyruvate (L-P)	U/L	297	2.7%	6	112
Beckman AU series Fosun \ Lactate to pyruvate (L-P)	U/L	317	3%	15	407
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	U/L	317	3.2%	14	339
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	U/L	330	12%	<5	<100
(LDH-I)乳酸脱氢酶同工酶1					
Roche Hitachi/Hitachi Systems OTHER \ OTHER	U/L	195	1.7%	<5	<100
(LDL-C)低密度脂蛋白胆固醇					
Beckman AU series Fosun \ OTHER	mmol/L	1.46	3.6%	8	211
Beckman AU series Beckman Coulter \ OTHER	mmol/L	1.6	2.4%	7	156
Roche Hitachi/Hitachi Systems Wako \ selected suppression	mmol/L	1.67	1.7%	6	114
Beckman AU series MAKER \ OTHER	mmol/L	1.63	4.7%	5	209
KHB ZY series KHB \ Direct Measure	mmol/L	1.11	6.4%	11	116
Roche cobas 4000/6000/8000 Roche \ Enzymatic	mmol/L	1.34	12%	<5	<100
(LIP)脂肪酶					
Roche Hitachi/Hitachi Systems OTHER \ OTHER	U/L	272	11.9%	12	271
Roche cobas 4000/6000/8000 DiaSys \ Enzymatic Cycling Assays Method	U/L	200	12%	<5	<100
(Lp(a))脂蛋白a					
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	mg/L	105	11.1%	7	177
Beckman AU series Fosun \ Immunoturbidimetry (ITA)	mg/L	73.7	6.3%	6	168
(Mg)镁					
Beckman AU series Fosun \ Xylidyl Blue	mmol/L	1.73	2%	14	338
Roche Hitachi/Hitachi Systems MAKER \ Xylidyl Blue	mmol/L	1.88	3.9%	12	390
Roche cobas 4000/6000/8000 Roche \ Xylidyl Blue	mmol/L	1.82	12%	<5	<100
(Na)钠					
Beckman AU series Fosun \ ISE (direct)	mmol/L	111	1.1%	8	222
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	110	0.7%	8	210
Roche Hitachi/Hitachi Systems Wako \ ISE (indirect)	mmol/L	114	0.8%	6	112
Roche Hitachi/Hitachi Systems OTHER \ OTHER	mmol/L	112	1.1%	5	<100
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	112	1.2%	24	573
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	113	12%	<5	<100
Beckman AU series Beckman Coulter \ dry chemistry	mmol/L	110	0%	<5	<100
Beckman AU series Genrui \ dry chemistry	mmol/L	106	0%	<5	<100
Mindray BS series Mindray \ dry chemistry	mmol/L	114	0%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ dry chemistry	mmol/L	114	0%	<5	<100
(P)磷					
Beckman AU series Fosun \ OTHER	mmol/L	2.63	1.9%	8	212
Beckman AU series Fosun \ Phosphomolybdate UV-rate	mmol/L	2.63	1.8%	7	173
Roche Hitachi/Hitachi Systems Wako \ Phosphomolybdate-UV-end point	mmol/L	2.92	1%	6	114
Beckman AU series Beckman Coulter \ Phosphomolybdate-UV-end point	mmol/L	2.64	1.6%	21	483

REF	580-2	6×5mL	EXP	2025-02-19	LOT	375593A	更新日期：	2024-05
6 / 7								

		批号：375593A			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000 Roche \ Phosphomolybdate UV-rate	mmol/L	2.69	12%	<5	<100
(PAB)前白蛋白					
Beckman AU series Fosun \ Immunoturbidimetry (ITA)	g/L	0.156	4.3%	8	211
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	g/L	0.169	8.4%	12	458
(PRO/ TP)(总) 蛋白					
Beckman AU series LEADMAN \ Biuret-endpoint	g/L	45.1	1.6%	7	167
Roche Hitachi/Hitachi Systems Wako \ Biuret-endpoint	g/L	47.2	0.9%	6	112
Beckman AU series Fosun \ Biuret-endpoint	g/L	45.7	1.8%	23	628
Beckman AU series MedicalSystem \ Biuret-endpoint	g/L	40.8	3.4%	<5	<100
Roche Hitachi/Hitachi Systems MAKER \ Biuret-endpoint	g/L	48.4	1%	<5	<100
Mindray BS series Mindray \ Biuret-endpoint	g/L	45.7	5.6%	<5	<100
Beckman AU series Beckman Coulter \ Biuret-endpoint	g/L	45.2	1.7%	16	353
KHB ZY series KHB \ Biuret-endpoint	g/L	45.3	5.1%	11	114
Roche cobas 4000/6000/8000 Roche \ Biuret-endpoint	g/L	46.4	12%	<5	<100
Beckman AU series Genrui \ Biuret-endpoint	g/L	41.9	0%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ Biuret-endpoint	g/L	48.2	0%	<5	<100
DiruiCS series DIRUI \ Biuret-endpoint	g/L	43.4	0%	<5	<100
(TBA)总胆汁酸					
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic Cycling Assays Method	μmol/L	118	4.5%	8	219
Beckman AU series Fosun \ Enzymatic Cycling Assays Method	μmol/L	111	5.6%	8	209
Beckman AU series KINGSBIO \ Enzymatic Cycling Assays Method	μmol/L	92.4	2.9%	<5	<100
Roche cobas 4000/6000/8000 DiaSys \ Enzymatic Cycling Assays Method	μmol/L	89.6	12.1%	<5	<100
(TBIL)总胆红素					
Beckman AU series LEADMAN \ Vanadate	μmol/L	87.9	2.4%	7	166
Roche Hitachi/Hitachi Systems MAKER \ OTHER	μmol/L	85.8	2.6%	7	160
Roche Hitachi/Hitachi Systems Wako \ Vanadate	μmol/L	84.9	0.6%	6	114
Beckman AU series Fosun \ Vanadate	μmol/L	83.5	2.9%	22	575
Beckman AU series MedicalSystem \ Diazotization	μmol/L	71.5	1.1%	<5	<100
Mindray BS series Mindray \ Diazotization	μmol/L	77.7	5.9%	<5	<100
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	88.5	1.5%	16	334
KHB ZY series KHB \ Vanadate	μmol/L	73.6	6.4%	11	113
Roche cobas 4000/6000/8000 DiaSys \ Diazotization	μmol/L	76.5	12%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ Diazotization	μmol/L	81.1	0%	<5	<100
DiruiCS series DIRUI \ Diazotization	μmol/L	86	0%	<5	<100
(TG)甘油三酯					
Beckman AU series Fosun \ OTHER	mmol/L	0.935	5.3%	8	213
Beckman AU series LEADMAN \ GPO-POD	mmol/L	1.11	1.9%	7	167
Roche Hitachi/Hitachi Systems Wako \ Free glycerol removed	mmol/L	1.12	1.2%	6	114
Beckman AU series MedicalSystem \ Acetyl - acetone photometry	mmol/L	0.887	6.4%	<5	<100
Mindray BS series Mindray \ Acetyl - acetone photometry	mmol/L	0.98	4.3%	<5	<100
Beckman AU series Beckman Coulter \ Acetyl - acetone photometry	mmol/L	1.11	8.9%	<5	<100
Beckman AU series Fosun \ GPO-POD	mmol/L	0.973	9.7%	14	332
KHB ZY series KHB \ GPO-POD	mmol/L	1.14	4.3%	11	109
Roche cobas 4000/6000/8000 Roche \ Free glycerol removed	mmol/L	1.16	12%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ Acetyl - acetone photometry	mmol/L	0.95	0%	<5	<100
DiruiCS series DIRUI \ Acetyl - acetone photometry	mmol/L	0.89	0%	<5	<100
Beckman AU series Genrui \ Acetyl - acetone photometry	mmol/L	1.18	0%	<5	<100

REF	580-2	6×5mL	EXP	2025-02-19	LOT	375593A	更新日期：	2024-05
7 / 7								

		批号：375593A			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
KHB ZY series YTZY \ Acetyl - acetone photometry	mmol/L	1.41	0%	<5	<100
(UA)尿酸					
Beckman AU series Fosun \ OTHER	μmol/L	513	1.2%	8	211
Beckman AU series Fosun \ Uricase-UV	μmol/L	528	1.9%	8	197
Beckman AU series Fosun \ Uricase colorimetry	μmol/L	513	1.3%	7	202
Beckman AU series LEADMAN \ Uricase-UV	μmol/L	582	1.9%	7	167
Roche Hitachi/Hitachi Systems Wako \ Uricase colorimetry	μmol/L	538	1.1%	6	114
Beckman AU series MedicalSystem \ Phosphotungstate colorimetry	μmol/L	472	1.7%	<5	<100
Beckman AU series Beckman Coulter \ dry chemistry	μmol/L	562	6.5%	<5	<100
Mindray BS series Mindray \ dry chemistry	μmol/L	598	2.4%	<5	<100
Beckman AU series Beckman Coulter \ Uricase-UV	μmol/L	580	1.1%	14	344
KHB ZY series KHB \ Uricase colorimetry	μmol/L	577	3.3%	11	113
Roche cobas 4000/6000/8000 Roche \ Uricase colorimetry	μmol/L	556	12%	<5	<100
Beckman AU series Genrui \ dry chemistry	μmol/L	540	0%	<5	<100
KHB ZY series YTZY \ dry chemistry	μmol/L	686	0%	<5	<100
Roche Hitachi/Hitachi Systems Hitachi \ dry chemistry	μmol/L	579	0%	<5	<100
DiruiCS series DIRUI \ dry chemistry	μmol/L	576	0%	<5	<100
(UIBC)不饱和铁结合力					
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	31.4	12%	<5	<100
(Urea)尿素					
Beckman AU series LEADMAN \ Urease-UV-rate	mmol/L	21	2.5%	7	166
Roche Hitachi/Hitachi Systems Wako \ Urease-UV-rate	mmol/L	21.6	2.8%	6	114
Beckman AU series Fosun \ Urease-UV-rate	mmol/L	21.3	1.9%	15	411
Beckman AU series Beckman Coulter \ Urease-UV-rate	mmol/L	21.7	2.2%	14	343
Roche cobas 4000/6000/8000 Roche \ Urease colorimetry	mmol/L	21	12%	<5	<100
(β2M)β-2-微球蛋白					
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3400	12.1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。