

质控品名： 特殊肿瘤标志物液体非定值质控品

REF	513-1	6×3mL	EXP	2025-02-04	LOT	30EF582	更新日期：	2024-05
1 / 3								

		批号：30EF582			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.24	5.1%	9	205
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.26	4%	28	593
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.7	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.55	10%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.11	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.5	10%	<5	<100
(CA 125)糖类抗原 125					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	21.5	3.1%	8	184
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	33.3	5.9%	29	621
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	27.8	10%	<5	<100
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.6	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	23.8	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	23.2	10%	<5	<100
(CA 15-3)糖类抗原 15-3					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	9.77	5.8%	9	201
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	16.3	9.4%	29	628
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.5	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.4	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.8	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	14.9	10%	<5	<100
(CA 19-9)糖类抗原 19-9					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.8	5.2%	9	192
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	35.9	10.1%	29	616
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.5	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	13.3	10%	<5	<100
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	25.2	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	26.3	10%	<5	<100
(CA 50)糖类抗原 50					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	8.47	7.7%	27	564
(CA 72-4)糖类抗原 72-4					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	1.78	3.8%	8	127
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	1.7	3.8%	5	<100
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	1.84	29.9%	27	555
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	3.1	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	2.69	10%	<5	<100
(CA242)糖类抗原 242					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	6.51	7.3%	27	563
Manual GLD \ Enzyme linked immunosorbent assays (ELISA)	U/mL	12.3	10%	<5	<100
(CEA)癌胚抗原					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.7	6.1%	9	206
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.82	6.7%	29	638
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.9	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.9	10%	<5	<100

质控品名： 特殊肿瘤标志物液体非定值质控品

REF	513-1	6×3mL	EXP	2025-02-04	LOT	30EF582	更新日期：	2024-05
2 / 3								

		批号：30EF582			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.5	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.82	10%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.45	3.4%	9	191
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.91	9.3%	9	139
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.23	9.5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.699	23.3%	19	422
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.01	10%	<5	<100
(Ferritin)铁蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	65.4	4.7%	29	620
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	ng/mL	69.1	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	62.9	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	52	10%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	47.5	10%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.21	3.9%	9	204
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.736	6.1%	29	635
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.99	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.03	10%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.56	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.03	10%	<5	<100
(hCG)人绒毛膜促性腺激素					
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	13.2	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Enzyme linked immunosorbent assays (ELISA)	IU/L	10.8	10%	<5	<100
(HE4)人附睾蛋白4					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	26.7	2.4%	7	128
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	22.9	6.4%	<5	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	23.7	7.8%	25	551
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	32.8	10%	<5	<100
(PSA)前列腺特异性抗原					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.7	5.9%	9	201
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.56	13.5%	29	620
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.74	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.38	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.6	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.24	10%	<5	<100
(SCC)鳞状细胞癌抗原					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.07	9.4%	9	193
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.901	10.5%	9	138
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.51	8.2%	19	482
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.22	10%	<5	<100
(TPA)组织多肽抗原					
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	U/L	380	10%	<5	<100

REF	513-1	6×3mL	EXP	2025-02-04	LOT	30EF582	更新日期：	2024-05
3 / 3								

		批号：30EF582			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(TPS)组织多肽特异抗原					
Manual OTHER \ Enzyme linked immunosorbent assays (ELISA)	U/L	16	10%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.012	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。