

REF	671-3	6×1mL	EXP	2025-12-27	LOT	4005B92	更新日期：	2024-04
1 / 1								

		批号：4005B92			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Sysmex CA 1500/6000/7000	Seconds	149	10.1%	<5	<100
Siemens Dade Actin(APTT) \ Light scattering					
Werfen ACL series	Seconds	105	10.5%	<5	<100
Werfen Hemosil SynthAsil APTT \ Light scattering					
Werfen ACL series	Seconds	97.8	10%	<5	<100
Werfen Hemosil APTT-SP \ Light scattering					
Stago STA series	Seconds	88.4	10%	<5	<100
Stago \ Magnetic Beads					
Sysmex CA 1500/6000/7000	Seconds	154	10%	<5	<100
Siemens Pathrombin SL(APTT) \ Light scattering					
(AT-III)抗凝血酶III					
Werfen ACL series	%	26	10%	<5	<100
Werfen Hemosil Liquid Antithrombin (AT□) \ Nephelometry					
Sysmex CA 1500/6000/7000	%	28	10.7%	<5	<100
Siemens Berichrom Antithrombin III (AT□) \ Nephelometry					
(D-D(DDU)) D-二聚体(DDU)					
Werfen ACL series	mg/L	1.11	10%	<5	<100
Wenfen Hemosil D-Dimer \ Nephelometry					
Werfen ACL series	mg/L	1.16	10%	<5	<100
Werfen Hemosil D-Dimer HS \ Nephelometry					
(D-D(FEU)) D-二聚体(FEU)					
Sysmex CA 1500/6000/7000	mg/L	3.82	10%	<5	<100
Siemens Innovance Antithrombin (AT□) \ Nephelometry					
(FIB)纤维蛋白原					
Stago STA series	g/L	1.41	10%	<5	<100
Stago \ Magnetic Beads					
Werfen ACL series	g/L	1.27	10.2%	<5	<100
Werfen Hemosil QFA Thrombin (FIB) \ Light scattering					
Werfen ACL series	g/L	1.15	10%	<5	<100
Werfen Hemosil Fib-C \ Light scattering					
Sysmex CA 1500/6000/7000	g/L	1.82	10%	<5	<100
Siemens Multifibren U (FIB) \ Light scattering					
(PT)血浆凝血酶原时间					
Werfen ACL series	Seconds	54.2	10%	<5	<100
Werfen Hemosil PT Recombiplastin 2G \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	44.5	10.1%	<5	<100
Siemens PT Innovin \ Light scattering					
Sysmex CA 1500/6000/7000	Seconds	51.5	10.1%	<5	<100
Siemens Thromborel S (PT) \ Light scattering					
Werfen ACL series	INR	4.61	10%	<5	<100
Werfen Hemosil PT Recombiplastin 2G \ Light scattering					
Stago STA series	Seconds	54	10%	<5	<100
Stago \ Magnetic Beads					
Werfen ACL series	INR	5.3	10.4%	<5	<100
Werfen Hemosil PT-Fib HS Plus \ Light scattering					
Werfen ACL series	Seconds	65.2	10%	<5	<100
Werfen Hemosil PT-Fib HS Plus \ Light scattering					
Sysmex CA 1500/6000/7000	INR	4.73	10.6%	<5	<100
Siemens Thromborel S (PT) \ Light scattering					
Werfen ACL series	INR	4.56	10.3%	<5	<100
Werfen Hemosil PT Recombiplastin \ Light scattering					
Werfen ACL series	Seconds	49.8	10%	<5	<100
Werfen Hemosil PT Recombiplastin \ Light scattering					
Sysmex CA 1500/6000/7000	INR	4.2	9.5%	<5	<100
Siemens PT Innovin \ Light scattering					
Stago STA series	INR	6.03	10%	<5	<100
Stago \ Magnetic Beads					

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。