



REF	667-2	6×4.5mL	EXP	2023-05-21	LOT	3FA19A4	更新日期：	2024-04
1 / 1								

		批号：3FA19A4			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(ESR)红细胞沉降率(血沉)					
OTHER OTHER \ Westergren	mm/hr	43.3	22.1%	<5	<100
OTHER Dedicated \ Westergren	mm/hr	74	25.2%	<5	110
Diesse Ves-Matic series Dedicated \ Westergren	mm/hr	70.3	11.2%	<5	<100
zonci zc series Dedicated \ Westergren	mm/hr	60.4	13.1%	<5	<100
Nanion Cardio Excyte series Dedicated \ Westergren	mm/hr	68.5	11.8%	<5	<100
Manual Dedicated \ Westergren	mm/hr	62.3	13.2%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。