

REF	581-1	6×5mL	EXP	2024-09-22	LOT	376B86D	更新日期：	2024-05
1 / 6								

		批号：376B86D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(a-HBDH)α-羟丁酸脱氢酶					
Abbott ARCHITECT ci series BIOSINO \ Enzymatic Cycling Assays Method	U/L	124	8%	12	254
(Alb)白蛋白					
KHB SA-6000 series KHB \ Bromocresol green	g/L	27.2	2.9%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Bromocresol green	g/L	43.7	1.9%	12	225
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	43.7	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Bromocresol green	g/L	45.2	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Bromocresol green	g/L	42.2	1.6%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Bromocresol green	g/L	36.5	15%	<5	<100
(ALP)碱性磷酸酶					
KHB SA-6000 series KHB \ Rate -AMP buffer	U/L	37	3.8%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Rate -AMP buffer	U/L	109	6.2%	12	226
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	81.6	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Rate -AMP buffer	U/L	90.7	2.1%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Rate -AMP buffer	U/L	89.5	15%	<5	<100
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	92.5	15%	<5	<100
(ALT/ SGPT)丙氨酸氨基转移酶					
KHB SA-6000 series KHB \ OTHER	U/L	34	4.2%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ UV-rate with P5P	U/L	38.9	5.7%	12	256
Beckman AU series Beckman Coulter \ UV-rate with P5P	U/L	43	15%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ UV-rate without P5P	U/L	42	15%	<5	<100
Roche cobas 4000/6000/8000 Roche \ UV-rate with P5P	U/L	41.1	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ UV-rate without P5P	U/L	54.6	3.5%	<5	<100
(AMY)淀粉酶					
Abbott ARCHITECT ci series BIOSINO \ PNP-G7	U/L	100	3.4%	12	255
Roche cobas 4000/6000/8000 Roche \ PNP-G7	U/L	99	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ PNP-G7	U/L	100	15%	<5	<100
Beckman AU series Beckman Coulter \ Iodine colorimetry	U/L	101	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ CNP_G3	U/L	103	1.1%	<5	<100
(Apo-A I)载脂蛋白A I					
Abbott ARCHITECT ci series BIOSINO \ immunoturbidimetric	g/L	1.78	5.7%	7	107
(Apo-B)载脂蛋白B					
Abbott ARCHITECT ci series BIOSINO \ immunoturbidimetric	g/L	0.774	8%	7	137
(ASO)抗链球菌溶血素 O 抗体					
OTHER OTHER \ OTHER	IU/mL	74.6	0.3%	<5	<100
(AST/ SGOT)天门冬氨酸氨基转移酶					
KHB SA-6000 series KHB \ OTHER	U/L	39	3.6%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ UV-rate with P5P	U/L	37.3	7.7%	12	256
Roche cobas 4000/6000/8000 Roche \ UV-rate with P5P	U/L	38.7	15%	<5	<100
Beckman AU series Beckman Coulter \ UV-rate without P5P	U/L	44	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ UV-rate without P5P	U/L	40.5	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ UV-rate without P5P	U/L	41.3	3.1%	<5	<100
(C3)补体 C3					
Abbott ARCHITECT ci series BIOSINO \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	138	3.2%	5	104
OTHER OTHER \ Immunonephelometry (INA)	mg/dL	125	3.4%	<5	<100

REF	581-1	6×5mL	EXP	2024-09-22	LOT	376B86D	更新日期：	2024-05
2 / 6								

批号：376B86D					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(C4)补体 C4					
Abbott ARCHITECT ci series BIOSINO \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	30	0.3%	5	105
OTHER OTHER \ OTHER	mg/dL	29.5	12%	<5	<100
(Ca)钙					
HORRON-H900 HORRON \ ISE (direct)	mmol/L	2.02	2.4%	5	<100
Abbott ARCHITECT ci series BIOSINO \ Arsenazo III	mmol/L	2.06	1.9%	12	255
Beckman AU series Beckman Coulter \ BAPTA	mmol/L	2.05	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Arsenazo III	mmol/L	2.11	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Arsenazo III	mmol/L	2.13	2.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ BAPTA	mmol/L	2.08	15%	<5	<100
(CHE)胆碱酯酶					
Abbott ARCHITECT ci series BIOSINO \ DTNB	U/L	7133	2%	5	119
(CHOL/ TC)总胆固醇					
KHB SA-6000 series KHB \ Cholesterol oxidase	mmol/L	2.31	2.4%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Cholesterol oxidase	mmol/L	4.19	1.6%	12	256
Roche cobas 4000/6000/8000 Roche \ Cholesterol oxidase	mmol/L	4.07	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Cholesterol oxidase	mmol/L	3.32	3.7%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Cholesterol oxidase	mmol/L	3.99	15%	<5	<100
Beckman AU series Beckman Coulter \ Cholesterol oxidase	mmol/L	3.99	15%	<5	<100
(CK)肌酸激酶					
Abbott ARCHITECT ci series BIOSINO \ Urease-UV-rate	U/L	73.1	11.5%	12	252
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	82	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Rate -AMP buffer	U/L	83.2	15%	<5	<100
Beckman AU series Beckman Coulter \ Urease-UV-rate	U/L	89	15%	<5	<100
(CK-MB)肌酸激酶-MB (活性)					
Abbott ARCHITECT ci series BIOSINO \ Immune suppression	U/L	12.4	11.6%	12	249
Roche cobas 4000/6000/8000 Roche \ Immune suppression	U/L	12.7	15%	<5	<100
Beckman AU series Beckman Coulter \ Immune suppression	U/L	17.6	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ DGKC	U/L	14.7	15%	<5	<100
(Cl)氯					
HORRON-H900 HORRON \ ISE (direct)	mmol/L	80	3.3%	5	<100
Abbott ARCHITECT ci series BIOSINO \ ISE (indirect)	mmol/L	119	1%	12	254
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	111	15%	<5	<100
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	113	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ ISE (indirect)	mmol/L	116	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ ISE (direct)	mmol/L	113	1.1%	<5	<100
Sysmex CA 50 sysmex \ Coulometric titration	mmol/L	70	0%	<5	<100
(Cr)肌酐					
KHB SA-6000 series KHB \ Enzymatic	μmol/L	73.5	2.9%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Picric acid-Blank Compensation	μmol/L	78.5	3.2%	12	256
Roche cobas 4000/6000/8000 Roche \ Enzymatic	μmol/L	79.9	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Enzymatic	μmol/L	75.4	1.7%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Enzymatic	μmol/L	78.9	15%	<5	<100
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	79.5	15%	<5	<100
(CRP)C反应蛋白					
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	0.67	15%	<5	<100

REF	581-1	6×5mL	EXP	2024-09-22	LOT	376B86D	更新日期：	2024-05
3 / 6								

批号：376B86D					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Cu)铜					
Beckman AU series Beckman Coulter \ ICP-MS	μmol/L	19.1	15%	<5	<100
(DBIL)直接胆红素					
KHB SA-6000 series KHB \ Vanadate	μmol/L	6.1	2.3%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Vanadate	μmol/L	5.5	6.7%	12	255
Siemens ADVIA 1800/2400 Siemens \ Vanadate	μmol/L	6.02	15%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Diazotization -Blank Compensation	μmol/L	5.8	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Diazotization	μmol/L	4.23	3.8%	<5	<100
Beckman AU series Beckman Coulter \ Diazotization -Blank Compensation	μmol/L	5.45	15%	<5	<100
(DGXN)地高辛					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.684	15%	<5	<100
(Fe)铁					
Abbott ARCHITECT ci series BIOSINO \ Atomic absorption spectroscopy (AAS)	μmol/L	18.9	2.4%	5	124
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	19.5	15%	<5	<100
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	20.9	15%	<5	<100
(GGT)L-γ-谷氨酰基转移酶					
KHB SA-6000 series KHB \ OTHER	U/L	31.5	6.7%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.8	5%	12	225
Siemens ADVIA 1800/2400 Siemens \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.3	15%	<5	<100
Roche cobas 4000/6000/8000 Roche \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	36.7	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.4	2.1%	<5	<100
Beckman AU series Beckman Coulter \ γ-glutamyl -4-nitroaniline	U/L	38.8	15%	<5	<100
(GLDH)谷氨酸脱氢酶					
Beckman AU series Beckman Coulter \ DGKC	U/L	6.2	15%	<5	<100
Roche cobas 4000/6000/8000 DiaSys \ Colorimetry	U/L	6.14	15%	<5	<100
(GLU)葡萄糖					
KHB SA-6000 series KHB \ Glucose oxidase (GOD)	mmol/L	3.73	1.1%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Hexokinase	mmol/L	4.18	3.8%	12	257
Roche cobas 4000/6000/8000 Roche \ Hexokinase	mmol/L	4.46	15%	<5	<100
Beckman AU series Beckman Coulter \ Dehydrogenase	mmol/L	4.41	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Dehydrogenase	mmol/L	4.41	1.3%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Hexokinase	mmol/L	4.32	15%	<5	<100
OTHER Uritest \ Dehydrogenase	mmol/L	4.28	0%	<5	<100
(HDL-C)高密度脂蛋白胆固醇					
KHB SA-6000 series KHB \ Direct Measure	mmol/L	0.905	14.8%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ selected suppression	mmol/L	1.42	7.5%	11	248
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	1.24	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Direct Measure	mmol/L	1.21	15%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Enzymatic	mmol/L	1.18	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Direct Measure	mmol/L	0.845	2.2%	<5	<100
(hs-CRP)超敏C反应蛋白					
OTHER OTHER \ Immunonephelometry (INA)	mg/L	9.45	1.3%	<5	<100
(IgA)免疫球蛋白 IgA					
OTHER OTHER \ Immunonephelometry (INA)	mg/dL	203	0.7%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Immunonephelometry (INA)	mg/dL	199	0%	<5	<100

REF	581-1	6×5mL	EXP	2024-09-22	LOT	376B86D	更新日期：	2024-05
4 / 6								

批号：376B86D					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(IgG)免疫球蛋白 IgG					
Abbott ARCHITECT ci series BIOSINO \ Immunonephelometry (INA)	mg/dL	919	6.2%	5	105
OTHER OTHER \ Immunonephelometry (INA)	mg/dL	1055	2%	<5	<100
(IgM)免疫球蛋白 IgM					
OTHER OTHER \ Immunonephelometry (INA)	mg/dL	104	2%	<5	<100
(K)钾					
HORRON-H900 HORRON \ ISE (direct)	mmol/L	2.51	7.3%	5	<100
Abbott ARCHITECT ci series BIOSINO \ ISE (indirect)	mmol/L	3.8	1.6%	12	255
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	3.72	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ ISE (indirect)	mmol/L	3.9	15%	<5	<100
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	3.73	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ ISE (direct)	mmol/L	3.81	1.7%	<5	<100
(LA)乳酸					
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	mmol/L	1.98	15%	<5	<100
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	2.05	15%	<5	<100
(LDH)乳酸脱氢酶					
Abbott ARCHITECT ci series BIOSINO \ Lactate to pyruvate (L-P)	U/L	128	5.8%	12	256
Siemens ADVIA 1800/2400 Siemens \ Lactate to pyruvate (L-P)	U/L	136	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Lactate to pyruvate (L-P)	U/L	141	1.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	U/L	149	15%	<5	<100
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	U/L	147	15%	<5	<100
(LDL-C)低密度脂蛋白胆固醇					
KHB SA-6000 series KHB \ Direct Measure	mmol/L	1.22	11.6%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ selected suppression	mmol/L	1.93	3.1%	11	248
Roche cobas 4000/6000/8000 Roche \ Enzymatic	mmol/L	2.45	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Direct Measure	mmol/L	2.19	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ SUR	mmol/L	2.19	2.6%	<5	<100
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	2.5	15%	<5	<100
(Li)锂					
Beckman AU series Beckman Coulter \ Colorimetry	mmol/L	0.79	15%	<5	<100
(LIP)脂肪酶					
Abbott ARCHITECT ci series BIOSINO \ Glucose oxidase (GOD)	U/L	42.7	17.7%	<5	<100
Beckman AU series Beckman Coulter \ Glucose oxidase (GOD)	U/L	48.5	15%	<5	<100
Roche cobas 4000/6000/8000 DiaSys \ Enzymatic Cycling Assays Method	U/L	44.2	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Enzymatic Cycling Assays Method	U/L	50.9	15%	<5	<100
(Mg)镁					
Abbott ARCHITECT ci series BIOSINO \ Calmagite colorimetry	mmol/L	0.897	3.5%	11	247
Roche cobas 4000/6000/8000 Roche \ Xylidyl Blue	mmol/L	0.94	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Xylidyl Blue	mmol/L	0.892	15%	<5	<100
Beckman AU series Beckman Coulter \ Xylidyl Blue	mmol/L	0.88	15%	<5	<100
(Na)钠					
HORRON-H900 HORRON \ ISE (direct)	mmol/L	116	3.1%	5	<100
Abbott ARCHITECT ci series BIOSINO \ ISE (indirect)	mmol/L	137	1.2%	12	255
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	139	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ ISE (indirect)	mmol/L	141	15%	<5	<100
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	141	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ ISE (direct)	mmol/L	136	0.5%	<5	<100

REF	581-1	6×5mL	EXP	2024-09-22	LOT	376B86D	更新日期：	2024-05
5 / 6								

批号：376B86D					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(P)磷					
Abbott ARCHITECT ci series BIOSINO \ Phosphomolybdate UV-rate	mmol/L	1.05	2.8%	12	256
Beckman AU series Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	1.01	15%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Phosphomolybdate UV-rate	mmol/L	0.99	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Phosphomolybdate UV-rate	mmol/L	1.08	3.1%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Phosphomolybdate UV-rate	mmol/L	1.03	15%	<5	<100
(PAMY)胰淀粉酶					
Beckman AU series Beckman Coulter \ Enzymatic	U/L	49	15%	<5	<100
(PRO/ TP)(总) 蛋白					
KHB SA-6000 series KHB \ OTHER	g/L	42.2	1.2%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Biuret-rate	g/L	68	2.6%	12	256
Siemens ADVIA 1800/2400 Siemens \ Biuret-endpoint	g/L	69.9	15%	<5	<100
Beckman AU series Beckman Coulter \ Biuret-endpoint	g/L	68.3	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Biuret-endpoint	g/L	65.2	1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Biuret-endpoint	g/L	70.4	15%	<5	<100
(TBA)总胆汁酸					
Abbott ARCHITECT ci series BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	14.4	9%	12	227
Roche cobas 4000/6000/8000 DiaSys \ Enzymatic Cycling Assays Method	μmol/L	14.6	15%	<5	<100
Beckman UniCel DxC 600/800 Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	21	15%	<5	<100
(TBIL)总胆红素					
KHB SA-6000 series KHB \ Vanadate	μmol/L	11.6	0%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Vanadate	μmol/L	17.1	5.2%	12	256
Siemens ADVIA 1800/2400 Siemens \ Vanadate	μmol/L	16.5	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Diazotization	μmol/L	17.1	1.8%	<5	<100
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	14	15%	<5	<100
Roche cobas 4000/6000/8000 DiaSys \ Diazotization	μmol/L	14.1	15%	<5	<100
(TG)甘油三酯					
KHB SA-6000 series KHB \ GPO-POD	mmol/L	0.825	7.7%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ GK-PK-LDH	mmol/L	1.08	4.7%	12	254
Beckman UniCel Dxl 600/800 Beckman Coulter \ GPO-POD	mmol/L	1.18	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ GPO-POD	mmol/L	1.33	3.6%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Free glycerol removed	mmol/L	1.11	15%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ GPO-POD	mmol/L	1.13	15%	<5	<100
(THEO)茶碱					
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μg/mL	12.2	15%	<5	<100
(TRF)转铁蛋白					
Abbott ARCHITECT ci series Dedicated \ Immunoturbidimetry (ITA)	mg/dL	24845	2.4%	5	124
(UA)尿酸					
Abbott ARCHITECT ci series BIOSINO \ Phosphotungstate colorimetry	μmol/L	308	1.7%	12	256
Roche cobas 4000/6000/8000 Roche \ Uricase colorimetry	μmol/L	299	15%	<5	<100
Siemens ADVIA 1800/2400 Reebio \ Uricase colorimetry	μmol/L	296	1.2%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Uricase colorimetry	μmol/L	309	15%	<5	<100
Beckman AU series Beckman Coulter \ Uricase colorimetry	μmol/L	296	15%	<5	<100
(UIBC)不饱和和铁结合力					
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	41.7	15%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Ferrozine colorimetry	μmol/L	41.8	15%	<5	<100

REF	581-1	6×5mL	EXP	2024-09-22	LOT	376B86D	更新日期：	2024-05
6 / 6								

		批号：376B86D			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Urea)尿素					
KHB SA-6000 series KHB \ OTHER	mmol/L	4.78	4.3%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Diacetyl monoxime	mmol/L	4.26	4.3%	12	253
Siemens ADVIA 1800/2400 Siemens \ Urease-UV-rate	mmol/L	4.39	15%	<5	<100
Beckman AU series Beckman Coulter \ Urease-UV-rate	mmol/L	4.55	15%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Urease colorimetry	mmol/L	4.43	15%	<5	<100
(Zn)锌					
Beckman AU series Beckman Coulter \ ICP-MS	μmol/L	25.8	15%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。