		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(ADA)腺苷脱氨酶					
Beckman AU series PRB \ Peroxidase	U/L	7.13	8.8%	5	<100
Beckman UniCel DxC 600/800 Reebio \ Peroxidase	U/L	8.4	7.8%	21	395
Roche cobas c 501/502/701/702 MedicalSystem \ Peroxidase	U/L	7.95	4.7%	19	598
Roche cobas c 501/502/701/702 WEGO \ Peroxidase	U/L	6.98	7.1%	10	307
(a-HBDH)α-羟丁酸脱氢酶					
KHB ZY series KHB \ DGKC	U/L	134	2.4%	8	<100
Roche Hitachi/Hitachi Systems MAKER \ DGKC	U/L	127	3.7%	8	208
Abbott ARCHITECT c series BIOSINO \ DGKC	U/L	136	2.6%	8	208
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic Cycling Assays Method	U/L	132	13.4%	7	309
Mindray BS series Mindray \ Enzymatic Cycling Assays Method	U/L	142	2.1%	5	<100
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic Cycling Assays Method	U/L	126	3.4%	<5	<100
Mindray BS series MAKER \ Enzymatic Cycling Assays Method	U/L	119	5.3%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	U/L	108	4%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic Cycling Assays Method	U/L	129	3%	28	1454
Abbott ARCHITECT ci series BIOSINO \ Enzymatic Cycling Assays Method	U/L	125	4.8%	23	648
Beckman UniCel DxC 600/800 DiaSys \ OTHER	U/L	99.2	6.8%	21	404
Beckman AU series chongqiong zybio \ OTHER	U/L	130	2%	<5	<100
Siemens ADVIA Chemistry XPT KHB \ DGKC	U/L	116	6.9%	17	441
Beckman AU series Iprocom \ Enzymatic Cycling Assays Method	U/L	137	4%	14	340
Beckman AU series HuaChenBio \ DGKC	U/L	126	2.9%	14	329
TOSHIBA TBA chongqiong zybio \ DGKC	U/L	128	3.4%	13	364
Roche Hitachi/Hitachi Systems chongqiong zybio \ DGKC	U/L	127	2.4%	12	347
Beckman AU series Dedicated \ Enzymatic Cycling Assays Method	U/L	123	4.5%	11	279
Roche cobas c 501/502/701/702 Roche \ OTHER	U/L	131	3.1%	10	307
(Alb)白蛋白					
KHB ZY series KHB \ Bromocresol green	g/L	46.4	3.4%	8	<100
Abbott ARCHITECT c series HuaChenBio \ Bromocresol green	g/L	43	0.9%	8	208
Ortho Vitros 3600/5600 Ortho \ dry chemistry	g/L	42.4	7.6%	7	137
Roche cobas 4000/6000/8000 SekiSui \ Bromocresol green	g/L	42.9	1.7%	6	<100
Siemens ADVIA 1800/2400 BIOSINO \ Bromocresol green	g/L	43.3	1.7%	42	2183
Siemens ADVIA Chemistry XPT Siemens \ Bromocresol green	g/L	44.8	1.5%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Bromocresol green	g/L	45.1	2.1%	35	907
Mindray BS series MAKER \ Bromocresol green	g/L	43.6	0.8%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Bromocresol green	g/L	43.7	1.2%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Bromocresol green	g/L	41.5	4.8%	23	560
Roche Hitachi/Hitachi Systems MedicalSystem \ Bromocresol green	g/L	42.5	1.9%	22	810
Beckman UniCel DxC 600/800 Beckman Coulter \ Bromocresol purple	g/L	41.5	1.5%	21	403
Roche Hitachi/Hitachi Systems KHB \ Bromocresol green	g/L	42.3	8.1%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Bromocresol green	g/L	44.1	1.7%	19	588
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	41.4	11.1%	17	450
Beckman AU series Reebio \ Bromocresol green	g/L	44.1	2.6%	14	339
Beckman AU series Autobio \ Bromocresol green	g/L	42.4	0.8%	14	330
Mindray BS series Mindray \ Bromocresol green	g/L	43.1	3.1%	13	<100
TOSHIBA TBA chongqiong zybio \ Bromocresol green	g/L	43.8	1.6%	13	364
(ALP)碱性磷酸酶					
KHB ZY series KHB \ Rate -AMP buffer	U/L	102	3%	9	<100

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	100	4.4%	6	<100
Roche cobas 4000/6000/8000 Roche \ Rate -other buffers	U/L	89.5	2.9%	6	104
Abbott ARCHITECT c series BIOSINO \ dry chemistry	U/L	102	2.4%	5	115
Siemens ADVIA 1800/2400 BIOSINO \ Rate -AMP buffer	U/L	104	2.7%	42	2180
Siemens ADVIA Chemistry XPT Siemens \ Rate -AMP buffer	U/L	98.3	5.4%	<5	<100
Abbott ARCHITECT c series BIOSINO \ Rate -AMP buffer	U/L	103	1.9%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Rate -AMP buffer	U/L	105	2.8%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Rate -AMP buffer	U/L	107	7.7%	23	648
Roche cobas c 501/502/701/702 Roche \ Rate -AMP buffer	U/L	92.1	4.1%	23	535
Roche Hitachi/Hitachi Systems MedicalSystem \ Rate -AMP buffer	U/L	114	8.5%	22	812
Beckman UniCel DxC 600/800 Beckman Coulter \ Rate -AMP buffer	U/L	109	4.7%	21	346
Roche Hitachi/Hitachi Systems KHB \ dry chemistry	U/L	95.4	3%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	U/L	107	2.3%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Rate -AMP buffer	U/L	90.3	4.2%	19	595
Siemens ADVIA Chemistry XPT KHB \ Rate -AMP buffer	U/L	124	7.9%	17	891
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	103	17.3%	16	412
Beckman AU series Iprocom \ Rate -AMP buffer	U/L	103	2.2%	14	343
Beckman AU series Autobio \ Rate -AMP buffer	U/L	98.1	4.6%	14	331
(ALT/ SGPT)丙氨酸氨基转移酶					
KHB ZY series KHB \ UV-rate with P5P	U/L	37.7	2.7%	9	<100
Beckman AU series Dedicated \ UV-rate with P5P	U/L	45.6	3%	9	232
Abbott ARCHITECT c series BIOSINO \ UV-rate without P5P	U/L	41.9	2.2%	8	203
Mindray BS series Mindray \ OTHER	U/L	48.5	1.6%	7	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	39	3.4%	6	<100
Roche cobas 4000/6000/8000 DiaSys \ UV-rate with P5P	U/L	41.5	3.6%	6	<100
Siemens ADVIA Chemistry XPT Siemens \ UV-rate without P5P	U/L	49	2.7%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ UV-rate with P5P	U/L	43.8	5%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ UV-rate without P5P	mmol/L	42.8	3.6%	28	1457
Roche cobas c 501/502/701/702 Roche \ UV-rate without P5P	U/L	39	4.9%	25	594
Abbott ARCHITECT ci series BIOSINO \ UV-rate with P5P	U/L	38.7	9%	23	573
Roche Hitachi/Hitachi Systems MedicalSystem \ UV-rate without P5P	U/L	46.2	7.7%	22	810
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	U/L	49.3	3%	21	403
Roche Hitachi/Hitachi Systems KHB \ dry chemistry	U/L	46.8	2.8%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ UV-rate without P5P	U/L	44.9	3.2%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	U/L	41.5	2.2%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ UV-rate without P5P	U/L	38.6	3.3%	19	592
Beckman AU series Beckman Coulter \ UV-rate without P5P	U/L	41.2	7%	16	386
Siemens ADVIA 1800/2400 BIOSINO \ UV-rate without P5P	U/L	42.6	3.7%	15	772
(ammo)血氨					
Siemens Dimension series MedicalSystem \ Enzymatic	mmol/L	0.037	18.9%	<5	<100
(AMY)淀粉酶					
KHB ZY series KHB \ CNP_G3	U/L	103	2.5%	9	<100
Beckman AU series Elikan \ PNP-G7	U/L	96.5	1.3%	8	268
Abbott ARCHITECT c series BIOSINO \ CNP_G3	U/L	104	1%	8	206
Roche cobas 4000/6000/8000 Roche \ PNP-G7	U/L	94.7	6.5%	7	152
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	82.5	7.6%	7	119
Siemens ADVIA Chemistry XPT Siemens \ PNP-G7	U/L	101	1.5%	<5	<100

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series MAKER \ PNP-G7	U/L	103	1%	<5	<100
Beckman AU series Beckman Coulter \ PNP-G7	U/L	105	3.6%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ CNP_G3	U/L	102	1.9%	<5	<100
Beckman AU series DiaSys \ PNP-G7	U/L	89.4	2.9%	<5	<100
Roche cobas c 501/502/701/702 Roche \ PNP-G7	U/L	98.8	2.8%	25	612
Abbott ARCHITECT ci series BIOSINO \ PNP-G7	U/L	102	2.9%	22	612
Beckman AU series OTHER \ OTHER	U/L	98.6	1.7%	<5	<100
Mindray BS series Mindray \ OTHER	U/L	102	2.3%	<5	<100
Siemens ADVIA Chemistry XPT Dedicated \ Iodine colorimetry	U/L	98.2	4.3%	17	888
Roche Hitachi/Hitachi Systems MedicalSystem \ PNP-G7	U/L	104	8%	16	557
Beckman UniCel DxC 600/800 KHB \ Iodine colorimetry	U/L	93.6	4.5%	15	281
Beckman AU series Iprocom \ PNP-G7	U/L	98.5	2.9%	14	344
Beckman AU series Autobio \ CNP_G3	U/L	111	2%	14	331
(Apo-A Ⅰ)载脂蛋白A Ⅰ					
KHB ZY series KHB \ immunoturbidimetric	g/L	1.75	7.3%	8	<100
Beckman AU series Beckman Coulter \ immunoturbidimetric	g/L	2.02	14.9%	6	<100
Siemens ADVIA 1800/2400 BIOSINO \ immunoturbidimetric	g/L	1.91	3.4%	42	2165
Roche cobas c 501/502/701/702 Roche \ immunoturbidimetric	g/L	1.84	5.1%	28	646
Beckman UniCel DxC 600/800 Beckman Coulter \ immunoturbidimetric	g/L	1.84	8.7%	21	298
Beckman AU series KINGSBIO \ OTHER	g/L	1.86	2.5%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ immunoturbidimetric	g/L	1.67	10.4%	19	673
Roche cobas c 501/502/701/702 Dedicated \ immunoturbidimetric	g/L	1.78	2.1%	19	581
Abbott ARCHITECT ci series BIOSINO \ immunoturbidimetric	g/L	1.72	4.1%	14	353
Beckman AU series Dedicated \ immunoturbidimetric	g/L	1.81	5.2%	10	236
Roche cobas 4000/6000/8000 Roche \ immunoturbidimetric	g/L	1.82	4.1%	10	211
Roche cobas c 501/502/701/702 DENUO \ immunoturbidimetric	g/L	1.91	2.4%	<5	<100
(Apo-B)载脂蛋白B					
KHB ZY series KHB \ immunoturbidimetric	g/L	0.706	5%	9	<100
Beckman AU series Beckman Coulter \ immunoturbidimetric	g/L	0.752	5.2%	6	<100
Siemens ADVIA 1800/2400 BIOSINO \ immunoturbidimetric	g/L	0.751	3.5%	42	2183
Roche cobas c 501/502/701/702 Roche \ immunoturbidimetric	g/L	0.711	5%	30	663
Beckman UniCel DxC 600/800 Beckman Coulter \ immunoturbidimetric	g/L	0.662	7.1%	20	367
Beckman AU series KINGSBIO \ OTHER	g/L	0.675	2.2%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ immunoturbidimetric	g/L	0.666	4%	19	674
Roche cobas c 501/502/701/702 Dedicated \ immunoturbidimetric	g/L	0.662	2.3%	19	585
Abbott ARCHITECT ci series BIOSINO \ immunoturbidimetric	g/L	0.808	5.3%	14	370
Beckman AU series Dedicated \ immunoturbidimetric	g/L	0.769	3.2%	11	303
Roche cobas 4000/6000/8000 Roche \ immunoturbidimetric	g/L	0.83	2.6%	10	264
Roche cobas c 501/502/701/702 DENUO \ immunoturbidimetric	g/L	0.704	0.9%	<5	<100
(AST/ SGOT)天门冬氨酸氨基转移酶					
KHB ZY series KHB \ UV-rate with P5P	U/L	36.6	4%	9	<100
Abbott ARCHITECT c series BIOSINO \ UV-rate without P5P	U/L	38.7	2.3%	8	205
Mindray BS series Mindray \ OTHER	U/L	47.2	4.3%	7	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	36.6	5.4%	6	<100
Roche cobas 4000/6000/8000 DiaSys \ UV-rate with P5P	U/L	38.8	3.8%	6	<100
Siemens ADVIA 1800/2400 BIOSINO \ UV-rate without P5P	U/L	40.6	5.5%	42	2193
Siemens ADVIA Chemistry XPT Siemens \ UV-rate without P5P	U/L	41.8	3.5%	<5	<100

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series MAKER \ OTHER	U/L	42.2	2.2%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ UV-rate with P5P	U/L	40.4	3.9%	<5	<100
Roche cobas c 501/502/701/702 Roche \ UV-rate without P5P	U/L	36.3	3.8%	24	465
Abbott ARCHITECT ci series BIOSINO \ UV-rate with P5P	U/L	35.5	7.9%	23	466
Roche Hitachi/Hitachi Systems MedicalSystem \ UV-rate without P5P	U/L	41.1	5.2%	21	751
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	U/L	44.3	3.4%	21	402
Roche Hitachi/Hitachi Systems KHB \ dry chemistry	U/L	41.6	4%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	U/L	39.6	2.9%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ UV-rate without P5P	U/L	36.2	3.5%	19	583
Beckman AU series Beckman Coulter \ UV-rate without P5P	U/L	41.9	6.4%	15	388
Beckman AU series Iprocom \ UV-rate with P5P	U/L	40	3.6%	14	341
Beckman AU series Autobio \ UV-rate with P5P	U/L	40.8	3.2%	14	330
(BUN)尿素氮					
KHB ZY series KHB \ Dehydrogenase	mmol/L	5.41	4.9%	9	<100
Mindray BS series Mindray \ Dehydrogenase	mmol/L	4.33	0.5%	7	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	3.82	3.8%	6	<100
Mindray BS series Mindray \ Urease colorimetry	mmol/L	4.48	1.9%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Urease-UV-rate	mmol/L	4.5	3.6%	34	897
Siemens ADVIA 1800/2400 BIOSINO \ Urease colorimetry	mmol/L	4.57	2.5%	34	1654
Mindray BS series MAKER \ Urease-UV-rate	mmol/L	4.79	1.2%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	mmol/L	4.6	3%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Urease colorimetry	mmol/L	4.76	5.7%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Urease-UV-rate	mmol/L	4.34	2.7%	21	751
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	mmol/L	4.73	5.2%	21	376
TOSHIBA TBA chongqiong zybio \ Urease-UV-rate	mmol/L	4.69	2.9%	13	354
Roche Hitachi/Hitachi Systems BIOSINO \ Urease-UV-rate	mmol/L	4.72	3.3%	12	656
Roche Hitachi/Hitachi Systems chongqiong zybio \ Urease-UV-rate	mmol/L	4.88	3.2%	12	349
Siemens Dimension series Siemens \ OTHER	mmol/L	4.48	3.2%	<5	<100
Roche cobas c 501/502/701/702 DENUO \ Urease-UV-rate	mmol/L	4.34	1.7%	<5	<100
(C1q)补体C1q					
Siemens ADVIA Chemistry XPT OTHER \ Immunonephelometry (INA)	mg/L	157	7.3%	17	894
(Ca)钙					
Roche cobas 4000/6000/8000 Roche \ BAPTA	mmol/L	2.06	7.1%	7	118
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	2.08	5.1%	6	<100
Mindray BS series Mindray \ Arsenazo III	mmol/L	2.1	1.1%	6	<100
Mindray BS series Mindray \ ICP-MS	mmol/L	1.83	1.3%	5	<100
Abbott ARCHITECT c series BIOSINO \ Arsenazo III	mmol/L	2.03	1%	5	114
TOSHIBA TBA chongqiong zybio \ Arsenazo III	mmol/L	2.19	1.4%	<5	<100
Siemens ADVIA Chemistry XPT Siemens \ Arsenazo III	mmol/L	2.07	2.1%	<5	<100
Abbott ARCHITECT c series BIOSINO \ OCPC colorimetry	mmol/L	2.13	2.1%	<5	<100
Mindray BS series MAKER \ Arsenazo III	mmol/L	2.13	1.6%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Arsenazo III	mmol/L	2.07	2.3%	23	612
Beckman UniCel DxC 600/800 Beckman Coulter \ ISE (indirect)	mmol/L	2.12	1.5%	21	401
Roche Hitachi/Hitachi Systems KHB \ Arsenazo III	mmol/L	2.16	9.2%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ MTB colorimetry	mmol/L	2.07	2.6%	19	689
Roche cobas c 501/502/701/702 Dedicated \ BAPTA	mmol/L	2.1	1.6%	19	585
Beckman AU series Beckman Coulter \ Arsenazo III	mmol/L	2.13	2.3%	16	336

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas c 501/502/701/702 Roche \ BAPTA	mmol/L	2.09	2%	15	364
Siemens ADVIA 1800/2400 BIOSINO \ OCPC colorimetry	mmol/L	2.16	2.5%	14	725
Siemens ADVIA 1800/2400 BIOSINO \ Arsenazo III	mmol/L	2.13	2.7%	14	717
Beckman AU series Reebio \ Arsenazo III	mmol/L	2.15	1.3%	14	337
(CG)甘胆酸					
Roche cobas 4000/6000/8000 MedicalSystem \ OTHER	mg/L	1.16	10.8%	<5	<100
Roche cobas c 501/502/701/702 MedicalSystem \ Latex agglutination (LA)	mg/L	1.72	16.1%	19	574
Siemens ADVIA Chemistry XPT Autobio \ Latex agglutination (LA)	mg/L	0.78	10.9%	<5	<100
(CHE)胆碱酯酶					
KHB ZY series KHB \ Butyrylthiocholine Method	U/L	7338	3.3%	9	<100
Beckman AU series Elikan \ Butyrylthiocholine Method	U/L	7453	1.9%	8	272
Beckman AU series Dedicated \ Butyrylthiocholine Method	U/L	6570	1.9%	8	230
Abbott ARCHITECT c series HuaChenBio \ Butyrylthiocholine Method	U/L	7535	2.3%	8	203
Roche Hitachi/Hitachi Systems MedicalSystem \ DTNB	U/L	7811	2.8%	7	301
Ortho Vitros 3600/5600 Ortho \ OTHER	U/L	6114	2%	6	<100
Beckman AU series Beckman Coulter \ Butyrylthiocholine Method	U/L	6641	3.3%	5	<100
Mindray BS series Mindray \ DTNB	U/L	5037	1.4%	<5	<100
Mindray BS series MAKER \ Butyrylthiocholine Method	U/L	6793	0.8%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ DTNB	U/L	7599	7.8%	23	645
Beckman UniCel DxC 600/800 KHB \ Butyrylthiocholine Method	U/L	6997	1.5%	21	402
Roche Hitachi/Hitachi Systems KHB \ Butyrylthiocholine Method	U/L	7413	2.7%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Butyrylthiocholine Method	U/L	7190	1.4%	16	381
TOSHIBA TBA chongqiong zybio \ Butyrylthiocholine Method	U/L	7096	2.4%	13	363
Roche Hitachi/Hitachi Systems chongqiong zybio \ Butyrylthiocholine Method	U/L	7084	2.4%	12	349
Roche cobas c 501/502/701/702 Roche \ DTNB	U/L	7239	1.9%	10	233
Abbott ARCHITECT c series Dedicated \ Butyrylthiocholine Method	U/L	5846	0.5%	<5	<100
Siemens Dimension series KUAKE \ OTHER	U/L	7643	2.4%	<5	<100
Roche cobas c 501/502/701/702 DENUO \ DTNB	U/L	7012	0.8%	<5	<100
(CHOL/ TC)总胆固醇					
KHB ZY series KHB \ Cholesterol oxidase	mmol/L	4.34	3.2%	9	<100
Abbott ARCHITECT c series BIOSINO \ Cholesterol oxidase	mmol/L	4.25	1.1%	8	208
Beckman AU series Beckman Coulter \ Cholesterol oxidase	mmol/L	3.93	19.5%	7	162
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	4.55	5.2%	6	<100
Roche cobas 4000/6000/8000 OTHER \ Cholesterol oxidase	mmol/L	4.11	1.4%	6	<100
Siemens ADVIA 1800/2400 BIOSINO \ Cholesterol oxidase	mmol/L	4.24	2.4%	42	2180
Siemens ADVIA Chemistry XPT Siemens \ Cholesterol oxidase	mmol/L	4.12	1.4%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Cholesterol oxidase	mmol/L	4.12	3.2%	34	863
Roche Hitachi/Hitachi Systems BSBE \ Cholesterol oxidase	mmol/L	4.13	1.5%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Cholesterol oxidase	mmol/L	4.2	2.5%	23	631
Roche Hitachi/Hitachi Systems MedicalSystem \ Cholesterol oxidase	mmol/L	4.11	3%	22	810
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	mmol/L	4.3	1.9%	21	402
Roche Hitachi/Hitachi Systems KHB \ Cholesterol oxidase	mmol/L	4.17	5.8%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Cholesterol oxidase	mmol/L	4.13	1.9%	19	592
Beckman AU series Autobio \ Cholesterol oxidase	mmol/L	4.27	1.5%	14	334
Beckman AU series Reebio \ Cholesterol oxidase	mmol/L	4.03	3.5%	14	326
TOSHIBA TBA chongqiong zybio \ Cholesterol oxidase	mmol/L	4.1	2.1%	13	359
Mindray BS series Mindray \ Cholesterol oxidase	mmol/L	4.1	3.2%	13	102

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems chongqiong zybio \ Cholesterol oxidase	mmol/L	4.13	1.8%	12	345
(CK)肌酸激酶					
KHB ZY series KHB \ OTHER	U/L	79.8	5.4%	9	<100
Abbott ARCHITECT c series BIOSINO \ DGKC	U/L	84.6	3.9%	8	206
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	83.5	7.3%	6	<100
Mindray BS series Mindray \ Rate -AMP buffer	U/L	74.9	4.2%	6	<100
Beckman AU series Beckman Coulter \ OTHER	U/L	79	6.2%	5	<100
Siemens ADVIA Chemistry XPT Siemens \ Rate -other buffers	U/L	81.4	5.2%	<5	<100
Mindray BS series MAKER \ OTHER	U/L	78.8	5%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Rate -AMP buffer	U/L	77	2.1%	<5	<100
Beckman AU series Beckman Coulter \ Rate -AMP buffer	U/L	72.7	5.8%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Urease-UV-rate	U/L	81.9	7.3%	23	609
Beckman UniCel DxC 600/800 DiaSys \ OTHER	U/L	87.4	5.3%	21	391
Roche Hitachi/Hitachi Systems KHB \ DGKC	U/L	98.8	7.2%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Rate -AMP buffer	U/L	81	4.8%	18	634
Siemens ADVIA Chemistry XPT Dedicated \ Rate -AMP buffer	U/L	93.5	8.1%	17	437
Roche cobas 4000/6000/8000 Roche \ Rate -AMP buffer	U/L	89.7	7.4%	17	401
Siemens ADVIA 1800/2400 BIOSINO \ Urease-UV-rate	U/L	86.9	4.3%	14	703
Siemens ADVIA 1800/2400 BIOSINO \ Rate -AMP buffer	U/L	87.3	4.6%	14	694
Beckman AU series Iprocom \ Rate -AMP buffer	U/L	85.1	2.6%	14	343
Beckman AU series Autobio \ DGKC	U/L	85.6	1.8%	14	329
(CK-MB)肌酸激酶-MB (活性)					
KHB ZY series KHB \ Immune suppression	U/L	11.4	9.2%	9	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ DGKC	U/L	17	11.5%	7	290
Ortho Vitros 3600/5600 Ortho \ OTHER	U/L	10.1	21.1%	6	<100
Beckman AU series Beckman Coulter \ Immune suppression	U/L	14.5	24.3%	<5	<100
Mindray BS series MAKER \ Enzyme immunoassay (EIA)	U/L	13.1	9.6%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ DGKC	U/L	14.1	11.6%	28	1111
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	U/L	13.7	12.2%	25	457
Abbott ARCHITECT ci series BIOSINO \ Immune suppression	U/L	14.1	11.4%	23	661
Beckman UniCel DxC 600/800 DiaSys \ OTHER	U/L	12.9	12.1%	21	401
Beckman AU series chongqiong zybio \ OTHER	U/L	18.4	7.6%	<5	<100
Beckman AU series Dedicated \ Immune suppression	U/L	12.7	13.4%	17	375
Roche cobas c 501/502/701/702 LEADMAN \ Immune suppression	U/L	12.7	7.8%	10	303
Roche cobas 4000/6000/8000 Roche \ DGKC	U/L	17.3	9%	10	266
Roche cobas 4000/6000/8000 Roche \ Immune suppression	U/L	12.5	12.8%	<5	<100
Abbott ARCHITECT c series Dedicated \ Immune suppression	U/L	11.8	2.8%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ DGKC	U/L	14.3	12.6%	<5	<100
Roche cobas c 501/502/701/702 DENUO \ Immunoturbidimetry (ITA)	U/L	9.93	10.4%	<5	<100
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	U/L	6	0%	<5	<100
(Cl)氯					
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	117	6.3%	7	138
Mindray BS series Mindray \ ISE (indirect)	mmol/L	117	0.9%	6	<100
Roche cobas 4000/6000/8000 OTHER \ ISE (indirect)	mmol/L	116	1.2%	6	<100
MZKL CORNLEY \ ISE (direct)	mmol/L	119	1.6%	6	<100
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	115	1.5%	59	1304
Abbott ARCHITECT c series Abbott \ Coulometric titration	mmol/L	120	0.9%	5	118

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
TOSHIBA TBA chongqiong zybio \ ISE (direct)	mmol/L	120	0.8%	<5	<100
Siemens ADVIA Chemistry XPT Siemens \ ISE (indirect)	mmol/L	118	0.5%	<5	<100
Abbott ARCHITECT c series Abbott \ ISE (indirect)	mmol/L	122	0.9%	<5	<100
Mindray BS series MAKER \ ISE (indirect)	mmol/L	118	0.7%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ ISE (indirect)	mmol/L	118	1.6%	28	1482
Abbott ARCHITECT ci series BIOSINO \ ISE (indirect)	mmol/L	119	1.6%	23	661
Roche Hitachi/Hitachi Systems Roche \ ISE (indirect)	mmol/L	114	1.4%	22	802
Beckman UniCel DxC 600/800 Beckman Coulter \ ISE (indirect)	mmol/L	113	1.4%	21	403
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	98	19.6%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ ISE (indirect)	mmol/L	116	1.4%	19	595
Roche Hitachi/Hitachi Systems Dedicated \ ICP-MS	mmol/L	113	3.9%	12	664
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	116	1.4%	11	313
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	119	1.5%	11	289
(CO2（HCO3-）)二氧化碳（碳酸氢根）					
Abbott ARCHITECT c series HuaChenBio \ Colorimetry	mmol/L	7.39	14.4%	7	191
Roche cobas c 501/502/701/702 OTHER \ PEPC	mmol/L	8.82	7.1%	5	118
Beckman AU series DiaSys \ PEPC	mmol/L	2.55	19.4%	<5	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	5	0%	<5	<100
(Cr)肌酐					
KHB ZY series KHB \ Enzymatic	μmol/L	88.4	2.8%	9	<100
Beckman AU series Dedicated \ Enzymatic	μmol/L	78.4	3.3%	9	234
Ortho Vitros 3600/5600 Ortho \ dry chemistry	μmol/L	89.8	4.7%	6	<100
Roche cobas 4000/6000/8000 OTHER \ Enzymatic	μmol/L	77.7	1.6%	6	<100
Abbott ARCHITECT c series BIOSINO \ dry chemistry	μmol/L	84.8	3%	5	118
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic	μmol/L	82.3	2.8%	42	2174
Beckman AU series Beckman Coulter \ Enzymatic	μmol/L	78.6	13.4%	<5	<100
Siemens ADVIA Chemistry XPT Siemens \ Enzymatic	μmol/L	80.9	1.9%	<5	<100
Abbott ARCHITECT c series BIOSINO \ PAP	μmol/L	85.4	1.1%	<5	<100
Mindray BS series MAKER \ Enzymatic	μmol/L	78.3	2.9%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	μmol/L	87.9	2.3%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Enzymatic	μmol/L	75.7	1.6%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Enzymatic	μmol/L	82.1	8.6%	25	564
Abbott ARCHITECT ci series BIOSINO \ Picric acid-Blank Compensation	μmol/L	78.2	3.6%	22	624
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic	μmol/L	81.7	2.8%	21	757
Beckman UniCel DxC 600/800 MedicalSystem \ Enzymatic	μmol/L	78.5	4.5%	21	404
Roche Hitachi/Hitachi Systems KHB \ dry chemistry	μmol/L	79.8	3.4%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Enzymatic	μmol/L	88.7	1.9%	19	584
Mindray BS series Mindray \ Enzymatic	μmol/L	77.5	8.6%	16	131
(CRP)C反应蛋白					
Beckman AU series DiaSys \ Immunoturbidimetry (ITA)	mg/dL	0.698	6.8%	8	216
Roche Hitachi/Hitachi Systems DiaSys \ Immunoturbidimetry (ITA)	mg/dL	0.761	3.1%	12	655
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	6.74	11%	<5	<100
Siemens Dimension series Siemens \ Immunoturbidimetry (ITA)	mg/dL	0.703	4.5%	<5	<100
Beckman AU series Dedicated \ Immunonephelometry (INA)	mg/dL	2.25	0%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	mg/dL	3.04	0%	<5	<100
(Cu)铜					
Beckman AU series Beckman Coulter \ ICP-MS	μmol/L	19.4	14.8%	<5	<100

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Cys C)半胱氨酸蛋白酶抑制剂C/ 胱抑素C					
Beckman AU series Szybio \ Immunoturbidimetry (ITA)	mg/L	0.541	4.7%	8	238
Beckman AU series PRB \ OTHER	mg/L	0.38	12%	<5	<100
Beckman UniCel DxC 600/800 KHB \ OTHER	mg/L	0.366	13.9%	21	391
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	mg/L	0.456	9.1%	21	391
Beckman AU series BSBE \ Immunonephelometry (INA)	mg/L	0.325	2.2%	<5	<100
Roche cobas c 501/502/701/702 DENUO \ Immunoturbidimetry (ITA)	mg/L	0.398	2.1%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	0.31	0%	<5	<100
(DBIL)直接胆红素					
KHB ZY series KHB \ Vanadate	μmol/L	7.77	9.5%	9	<100
Siemens ADVIA Chemistry XPT DiaSys \ Vanadate	μmol/L	3.31	5.7%	9	429
Beckman AU series Beckman Coulter \ Diazotization -Blank Compensation	μmol/L	3.99	13.7%	9	301
Roche cobas 4000/6000/8000 Wako \ Vanadate	μmol/L	5.94	3%	6	<100
Abbott ARCHITECT c series HuaChenBio \ Diazotization	μmol/L	6.51	2.7%	5	118
Siemens ADVIA 1800/2400 BIOSINO \ Vanadate	μmol/L	5.34	5.7%	42	2184
Siemens ADVIA Chemistry XPT Siemens \ Vanadate	μmol/L	6.35	2.7%	<5	<100
Abbott ARCHITECT c series HuaChenBio \ Vanadate	μmol/L	6.45	1.3%	<5	<100
Beckman AU series Beckman Coulter \ Diazotization -No blank Compensation	μmol/L	3.88	5.5%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Diazotization	μmol/L	6.68	2.6%	<5	<100
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	4.44	5.7%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Vanadate	μmol/L	5.35	6%	23	615
Roche Hitachi/Hitachi Systems Wako \ Vanadate	μmol/L	5.44	5.1%	22	778
Roche Hitachi/Hitachi Systems KHB \ Diazotization	μmol/L	7.32	2.6%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	μmol/L	3.96	5.2%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Diazotization -No blank Compensation	μmol/L	5.89	4.4%	19	583
Mindray BS series Mindray \ Vanadate	μmol/L	6.68	11.5%	18	187
Beckman AU series Dedicated \ sodium nitrite oxidase method	μmol/L	4.87	6.1%	16	388
Roche cobas c 501/502/701/702 Roche \ Diazotization	μmol/L	6.09	5.5%	14	348
(DGXN)地高辛					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.672	10.5%	<5	<100
(Fe)铁					
Abbott ARCHITECT ci series BIOSINO \ Atomic absorption spectroscopy (AAS)	μmol/L	19.3	2.5%	7	163
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	19.5	11%	7	122
Siemens ADVIA 1800/2400 BIOSINO \ Ferrozine colorimetry	μmol/L	18.6	3.8%	42	2170
Beckman AU series Beckman Coulter \ Ferrozine colorimetry	μmol/L	18.4	19%	<5	100
Roche cobas c 501/502/701/702 Roche \ Ferrozine colorimetry	μmol/L	18.5	3.2%	15	357
Beckman AU series Autobio \ Ferrozine colorimetry	μmol/L	20.1	3.8%	14	330
Roche Hitachi/Hitachi Systems Hitachi \ Ferrozine colorimetry	μmol/L	20.9	2.7%	11	298
(GGT)L-γ-谷氨酰基转移酶					
KHB ZY series KHB \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.8	2.1%	9	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	44	2.9%	6	<100
Roche cobas 4000/6000/8000 DiaSys \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.6	2.4%	6	<100
Abbott ARCHITECT c series BIOSINO \ dry chemistry	U/L	40.5	1.6%	5	117
Siemens ADVIA 1800/2400 BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	38.1	2.9%	42	2180
Siemens ADVIA Chemistry XPT Siemens \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	38.6	4.3%	<5	<100
Abbott ARCHITECT c series BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	41	1.3%	<5	<100
Roche cobas c 501/502/701/702 Roche \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.7	5.3%	25	572

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT ci series BIOSINO \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	38.6	4.9%	23	623
Roche Hitachi/Hitachi Systems MedicalSystem \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	39.2	3.2%	21	747
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	U/L	28.9	5.2%	21	403
Roche Hitachi/Hitachi Systems KHB \ dry chemistry	U/L	37.4	2.4%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	U/L	38.7	1.5%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.5	3.2%	19	595
Siemens ADVIA Chemistry XPT Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	38.2	10.2%	17	874
Beckman AU series Beckman Coulter \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	37.7	7.5%	15	383
Mindray BS series Mindray \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	36.3	9.9%	15	143
Beckman AU series Iprocom \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	36.8	2.3%	14	342
Beckman AU series Autobio \ γ-glutamyl -4-nitroaniline	U/L	35.6	4.6%	14	331
(GLDH)谷氨酸脱氢酶					
Roche cobas 4000/6000/8000 DiaSys \ Colorimetry	U/L	6.09	11.8%	<5	<100
Beckman AU series Beckman Coulter \ DGKC	U/L	6.22	11.7%	<5	<100
(GLU)葡萄糖					
KHB ZY series KHB \ Glucose oxidase (GOD)	mmol/L	4.13	3.6%	9	<100
Mindray BS series Mindray \ Hexokinase	mmol/L	4.07	7.5%	8	107
Mindray BS series Mindray \ Glucose oxidase (GOD)	mmol/L	4.07	1.6%	7	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	4.29	1.5%	6	<100
Abbott ARCHITECT c series BIOSINO \ Dehydrogenase	mmol/L	4.34	0.7%	5	117
Siemens ADVIA 1800/2400 BIOSINO \ Hexokinase	mmol/L	4.34	2.6%	42	2187
Siemens ADVIA Chemistry XPT Siemens \ Hexokinase	mmol/L	4.3	3.3%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Hexokinase	mmol/L	4.19	2.9%	34	751
Abbott ARCHITECT c series BIOSINO \ Hexokinase	mmol/L	4.3	0.7%	<5	<100
Mindray BS series MAKER \ Hexokinase	mmol/L	4.17	0.8%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Glucose oxidase (GOD)	mmol/L	4.07	2.1%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Hexokinase	mmol/L	4.14	2.6%	23	635
Roche Hitachi/Hitachi Systems MedicalSystem \ Glucose oxidase (GOD)	mmol/L	4.03	1.8%	21	751
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	mmol/L	4.23	2.1%	21	405
Roche Hitachi/Hitachi Systems KHB \ Glucose oxidase (GOD)	mmol/L	4.22	5%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Hexokinase	mmol/L	4.13	1.5%	19	589
Roche cobas 4000/6000/8000 Roche \ Hexokinase	mmol/L	4.19	4.7%	17	372
Beckman AU series Beckman Coulter \ Hexokinase	mmol/L	4.29	2.8%	16	342
Beckman AU series Reebio \ Hexokinase	mmol/L	4.46	2.3%	14	337
(HCY)同型半胱氨酸					
KHB ZY series KHB \ Enzymatic Cycling Assays Method	μmol/L	15.7	9.6%	8	<100
Beckman AU series Szybio \ Enzymatic Cycling Assays Method	μmol/L	13.3	3.8%	8	274
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	μmol/L	11.6	12.4%	8	205
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic Cycling Assays Method	μmol/L	11.9	4.7%	7	299
Roche cobas c 501/502/701/702 WEGO \ Enzymatic Cycling Assays Method	μmol/L	9.76	8.1%	7	204
Beckman AU series REEGEN \ Enzymatic Cycling Assays Method	μmol/L	12.8	4.4%	5	<100
Abbott ARCHITECT c series BJBJ \ Enzymatic Cycling Assays Method	μmol/L	11.9	3.5%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Immunoturbidimetry (ITA)	μmol/L	12.1	3.7%	27	1393
Roche cobas c 501/502/701/702 Roche \ Enzymatic Cycling Assays Method	μmol/L	13.7	11.4%	25	413
Beckman UniCel DxC 600/800 REEGEN \ Enzymatic Cycling Assays Method	μmol/L	13.4	5.1%	21	403
Beckman AU series OTHER \ OTHER	μmol/L	10.4	2%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	12.3	3.5%	14	720

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas c 501/502/701/702 DENUO \ Enzymatic Cycling Assays Method	μmol/L	12.9	6.1%	<5	<100
(HDL-C)高密度脂蛋白胆固醇					
KHB ZY series KHB \ Direct Measure	mmol/L	1.23	11.2%	9	<100
Beckman AU series Elikan \ Direct Measure	mmol/L	1.33	2.1%	8	271
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	1.39	5.4%	6	<100
Roche cobas 4000/6000/8000 OTHER \ OTHER	mmol/L	1.08	4.9%	6	<100
Beckman AU series Beckman Coulter \ Direct Measure	mmol/L	1.17	4.9%	5	<100
Abbott ARCHITECT c series BIOSINO \ Direct Measure	mmol/L	1.64	1.4%	5	117
Siemens ADVIA 1800/2400 BIOSINO \ Direct Measure	mmol/L	1.51	4.1%	42	2174
Siemens ADVIA Chemistry XPT Siemens \ Direct Measure	mmol/L	1.28	2.7%	<5	<100
Abbott ARCHITECT c series BIOSINO \ Enzymatic	mmol/L	1.55	2.9%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Enzymatic	mmol/L	1.29	2.3%	<5	<100
Beckman AU series OTHER \ Enzymatic	mmol/L	1.19	3.9%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ selected suppression	mmol/L	1.48	3.8%	23	631
Beckman UniCel DxC 600/800 LEADMAN \ selected suppression	mmol/L	1.44	4.7%	21	401
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	0.964	30.1%	<5	<100
Roche Hitachi/Hitachi Systems KHB \ Direct Measure	mmol/L	1.29	1.9%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic	mmol/L	1.56	2.9%	19	678
Roche cobas c 501/502/701/702 Dedicated \ Direct Measure	mmol/L	1.2	2.4%	19	594
Roche cobas c 501/502/701/702 Roche \ selected suppression	mmol/L	1.2	4%	15	349
Beckman AU series Reebio \ Direct Measure	mmol/L	1.42	3.8%	14	339
(hs-CRP)超敏C反应蛋白					
Beckman AU series PRB \ Immunoturbidimetry (ITA)	mg/L	6.95	9.7%	5	<100
Beckman UniCel DxC 600/800 DiaSys \ Immunoturbidimetry (ITA)	mg/L	8.55	8.4%	21	404
Abbott ARCHITECT ci series BIOSINO \ Immunoturbidimetry (ITA)	mg/L	6.2	1.9%	14	390
(IV-Col)IV型胶原					
Beckman AU series Bio-assay \ OTHER	μg/L	40.1	28%	<5	<100
(K)钾					
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	3.86	5.1%	7	113
Roche cobas 4000/6000/8000 OTHER \ ISE (indirect)	mmol/L	3.69	1%	6	<100
Mindray BS series Mindray \ ISE (indirect)	mmol/L	3.86	0.9%	6	<100
MZKL CORNLEY \ ISE (direct)	mmol/L	4.12	0.9%	6	<100
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	3.8	1.9%	59	1439
Abbott ARCHITECT c series Abbott \ dry chemistry	mmol/L	3.75	1%	5	118
TOSHIBA TBA chongqiong zybio \ ISE (direct)	mmol/L	3.77	0.9%	<5	<100
Siemens ADVIA Chemistry XPT Siemens \ ISE (indirect)	mmol/L	3.88	1.1%	<5	<100
Abbott ARCHITECT c series Abbott \ ISE (indirect)	mmol/L	3.81	1.2%	<5	<100
Mindray BS series MAKER \ ISE (indirect)	mmol/L	3.85	0.6%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ ISE (indirect)	mmol/L	3.89	2.7%	28	1469
Abbott ARCHITECT ci series BIOSINO \ ISE (indirect)	mmol/L	3.75	1.8%	23	631
Roche Hitachi/Hitachi Systems Roche \ ISE (indirect)	mmol/L	3.77	2.9%	22	809
Beckman UniCel DxC 600/800 Beckman Coulter \ ISE (indirect)	mmol/L	3.79	1.9%	21	402
Beckman AU series Beckman Coulter \ ISE (direct)	mmol/L	3.56	11.5%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ ISE (indirect)	mmol/L	3.9	0.8%	19	596
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	3.75	1.9%	13	350
Roche Hitachi/Hitachi Systems Dedicated \ ICP-MS	mmol/L	3.89	4.1%	12	667
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	4	4.2%	11	309

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(LA)乳酸					
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	1.96	5%	6	<100
Roche cobas c 501/502/701/702 LEADMAN \ OTHER	mmol/L	2.14	2.4%	10	308
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	mmol/L	1.99	11.1%	<5	<100
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	mmol/L	1.98	10.6%	<5	<100
Siemens Dimension series Siemens \ OTHER	mmol/L	2.25	3.4%	<5	<100
(LDH)乳酸脱氢酶					
KHB ZY series KHB \ Lactate to pyruvate (L-P)	U/L	141	4.6%	9	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	U/L	166	3.1%	6	<100
Beckman AU series Beckman Coulter \ OTHER	U/L	140	4.2%	5	<100
Beckman AU series Beckman Coulter \ Lactate to pyruvate (L-P)	U/L	154	5.9%	5	<100
Abbott ARCHITECT c series BIOSINO \ dry chemistry	U/L	143	2.2%	5	114
Siemens ADVIA Chemistry XPT Siemens \ Lactate to pyruvate (L-P)	U/L	151	1.9%	<5	<100
Abbott ARCHITECT c series BIOSINO \ Lactate to pyruvate (L-P)	U/L	144	1.7%	<5	<100
Mindray BS series MAKER \ Lactate to pyruvate (L-P)	U/L	142	2.7%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Lactate to pyruvate (L-P)	U/L	144	3.2%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Lactate to pyruvate (L-P)	U/L	149	3.2%	28	1456
Roche cobas c 501/502/701/702 Roche \ Lactate to pyruvate (L-P)	U/L	142	2.9%	25	535
Abbott ARCHITECT ci series BIOSINO \ Lactate to pyruvate (L-P)	U/L	135	4.7%	23	644
Roche Hitachi/Hitachi Systems MedicalSystem \ Lactate to pyruvate (L-P)	U/L	143	9.2%	22	791
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	U/L	114	4.4%	21	392
Roche Hitachi/Hitachi Systems KHB \ dry chemistry	U/L	150	3.5%	<5	<100
Siemens ADVIA Chemistry XPT Dedicated \ Lactate to pyruvate (L-P)	U/L	138	6.3%	17	892
Roche cobas 4000/6000/8000 Roche \ Lactate to pyruvate (L-P)	U/L	145	4.9%	17	390
Beckman AU series Iprocom \ Lactate to pyruvate (L-P)	U/L	143	3.2%	14	344
Beckman AU series Autobio \ Lactate to pyruvate (L-P)	U/L	142	1.3%	14	330
(LDL-C)低密度脂蛋白胆固醇					
KHB ZY series KHB \ Direct Measure	mmol/L	1.88	6.3%	8	<100
Beckman AU series OTHER \ Direct Measure	mmol/L	2.31	2.1%	8	272
Abbott ARCHITECT c series BIOSINO \ SUR	mmol/L	2.29	3.7%	8	203
Roche cobas 4000/6000/8000 OTHER \ SUR	mmol/L	2.43	1.3%	6	<100
Beckman AU series Beckman Coulter \ Direct Measure	mmol/L	2.67	3.3%	5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Direct Measure	mmol/L	2.1	3%	42	2181
Siemens ADVIA Chemistry XPT Siemens \ Direct Measure	mmol/L	2.53	3.4%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Enzymatic	mmol/L	2.34	1.9%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ selected suppression	mmol/L	1.98	4%	23	631
Beckman UniCel DxC 600/800 LEADMAN \ SUR	mmol/L	2.17	4.5%	21	400
Beckman AU series Beckman Coulter \ Enzymatic	mmol/L	1.97	28.8%	<5	<100
Beckman AU series OTHER \ SUR	mmol/L	2.56	3.2%	<5	<100
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic	mmol/L	2.23	8.4%	19	666
Roche cobas c 501/502/701/702 Dedicated \ Direct Measure	mmol/L	2.41	2%	19	598
Roche cobas c 501/502/701/702 Roche \ selected suppression	mmol/L	2.44	2.5%	15	364
Beckman AU series Iprocom \ selected suppression	mmol/L	2.36	4.2%	14	343
Beckman AU series Autobio \ SUR	mmol/L	2.06	2.5%	14	333
TOSHIBA TBA chongqiong zybio \ Direct Measure	mmol/L	2.95	3.8%	13	365
Roche Hitachi/Hitachi Systems chongqiong zybio \ Enzymatic	mmol/L	3.05	2.9%	12	347

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Li)锂					
Beckman AU series Beckman Coulter \ Colorimetry	mmol/L	0.76	14.5%	<5	<100
(LIP)脂肪酶					
Abbott ARCHITECT ci series BIOSINO \ Glucose oxidase (GOD)	U/L	36.8	12.3%	7	195
Beckman AU series G-CELL \ Colorimetry	U/L	74.5	5.7%	<5	<100
Beckman AU series KINGSBIO \ OTHER	U/L	45.2	9.9%	<5	<100
Siemens ADVIA Chemistry XPT KHB \ Enzymatic Cycling Assays Method	U/L	55.3	22%	17	872
Roche cobas c 501/502/701/702 Roche \ OTHER	U/L	51.7	4.5%	10	302
Roche cobas 4000/6000/8000 Roche \ Enzymatic Cycling Assays Method	U/L	47.6	4.2%	10	197
Beckman AU series Beckman Coulter \ Glucose oxidase (GOD)	U/L	46.7	10.1%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Colorimetry	U/L	51.4	2.5%	<5	<100
Roche cobas 4000/6000/8000 DiaSys \ Enzymatic Cycling Assays Method	U/L	43.9	11.1%	<5	<100
(Lp(a))脂蛋白a					
Mindray BS series Mindray \ Immunonephelometry (INA)	mg/L	52.2	4.5%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	mg/L	69.5	10.9%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	68.7	38.1%	<5	<100
(Lp-PLA2)脂蛋白磷脂酶A2					
Beckman AU series chongqiong zybio \ Rate -other buffers	IU/L	408	11.1%	<5	<100
(Mg)镁					
Abbott ARCHITECT c series BIOSINO \ Arsenazo □	mmol/L	0.88	2.3%	8	207
Beckman AU series Beckman Coulter \ Xylidyl Blue	mmol/L	0.896	7.6%	8	112
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	0.965	3.1%	6	<100
Roche cobas 4000/6000/8000 SekiSui \ Xylidyl Blue	mmol/L	0.937	2.8%	6	<100
Siemens ADVIA 1800/2400 BIOSINO \ Xylidyl Blue	mmol/L	0.889	3.5%	28	1453
Roche cobas c 501/502/701/702 Roche \ Xylidyl Blue	mmol/L	0.9	4%	25	658
Abbott ARCHITECT ci series BIOSINO \ Calmagite colorimetry	mmol/L	0.882	5.6%	23	648
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	mmol/L	0.917	4.5%	21	402
Beckman AU series OTHER \ Xylidyl Blue	mmol/L	0.843	1.4%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ Xylidyl Blue	mmol/L	0.947	4.4%	19	636
Roche cobas c 501/502/701/702 Dedicated \ Calmagite colorimetry	mmol/L	0.876	2.4%	19	602
TOSHIBA TBA chongqiong zybio \ Xylidyl Blue	mmol/L	0.93	3.6%	13	363
Roche Hitachi/Hitachi Systems chongqiong zybio \ Xylidyl Blue	mmol/L	0.941	5.2%	12	338
Roche Hitachi/Hitachi Systems Hitachi \ Arsenazo □	mmol/L	0.933	2.5%	11	276
Roche cobas c 501/502/701/702 Roche \ Calmagite colorimetry	mmol/L	0.916	2.6%	10	233
Roche cobas 4000/6000/8000 Roche \ ICP-MS	mmol/L	0.895	3.4%	10	226
Siemens ADVIA 1800/2400 Siemens \ Xylidyl Blue	mmol/L	0.85	10.6%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Xylidyl Blue	mmol/L	0.96	12.5%	<5	<100
Roche cobas c 501/502/701/702 DENUO \ Calmagite colorimetry	mmol/L	0.918	2.1%	<5	<100
(Na)钠					
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	141	5.3%	7	119
Beckman AU series Beckman Coulter \ ISE (indirect)	mmol/L	137	2.1%	60	1423
Mindray BS series Mindray \ ISE (indirect)	mmol/L	138	0.9%	6	<100
Roche cobas 4000/6000/8000 OTHER \ ISE (indirect)	mmol/L	138	1%	6	<100
MZKL CORNLEY \ ISE (direct)	mmol/L	148	0.9%	6	<100
Abbott ARCHITECT c series Abbott \ dry chemistry	mmol/L	136	0.8%	5	118
Roche Hitachi/Hitachi Systems Hitachi \ ISE (indirect)	mmol/L	142	1%	<5	<100
TOSHIBA TBA chongqiong zybio \ ISE (direct)	mmol/L	136	0.8%	<5	<100

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Chemistry XPT Siemens \ ISE (indirect)	mmol/L	139	0.8%	<5	<100
Abbott ARCHITECT c series Abbott \ ISE (indirect)	mmol/L	138	0.9%	<5	<100
Mindray BS series MAKER \ ISE (indirect)	mmol/L	138	0.6%	<5	<100
Siemens ADVIA 1800/2400 Dedicated \ ISE (indirect)	mmol/L	139	1.5%	28	1472
Abbott ARCHITECT ci series BIOSINO \ ISE (indirect)	mmol/L	135	1.4%	23	632
Roche Hitachi/Hitachi Systems Roche \ ISE (indirect)	mmol/L	137	1.5%	22	800
Beckman UniCel DxC 600/800 Beckman Coulter \ ISE (indirect)	mmol/L	137	1%	21	401
Roche cobas c 501/502/701/702 Dedicated \ ISE (indirect)	mmol/L	140	0.8%	19	593
Roche Hitachi/Hitachi Systems Dedicated \ ICP-MS	mmol/L	139	4%	12	668
Roche cobas 4000/6000/8000 Roche \ ISE (indirect)	mmol/L	139	3.5%	11	332
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	137	2%	11	289
(NEFA)游离脂肪酸					
Siemens ADVIA 1800/2400 OTHER \ ACS - ACOD	mmol/L	0.539	5%	42	2174
Beckman AU series KINGSBIO \ OTHER	mmol/L	0.562	2%	<5	<100
(P)磷					
Beckman AU series Elikan \ Molybdenum blue colorimetry	mmol/L	1.01	2.2%	9	284
Roche cobas 4000/6000/8000 Roche \ Molybdenum blue colorimetry	mmol/L	0.973	2.4%	6	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	1.14	3.3%	6	<100
Mindray BS series Mindray \ Phosphomolybdate-UV-end point	mmol/L	1.05	1.3%	6	<100
Mindray BS series Mindray \ ICP-MS	mmol/L	0.693	2.2%	5	<100
Beckman AU series Beckman Coulter \ Phosphomolybdate-UV-end point	mmol/L	1.04	7.4%	5	<100
Abbott ARCHITECT c series BIOSINO \ dry chemistry	mmol/L	1.02	1%	5	110
Siemens ADVIA Chemistry XPT Siemens \ Phosphomolybdate-UV-end point	mmol/L	1.08	2.3%	<5	<100
Abbott ARCHITECT c series BIOSINO \ Phosphomolybdate-UV-end point	mmol/L	1.01	1.4%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	mmol/L	1.06	3.5%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ OTHER	mmol/L	1.01	4.3%	<5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Phosphomolybdate-UV-end point	mmol/L	0.99	4.9%	28	1449
Abbott ARCHITECT ci series BIOSINO \ Phosphomolybdate UV-rate	mmol/L	1.02	3.3%	23	631
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	mmol/L	1.07	3%	21	400
Beckman AU series Beckman Coulter \ Phosphomolybdate UV-rate	mmol/L	0.917	23.3%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ Phosphomolybdate-UV-end point	mmol/L	0.964	5.4%	19	695
Roche cobas c 501/502/701/702 Dedicated \ Phosphomolybdate-UV-end point	mmol/L	1.02	2.4%	19	580
Roche cobas c 501/502/701/702 Roche \ Molybdenum blue colorimetry	mmol/L	0.998	2.5%	15	362
Beckman AU series Autobio \ Phosphomolybdate-UV-end point	mmol/L	0.93	6.2%	14	331
(PAB)前白蛋白					
Roche cobas 4000/6000/8000 RD \ Immunoturbidimetry (ITA)	g/L	0.237	3%	6	<100
Beckman UniCel DxC 600/800 MedicalSystem \ Immunonephelometry (INA)	g/L	0.243	3.3%	6	122
Roche cobas c 501/502/701/702 LEADMAN \ OTHER	g/L	0.244	2.4%	<5	<100
Beckman AU series WEGO \ Immunoturbidimetry (ITA)	g/L	0.246	2.5%	<5	<100
Beckman UniCel DxC 600/800 MedicalSystem \ OTHER	g/L	0.244	3.8%	15	280
(PAMY)胰淀粉酶					
Siemens ADVIA Chemistry XPT Dedicated \ OTHER	U/L	49.9	4.8%	17	881
Beckman AU series Beckman Coulter \ Enzymatic	U/L	48.3	11%	<5	<100
(PRO/ TP)(总) 蛋白					
KHB ZY series KHB \ Biuret-endpoint	g/L	72	4%	9	<100
Abbott ARCHITECT c series BIOSINO \ Biuret-endpoint	g/L	69.9	0.7%	8	207
Ortho Vitros 3600/5600 Ortho \ dry chemistry	g/L	70.4	2.9%	6	<100

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000 SekiSui \ Biuret-rate	g/L	64.9	1.3%	6	<100
Siemens ADVIA 1800/2400 BIOSINO \ Biuret-rate	g/L	70.1	2.9%	42	2173
Siemens ADVIA Chemistry XPT Siemens \ Biuret-endpoint	g/L	70	1.3%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Biuret-endpoint	g/L	67.5	1.8%	35	915
Mindray BS series MAKER \ Biuret-endpoint	g/L	66.1	3.4%	<5	<100
Beckman AU series Beckman Coulter \ Biuret-rate	g/L	68.4	1.8%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Biuret-endpoint	g/L	69.1	1.1%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Biuret-rate	g/L	66.5	2%	23	553
Roche Hitachi/Hitachi Systems MedicalSystem \ Biuret-endpoint	g/L	67.7	2%	22	803
Beckman UniCel DxC 600/800 Beckman Coulter \ Biuret-rate	g/L	68.6	1.5%	21	401
Roche Hitachi/Hitachi Systems KHB \ Biuret-endpoint	g/L	64.3	2.5%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Biuret-endpoint	g/L	67.4	1.6%	19	617
Beckman AU series Reebio \ Biuret-endpoint	g/L	68.5	1%	14	334
Beckman AU series Autobio \ Biuret-endpoint	g/L	64.6	0.9%	14	329
TOSHIBA TBA chongqiong zybio \ Biuret-endpoint	g/L	67.8	1.3%	13	363
Beckman AU series Beckman Coulter \ Biuret-endpoint	g/L	69.2	3.2%	13	355
(PROG)孕酮					
Ortho Vitros 3600/5600 Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	72.1	10%	<5	<100
(RBP)视黄醇结合蛋白					
Beckman AU series OTHER \ Immunoturbidimetry (ITA)	mg/dL	4.91	14%	<5	<100
Roche cobas c 501/502/701/702 WEGO \ OTHER	mg/dL	3.63	4.3%	10	306
(SA)唾液酸					
Beckman UniCel DxC 600/800 DongOu \ Enzymatic	mg/L	540	2.8%	21	403
(TBA)总胆汁酸					
KHB ZY series KHB \ Enzymatic Cycling Assays Method	μmol/L	16.6	5.2%	9	<100
Roche Hitachi/Hitachi Systems MAKER \ Enzymatic Cycling Assays Method	μmol/L	19.5	8%	9	227
Roche Hitachi/Hitachi Systems MedicalSystem \ Enzymatic Cycling Assays Method	μmol/L	12	12.4%	8	337
Beckman AU series Szybio \ Enzymatic Cycling Assays Method	μmol/L	12.3	6.9%	8	268
Abbott ARCHITECT c series HuaChenBio \ Enzymatic Cycling Assays Method	μmol/L	20.5	3.2%	8	201
Beckman AU series Beckman Coulter \ Enzymatic Cycling Assays Method	μmol/L	13.1	8.9%	6	<100
Mindray BS series Mindray \ Enzymatic Cycling Assays Method	μmol/L	11.1	3.1%	5	<100
Siemens ADVIA 1800/2400 BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	16.8	4.8%	42	2164
Roche cobas 4000/6000/8000 SekiSui \ Enzymatic Cycling Assays Method	μmol/L	13.9	2.9%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Enzymatic Cycling Assays Method	μmol/L	17.6	7.5%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Enzymatic Cycling Assays Method	μmol/L	14.2	11%	23	556
Roche cobas c 501/502/701/702 Roche \ Enzymatic Cycling Assays Method	μmol/L	19.5	4.6%	23	536
Beckman UniCel DxC 600/800 DiaSys \ Enzymatic Cycling Assays Method	μmol/L	14	5.9%	21	397
Roche Hitachi/Hitachi Systems KHB \ Enzymatic Cycling Assays Method	μmol/L	16.7	3.8%	<5	<100
Siemens ADVIA Chemistry XPT KHB \ Enzymatic Cycling Assays Method	μmol/L	15.6	10.2%	17	871
TOSHIBA TBA chongqiong zybio \ Enzymatic Cycling Assays Method	μmol/L	17.7	5.1%	13	360
Beckman AU series Dedicated \ Enzymatic Cycling Assays Method	μmol/L	15.3	11.8%	13	338
Roche Hitachi/Hitachi Systems chongqiong zybio \ Enzymatic Cycling Assays Method	μmol/L	18.6	7.7%	12	324
Roche cobas c 501/502/701/702 LEADMAN \ Enzymatic Cycling Assays Method	μmol/L	18.4	5.2%	10	305
(TBIL)总胆红素					
KHB ZY series KHB \ Vanadate	μmol/L	14.3	10.9%	9	<100
Roche Hitachi/Hitachi Systems MAKER \ sodium nitrite oxidase method	μmol/L	13.5	2.7%	8	162
Ortho Vitros 3600/5600 Ortho \ dry chemistry	μmol/L	17.5	6.6%	6	<100

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000 Wako \ Vanadate	μmol/L	14.6	4.1%	6	<100
Abbott ARCHITECT c series HuaChenBio \ Diazotization	μmol/L	14.6	2%	5	117
Siemens ADVIA 1800/2400 BIOSINO \ Vanadate	μmol/L	18.5	4.1%	42	2188
Siemens ADVIA Chemistry XPT Siemens \ Vanadate	μmol/L	16.5	2.8%	<5	<100
Abbott ARCHITECT c series HuaChenBio \ Vanadate	μmol/L	15.2	1.9%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Diazotization	μmol/L	15.4	2.5%	<5	<100
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	18.4	2.6%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Vanadate	μmol/L	17.2	6.3%	23	629
Roche Hitachi/Hitachi Systems Wako \ Vanadate	μmol/L	14.3	3.7%	22	779
Roche cobas c 501/502/701/702 Roche \ Diazotization -No blank Compensation	μmol/L	12.5	5.8%	22	582
Beckman UniCel DxC 600/800 Beckman Coulter \ Diazotization	μmol/L	19.2	11%	21	398
Beckman AU series Beckman Coulter \ Diazotization	μmol/L	15.4	9.5%	<5	<100
Roche Hitachi/Hitachi Systems KHB \ Diazotization	μmol/L	23.8	18.2%	<5	<100
Beckman AU series Beckman Coulter \ OTHER	μmol/L	16.5	3.9%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Diazotization -Blank Compensation	μmol/L	12.4	12.4%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Diazotization	μmol/L	12.7	3.2%	19	590
(TES)睾酮					
Ortho Vitros 3600/5600 Ortho \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.26	10%	<5	<100
(TG)甘油三酯					
KHB ZY series KHB \ GPO-POD	mmol/L	1.1	5%	9	<100
Beckman AU series OTHER \ GPO-POD	mmol/L	1.1	2%	8	240
Ortho Vitros 3600/5600 Ortho \ dry chemistry	mmol/L	1.25	2.5%	6	<100
Beckman AU series Beckman Coulter \ GPO-POD	mmol/L	0.903	29.3%	6	133
Abbott ARCHITECT c series BIOSINO \ Acetyl - acetone photometry	mmol/L	1.14	1.1%	5	118
Siemens ADVIA 1800/2400 BIOSINO \ GPO-POD	mmol/L	1.07	2%	42	2165
Siemens ADVIA Chemistry XPT Siemens \ GPO-POD	mmol/L	1.14	2.9%	<5	<100
Abbott ARCHITECT c series BIOSINO \ GPO-POD	mmol/L	1.14	1.6%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ GK-PK-LDH	mmol/L	1.1	4%	<5	<100
Beckman AU series OTHER \ Free glycerol removed	mmol/L	1.1	6.4%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ GK-PK-LDH	mmol/L	1.1	5.6%	23	661
Roche Hitachi/Hitachi Systems MedicalSystem \ GPO-POD	mmol/L	1.22	4.9%	22	809
Beckman UniCel DxC 600/800 Beckman Coulter \ OTHER	mmol/L	1.22	2.6%	21	402
Roche Hitachi/Hitachi Systems KHB \ Acetyl - acetone photometry	mmol/L	1.1	5.1%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ GPO-POD	mmol/L	1.16	1.9%	19	581
Mindray BS series Mindray \ GPO-POD	mmol/L	0.961	17.3%	16	142
Roche cobas c 501/502/701/702 Roche \ GPO-POD	mmol/L	1.13	5.8%	15	365
Beckman AU series Reebio \ GPO-POD	mmol/L	1.15	2.4%	14	334
Beckman AU series Autobio \ GPO-POD	mmol/L	1.08	1.6%	14	331
(THEO)茶碱					
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	μg/mL	12.2	10.8%	<5	<100
(UA)尿酸					
KHB ZY series KHB \ Uricase colorimetry	μmol/L	313	2.5%	9	<100
Mindray BS series Mindray \ dry chemistry	μmol/L	294	2%	7	<100
Ortho Vitros 3600/5600 Ortho \ dry chemistry	μmol/L	302	1.4%	6	<100
Mindray BS series Mindray \ Uricase colorimetry	μmol/L	287	0.9%	6	<100
Abbott ARCHITECT c series BIOSINO \ dry chemistry	μmol/L	308	0.9%	5	117
Siemens ADVIA 1800/2400 BIOSINO \ Uricase colorimetry	μmol/L	298	2.8%	42	2164

		批号：376B86F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Siemens ADVIA Chemistry XPT Siemens \ Uricase colorimetry	μmol/L	307	1.2%	<5	<100
Beckman AU series Beckman Coulter \ Uricase colorimetry	μmol/L	273	18.1%	<5	104
Roche cobas c 501/502/701/702 Roche \ Uricase colorimetry	μmol/L	285	2.4%	35	917
Abbott ARCHITECT c series BIOSINO \ Uricase colorimetry	μmol/L	312	0.8%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Phosphotungstate colorimetry	μmol/L	292	1.1%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Phosphotungstate colorimetry	μmol/L	298	2.6%	23	661
Roche Hitachi/Hitachi Systems MedicalSystem \ Uricase colorimetry	μmol/L	303	4%	22	809
Roche Hitachi/Hitachi Systems KHB \ dry chemistry	μmol/L	300	2.1%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Uricase colorimetry	μmol/L	288	1.8%	19	588
Siemens ADVIA Chemistry XPT Dedicated \ Phosphotungstate colorimetry	μmol/L	300	2.4%	17	887
Roche cobas 4000/6000/8000 Roche \ Uricase colorimetry	μmol/L	284	3.8%	17	368
Beckman AU series Reebio \ Uricase colorimetry	μmol/L	277	1.6%	14	338
Beckman AU series Autobio \ Uricase colorimetry	μmol/L	301	1.1%	14	331
(UIBC)不饱和铁结合力					
Siemens ADVIA 1800/2400 DiaSys \ Ferrozine colorimetry	μmol/L	48.2	3%	14	707
Beckman UniCel DxI 600/800 Beckman Coulter \ Ferrozine colorimetry	μmol/L	42.8	10%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Ferrozine colorimetry	μmol/L	43.6	11%	<5	<100
(Urea)尿素					
Beckman AU series OTHER \ Urease colorimetry	mmol/L	4.46	2.5%	8	274
Mindray BS series Mindray \ Urease-UV-rate	mmol/L	4.58	1.2%	6	<100
Roche cobas 4000/6000/8000 OTHER \ Urease-UV-rate	mmol/L	4.55	1.8%	6	<100
Abbott ARCHITECT c series BIOSINO \ Diacetyl monoxime	mmol/L	4.57	2%	5	117
Siemens ADVIA Chemistry XPT Siemens \ Urease-UV-rate	mmol/L	4.81	2.1%	<5	<100
Abbott ARCHITECT c series BIOSINO \ Urease colorimetry	mmol/L	4.62	1.5%	<5	<100
Beckman AU series Beckman Coulter \ Urease colorimetry	mmol/L	4.62	3%	<5	<100
Roche Hitachi/Hitachi Systems BSBE \ Urease colorimetry	mmol/L	4.76	5.8%	<5	<100
Beckman AU series OTHER \ Urease-UV-rate	mmol/L	4.62	2.4%	<5	<100
Abbott ARCHITECT ci series BIOSINO \ Diacetyl monoxime	mmol/L	4.19	4.3%	23	424
Roche Hitachi/Hitachi Systems KHB \ Diacetyl monoxime	mmol/L	4.48	1.9%	<5	<100
Roche cobas c 501/502/701/702 Dedicated \ Urease-UV-rate	mmol/L	4.7	2.8%	19	598
Beckman AU series Dedicated \ Urease colorimetry	mmol/L	4.72	4.4%	14	343
Beckman AU series Autobio \ Urease colorimetry	mmol/L	4.55	2.7%	14	330
Beckman AU series Reebio \ Urease-UV-rate	mmol/L	4.56	1.7%	14	329
Roche cobas 4000/6000/8000 Roche \ Urease colorimetry	mmol/L	5.19	10.7%	11	311
Roche Hitachi/Hitachi Systems MAKER \ Urease-UV-rate	mmol/L	4.95	3.4%	11	294
Siemens ADVIA 1800/2400 BIOSINO \ OTHER	mmol/L	4.49	2.5%	10	522
Siemens ADVIA 1800/2400 Siemens \ Urease-UV-rate	mmol/L	4.37	10.1%	<5	<100
(Zn)锌					
Beckman AU series Beckman Coulter \ ICP-MS	μmol/L	25.3	11.1%	<5	<100
(β2M)β-2-微球蛋白					
Beckman AU series PRB \ Immunonephelometry (INA)	ng/mL	1301	4.7%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	ng/mL	1472	18.1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。