

REF	767-2	6×4mL	EXP	2025-03-22	LOT	492AFB7	更新日期：	2024-05
1 / 1								

		批号：492AFB7			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(CsA)环孢霉素A					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	178	11%	11	301
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	234	10.3%	<5	<100
Siemens Viva-E Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	254	10.2%	<5	<100
(FK-506)他克莫司					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.7	8.8%	11	306
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	9.36	11.8%	<5	<100
Siemens Viva-E Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.2	11.8%	<5	<100
(SRL/ RAPA)西罗莫司/ 雷帕霉素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12	10.1%	9	111
Siemens Viva-E Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.6	9.6%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.2	7.9%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。