

质控品名： 特定蛋白液体非定值质控品

REF	521-2	6×3mL	EXP	2024-04-04	LOT	31B51F8	更新日期：	2024-04
1 / 6								

		批号：31B51F8			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Alb)白蛋白					
Siemens ADVIA Centaur CP Dedicated \ Bromocresol purple	g/L	33.1	1.7%	<5	<100
Roche Hitachi/Hitachi Systems Roche \ Bromocresol purple	g/L	33.2	11.1%	<5	<100
(ASO)抗链球菌溶血素 O 抗体					
Roche Hitachi/Hitachi Systems BSBE \ Immunoturbidimetry (ITA)	IU/mL	224	2%	9	227
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	IU/mL	271	2.3%	9	173
Siemens BN ProSpec Siemens \ Immunonephelometry (INA)	IU/mL	217	2.4%	5	145
Roche cobas c 501/502/701/702 Dedicated \ Immunoturbidimetry (ITA)	IU/mL	276	3.4%	5	144
Beckman Image 800 Beckman Coulter \ Immunonephelometry (INA)	IU/mL	148	8.9%	46	987
Beckman Image 800 OTHER \ Immunonephelometry (INA)	IU/mL	131	5.3%	41	1225
Siemens ADVIA 1800/2400 DiaSys \ Immunoturbidimetry (ITA)	IU/mL	234	7.2%	28	1108
Beckman Image 800 Dedicated \ Immunonephelometry (INA)	IU/mL	136	7.9%	21	550
Mindray BS series Mindray \ Immunonephelometry (INA)	IU/mL	261	11.5%	21	373
Beckman Image 800 Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	146	20.7%	20	436
Beckman AU series HuaChenBio \ Immunonephelometry (INA)	IU/mL	216	1.5%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ Immunoturbidimetry (ITA)	IU/mL	253	1.6%	<5	<100
Mindray BS series Whitman \ Immunoturbidimetry (ITA)	IU/mL	165	0.9%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Immunoturbidimetry (ITA)	IU/mL	252	1.9%	<5	133
TOSHIBA TBA MedicalSystem \ Immunonephelometry (INA)	IU/mL	237	8.4%	18	250
Siemens BN II Siemens \ OTHER	IU/mL	241	6.4%	16	419
Beckman Image 800 Beckman Coulter \ OTHER	IU/mL	126	12.4%	14	307
Abbott ARCHITECT c series BSBE \ OTHER	IU/mL	208	11.7%	11	213
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	239	11.4%	10	225
(C3)补体 C3					
Roche Hitachi/Hitachi Systems BSBE \ Immunoturbidimetry (ITA)	mg/dL	150	3.8%	9	221
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	144	3.8%	9	208
Roche cobas c 501/502/701/702 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	151	3.3%	5	142
Beckman Image 800 OTHER \ Immunonephelometry (INA)	mg/dL	131	4.2%	43	1257
Beckman Image 800 Dedicated \ Immunonephelometry (INA)	mg/dL	131	5.3%	<5	<100
Beckman Image 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	135	4%	34	676
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	141	1.3%	<5	<100
Siemens ADVIA 1800/2400 DiaSys \ Immunoturbidimetry (ITA)	mg/dL	138	3.8%	28	1121
Biostec A15/A25 Biostec \ Immunonephelometry (INA)	mg/dL	137	2.6%	25	681
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	mg/dL	156	6.2%	25	383
Beckman Image 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	130	5.3%	20	462
Siemens ADVIA 1800/2400 Siemens \ Immunoturbidimetry (ITA)	mg/dL	135	2.3%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ Immunoturbidimetry (ITA)	mg/dL	140	2.4%	<5	<100
Roche Hitachi/Hitachi Systems YongHe SUN \ Immunoturbidimetry (ITA)	mg/dL	130	3.1%	<5	<100
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	129	10.1%	18	243
Beckman Image 800 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	139	4.2%	17	495
Siemens BN ProSpec Siemens \ Immunonephelometry (INA)	mg/dL	132	5.1%	16	471
Siemens BN II Siemens \ OTHER	mg/dL	158	8.9%	16	438
TOSHIBA TBA LEADMAN \ Immunoturbidimetry (ITA)	mg/dL	156	6.9%	16	342
(C4)补体 C4					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	39.2	8.4%	9	214
Roche Hitachi/Hitachi Systems BSBE \ Immunoturbidimetry (ITA)	mg/dL	38.5	2.1%	9	210
Roche cobas c 501/502/701/702 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	31.2	4.1%	5	143

REF	521-2	6×3mL	EXP	2024-04-04	LOT	31B51F8	更新日期：	2024-04
2 / 6								

		批号：31B51F8			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman Image 800 OTHER \ Immunonephelometry (INA)	mg/dL	37.2	6%	43	1255
Beckman Image 800 Dedicated \ Immunonephelometry (INA)	mg/dL	37.5	6.2%	<5	<100
Beckman Image 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	35.9	4.9%	34	678
Siemens ADVIA 1800/2400 DiaSys \ Immunoturbidimetry (ITA)	mg/dL	32.3	3.3%	28	1125
Biostec A15/A25 Biostec \ Immunonephelometry (INA)	mg/dL	42.1	5.1%	25	682
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	mg/dL	38.5	3.2%	25	403
Beckman Image 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	35.9	5.9%	20	456
Siemens ADVIA 1800/2400 Siemens \ Immunoturbidimetry (ITA)	mg/dL	36.2	7.3%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ Immunoturbidimetry (ITA)	mg/dL	40.3	1.9%	<5	<100
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	30.8	1.9%	<5	<100
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	34.2	9.7%	18	255
Beckman Image 800 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	37.6	3.9%	17	496
Siemens BN ProSpec Siemens \ Immunonephelometry (INA)	mg/dL	40.9	7.7%	16	470
Siemens BN II Siemens \ OTHER	mg/dL	35.4	7.7%	16	435
Beckman Image 800 Beckman Coulter \ OTHER	mg/dL	37.7	5.7%	14	303
TOSHIBA TBA LEADMAN \ Immunoturbidimetry (ITA)	mg/dL	34.1	11%	14	299
(Cer)铜蓝蛋白					
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	g/L	0.337	11%	9	219
Beckman Image 800 OTHER \ Immunonephelometry (INA)	g/L	0.302	8%	21	612
Siemens BN II Siemens \ OTHER	g/L	0.477	8.5%	16	427
Siemens BN II Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	g/L	0.321	10%	<5	<100
Beckman Image 800 Dedicated \ Immunonephelometry (INA)	g/L	0.295	6.5%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	g/L	0.367	0%	<5	<100
(CRP)C反应蛋白					
Beckman Image 800 OTHER \ Immunonephelometry (INA)	mg/dL	2.24	3.9%	43	1254
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	mg/dL	2.81	4.3%	<5	<100
Roche Hitachi/Hitachi Systems chongqiong zybio \ Immunoturbidimetry (ITA)	mg/dL	2.76	7.2%	<5	<100
Beckman Image 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	2.45	4.8%	34	676
Siemens ADVIA 1800/2400 Siemens \ Immunoturbidimetry (ITA)	mg/dL	2.61	3.6%	<5	109
Beckman Image 800 Dedicated \ Immunonephelometry (INA)	mg/dL	2.39	6.6%	19	525
Siemens BN II Siemens \ OTHER	mg/dL	2.43	8.7%	16	423
TOSHIBA TBA BSBE \ Immunoturbidimetry (ITA)	mg/dL	3.07	9.2%	16	333
Beckman Image 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	2.31	5.4%	12	271
NIHON KOHDEN MEK Series MAKER \ Immunoturbidimetry (ITA)	mg/dL	1.87	2.8%	<5	<100
Beckman AU series KHB \ Immunoturbidimetry (ITA)	mg/dL	2.08	9.6%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	2.84	1.1%	<5	<100
Beckman Image 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	2.09	10.5%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	2.63	12.2%	<5	<100
(Ferritin)铁蛋白					
Roche Hitachi/Hitachi Systems MedicalSystem \ Immunoturbidimetry (ITA)	ng/mL	142	3.1%	<5	<100
Roche Hitachi/Hitachi Systems LEADMAN \ Immunoturbidimetry (ITA)	ng/mL	117	2.7%	17	338
Siemens BN II Siemens \ OTHER	ng/mL	131	8.8%	16	430
(Free Kappa Light Chain)游离轻链κ					
Beckman Image 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	10900	9.8%	14	268
Beckman Image 800 Beckman Coulter \ Immunonephelometry (INA)	mg/L	9798	9.6%	12	238
(Free Lambda Light Chain)游离轻链λ					
Beckman Image 800 Beckman Coulter \ Immunonephelometry (INA)	mg/L	5497	10.5%	26	499

质控品名： 特定蛋白液体非定值质控品

REF	521-2	6×3mL	EXP	2024-04-04	LOT	31B51F8	更新日期：	2024-04
3 / 6								

		批号：31B51F8			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(Hap (HPT))触珠蛋白 (结合珠蛋白)					
Siemens BN II Siemens \ OTHER	mg/L	1432	10.3%	16	437
Siemens BN II Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mg/L	1369	9.9%	<5	<100
(HbA1c)糖化血红蛋白A1c					
Roche cobas c 501/502/701/702 Dedicated \ immunoturbidimetric	%	5.77	1.3%	5	146
(hs-CRP)超敏C反应蛋白					
Abbott ARCHITECT c series BSBE \ OTHER	mg/L	29	8.4%	22	455
Siemens ADVIA Centaur CP OTHER \ Immunoturbidimetry (ITA)	mg/L	27.3	13.5%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/L	25.7	9.6%	12	274
Siemens ADVIA Centaur CP Dedicated \ Immunoturbidimetry (ITA)	mg/L	25	2.2%	<5	<100
Mindray BC-5000 Series Mindray \ Immunonephelometry (INA)	mg/L	27.5	4.1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	mg/L	32.1	0%	<5	<100
(IgA)免疫球蛋白 IgA					
Roche Hitachi/Hitachi Systems BSBE \ Immunoturbidimetry (ITA)	mg/dL	265	2.4%	9	223
Beckman Immage 800 Dedicated \ Immunonephelometry (INA)	mg/dL	228	16.4%	5	<100
Roche cobas c 501/502/701/702 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	239	2.4%	5	141
Beckman Immage 800 OTHER \ Immunonephelometry (INA)	mg/dL	203	4.7%	43	1255
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	226	4.6%	34	675
Siemens ADVIA 1800/2400 DiaSys \ Immunoturbidimetry (ITA)	mg/dL	240	5.9%	28	1126
Biostec A15/A25 Biostec \ Immunonephelometry (INA)	mg/dL	214	5.1%	25	683
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	mg/dL	273	5.8%	24	407
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	217	5.2%	20	458
Siemens ADVIA 1800/2400 Siemens \ Immunoturbidimetry (ITA)	mg/dL	231	1.2%	<5	<100
Mindray BS series MAKER \ Immunoturbidimetry (ITA)	mg/dL	118	24.6%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ Immunoturbidimetry (ITA)	mg/dL	260	1.5%	<5	<100
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	231	1.8%	<5	<100
Beckman Immage 800 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	229	4.9%	17	493
Siemens BN II Siemens \ OTHER	mg/dL	277	8.5%	16	431
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	192	7.5%	16	254
Siemens BN ProSpec Siemens \ Immunonephelometry (INA)	mg/dL	248	8.9%	14	398
Beckman Immage 800 Beckman Coulter \ OTHER	mg/dL	209	5.9%	14	309
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	mg/dL	202	8.5%	10	303
(IgE)免疫球蛋白 IgE					
Roche cobas c 501/502/701/702 MedicalSystem \ Immunoturbidimetry (ITA)	IU/mL	41.9	10.8%	5	148
Siemens BN II Siemens \ OTHER	IU/mL	52.1	8.1%	16	437
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	39.7	9.3%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	38.2	9.9%	<5	<100
(IgG)免疫球蛋白 IgG					
Roche Hitachi/Hitachi Systems BSBE \ Immunoturbidimetry (ITA)	mg/dL	1408	5.6%	9	226
Roche Hitachi/Hitachi Systems MAKER \ Immunoturbidimetry (ITA)	mg/dL	1459	16.2%	7	217
Roche Hitachi/Hitachi Systems DongOu \ Immunoturbidimetry (ITA)	mg/dL	905	2.2%	7	132
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	1303	4.2%	50	1024
Roche cobas c 501/502/701/702 Dedicated \ Immunoturbidimetry (ITA)	mg/dL	1344	2.1%	5	141
Beckman Immage 800 OTHER \ Immunonephelometry (INA)	mg/dL	1213	5.6%	43	1249
Beckman Immage 800 Dedicated \ Immunonephelometry (INA)	mg/dL	1234	5%	<5	<100
Siemens ADVIA 1800/2400 DiaSys \ Immunoturbidimetry (ITA)	mg/dL	1276	3.3%	28	1123
Biostec A15/A25 Biostec \ Immunonephelometry (INA)	mg/dL	1251	6.6%	25	683

质控品名： 特定蛋白液体非定值质控品

REF	521-2	6×3mL	EXP	2024-04-04	LOT	31B51F8	更新日期：	2024-04
4 / 6								

		批号：31B51F8			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Mindray BS series					
Mindray \ Immunoturbidimetry (ITA)	mg/dL	1409	8.5%	23	300
Beckman Image 800					
Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	1203	6.7%	20	460
Siemens ADVIA 1800/2400					
Siemens \ Immunoturbidimetry (ITA)	mg/dL	1235	1.3%	<5	<100
Mindray BS series					
MAKER \ Immunoturbidimetry (ITA)	mg/dL	745	5.5%	<5	<100
Siemens ADVIA Centaur CP					
Dedicated \ Immunoturbidimetry (ITA)	mg/dL	1325	3.6%	<5	<100
Abbott ARCHITECT c series					
BSBE \ OTHER	mg/dL	1399	3.1%	<5	<100
Beckman Image 800					
Dedicated \ Immunoturbidimetry (ITA)	mg/dL	1311	4.5%	17	496
Siemens BN ProSpec					
Siemens \ Immunonephelometry (INA)	mg/dL	1320	4.1%	16	453
Siemens BN II					
Siemens \ OTHER	mg/dL	1409	8.2%	16	437
TOSHIBA TBA					
MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	1208	14.2%	15	163
(IgM)免疫球蛋白 IgM					
Roche Hitachi/Hitachi Systems					
BSBE \ Immunoturbidimetry (ITA)	mg/dL	176	4.2%	9	225
Roche cobas 4000/6000/8000					
Roche \ Immunonephelometry (INA)	mg/dL	162	9.4%	9	213
Siemens BN ProSpec					
Siemens \ Immunonephelometry (INA)	mg/dL	203	3.1%	8	223
Roche cobas c 501/502/701/702					
Dedicated \ Immunoturbidimetry (ITA)	mg/dL	164	2.8%	5	141
Beckman Image 800					
OTHER \ Immunonephelometry (INA)	mg/dL	165	3.4%	43	1252
Beckman Image 800					
Beckman Coulter \ Immunonephelometry (INA)	mg/dL	165	3%	34	679
Beckman Image 800					
Dedicated \ Immunonephelometry (INA)	mg/dL	168	6.6%	<5	<100
Siemens ADVIA 1800/2400					
DiaSys \ Immunoturbidimetry (ITA)	mg/dL	160	2.3%	28	1117
Biostec A15/A25					
Biostec \ Immunonephelometry (INA)	mg/dL	128	6.9%	25	683
Mindray BS series					
Mindray \ Immunoturbidimetry (ITA)	mg/dL	156	6.9%	23	383
Beckman Image 800					
Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	162	6.6%	20	460
Siemens ADVIA 1800/2400					
Siemens \ Immunoturbidimetry (ITA)	mg/dL	170	1.6%	<5	<100
Siemens ADVIA Centaur CP					
Dedicated \ Immunoturbidimetry (ITA)	mg/dL	172	1.7%	<5	<100
Mindray BS series					
MAKER \ Immunoturbidimetry (ITA)	mg/dL	78.2	0.7%	<5	<100
Abbott ARCHITECT c series					
BSBE \ OTHER	mg/dL	196	1.3%	<5	<100
Beckman Image 800					
Dedicated \ Immunoturbidimetry (ITA)	mg/dL	170	2.7%	17	497
Siemens BN II					
Siemens \ OTHER	mg/dL	184	8%	16	438
TOSHIBA TBA					
MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	143	5.4%	15	217
Beckman Image 800					
Beckman Coulter \ OTHER	mg/dL	165	4.7%	14	306
(Kappa Light Chain)κ轻链抗体					
Beckman Image 800					
OTHER \ Immunonephelometry (INA)	g/L	9.39	5.7%	43	1255
Beckman Image 800					
Dedicated \ Immunonephelometry (INA)	g/L	9.56	7.2%	<5	<100
Beckman Image 800					
Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	9.34	10.3%	<5	<100
Siemens BN II					
Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	9.34	10.5%	<5	<100
(Lambda Light Chain)λ 轻链抗体					
Beckman Image 800					
OTHER \ Immunonephelometry (INA)	g/L	4.61	5.4%	43	1255
Beckman Image 800					
Dedicated \ Immunonephelometry (INA)	g/L	4.57	4.6%	<5	<100
Siemens BN II					
Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	4.32	11.1%	<5	<100
Beckman Image 800					
Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	4.78	10.7%	<5	<100
(MA)微量白蛋白					
Roche cobas c 501/502/701/702					
Roche \ Immunoturbidimetry (ITA)	mg/L	34.9	8%	5	142
(PAB)前白蛋白					
Roche Hitachi/Hitachi Systems					
CONDOR-TECO \ Immunoturbidimetry (ITA)	g/L	0.333	9.6%	6	174
Roche cobas 4000/6000/8000					
Roche \ Immunonephelometry (INA)	g/L	0.317	2.8%	5	<100
Roche Hitachi/Hitachi Systems					
CONDOR-TECO \ Immunoturbidimetry (ITA)	%	36.6	6.6%	5	134
Roche Hitachi/Hitachi Systems					
BSBE \ Immunoturbidimetry (ITA)	g/L	0.327	1.8%	34	629

质控品名： 特定蛋白液体非定值质控品

REF	521-2	6×3mL	EXP	2024-04-04	LOT	31B51F8	更新日期：	2024-04
5 / 6								

		批号：31B51F8			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Roche Hitachi/Hitachi Systems MedicalSystem \ Immunoturbidimetry (ITA)	g/L	0.315	5%	<5	<100
Siemens ADVIA 1800/2400 Szybio \ Immunoturbidimetry (ITA)	g/L	0.315	2.9%	<5	115
Beckman AU series HuaChenBio \ Immunonephelometry (INA)	g/L	0.323	3.7%	14	<100
Beckman AU series KINGSBIO \ Immunoturbidimetry (ITA)	g/L	0.318	2.2%	13	369
Siemens ADVIA 1800/2400 Siemens \ Immunoturbidimetry (ITA)	g/L	0.327	4.5%	<5	<100
Beckman Image 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	0.377	9%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	g/L	0.265	10.6%	<5	<100
NIHON KOHDEN MEK Series MAKER \ Immunoturbidimetry (ITA)	g/L	0.294	2.6%	<5	<100
(RBP)视黄醇结合蛋白					
Siemens ADVIA 1800/2400 BIOSINO \ Immunoturbidimetry (ITA)	mg/dL	4.9	11.8%	30	1235
Roche Hitachi/Hitachi Systems MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	4.4	4.2%	<5	<100
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	4.43	10.4%	<5	<100
Beckman AU series chongqiong zybio \ Immunoturbidimetry (ITA)	mg/dL	4.73	2.3%	<5	<100
(RF)类风湿因子					
Roche Hitachi/Hitachi Systems BSBE \ Immunoturbidimetry (ITA)	IU/mL	85	1.9%	9	226
Beckman Image 800 Beckman Coulter \ Immunonephelometry (INA)	IU/mL	59.6	7.9%	50	1032
Beckman AU series HuaChenBio \ Immunonephelometry (INA)	IU/mL	67.3	1.5%	5	<100
Roche cobas c 501/502/701/702 Dedicated \ Immunoturbidimetry (ITA)	IU/mL	65.7	1.8%	5	141
Beckman Image 800 OTHER \ Immunonephelometry (INA)	IU/mL	57.8	7.3%	41	1224
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	IU/mL	81.1	2.5%	<5	<100
Abbott ARCHITECT c series BSBE \ OTHER	IU/mL	97.1	5.2%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ Immunoturbidimetry (ITA)	IU/mL	58.8	9.1%	<5	<100
Siemens ADVIA 1800/2400 DiaSys \ Immunoturbidimetry (ITA)	IU/mL	60.3	4.5%	28	1117
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	IU/mL	55.6	15.3%	25	310
Beckman Image 800 Dedicated \ Immunonephelometry (INA)	IU/mL	60	4%	21	549
Roche cobas c 501/502/701/702 Roche \ Immunonephelometry (INA)	IU/mL	32.5	24.8%	<5	<100
Mindray BS series Whitman \ Immunoturbidimetry (ITA)	IU/mL	38.2	1.1%	<5	<100
Siemens ADVIA Centaur CP Dedicated \ Immunoturbidimetry (ITA)	IU/mL	74.5	1.8%	<5	<100
Siemens ADVIA 1800/2400 Siemens \ Immunoturbidimetry (ITA)	IU/mL	71.1	1.5%	<5	134
Siemens BN ProSpec Siemens \ Immunonephelometry (INA)	IU/mL	41.9	4.1%	16	459
Siemens BN II Siemens \ OTHER	IU/mL	56	8.7%	16	423
Beckman Image 800 Beckman Coulter \ OTHER	IU/mL	58.5	7.4%	14	296
TOSHIBA TBA BSBE \ Immunoturbidimetry (ITA)	IU/mL	90.9	5.4%	14	293
(TRF)转铁蛋白					
Roche cobas c 501/502/701/702 MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	171	9.4%	5	142
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	289	1%	<5	<100
Roche Hitachi/Hitachi Systems LEADMAN \ Immunoturbidimetry (ITA)	mg/dL	267	2.8%	17	338
Siemens BN II Siemens \ OTHER	mg/dL	263	10.6%	16	439
Siemens ADVIA 1800/2400 DiaSys \ Immunoturbidimetry (ITA)	mg/dL	246	5.4%	14	554
Mindray BC-2000 series Mindray \ Immunoturbidimetry (ITA)	mg/dL	254	1%	<5	<100
Beckman Image 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	235	9.8%	<5	<100
(α-1-AG)α-1-酸性糖蛋白					
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	g/L	0.865	9.9%	<5	<100
(α-1-AT)α-1-抗胰蛋白酶					
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	123	11.4%	<5	<100
(α-1-G)α-1-球蛋白					
TOSHIBA TBA LEADMAN \ Capillary Method	%	22.2	2.2%	<5	103

REF	521-2	6×3mL	EXP	2024-04-04	LOT	31B51F8	更新日期：	2024-04
6 / 6								

		批号: 31B51F8			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(α1M)α1微球蛋白					
Roche cobas c 501/502/701/702 Roche \ immunoturbidimetric	mg/L	22.3	15.4%	<5	<100
(β-2-G)β-2-球蛋白					
TOSHIBA TBA LEADMAN \ Capillary Method	%	3.94	4.3%	6	110
(β2M)β-2-微球蛋白					
Roche Hitachi/Hitachi Systems MedicalSystem \ Immunoturbidimetry (ITA)	ng/mL	3390	3.3%	9	239
Roche Hitachi/Hitachi Systems BSBE \ Immunoturbidimetry (ITA)	ng/mL	3390	2.7%	34	629
Abbott ARCHITECT c series BSBE \ OTHER	ng/mL	3654	5%	22	454
Mindray BS series Mindray \ OTHER	ng/mL	3904	5.5%	20	332
TOSHIBA TBA MedicalSystem \ OTHER	ng/mL	3471	9.1%	18	255
Beckman AU series chongqiong zybion \ Immunoturbidimetry (ITA)	ng/mL	3556	3%	14	391
TOSHIBA TBA OTHER \ Immunoturbidimetry (ITA)	ng/mL	3966	4.5%	12	234
NIHON KOHDEN MEK Series MAKER \ Immunoturbidimetry (ITA)	ng/mL	3809	1.9%	<5	<100
Beckman AU series HuaChenBio \ Immunonephelometry (INA)	ng/mL	4025	1.1%	<5	<100
Beckman Image 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4300	14%	<5	<100
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3600	11.1%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	ng/mL	3655	0%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。