报告类别： **汇总实验室数据报告**

质控品名： **生化与激素非定值质控品**

| REF | 581-2 | 6×5mL | EXP | 2025-10-06 | LOT | 376DFE3 | 更新日期： | 2024-05 |
|---|---|---|---|---|---|---|---|---|



| | | 批号：376DFE3 | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| (Alb)白蛋白 | | | | | |
| Mindray BS series<br>chongqiong zybio \ Bromocresol green | g/L | 44.7 | 2.3% | 9 | 418 |
| Beckman AU series<br>BIOSINO \ Bromocresol green | g/L | 45.6 | 1.2% | 9 | 412 |
| Beckman AU series<br>Beckman Coulter \ Bromocresol green | g/L | 44.7 | 3.1% | <5 | 132 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Bromocresol green | g/L | 46.5 | 2.1% | 21 | 914 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Bromocresol green | g/L | 46.5 | 0.8% | <5 | <100 |
| Mindray BS series<br>Dedicated \ Bromocresol green | g/L | 43.8 | 1.9% | 15 | 504 |
| Beckman AU series<br>OTHER \ Bromocresol green | g/L | 45.2 | 1.9% | 13 | 1234 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | g/L | 47.7 | 0.6% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Bromocresol green | g/L | 49 | 11.9% | <5 | 100 |
| (ALP)碱性磷酸酶 | | | | | |
| Beckman AU series<br>MAKER \ Rate -AMP buffer | U/L | 510 | 3% | 9 | 422 |
| Mindray BS series<br>Elikan \ Rate -other buffers | U/L | 487 | 3.5% | 6 | 274 |
| Mindray BS series<br>chongqiong zybio \ Rate -other buffers | U/L | 481 | 3.1% | <5 | 144 |
| Beckman AU series<br>Beckman Coulter \ Rate -AMP buffer | U/L | 501 | 1.3% | <5 | 136 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Rate -AMP buffer | U/L | 475 | 5.9% | 21 | 922 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Rate -AMP buffer | U/L | 536 | 3% | <5 | <100 |
| Mindray BS series<br>Dedicated \ Rate -AMP buffer | U/L | 499 | 2.1% | 15 | 516 |
| Beckman AU series<br>OTHER \ Rate -AMP buffer | U/L | 492 | 2.2% | 13 | 1038 |
| Roche cobas 4000/6000/8000<br>Roche \ Rate -AMP buffer | U/L | 440 | 11.9% | <5 | <100 |
| (ALT/ SGPT)丙氨酸氨基转移酶 | | | | | |
| Beckman AU series<br>BIOSINO \ UV-rate without P5P | U/L | 174 | 3.1% | 9 | 450 |
| Mindray BS series<br>Elikan \ UV-rate without P5P | U/L | 185 | 3.9% | 6 | 270 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ OTHER | U/L | 178 | 3% | 5 | 232 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ OTHER | U/L | 165 | 2.2% | 44 | 2426 |
| Mindray BS series<br>chongqiong zybio \ UV-rate without P5P | U/L | 175 | 1.1% | <5 | 146 |
| Beckman AU series<br>Beckman Coulter \ UV-rate without P5P | U/L | 185 | 2.5% | <5 | 120 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ UV-rate without P5P | U/L | 176 | 5.2% | 16 | 624 |
| Mindray BS series<br>Dedicated \ UV-rate without P5P | U/L | 174 | 3.8% | 14 | 412 |
| Beckman AU series<br>OTHER \ UV-rate without P5P | U/L | 175 | 3.7% | 13 | 1212 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | U/L | 171 | 9.9% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ UV-rate with P5P | U/L | 165 | 11.8% | <5 | <100 |
| (AMY)淀粉酶 | | | | | |
| Mindray BS series<br>Elikan \ OTHER | U/L | 858 | 2.9% | 9 | 416 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ PNP-G7 | U/L | 880 | 3.3% | 7 | 290 |
| Beckman AU series<br>MAKER \ PNP-G7 | U/L | 806 | 1.4% | 5 | 128 |
| Roche Hitachi/Hitachi Systems<br>Wako \ CNP_G3 | U/L | 949 | 0.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ OTHER | U/L | 819 | 1.2% | <5 | <100 |
| Mindray BS series<br>Dedicated \ PNP-G7 | U/L | 817 | 2.9% | 15 | 596 |
| Beckman AU series<br>Elikan \ PNP-G7 | U/L | 827 | 1.7% | 14 | 1270 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | U/L | 412 | 0.4% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ PNP-G7 | U/L | 774 | 11.8% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ PNP-G7 | U/L | 987 | 0.7% | <5 | <100 |
| (Apo-AⅠ)载脂蛋白AⅠ | | | | | |
| Beckman AU series<br>KINGSBIO \ immunoturbidimetric | g/L | 2.2 | 5.7% | 9 | 442 |
| Mindray BS series<br>OTHER \ immunoturbidimetric | g/L | 2.12 | 3.4% | 9 | 416 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ immunoturbidimetric | g/L | 3.03 | 3.2% | <5 | 134 |

| | | 批号：376DFE3 | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ immunoturbidimetric | g/L | 2.41 | 2.6% | 15 | 728 |
| Mindray BS series<br>Dedicated \ immunoturbidimetric | g/L | 2.29 | 4.5% | 15 | 484 |
| Roche cobas 4000/6000/8000<br>SekiSui \ immunoturbidimetric | g/L | 1.72 | 11.9% | <5 | <100 |
| (Apo-B)载脂蛋白B | | | | | |
| Mindray BS series<br>OTHER \ immunoturbidimetric | g/L | 1.14 | 4.4% | 9 | 416 |
| Beckman AU series<br>KINGSBIO \ immunoturbidimetric | g/L | 0.942 | 6.5% | 8 | 398 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ immunoturbidimetric | g/L | 0.982 | 2.9% | 15 | 724 |
| Mindray BS series<br>Dedicated \ immunoturbidimetric | g/L | 1.17 | 4.9% | 14 | 524 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ immunoturbidimetric | g/L | 1 | 3.5% | 11 | 530 |
| Roche cobas 4000/6000/8000<br>SekiSui \ immunoturbidimetric | g/L | 1.01 | 11.9% | <5 | <100 |
| (AST/ SGOT)天门冬氨酸氨基转移酶 | | | | | |
| Beckman AU series<br>BIOSINO \ UV-rate without P5P | U/L | 239 | 2.6% | 9 | 444 |
| Mindray BS series<br>Elikan \ UV-rate without P5P | U/L | 237 | 4.1% | 6 | 272 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ OTHER | U/L | 221 | 4.1% | 5 | 222 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ OTHER | U/L | 229 | 3.9% | 30 | 1444 |
| Mindray BS series<br>chongqiong zybio \ UV-rate without P5P | U/L | 230 | 1.8% | <5 | 146 |
| Beckman AU series<br>Beckman Coulter \ UV-rate without P5P | U/L | 249 | 5.3% | <5 | 120 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ UV-rate without P5P | U/L | 228 | 4.2% | 16 | 648 |
| Mindray BS series<br>Dedicated \ UV-rate without P5P | U/L | 237 | 4.4% | 15 | 582 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ UV-rate with P5P | U/L | 229 | 3.5% | 14 | 912 |
| Beckman AU series<br>OTHER \ UV-rate without P5P | U/L | 238 | 5.4% | 13 | 1250 |
| Roche cobas 4000/6000/8000<br>Roche \ UV-rate with P5P | U/L | 225 | 11.9% | <5 | <100 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | U/L | 258 | 4.2% | <5 | <100 |
| (Ca)钙 | | | | | |
| Beckman AU series<br>MAKER \ Arsenazo III | mmol/L | 2.73 | 2% | 7 | 310 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Arsenazo III | mmol/L | 2.87 | 2.7% | 5 | 238 |
| Roche Hitachi/Hitachi Systems<br>Health World \ Arsenazo III | mmol/L | 2.78 | 1.6% | <5 | 112 |
| Roche Hitachi/Hitachi Systems<br>Wako \ OTHER | mmol/L | 2.77 | 1.3% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ Arsenazo III | mmol/L | 2.9 | 9.7% | <5 | <100 |
| Mindray BS series<br>Dedicated \ Arsenazo III | mmol/L | 2.84 | 3.7% | 15 | 516 |
| Beckman AU series<br>OTHER \ Arsenazo III | mmol/L | 2.78 | 1.8% | 13 | 1098 |
| Roche cobas 4000/6000/8000<br>Roche \ BAPTA | mmol/L | 2.74 | 11.9% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ ISE (direct) | mmol/L | 101 | 1.7% | <5 | <100 |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Arsenazo III | mmol/L | 2.64 | 1.8% | <5 | <100 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | mmol/L | 2.76 | 0.2% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ OTHER | mmol/L | 2.8 | 1.8% | <5 | <100 |
| (CHOL/ TC)总胆固醇 | | | | | |
| Beckman AU series<br>LEADMAN \ Cholesterol oxidase | mmol/L | 5.33 | 2.4% | 9 | 442 |
| Mindray BS series<br>Elikan \ Cholesterol oxidase | mmol/L | 5.21 | 2.1% | 6 | 270 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Cholesterol oxidase | mmol/L | 5.56 | 1% | 5 | 190 |
| Mindray BS series<br>OTHER \ Cholesterol oxidase | mmol/L | 5.28 | 1.4% | <5 | 146 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Cholesterol oxidase | mmol/L | 5.43 | 1.6% | 20 | 1280 |
| Mindray BS series<br>Dedicated \ Cholesterol oxidase | mmol/L | 5.21 | 2% | 15 | 544 |
| Beckman AU series<br>OTHER \ Cholesterol oxidase | mmol/L | 5.31 | 2.4% | 14 | 1320 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ OTHER | mmol/L | 5.56 | 2.1% | 11 | 566 |
| Roche cobas 4000/6000/8000<br>Roche \ Cholesterol oxidase | mmol/L | 5.26 | 11.9% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ Cholesterol oxidase | mmol/L | 5.12 | 2.5% | <5 | <100 |

| | | 批号：376DFE3 | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | mmol/L | 1.72 | 0.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ OTHER | mmol/L | 5.54 | 1.6% | <5 | <100 |
| (CK)肌酸激酶 | | | | | |
| Roche Hitachi/Hitachi Systems<br>MAKER \ OTHER | U/L | 645 | 3.2% | 8 | 318 |
| Mindray BS series<br>Elikan \ Rate -other buffers | U/L | 664 | 3.4% | 6 | 268 |
| Beckman AU series<br>LEADMAN \ Rate -AMP buffer | U/L | 641 | 1.7% | 6 | 196 |
| Mindray BS series<br>chongqiong zybio \ Rate -other buffers | U/L | 674 | 1.8% | <5 | 146 |
| Beckman AU series<br>OTHER \ Rate -other buffers | U/L | 692 | 6.2% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Rate -AMP buffer | U/L | 636 | 8.5% | 16 | 394 |
| Mindray BS series<br>Dedicated \ Rate -AMP buffer | U/L | 637 | 5.1% | 15 | 502 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | U/L | 498 | 2.5% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Rate -AMP buffer | U/L | 658 | 11.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ Rate -AMP buffer | U/L | 702 | 0.9% | <5 | <100 |
| (CK-MB mass)肌酸激酶-MB质量 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 21.4 | 12% | <5 | <100 |
| (CK-MB)肌酸激酶-MB (活性) | | | | | |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Immune suppression | U/L | 45.1 | 4.2% | 15 | 772 |
| Mindray BS series<br>Dedicated \ Immune suppression | U/L | 50.7 | 8.3% | 15 | 564 |
| Roche cobas 4000/6000/8000<br>Roche \ Immune suppression | U/L | 49.5 | 11.9% | <5 | <100 |
| (Cl)氯 | | | | | |
| Roche Hitachi/Hitachi Systems<br>Hitachi \ ISE (indirect) | mmol/L | 118 | 1.1% | 6 | 192 |
| Beckman AU series<br>Beckman Coulter \ ISE (indirect) | mmol/L | 114 | 1.4% | 23 | 1618 |
| Roche Hitachi/Hitachi Systems<br>Dedicated \ ISE (direct) | mmol/L | 127 | 6.1% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>Wako \ ISE (indirect) | mmol/L | 115 | 1% | 15 | 724 |
| Mindray BS series<br>Dedicated \ OTHER | mmol/L | 116 | 1.6% | 15 | 556 |
| Beckman AU series<br>Beckman Coulter \ ISE (direct) | mmol/L | 101 | 1.4% | <5 | <100 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | mmol/L | 119 | 1.5% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ ISE (indirect) | mmol/L | 112 | 11.9% | <5 | <100 |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ ISE (indirect) | mmol/L | 117 | 3.7% | <5 | <100 |
| (CO2（HCO3-）)二氧化碳（碳酸氢根） | | | | | |
| Roche Hitachi/Hitachi Systems<br>MAKER \ PEPC | mmol/L | 7.3 | 11.6% | 5 | 206 |
| Mindray BS series<br>Dedicated \ PEPC | mmol/L | 8.2 | 11.7% | 15 | 562 |
| Roche cobas c 501/502/701/702<br>Roche \ Colorimetry | mmol/L | 9.8 | 11.9% | <5 | <100 |
| (Cr)肌酐 | | | | | |
| Beckman AU series<br>MAKER \ Enzymatic | μmol/L | 461 | 2.6% | 8 | 340 |
| Mindray BS series<br>simes-sikma \ OTHER | μmol/L | 474 | 1.9% | 7 | 284 |
| Mindray BS series<br>MAKER \ Enzymatic | μmol/L | 507 | 1.9% | <5 | 134 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ OTHER | μmol/L | 465 | 1.3% | 29 | 1576 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Enzymatic | μmol/L | 465 | 1.8% | 15 | 720 |
| Mindray BS series<br>Dedicated \ Picric acid-Blank Compensation | μmol/L | 467 | 2.4% | 15 | 476 |
| Beckman AU series<br>OTHER \ Enzymatic | μmol/L | 471 | 1.9% | 14 | 1186 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ OTHER | μmol/L | 465 | 1.4% | 11 | 542 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Enzymatic | μmol/L | 492 | 2.1% | 10 | 420 |
| Beckman AU series<br>Beckman Coulter \ Enzymatic | μmol/L | 424 | 0.7% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Enzymatic | μmol/L | 476 | 11.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ OTHER | μmol/L | 464 | 1.1% | <5 | <100 |

| | | 批号：376DFE3 | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| (CRP)C反应蛋白 | | | | | |
| Beckman AU series<br>DiaSys \ Immunoturbidimetry (ITA) | mg/dL | 0.423 | 16.6% | <5 | <100 |
| Beckman Immage 800<br>Beckman Coulter \ Immunonephelometry (INA) | mg/dL | 0.29 | 12% | <5 | <100 |
| (DBIL)直接胆红素 | | | | | |
| Mindray BS series<br>Elikan \ Vanadate | μmol/L | 45.9 | 4.3% | 9 | 420 |
| Beckman AU series<br>MAKER \ Vanadate | μmol/L | 45.8 | 3.9% | 9 | 408 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ sodium nitrite oxidase method | μmol/L | 48.4 | 3.3% | 5 | 182 |
| Beckman AU series<br>Beckman Coulter \ Diazotization -Blank Compensation | μmol/L | 29.1 | 4.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Vanadate | μmol/L | 41.8 | 5.4% | 16 | 742 |
| Mindray BS series<br>Dedicated \ Vanadate | μmol/L | 42.3 | 4.3% | 15 | 608 |
| Beckman AU series<br>OTHER \ Vanadate | μmol/L | 44.6 | 2.7% | 13 | 1096 |
| Beckman AU series<br>Beckman Coulter \ Diazotization | μmol/L | 35.6 | 3% | <5 | <100 |
| OTHER<br>BSBE \ OTHER | μmol/L | 41.1 | 2.4% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Diazotization -Blank Compensation | μmol/L | 41.2 | 11.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Vanadate | μmol/L | 40.4 | 3.3% | <5 | <100 |
| (DGXN)地高辛 | | | | | |
| Roche cobas e411<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 0.311 | 12% | <5 | <100 |
| (E2)雌二醇 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | pmol/L | 235 | 12% | <5 | <100 |
| (FA)叶酸 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 6.16 | 12% | <5 | <100 |
| (Fe)铁 | | | | | |
| Roche Hitachi/Hitachi Systems<br>Wako \ Bathophenanthroline | μmol/L | 33.7 | 1.7% | 5 | 180 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Ferrozine colorimetry | μmol/L | 30.1 | 1.6% | 15 | 734 |
| Mindray BS series<br>Dedicated \ OTHER | μmol/L | 30.7 | 3.4% | 15 | 516 |
| Beckman AU series<br>Beckman Coulter \ Ferrozine colorimetry | μmol/L | 25.6 | 2% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Ferrozine colorimetry | μmol/L | 31 | 11.9% | <5 | 100 |
| (FT3)游离T3 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | pmol/L | 4.16 | 12% | <5 | <100 |
| (FT4 )游离T4 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | pmol/L | 17.2 | 12% | <5 | <100 |
| (GGT)L-γ-谷氨酰基转移酶 | | | | | |
| Beckman AU series<br>MAKER \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 201 | 2.2% | 9 | 452 |
| Mindray BS series<br>Elikan \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 195 | 1.9% | 6 | 276 |
| Beckman AU series<br>Beckman Coulter \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 189 | 6% | <5 | 160 |
| Mindray BS series<br>chongqiong zybio \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 195 | 4.3% | <5 | 146 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 202 | 2.5% | 21 | 994 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 187 | 1.5% | <5 | <100 |
| Mindray BS series<br>Dedicated \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 199 | 1.9% | 15 | 586 |
| Beckman AU series<br>OTHER \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 196 | 1.4% | 13 | 1248 |
| Roche cobas 4000/6000/8000<br>Roche \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 196 | 11.9% | <5 | <100 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | U/L | 294 | 0.2% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ γ-glutamyl -3-carboxy-4-nitroaniline | U/L | 205 | 0.8% | <5 | <100 |
| (GLDH)谷氨酸脱氢酶 | | | | | |
| Roche cobas 4000/6000/8000<br>DiaSys \ Colorimetry | U/L | 8.65 | 12% | <5 | <100 |
| (GLU)葡萄糖 | | | | | |
| Beckman AU series<br>BJBJ \ Glucose oxidase (GOD) | mmol/L | 14.6 | 2.1% | 9 | 446 |

| | | 批号：376DFE3 | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| Mindray BS series<br>Elikan \ Glucose oxidase (GOD) | mmol/L | 14.7 | 2% | 6 | 272 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ OTHER | mmol/L | 14.8 | 2.3% | 30 | 1432 |
| Mindray BS series<br>OTHER \ Glucose oxidase (GOD) | mmol/L | 14.1 | 1.2% | <5 | 146 |
| Beckman AU series<br>Beckman Coulter \ Hexokinase | mmol/L | 15.2 | 1.8% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Hexokinase | mmol/L | 14.8 | 2.4% | 16 | 936 |
| Mindray BS series<br>Dedicated \ Glucose oxidase (GOD) | mmol/L | 14.1 | 3.5% | 15 | 528 |
| Beckman AU series<br>OTHER \ Hexokinase | mmol/L | 14.6 | 2.7% | 13 | 1156 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Glucose oxidase (GOD) | mmol/L | 14.4 | 1.9% | 11 | 578 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Hexokinase | mmol/L | 14.3 | 4.3% | 10 | 424 |
| Roche cobas 4000/6000/8000<br>Roche \ Hexokinase | mmol/L | 14.7 | 11.9% | <5 | <100 |
| OTHER<br>BSBE \ Hexokinase | mmol/L | 14.1 | 1.7% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ Glucose oxidase (GOD) | mmol/L | 15 | 1.7% | <5 | <100 |
| (HDL-C)高密度脂蛋白胆固醇 | | | | | |
| Beckman AU series<br>LEADMAN \ selected suppression | mmol/L | 1.16 | 4.4% | 9 | 438 |
| Mindray BS series<br>OTHER \ selected suppression | mmol/L | 1.46 | 3% | 9 | 414 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ selected suppression | mmol/L | 1.32 | 4.5% | 44 | 2292 |
| Beckman AU series<br>Elikan \ Direct Measure | mmol/L | 1.3 | 2.5% | <5 | 182 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Direct Measure | mmol/L | 1.25 | 5.3% | 16 | 620 |
| Mindray BS series<br>Dedicated \ Direct Measure | mmol/L | 1.22 | 3.9% | 15 | 514 |
| Beckman AU series<br>OTHER \ Direct Measure | mmol/L | 1.48 | 5.4% | 11 | 868 |
| Roche cobas 4000/6000/8000<br>Roche \ Enzymatic | mmol/L | 1.04 | 11.9% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ Enzymatic | mmol/L | 1.55 | 3.7% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ Direct Measure | mmol/L | 1.26 | 2% | <5 | <100 |
| (IgA )免疫球蛋白 IgA | | | | | |
| Beckman Immage 800<br>Beckman Coulter \ Immunoturbidimetry (ITA) | mg/dL | 249 | 12% | <5 | <100 |
| (IgG)免疫球蛋白 IgG | | | | | |
| Beckman Immage 800<br>Beckman Coulter \ Immunoturbidimetry (ITA) | mg/dL | 1040 | 12% | <5 | <100 |
| (IgM)免疫球蛋白 IgM | | | | | |
| Beckman Immage 800<br>Beckman Coulter \ Immunoturbidimetry (ITA) | mg/dL | 110 | 12% | <5 | <100 |
| (K)钾 | | | | | |
| Roche Hitachi/Hitachi Systems<br>Hitachi \ ISE (indirect) | mmol/L | 7.3 | 1.4% | 5 | 174 |
| Roche Hitachi/Hitachi Systems<br>Dedicated \ ISE (direct) | mmol/L | 6.96 | 3.5% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ ISE (indirect) | mmol/L | 7.05 | 1.9% | 23 | 1656 |
| Roche Hitachi/Hitachi Systems<br>Wako \ ISE (indirect) | mmol/L | 7.11 | 0.9% | 15 | 720 |
| Mindray BS series<br>Dedicated \ OTHER | mmol/L | 7.21 | 1.8% | 15 | 560 |
| Beckman AU series<br>Beckman Coulter \ ISE (direct) | mmol/L | 1.57 | 1.4% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ ISE (indirect) | mmol/L | 7.1 | 11.9% | <5 | <100 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | mmol/L | 7.21 | 0.1% | <5 | <100 |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ ISE (indirect) | mmol/L | 6.94 | 1.1% | <5 | <100 |
| (LA)乳酸 | | | | | |
| Roche cobas 4000/6000/8000<br>Roche \ Lactate to pyruvate (L-P) | mmol/L | 7.15 | 12% | <5 | <100 |
| (LDH)乳酸脱氢酶 | | | | | |
| Beckman AU series<br>LEADMAN \ Lactate to pyruvate (L-P) | U/L | 470 | 2% | 9 | 420 |
| Mindray BS series<br>Elikan \ Lactate to pyruvate (L-P) | U/L | 475 | 2.7% | 6 | 274 |
| Beckman AU series<br>OTHER \ Lactate to pyruvate (L-P) | U/L | 479 | 4.4% | <5 | 218 |
| Mindray BS series<br>chongqiong zybio \ Lactate to pyruvate (L-P) | U/L | 483 | 1.7% | <5 | 146 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Lactate to pyruvate (L-P) | U/L | 449 | 4% | 21 | 906 |

| | | 批号：376DFE3 | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| OTHER<br>BSBE \ Lactate to pyruvate (L-P) | U/L | 440 | 1.8% | <5 | <100 |
| Mindray BS series<br>Dedicated \ Lactate to pyruvate (L-P) | U/L | 479 | 2.4% | 15 | 570 |
| Beckman AU series<br>Beckman Coulter \ Lactate to pyruvate (L-P) | U/L | 455 | 0.3% | <5 | <100 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | U/L | 485 | 0.2% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Lactate to pyruvate (L-P) | U/L | 484 | 11.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ OTHER | U/L | 475 | 0.8% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ Lactate to pyruvate (L-P) | U/L | 445 | 1.7% | <5 | <100 |
| (LDL-C)低密度脂蛋白胆固醇 | | | | | |
| Mindray BS series<br>OTHER \ SUR | mmol/L | 3.07 | 2.3% | 9 | 420 |
| Beckman AU series<br>LEADMAN \ selected suppression | mmol/L | 3.19 | 1.9% | 9 | 414 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ SUR | mmol/L | 3.16 | 2.4% | 44 | 2336 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Direct Measure | mmol/L | 3.2 | 4.1% | 16 | 716 |
| Mindray BS series<br>Dedicated \ Direct Measure | mmol/L | 3.3 | 3.9% | 15 | 568 |
| Beckman AU series<br>OTHER \ Direct Measure | mmol/L | 3.47 | 2.6% | 14 | 1188 |
| Beckman AU series<br>Beckman Coulter \ Enzymatic | mmol/L | 3.23 | 1.3% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Enzymatic | mmol/L | 3.25 | 11.9% | <5 | <100 |
| OTHER<br>BSBE \ SUR | mmol/L | 2.72 | 0.7% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ Direct Measure | mmol/L | 3.28 | 2.1% | <5 | <100 |
| (LIP)脂肪酶 | | | | | |
| Mindray BS series<br>Autobio \ OTHER | U/L | 89.1 | 4.3% | 9 | 211 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ OTHER | U/L | 107 | 12.3% | 9 | 171 |
| Beckman AU series<br>OTHER \ OTHER | U/L | 104 | 5.3% | 6 | 301 |
| Beckman AU series<br>Reebio \ OTHER | U/L | 78.1 | 17.1% | 5 | 145 |
| Roche cobas 4000/6000/8000<br>DiaSys \ Enzymatic Cycling Assays Method | U/L | 115 | 12% | <5 | <100 |
| Mindray BS series<br>OTHER \ OTHER | U/L | 93.8 | 3% | <5 | <100 |
| (Mg)镁 | | | | | |
| Beckman AU series<br>Elikan \ Xylidyl Blue | mmol/L | 1.28 | 4.6% | 6 | 600 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Xylidyl Blue | mmol/L | 1.27 | 7.9% | 5 | 228 |
| Roche Hitachi/Hitachi Systems<br>Health World \ Arsenazo □ | mmol/L | 1.47 | 2.1% | <5 | 110 |
| Roche Hitachi/Hitachi Systems<br>Wako \ OTHER | mmol/L | 1.46 | 1.9% | <5 | <100 |
| Mindray BS series<br>Dedicated \ Xylidyl Blue | mmol/L | 1.49 | 2.9% | 15 | 576 |
| Roche cobas 4000/6000/8000<br>Roche \ Xylidyl Blue | mmol/L | 1.41 | 11.9% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ Xylidyl Blue | mmol/L | 1.47 | 2.9% | <5 | <100 |
| (Na)钠 | | | | | |
| Roche Hitachi/Hitachi Systems<br>Hitachi \ ISE (indirect) | mmol/L | 158 | 1.2% | 5 | 184 |
| Beckman AU series<br>Beckman Coulter \ ISE (indirect) | mmol/L | 153 | 1.3% | 22 | 1804 |
| Roche Hitachi/Hitachi Systems<br>Dedicated \ ISE (indirect) | mmol/L | 156 | 2.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>Wako \ ISE (indirect) | mmol/L | 155 | 0.8% | 15 | 720 |
| Mindray BS series<br>Dedicated \ OTHER | mmol/L | 154 | 1.7% | 15 | 548 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | mmol/L | 158 | 0.4% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ ISE (indirect) | mmol/L | 152 | 11.9% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ ISE (direct) | mmol/L | 143 | 1% | <5 | <100 |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ ISE (indirect) | mmol/L | 152 | 1.1% | <5 | <100 |
| OTHER<br>ABX \ ISE (indirect) | mmol/L | 155 | 0% | <5 | <100 |
| (NEFA)游离脂肪酸 | | | | | |
| Roche cobas 4000/6000/8000<br>DiaSys \ ACS - ACOD | mmol/L | 1.54 | 12% | <5 | <100 |

| | | 批号：376DFE3 | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| **(P)磷** | | | | | |
| Beckman AU series<br>LEADMAN \ Phosphomolybdate-UV-end point | mmol/L | 2.82 | 2.4% | 9 | 420 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Phosphomolybdate UV-rate | mmol/L | 2.66 | 4.7% | 5 | 236 |
| Roche Hitachi/Hitachi Systems<br>Wako \ Phosphomolybdate-UV-end point | mmol/L | 2.84 | 1% | 5 | 186 |
| Mindray BS series<br>Dedicated \ Phosphomolybdate-UV-end point | mmol/L | 2.81 | 3.4% | 15 | 458 |
| Beckman AU series<br>Elikan \ Molybdenum blue colorimetry | mmol/L | 2.73 | 2.5% | 14 | 1266 |
| Roche cobas 4000/6000/8000<br>Roche \ Phosphomolybdate UV-rate | mmol/L | 2.76 | 11.9% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ Phosphomolybdate UV-rate | mmol/L | 2.82 | 2.9% | <5 | <100 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | mmol/L | 2.92 | 2% | <5 | <100 |
| Siemens ADVIA Centaur XP/XPT<br>Siemens \ Phosphomolybdate-UV-end point | mmol/L | 2.72 | 4.2% | <5 | <100 |
| **(PRO/ TP)(总) 蛋白** | | | | | |
| Beckman AU series<br>BIOSINO \ Biuret-endpoint | g/L | 71.4 | 1.8% | 9 | 426 |
| Mindray BS series<br>chongqiong zybio \ Biuret-endpoint | g/L | 71.7 | 1.4% | 9 | 422 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Biuret-endpoint | g/L | 72.4 | 1.4% | 5 | 190 |
| Beckman AU series<br>Beckman Coulter \ Biuret-endpoint | g/L | 69.7 | 1.2% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Biuret-endpoint | g/L | 72 | 1.1% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Biuret-rate | g/L | 74.9 | 3.5% | 16 | 718 |
| Mindray BS series<br>Dedicated \ Biuret-rate | g/L | 73.9 | 3.1% | 14 | 368 |
| Beckman AU series<br>OTHER \ Biuret-endpoint | g/L | 69.2 | 1.6% | 13 | 1258 |
| Mindray BS series<br>Elikan \ Biuret-endpoint | g/L | 71.8 | 0.6% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Biuret-endpoint | g/L | 71.9 | 11.9% | <5 | <100 |
| OTHER<br>BSBE \ Biuret-endpoint | g/L | 74.2 | 0% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ Biuret-rate | g/L | 76.8 | 0.9% | <5 | <100 |
| **(PROG)孕酮** | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 2.31 | 12% | <5 | <100 |
| **(TBA)总胆汁酸** | | | | | |
| Beckman AU series<br>BIOSINO \ Enzymatic Cycling Assays Method | μmol/L | 66.5 | 4.5% | 8 | 396 |
| Beckman AU series<br>OTHER \ Enzymatic Cycling Assays Method | μmol/L | 60.9 | 4% | <5 | 322 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Enzymatic Cycling Assays Method | μmol/L | 70.2 | 4.1% | 15 | 766 |
| Mindray BS series<br>Dedicated \ Enzymatic Cycling Assays Method | μmol/L | 55.3 | 4.4% | 15 | 416 |
| Beckman AU series<br>Szybio \ Enzymatic Cycling Assays Method | μmol/L | 64.7 | 5% | 10 | 912 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Enzymatic Cycling Assays Method | μmol/L | 73 | 5.3% | 10 | 342 |
| Roche cobas 4000/6000/8000<br>DiaSys \ Enzymatic Cycling Assays Method | μmol/L | 63 | 11.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ Enzymatic Cycling Assays Method | μmol/L | 71.2 | 3.2% | <5 | <100 |
| **(TBIL)总胆红素** | | | | | |
| Beckman AU series<br>MAKER \ Vanadate | μmol/L | 119 | 3% | 9 | 452 |
| Mindray BS series<br>Elikan \ Vanadate | μmol/L | 127 | 1.8% | 6 | 276 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ sodium nitrite oxidase method | μmol/L | 120 | 3.5% | 5 | 190 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Vanadate | μmol/L | 141 | 1.6% | 44 | 2082 |
| Mindray BS series<br>chongqiong zybio \ Vanadate | μmol/L | 124 | 1.7% | <5 | 144 |
| Beckman AU series<br>Beckman Coulter \ Diazotization -Blank Compensation | μmol/L | 128 | 1.5% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Vanadate | μmol/L | 115 | 3.8% | 15 | 628 |
| Mindray BS series<br>Dedicated \ Vanadate | μmol/L | 115 | 2.7% | 15 | 542 |
| Beckman AU series<br>OTHER \ Vanadate | μmol/L | 122 | 2.7% | 13 | 1280 |
| Beckman AU series<br>Beckman Coulter \ Diazotization | μmol/L | 124 | 0.8% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ OTHER | μmol/L | 119 | 0.1% | <5 | <100 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | μmol/L | 134 | 2.1% | <5 | <100 |

| | | 批号：376DFE3 | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| Roche cobas 4000/6000/8000<br>DiaSys \ Diazotization | μmol/L | 113 | 11.9% | <5 | <100 |
| (TES)睾酮 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 1.7 | 12% | <5 | <100 |
| (TG)甘油三酯 | | | | | |
| Beckman AU series<br>LEADMAN \ GPO-POD | mmol/L | 1.93 | 4.6% | 9 | 318 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ OTHER | mmol/L | 2.38 | 4.3% | 9 | 112 |
| Mindray BS series<br>Elikan \ GPO-POD | mmol/L | 2.12 | 3.2% | 6 | 276 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ GPO-POD | mmol/L | 2.21 | 1.5% | 5 | 190 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ GPO-POD | mmol/L | 1.9 | 2.2% | 33 | 1524 |
| Mindray BS series<br>OTHER \ GPO-POD | mmol/L | 2.06 | 1.9% | <5 | 146 |
| Mindray BS series<br>Dedicated \ GPO-POD | mmol/L | 1.99 | 2.7% | 15 | 500 |
| Beckman AU series<br>OTHER \ GPO-POD | mmol/L | 2.2 | 2.5% | 14 | 1282 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ OTHER | mmol/L | 1.92 | 1.2% | 12 | 828 |
| Roche cobas 4000/6000/8000<br>Roche \ Free glycerol removed | mmol/L | 2.14 | 11.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ OTHER | mmol/L | 2.44 | 1.5% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ GPO-POD | mmol/L | 3.58 | 3.1% | <5 | 100 |
| (TIBC)总铁结合力 | | | | | |
| Roche cobas 4000/6000/8000<br>Roche \ Ferrozine colorimetry | μmol/L | 61.4 | 12% | <5 | <100 |
| (TRF)转铁蛋白 | | | | | |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Immunoturbidimetry (ITA) | mg/dL | 345 | 4.2% | 15 | 365 |
| Roche cobas 4000/6000/8000<br>Roche \ Immunoturbidimetry (ITA) | mg/dL | 290 | 12% | <5 | <100 |
| (TSH)促甲状腺素 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | mIU/L | 1.88 | 12% | <5 | <100 |
| (TT3)总三碘甲状原氨酸 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 1.07 | 12% | <5 | <100 |
| (TT4)总T4 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | ng/mL | 76.8 | 12% | <5 | <100 |
| (UA)尿酸 | | | | | |
| Beckman AU series<br>MAKER \ Uricase colorimetry | μmol/L | 483 | 2.1% | 9 | 432 |
| Mindray BS series<br>simes-sikma \ OTHER | μmol/L | 480 | 1.5% | 9 | 416 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Uricase-UV | μmol/L | 481 | 1.1% | 5 | 188 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Uricase colorimetry | μmol/L | 488 | 1.2% | 44 | 2374 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Uricase colorimetry | μmol/L | 491 | 1.8% | 16 | 722 |
| Mindray BS series<br>Dedicated \ Uricase-UV | μmol/L | 484 | 2.8% | 15 | 452 |
| Beckman AU series<br>OTHER \ Uricase colorimetry | μmol/L | 495 | 2.4% | 14 | 1232 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | μmol/L | 472 | 0.5% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ Uricase colorimetry | μmol/L | 501 | 0.4% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Uricase colorimetry | μmol/L | 469 | 11.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ Uricase colorimetry | μmol/L | 491 | 1% | <5 | <100 |
| (UIBC)不饱和铁结合力 | | | | | |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ Ferrozine colorimetry | μmol/L | 29.5 | 3.6% | 15 | 366 |
| Roche cobas 4000/6000/8000<br>Roche \ Ferrozine colorimetry | μmol/L | 30.4 | 12% | <5 | <100 |
| (Urea)尿素 | | | | | |
| Mindray BS series<br>simes-sikma \ OTHER | mmol/L | 20.7 | 3.2% | 9 | 418 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ Urease-UV-rate | mmol/L | 20.8 | 2% | 5 | 186 |
| Roche Hitachi/Hitachi Systems<br>LEADMAN \ OTHER | mmol/L | 22.3 | 2.3% | 35 | 1796 |
| Roche Hitachi/Hitachi Systems<br>MAKER \ OTHER | mmol/L | 20.2 | 3.8% | 16 | 792 |

| | | 批号：376DFE3 | | | |
|---|---|---|---|---|---|
| 项目\仪器\试剂\方法 | 单位 | 均值 \ CV | | Lab \ 点数 | |
| Mindray BS series<br>Dedicated \ Urease-UV-rate | mmol/L | 22.2 | 5.3% | 15 | 568 |
| Beckman AU series<br>OTHER \ Urease colorimetry | mmol/L | 21.9 | 2.8% | 14 | 970 |
| Ortho Vitros 3600/5600<br>Ortho \ dry chemistry | mmol/L | 17.2 | 1.3% | <5 | <100 |
| Beckman AU series<br>Beckman Coulter \ Urease colorimetry | mmol/L | 17.6 | 2.7% | <5 | <100 |
| Roche cobas 4000/6000/8000<br>Roche \ Urease colorimetry | mmol/L | 20.9 | 11.9% | <5 | <100 |
| Roche Hitachi/Hitachi Systems<br>BIOBASE \ OTHER | mmol/L | 21.4 | 1.5% | <5 | <100 |
| (VB12)维生素B12 | | | | | |
| Roche cobas E601/602<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | pg/mL | 273 | 12% | <5 | <100 |
| (β-hCG)人绒毛膜促性腺激素β亚基 | | | | | |
| Roche cobas 4000/6000/8000<br>Roche \ Electro-Chemiluminescence Immunoassay (ECLIA) | IU/mL | 0.003 | 12% | <5 | <100 |

**注：**此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。