

REF	671-2	6×1mL	EXP	2024-12-31	LOT	400341F	更新日期：	2024-04
1 / 3								

		批号：400341F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(APTT)活化部分凝血活酶时间					
Werfen ACL series Dedicated \ Light scattering	Seconds	58.1	4.2%	<5	<100
Werfen ACL series Werfen \ Magnetic Beads	Seconds	57.7	10.8%	<5	118
Stago STA series Stago \ Magnetic Beads	Seconds	60.3	34.3%	<5	111
Sysmex CA 1500/6000/7000 Siemens Dade Actin FS(APTT) \ Light scattering	Seconds	55.7	31.6%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Pathrombin SL(APTT) \ Light scattering	Seconds	61.1	45.1%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	66.1	8.7%	<5	<100
Sysmex CS 5100 Dedicated \ Light scattering	Seconds	73.3	4.2%	12	115
Sysmex CS-1300 sysmex \ Light scattering	Seconds	114	1.5%	<5	<100
Werfen ACL series Werfen Hemosil SynthAsil APTT \ Light scattering	Seconds	31	11.9%	<5	<100
Werfen ACL series Werfen Hemosil APTT-SP \ Light scattering	Seconds	28	11.8%	<5	<100
Sysmex CS 5100 Siemens \ Light scattering	Seconds	103	4.5%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Dade Actin(APTT) \ Light scattering	Seconds	31	11.9%	<5	<100
Sysmex CS-2500 sysmex \ Light scattering	Seconds	107	3.8%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	70.7	14.2%	<5	116
(AT-III)抗凝血酶III					
Werfen ACL series Werfen \ Nephelometry	%	44.6	9.8%	<5	106
Sysmex CS 5100 Dedicated \ Nephelometry	%	46.2	7.8%	12	115
Sysmex CS 5100 sysmex \ Nephelometry	%	41.2	11%	<5	<100
Sysmex CS 5100 Siemens \ Nephelometry	%	39.1	13.9%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	%	46	12.5%	<5	<100
Werfen ACL series Werfen HemosIL Liquid Antithrombin (AT□) \ Nephelometry	%	98	1.1%	<5	<100
Werfen ACL series Werfen IL Antithrombin III Chromogenic (AT□) \ Nephelometry	%	99	1.1%	<5	<100
Sysmex CS-1300 sysmex \ Nephelometry	%	30.3	3.1%	<5	<100
Stago STA series Stago \ Magnetic Beads	%	37	12%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Innovance Antithrombin (AT□) \ Nephelometry	%	101	1.2%	<5	<100
Sysmex CS-2500 sysmex \ Nephelometry	%	45.8	3.3%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Berichrom Antithrombin III (AT□) \ Nephelometry	%	98	1.1%	<5	<100
(D-D(DDU)) D-二聚体(DDU)					
Sysmex CS 5100 Dedicated \ Nephelometry	mg/L	3.13	9.6%	9	<100
Werfen ACL series Dedicated \ Nephelometry	mg/L	2.09	2.7%	<5	<100
Werfen ACL series Werfen \ Nephelometry	mg/L	1.76	31.3%	<5	161
Sysmex CS 5100 sysmex \ Nephelometry	mg/L	2.61	10%	<5	<100
Werfen ACL series Werfen HemosIL D-Dimer \ Nephelometry	mg/L	0.599	11.9%	<5	<100
Werfen ACL series Werfen HemosIL D-Dimer HS \ Nephelometry	mg/L	0.502	12%	<5	<100
Sysmex CS 5100 SekiSui \ Nephelometry	mg/L	3.27	12%	<5	<100
Sysmex CS-1300 sysmex \ Nephelometry	mg/L	2.36	5.3%	<5	<100
(D-D(FEU)) D-二聚体(FEU)					
Stago STA series Stago \ Nephelometry	mg/L	1.27	32.3%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Nephelometry	mg/L	2.65	33.9%	<5	<100
Sysmex CS 5100 Siemens \ Nephelometry	mg/L	3.4	5.2%	<5	<100
(FDP)纤维蛋白(原)降解产物					
Sysmex CS 5100 Dedicated \ Nephelometry	mg/L	10	8.1%	9	<100
Werfen ACL series Dedicated \ Nephelometry	mg/L	4.55	6.6%	<5	<100
Werfen ACL series Werfen \ Nephelometry	mg/L	0.005	11.9%	<5	<100
Sysmex CS 5100 SekiSui \ Nephelometry	mg/L	5.5	12%	<5	<100
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	6.95	11.9%	<5	<100

REF	671-2	6×1mL	EXP	2024-12-31	LOT	400341F	更新日期：	2024-04
2 / 3								

		批号：400341F			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Sysmex CS 5100 sysmex \ Nephelometry	mg/L	7.05	27.4%	<5	145
(FIB)纤维蛋白原					
Werfen ACL series Dedicated \ Light scattering	g/L	2.33	7.1%	<5	<100
Stago STA series Stago \ Magnetic Beads	g/L	2.65	36.1%	<5	<100
Werfen ACL series Werfen \ Magnetic Beads	g/L	2.12	10.4%	<5	108
Werfen ACL series Werfen \ Light scattering	g/L	2.19	8.9%	<5	<100
Sysmex CS 5100 Dedicated \ Light scattering	g/L	2.08	4.6%	12	114
Werfen ACL series Werfen HemosIL QFA Thrombin (FIB) \ Light scattering	g/L	3.8	11.8%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	g/L	1.98	11.9%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Thrombin (FIB) \ Light scattering	g/L	3.6	11.9%	<5	<100
Sysmex CS-1300 sysmex \ Light scattering	g/L	182	2.2%	<5	<100
Werfen ACL series Werfen HemosIL Fib-C \ Light scattering	g/L	3.68	12%	<5	<100
Sysmex CS 5100 Siemens \ Light scattering	g/L	2.29	11.9%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Multifibren U (FIB) \ Light scattering	g/L	3.91	11.8%	<5	<100
Sysmex CS-2500 sysmex \ Light scattering	g/L	2.1	4.5%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	g/L	1.91	11.2%	<5	117
(PT INR)PT国际标准化比值					
Werfen ACL series Werfen \ Calculate	INR	3.12	18.1%	5	164
Werfen ACL series Dedicated \ Calculate	INR	2.97	4.8%	<5	<100
Stago STA series Stago \ Calculate	INR	2.25	40.6%	<5	<100
Sysmex CS 5100 Siemens \ Calculate	INR	2.72	9.2%	<5	<100
Sysmex CS 5100 Dedicated \ Calculate	INR	2.8	4.8%	12	117
Sysmex CS 5100 sysmex \ Calculate	INR	2.75	0.5%	<5	<100
Sysmex CA 1500/6000/7000 Siemens Thromborel S (PT) \ Calculate	INR	1.05	11.4%	<5	<100
Sysmex CA 1500/6000/7000 Siemens PT Innovin \ Calculate	INR	1.13	10.6%	<5	<100
Werfen ACL series Werfen Hemosil PT-Fib HS Plus \ Calculate	INR	1.05	12.4%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin \ Calculate	INR	1.03	11.7%	<5	<100
Sysmex CS-2500 sysmex \ Calculate	INR	2.55	3.3%	<5	<100
Stago STA series STA-Neoplastine CI Plus \ Calculate	INR	2.52	2.5%	<5	<100
Werfen ACL series Werfen Hemosil PT Recombiplastin 2G \ Calculate	INR	1.01	11.9%	<5	<100
(PT)血浆凝血酶原时间					
Werfen ACL series Dedicated \ Light scattering	Seconds	30.5	4.6%	<5	<100
Werfen ACL series Werfen \ Magnetic Beads	Seconds	31	8.6%	<5	117
Werfen ACL series Werfen \ Light scattering	Seconds	32.5	8%	<5	<100
Sysmex CS 5100 Dedicated \ Light scattering	Seconds	32.3	4.5%	12	118
Sysmex CS 5100 sysmex \ Light scattering	Seconds	31.6	6.6%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	29.6	12%	<5	<100
Sysmex CS-2500 sysmex \ Light scattering	Seconds	29.3	3.3%	<5	<100
Sysmex CS-1300 sysmex \ Light scattering	Seconds	30.4	4.8%	<5	<100
Sysmex CS 5100 Siemens \ Light scattering	Seconds	29.7	2.7%	<5	<100
Stago STA series STA-Neoplastine CI Plus \ Magnetic Beads	Seconds	26.8	2.3%	<5	<100
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	32.1	11.6%	<5	<100
(TT)凝血酶时间					
Werfen ACL series Dedicated \ Light scattering	Seconds	20.1	4.6%	<5	<100
Werfen ACL series Werfen \ Magnetic Beads	Seconds	18	9.5%	<5	<100
Werfen ACL series Werfen \ Light scattering	Seconds	17	14.2%	<5	<100
Sysmex CS 5100 sysmex \ Light scattering	Seconds	16.4	13.5%	<5	142

报告类别： 汇总实验室数据报告



质控品名： 凝血七项非定值质控品

REF	671-2	6×1mL	EXP	2024-12-31	LOT	400341F	更新日期：	2024-04
3 / 3								

		批号：400341F			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Sysmex CS 5100 Dedicated \ Light scattering	Seconds	16.5	3.4%	12	115
Sysmex CS 5100 Siemens \ Light scattering	Seconds	19.3	1.2%	<5	<100
Sysmex CS-2500 sysmex \ Light scattering	Seconds	17.4	3.7%	<5	<100
Stago STA series Stago \ Magnetic Beads	Seconds	16	12%	<5	<100
Sysmex CS-1300 sysmex \ Light scattering	Seconds	20.6	1.1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。