

<b>REF</b> 510-2	6x3mL	<b>EXP</b> 2025-09-26	<b>LOT</b> 30A88BA	更新日期： 2020-12 1 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30A88BA				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
<b>(AFP)甲胎蛋白</b>				
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	83.2	67.1-99.3	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	70.5	56.1-84.9	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	60.3	48.1-72.5	
<b>(CA 125)糖类抗原 125</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	179	138-220	
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	IU/mL	276	221-331	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	229	184-274	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	288	232-344	
<b>(CA 15-3)糖类抗原 15-3</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	69.3	54.5-84.1	
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	IU/mL	95.3	76.4-114	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	96.3	77-116	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	104	80.4-128	
<b>(CA 19-9)糖类抗原 19-9</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	76.8	60.5-93.1	
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	IU/mL	187	152-222	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	178	142-214	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	398	318-478	
<b>(CA 72-4)糖类抗原 72-4</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	8.23	6.21-10.3	
<b>(CEA)癌胚抗原</b>				
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	31.3	25.3-37.3	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	35.7	27.9-43.5	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	26.5	21.7-31.3	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	46.3	37.3-55.3	
<b>(Cyfra 21-1)细胞角蛋白19片段</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.3	8.08-12.5	
<b>(Ferritin)铁蛋白</b>				
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	µg/L	268	219-317	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	µg/L	352	278-426	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	µg/L	306	250-362	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	µg/L	409	329-489	
<b>(F-PSA)游离前列腺特异性抗原</b>				
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	6.03	4.66-7.4	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.86	3.89-5.83	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.12	4.16-6.08	

<b>REF</b> 510-2	6x3mL	<b>EXP</b> 2025-09-26	<b>LOT</b> 30A88BA	更新日期： 2020-12 2 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30A88BA				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	4.23	3.36-5.1	
<b>(hCG)人绒毛膜促性腺激素</b>				
Siemens ADVIA Centaur XP Siemens \ Enzyme linked immunosorbent assays (ELISA)	IU/L	98.6	78.8-118	
<b>(PSA)前列腺特异性抗原</b>				
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	19.3	15.4-23.2	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.3	14.7-21.9	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	16.2	13-19.4	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	17.9	14.6-21.2	
<b>(SCC)鳞状细胞癌抗原</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.85	6.98-10.7	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	12.8	10.2-15.4	
<b>(TPA)组织多肽抗原</b>				
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	U/L	3820	3040-4590	
<b>(TPS)组织多肽特异抗原</b>				
Manual OTHER \ Enzyme linked immunosorbent assays (ELISA)	U/L	104	86-122	
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	157	129-185	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。