

REF	671-2	6×1mL	EXP	2024-07-28	LOT	4003419	更新日期：	2022-03
1 / 1								

批号：4003419			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(APTT)活化部分凝血活酶时间			
Sysmex CS 5100 Siemens \ Light scattering	Seconds	93.4	71-116
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	95.3	72.4-118
Stago STA series Dedicated \ Magnetic Beads	Seconds	80.4	61.1-99.7
IL ACL series IL \ Light scattering	Seconds	64.3	48.9-79.7
(AT-III)抗凝血酶III			
IL ACL series Dedicated \ Nephelometry	%	43.6	33.1-54.1
Sysmex CA 1500/6000/7000 Siemens \ Nephelometry	%	44.3	33.7-54.9
(D-D(DDU)) D-二聚体(DDU)			
IL ACL series Dedicated \ Nephelometry	mg/L	0.745	0.567-0.923
(D-D(FEU)) D-二聚体(FEU)			
Stago STA series Dedicated \ Latex agglutination (LA)	mg/L	1.67	1.27-2.07
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	3.31	2.52-4.1
Sysmex CS 5100 Dedicated \ Nephelometry	mg/L	2.93	2.23-3.63
(FDP)纤维蛋白(原)降解产物			
IL ACL series SekiSui \ Nephelometry	mg/L	4.23	3.21-5.25
Stago STA series Dedicated \ Nephelometry	mg/L	5.52	4.2-6.84
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	5.63	4.28-6.98
(FIB)纤维蛋白原			
IL ACL series IL \ Light scattering	g/L	2.24	1.7-2.78
Stago STA series Dedicated \ Magnetic Beads	g/L	2.34	1.78-2.9
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	g/L	2.15	1.63-2.67
Sysmex CS 5100 Siemens \ Light scattering	g/L	1.86	1.41-2.31
(PT INR)PT国际标准化比值			
Sysmex CA 1500/6000/7000 Siemens \ Calculate	INR	2.68	2.04-3.32
IL ACL series IL \ Calculate	INR	2.79	2.12-3.46
Stago STA series Stago \ Calculate	INR	2.76	2.1-3.42
(PT)血浆凝血酶原时间			
IL ACL series IL \ Light scattering	Seconds	32.4	24.6-40.2
Stago STA series Dedicated \ Magnetic Beads	Seconds	29.5	22.4-36.6
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	31.3	23.8-38.8
Sysmex CS 5100 Siemens \ Light scattering	Seconds	33.5	25.5-41.5
(TT)凝血酶时间			
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	17.3	13.1-21.5
Stago STA series Dedicated \ Magnetic Beads	Seconds	16.3	12.4-20.2
IL ACL series SUNBIO \ Light scattering	Seconds	17.4	13.2-21.6

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。