

REF	581-1	6×5mL	EXP	2025-09-03	LOT	376B8D1	更新日期：	2022-09
1 / 2								

批号：376B8D1			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(a-HBDH)α-羟丁酸脱氢酶			
Beckman AU series MedicalSystem \ Enzymatic Cycling Assays Method	U/L	160	122-198
(Alb)白蛋白			
Beckman AU series MedicalSystem \ Bromocresol green	μmol/L	402	306-498
Beckman AU series MedicalSystem \ Bromocresol green	g/L	26.5	20.1-32.9
(ALP)碱性磷酸酶			
Beckman AU series MedicalSystem \ Rate -AMP buffer	U/L	47.9	36.4-59.4
(ALT/ SGPT)丙氨酸氨基转移酶			
Beckman AU series MedicalSystem \ UV with P5P	U/L	38.9	29.6-48.2
(AMY)淀粉酶			
Beckman AU series MedicalSystem \ PNP-G7	U/L	174	132-216
(AST/ SGOT)天门冬氨酸氨基转移酶			
Beckman AU series MedicalSystem \ UV with P5P	U/L	37.3	28.3-46.3
(Ca)钙			
Beckman AU series MedicalSystem \ BAPTA	mmol/L	1.85	1.41-2.29
(CHE)胆碱酯酶			
Beckman AU series MedicalSystem \ Photometry	U/L	5000	3790-6200
(CHOL/ TC)总胆固醇			
Beckman AU series MedicalSystem \ Cholesterol oxidase	mmol/L	2.41	1.83-2.99
(CK)肌酸激酶			
Beckman AU series MedicalSystem \ Rate -AMP buffer	U/L	144	110-178
(CK-MB)肌酸激酶-MB (活性)			
Beckman AU series MedicalSystem \ Immune suppression	U/L	16.9	12.8-21
(Cl)氯			
Beckman AU series MedicalSystem \ ISE (indirect)	mmol/L	88.7	67.4-110
(CO2 (HCO3-))二氧化碳 (碳酸氢根)			
Beckman AU series MedicalSystem \ Colorimetry	mmol/L	4.27	3.25-5.29
(Cr)肌酐			
Beckman AU series MedicalSystem \ Enzymatic	μmol/L	73.6	55.9-91.3
(DBIL)直接胆红素			
Beckman AU series MedicalSystem \ Diazotization -Blank Compensation	μmol/L	4.6	3.5-5.7
(Fe)铁			
Beckman AU series MedicalSystem \ Ferrozine colorimetry	μmol/L	16.2	12.3-20.1
(GGT)L-γ-谷氨酰基转移酶			
Beckman AU series MedicalSystem \ γ-glutamyl -3-carboxy-4-nitroaniline	U/L	29.8	22.6-37
(GLU)葡萄糖			
Beckman AU series MedicalSystem \ Hexokinase	mmol/L	3.66	2.78-4.54
(HDL-C)高密度脂蛋白胆固醇			
Beckman AU series MedicalSystem \ Enzymatic	mmol/L	0.98	0.745-1.22
(K)钾			
Beckman AU series MedicalSystem \ ISE (indirect)	mmol/L	2.15	1.63-2.67
(LA)乳酸			
Beckman AU series MedicalSystem \ Lactate to pyruvate (L-P)	mmol/L	1.57	1.19-1.95
(LDH)乳酸脱氢酶			
Beckman AU series MedicalSystem \ Lactate to pyruvate (L-P)	U/L	144	109-179

REF	581-1	6×5mL	EXP	2025-09-03	LOT	376B8D1	更新日期：	2022-09
								2 / 2

批号：376B8D1			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(LDL-C)低密度脂蛋白胆固醇			
Beckman AU series MedicalSystem \ Enzymatic	mmol/L	1.88	1.43-2.33
(Li)锂			
Beckman AU series BIOSINO \ Colorimetry	mmol/L	0.8	0.608-0.992
(LIP)脂肪酶			
Beckman AU series MedicalSystem \ Enzymatic Cycling Assays Method	U/L	38.4	29.2-47.6
(Mg)镁			
Beckman AU series MedicalSystem \ Xylidyl Blue	mmol/L	0.671	0.51-0.832
(Na)钠			
Beckman AU series MedicalSystem \ ISE (indirect)	mmol/L	117	88.9-145
(P)磷			
Beckman AU series MedicalSystem \ Phosphomolybdate UV-rate	mmol/L	0.82	0.623-1.02
(PRO/ TP)(总) 蛋白			
Beckman AU series MedicalSystem \ Biuret-endpoint	g/L	39.8	30.2-49.4
(TBA)总胆汁酸			
Beckman AU series MedicalSystem \ Enzymatic Cycling Assays Method	μmol/L	25.3	19.2-31.4
(TBIL)总胆红素			
Beckman AU series MedicalSystem \ Diazotization	μmol/L	16.7	12.7-20.7
(TG)甘油三酯			
Beckman AU series MedicalSystem \ Free glycerol removed	mmol/L	0.82	0.623-1.02
(UA)尿酸			
Beckman AU series MedicalSystem \ Uricase colorimetry	μmol/L	174	132-216
(Urea)尿素			
Beckman AU series MedicalSystem \ Urease colorimetry	mmol/L	5.09	3.87-6.31

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。