报告类别： **示范实验室数据报告**

质控品名： **特定蛋白液体非定值质控品**

REF	521-3	6×3mL	EXP	2021-07-15	LOT	31B790B	更新日期：	2018-10

		批号：31B790B	
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(a1-AG)a1 酸性糖蛋白			
Siemens BN II Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mg/L	1409	1,187-1,631
(ALB)白蛋白			
Roche Hitachi Systems Dedicated \ Biuret-rate	g/dL	5.43	4.45-6.41
(ASO)抗链球菌溶血素 O 抗体			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	226	188-264
(C3)补体 C3			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	192	160-224
(C4)补体 C4			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	49.8	40.2-59.4
(Cer)铜蓝蛋白			
Siemens BN II Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mg/L	428	362-494
(CRP)C反应蛋白			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	6.98	5.66-8.30
(Hap)触珠蛋白			
Siemens BN II Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mg/L	2209	1,841-2,577
(IgA)免疫球蛋白 IgA			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	367	301-433
(IgE)免疫球蛋白 IgE			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	104	86-122
(IgG)免疫球蛋白 IgG			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	1806	1,508-2,104
(IgM)免疫球蛋白 IgM			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	243	199-287
(Kappa Light Chain)k 轻链抗体			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	1687	1,401-1,973
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	532	430-634
(Lambda Light Chain)λ 轻链抗体			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	809	657-961
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	284	232-336
(PAB)前白蛋白			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	57.3	47.5-67.1
(RBP)视黄醇结合蛋白			
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	9.42	7.82-11.02
(RF)类风湿因子			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	112	90-134
(TRF)转铁蛋白			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	372	306-438

报告类别： **示范实验室数据报告**

质控品名： **特定蛋白液体非定值质控品**

REF	521-3	6×3mL	EXP	2021-07-15	LOT	31B790B	更新日期：	2018-10

		批号：31B790B	
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(β2M)β-2-微球蛋白			
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	0.67	0.55-0.79
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	0.53	0.41-0.65

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。