报告类别： **示范实验室数据报告**

质控品名： **凝血七项非定值质控品**

REF	671-3	6×1mL	EXP	2023-03-05	LOT	4005A69	更新日期：	2021-04

		批号：4005A69	
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(APTT)活化部分凝血活酶时间			
Sysmex CS 2000i/2100i sysmex \ Light scattering	Seconds	116	88.2-144
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	154	126-182
IL ACL series IL \ Light scattering	Seconds	96.4	76.6-116
Stago STA series Dedicated \ Magnetic Beads	Seconds	105	83-127
(AT-III)抗凝血酶III			
IL ACL series IL \ Nephelometry	%	24.6	19.8-29.4
Sysmex CA 1500/6000/7000 Siemens \ Nephelometry	%	26.7	21.3-32.1
Sysmex CS 2000i/2100i sysmex \ Nephelometry	%	30	22.8-37.2
(D-D(DDU)) D-二聚体(DDU)			
IL ACL series IL \ Nephelometry	mg/L	1.39	1.13-1.65
(D-D(FEU)) D-二聚体(FEU)			
Sysmex CA 1500/6000/7000 Siemens \ Nephelometry	mg/L	5.63	4.61-6.65
Sysmex CS 2000i/2100i sysmex \ Nephelometry	mg/L	5.31	4.04-6.58
(FDP)纤维蛋白(原)降解产物			
Stago STA series Dedicated \ Magnetic Beads	mg/L	8.29	6.47-10.1
IL ACL series IL \ Nephelometry	mg/L	6.33	4.97-7.69
Sysmex CA 1500/6000/7000 SekiSui \ Nephelometry	mg/L	9.38	7.62-11.1
Sysmex CS 2000i/2100i sysmex \ Nephelometry	mg/L	10.9	8.28-13.5
(FIB)纤维蛋白原			
IL ACL series IL \ Light scattering	g/L	1.28	1.02-1.54
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	g/L	1.29	1.07-1.51
Sysmex CS 2000i/2100i sysmex \ Light scattering	g/L	1.45	1.1-1.8
(PT INR)PT国际标准化比值			
Sysmex CS 2000i/2100i sysmex \ Calculate	INR	4.25	3.23-5.27
IL ACL series IL \ Calculate	INR	3.99	3.23-4.75
(PT)血浆凝血酶原时间			
IL ACL series IL \ Light scattering	Seconds	47.7	38.3-57.1
Stago STA series Dedicated \ Magnetic Beads	Seconds	43.6	35-52.2
Sysmex CA 1500/6000/7000 Siemens \ Light scattering	Seconds	44.8	36.4-53.2
Sysmex CS 2000i/2100i sysmex \ Light scattering	Seconds	45.7	34.7-56.7
(TT)凝血酶时间			
Sysmex CS 2000i/2100i sysmex \ Light scattering	Seconds	15.9	12.1-19.7
Sysmex CA 1500/6000/7000 Dedicated \ Light scattering	Seconds	17.2	13.6-20.8
IL ACL series IL \ Light scattering	Seconds	17.5	13.9-21.1

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。