报告类别： **示范实验室数据报告**

质控品名： **尿液定量生化液体非定值质控品**

REF	586-2	6×8mL	EXP	2022-03-18	LOT	37E7FD0	更新日期：	2019-06

		批号：37E7FD0	
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(AMY)淀粉酶			
Beckman AU series Dedicated \ PNP-G5	U/L	199	167-231
Roche cobas c 501/502/701 Dedicated \ PNP-G7	U/L	194	160-228
(BUN)尿素氮			
Beckman AU series Dedicated \ Dehydrogenase	mg/dL	839	673-1,005
(Ca)钙			
Beckman AU series Dedicated \ ISE (direct)	mg/dL	11.7	9.5-13.9
(Cl)氯			
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	179	145-213
(Cortisol)皮质醇			
Beckman Access Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ug/dL	25.7	20.9-30.5
(CRE)肌酐			
Beckman AU series Dedicated \ Enzymatic	mg/dL	163	131-195
Roche cobas c 501/502/701 MedicalSystem \ Enzymatic	umol/L	1534	1,228-1,840
(GLU)葡萄糖			
Beckman AU series Dedicated \ Hexokinase	mg/dL	277	225-329
Roche cobas c 501/502/701 Dedicated \ Hexokinase	mmol/L	15.8	13.6-18.0
(hCG)人绒毛膜促性腺激素			
Beckman Access Dedicated \ Enzyme linked immunosorbent assays (ELISA)	mIU/mL	640	514-766
(K)钾			
Beckman AU series Dedicated \ ISE (indirect)	mEq/L	78.2	63.2-93.2
(MA)微量白蛋白			
Beckman AU series Dedicated \ Biuret-endpoint	mg/L	59.5	48.9-70.1
(Mg)镁			
Beckman AU series Dedicated \ Calmagite colorimetry	mg/dL	12.6	10.2-15.0
(Na)钠			
Beckman AU series Dedicated \ ISE (indirect)	mmol/L	166	134-198
(P)磷			
Beckman AU series Dedicated \ Molybdenum blue colorimetry	mg/dL	47.3	44.7-49.9
(PRO/ TP)(总) 蛋白			
Beckman AU series Dedicated \ Biuret-rate	mg/dL	69.4	56.8-82.0
(UA)尿酸			
Roche cobas c 501/502/701 Dedicated \ Uricase-UV	mg/dL	14.7	12.1-17.3

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。