



REF	671-2	6×1mL	EXP	2024-04-22	LOT	40033BE	更新日期：	2022-05
1 / 1								

批号：40033BE			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(APTT)活化部分凝血活酶时间			
Sysmex CS 5100	Seconds	93.4	74.8-112
Siemens \ Light scattering			
Sysmex CA 1500/6000/7000	Seconds	102	84-120
Siemens \ Light scattering			
Succeeder SF series	Seconds	69.5	52.8-86.2
Succeeder \ Magnetic Beads			
Stago STA series	Seconds	80.4	63.8-97
Dedicated \ Magnetic Beads			
IL ACL series	Seconds	64.3	51.1-77.5
IL \ Light scattering			
(AT-III)抗凝血酶III			
IL ACL series	%	43.6	35.2-52
Dedicated \ Nephelometry			
(D-D(DDU)) D-二聚体(DDU)			
Succeeder SF series	mg/L	1.45	1.1-1.8
Succeeder \ Magnetic Beads			
IL ACL series	mg/L	0.745	0.599-0.891
Dedicated \ Nephelometry			
(D-D(FEU)) D-二聚体(FEU)			
Sysmex CA 1500/6000/7000	mg/L	3.31	2.69-3.93
SekiSui \ Nephelometry			
Sysmex CS 5100	mg/L	2.93	2.35-3.51
Dedicated \ Nephelometry			
Stago STA series	mg/L	1.67	1.33-2.01
Dedicated \ Latex agglutination (LA)			
(FDP)纤维蛋白(原)降解产物			
IL ACL series	mg/L	4.23	3.39-5.07
SekiSui \ Nephelometry			
Stago STA series	mg/L	5.52	4.54-6.5
Dedicated \ Nephelometry			
Sysmex CA 1500/6000/7000	mg/L	5.63	4.59-6.67
SekiSui \ Nephelometry			
(FIB)纤维蛋白原			
IL ACL series	g/L	2.24	1.8-2.68
IL \ Light scattering			
Stago STA series	g/L	2.34	1.92-2.76
Dedicated \ Magnetic Beads			
Sysmex CA 1500/6000/7000	g/L	2.15	1.79-2.51
Siemens \ Light scattering			
Succeeder SF series	g/L	2.22	1.69-2.75
Succeeder \ Magnetic Beads			
Sysmex CS 5100	g/L	1.86	1.54-2.18
Siemens \ Light scattering			
(PT INR)PT国际标准化比值			
Succeeder SF series	INR	2.62	1.99-3.25
Succeeder \ Magnetic Beads			
(PT)血浆凝血酶原时间			
IL ACL series	Seconds	32.4	26-38.8
IL \ Light scattering			
Stago STA series	Seconds	29.5	23.7-35.3
Dedicated \ Magnetic Beads			
Succeeder SF series	Seconds	32.3	24.5-40.1
Succeeder \ Magnetic Beads			
Sysmex CA 1500/6000/7000	Seconds	31.3	25.1-37.5
Siemens \ Light scattering			
Sysmex CS 5100	Seconds	33.5	26.9-40.1
Siemens \ Light scattering			
(TT)凝血酶时间			
Sysmex CA 1500/6000/7000	Seconds	17.3	13.9-20.7
Dedicated \ Light scattering			
Stago STA series	Seconds	16.3	13.1-19.5
Dedicated \ Magnetic Beads			
Succeeder SF series	Seconds	14.1	10.7-17.5
Succeeder \ Magnetic Beads			
IL ACL series	Seconds	17.4	14.2-20.6
SUNBIO \ Light scattering			

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。