

REF

600-2

30×1.7mL

EXP

2026-02-15

LOT

393DDBB

更新日期：

2022-04

1 / 1

批号：393DDBB

项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
(Ca)钙			
Radiometer ABL Series Dedicated \ ISE (indirect)	mmol/L	1.05	0.83-1.27
Roche cobas b 121/123/221 Dedicated \ ISE (indirect)	mmol/L	1.02	0.8-1.24
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	1.08	0.84-1.32
(Cl)氯			
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	94.5	76.3-113
IL Gem Premier 3000 series Dedicated \ ISE (direct)	mmol/L	92.5	74.3-111
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	93.6	75.4-112
(K)钾			
IL Gem Premier 3000 series Dedicated \ ISE (direct)	mmol/L	3.88	3.1-4.66
Siemens RAPIDPoint 400/500 Dedicated \ ISE (direct)	mmol/L	3.97	3.19-4.75
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	3.82	3.06-4.58
(Na)钠			
Radiometer ABL Series Dedicated \ ISE (direct)	mmol/L	140	116-164
Radiometer AQT Series Dedicated \ ISE (direct)	mmol/L	141	117-165
Siemens RAPIDPoint 400/500 Siemens \ ISE (indirect)	mmol/L	142	120-164
IL Gem Premier 3000 series Dedicated \ ICP-MS	mmol/L	145	121-169
(PCO2)二氧化碳分压			
Siemens RAPIDPoint 400/500 Siemens \ Glucose oxidase (oxygen electrode)	mmHg	31.3	24.8-37.8
Nova Stat Profile pHox series Dedicated \ Electrode	mmHg	31.8	25.9-37.7
IL Gem Premier 3000 series Dedicated \ Electrode	mmHg	34.7	27.8-41.6
Roche cobas b 121/123/221 Dedicated \ Electrode	mmHg	37.2	29.8-44.6
Radiometer AQT Series Dedicated \ Electrode	mmHg	39.6	31.8-47.4
Radiometer ABL Series Dedicated \ Electrode	mmHg	38.5	30.7-46.3
(pH)酸碱度			
Roche cobas b 121/123/221 Dedicated \ Electrode	No Units	7.49	6.07-8.91
Siemens RAPIDPoint 400/500 Siemens \ Electrode	No Units	7.57	6.19-8.95
IL Gem Premier 3000 series Dedicated \ Electrode	No Units	7.47	6.09-8.85
Radiometer AQT Series Dedicated \ Electrode	No Units	7.49	6.11-8.87
Radiometer ABL Series Dedicated \ Electrode	No Units	7.53	6.17-8.89
(PO2)氧分压			
Radiometer AQT Series Dedicated \ Electrode	mmHg	101	79-123
IL Gem Premier 3000 series Dedicated \ Electrode	mmHg	94.8	76-114
Siemens RAPIDPoint 400/500 Dedicated \ Electrode	mmHg	122	97.9-146
Roche cobas b 121/123/221 Dedicated \ Electrode	mmHg	99.6	80.4-119
Nova Stat Profile pHox series Dedicated \ Electrode	mmHg	121	96.9-145
Radiometer ABL Series Dedicated \ Electrode	mmHg	109	87-131

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。